

regular servi**CEs**



Knowledge

Feature Publication
HOT-TOPIC

PROPOSAL

2025/0404 (COD)

**FOR AMENDING REGULATION (EU) 2017/745 OF THE
EUROPEAN PARLIAMENT AND OF THE COUNCIL**

on Medical Devices

English

Revision J

You can find more feature publications here:

www.regular-services.com



regular services rs GmbH
Landsberger Straße 155
80687 München
Germany

fon +49 (0)89 45 20 676 – 0
mail info@regular-services.com
www.regular-services.com

Note

This Document is for information purposes only.

Every care has been taken in preparing this document, regular services gives no warranties of whatever nature in respect of this documents, including but not limited to the accuracy or completeness of any information, facts and/ or opinions contained therein. Authors cannot be held liable for the use of and reliance of the information in this document.

History

Revision	Description of change or change identifier
A	Initial version
B	Corrigendum, OJ, L 117, 03.05.2019
C	Corrigendum, OJ, L 334, 27.12.2019
D	Not used
E	Amended by Regulation (EU) 2020/561, Official Journal, L130, 24.04.2020 Amended by Commission Delegated Regulation (EU) 2023/502, Official Journal, L70, 08.03.2023 Amended by Regulation (EU) 2023/607, Official Journal, L80, 20.03.2023
F	Change of legal form from Regular Services GbR to regular services rs GmbH
G	Amended by Regulation (EU) 2024/1860, Official Journal of the European Union, L1860, 09.07.2024
H	PROPOSAL 2025/0404 (COD)
J	PROPOSAL 2025/0404 (COD) update

Introduction

This **regular services Feature Publication** compares the present valid European Regulation 2017/745 (MDR), Revision 09.07.2024, with proposed amendments outlined in PROPOSAL 2025/0404 (COD).

The following colors have been used to illustrate the scope of proposed change:

Red: **New text**

~~Red~~: **Deleted text**

Blue [Green]: **Changed text** [previous text]

Black on Grey: **Unchanged text moved** to other spot

Note

The below listed sections of PROPOSAL 2025/0404 (COD) are not addressed by this Feature Publication. For details please refer to PROPOSAL 2025/0404 (COD).

EXPLANATORY MEMORANDUM

1. **CONTEXT OF THE PROPOSAL**
2. **LEGAL BASIS, SUBSIDIARITY AND PROPORTIONALITY**
3. **RESULTS OF EX-POST EVALUATIONS, STAKEHOLDER CONSULTATIONS AND IMPACT ASSESSMENTS**
4. **BUDGETARY IMPLICATIONS**
5. **OTHER ELEMENTS**

Contents

Chapter I Scope and definitions	11
Article 1 Subject matter and scope	11
Article 2 Definitions	13
Article 3 Amendment and implementation of certain definitions	20
Article 4 Regulatory status of products	20
Article 4a Opinion on and determination of the regulatory status of a product	21
Chapter II Making available on the market and putting into service of devices, obligations of economic operators, reprocessing, CE marking, free movement	22
Article 5 Placing on the market and putting into service	22
Article 6 Distance sales	23
Article 7 Claims	24
Article 8 Use of harmonised standards	24
Article 9 Common specifications	25
Article 10 General obligations of manufacturers	25
Article 10a Obligations in case of interruption or discontinuation of supply of certain devices	28
Article 11 Authorised representative	29
Article 12 Change of authorised representative	30
Article 13 General obligations of importers	30
Article 14 General obligations of distributors	31
Article 15 Person responsible for regulatory compliance	32
Article 16 Cases in which obligations of manufacturers apply to importers, distributors or other persons	33
Article 17 Single-use devices and their reprocessing of devices that are not for single use	34
Article 18 Implant card and information to be supplied to the patient with an implanted device	36
Article 19 EU declaration of conformity	36
Article 20 CE marking of conformity	37
Article 21 Devices for special purposes	37
Article 22 Systems and procedure packs	38
Article 23 Parts and components	39
Article 24 Free movement	39
Chapter III Identification and traceability of devices, registration of devices and of economic operators, summary of safety and clinical performance, European database on medical devices	39
Article 25 Identification within the supply chain	39
Article 26 Medical devices nomenclature	39
Article 27 Unique Device Identification system	40
Article 28 UDI database	42
Article 29 Registration of devices and systems or procedure packs	42
Article 30 Electronic system for registration of economic operators	43
Article 31 Registration of economic operators [manufacturers, authorised representatives and importers]	43
Article 32 Summary of safety and clinical performance	44
Article 33 European database on medical devices	45
Article 34 Functionality of Eudamed	46
Chapter IV Notified bodies	46
Article 35 Authorities responsible for notified bodies	46
Article 36 Requirements relating to notified bodies	47
Article 37 Subsidiaries and subcontracting	48
Article 38 Application by conformity assessment bodies for designation	48
Article 39 Assessment of the application	49

Article 40	Nomination of experts for joint assessment of applications for notification	52
Article 40a	Funding of activities related to designation and monitoring of notified bodies	52
Article 41	Language requirements	52
Article 42	Designation and notification procedure	52
Article 43	Identification number and list of notified bodies	54
Article 44	Monitoring and re-assessment of notified bodies	54
Article 45	Review of notified body assessment of technical documentation and clinical evaluation documentation	56
Article 46	Changes to designations and notifications	57
Article 47	Challenge to the competence of notified bodies	59
Article 48	Peer review and exchange of experience between authorities responsible for notified bodies	60
Article 49	Coordination of notified bodies	60
Article 50	Access to notified bodies and fees [List of standard fees]	61
Chapter V	Classification and conformity assessment	62
SECTION 1	CLASSIFICATION	62
Article 51	Classification of devices	62
Article 51a	Classification in the event of a dispute between manufacturer and notified body	63
Article 51b	Challenges to the classification of CE marked devices	63
SECTION 2	CONFORMITY ASSESSMENT	64
Article 52	Conformity assessment procedures	64
Article 52a	Conformity assessment of breakthrough devices and of orphan devices	66
Article 52b	Digitalisation of technical documentation, conformity assessment and reports	67
Article 53	Involvement of notified bodies in conformity assessment procedures	67
Article 54	Clinical evaluation consultation procedure for certain high-risk [class III and class IIb] devices	68
Article 55	Mechanism for scrutiny of conformity assessments of certain class III and class IIb devices	69
Article 56	Certificates of conformity	70
Article 57	Electronic system on notified bodies and on certificates of conformity	70
Article 58	Voluntary change of notified body	71
Article 59	Derogation from the conformity assessment procedures	71
Article 59a	Derogations from certain requirements in the event of a serious cross-border threat to health, disaster or crisis	72
Article 59b	Regulatory sandboxes at national level	73
Article 59c	Union regulatory sandboxes	74
Article 60	Certificate of free sale	74
Chapter VI	Clinical evaluation and clinical investigations	75
Article 61	Clinical evaluation	75
Article 62	General requirements regarding clinical investigations conducted to demonstrate conformity of devices	77
Article 63	Informed consent	79
Article 64	Clinical investigations on incapacitated subjects	80
Article 65	Clinical investigations on minors	81
Article 66	Clinical investigations on pregnant or breastfeeding women	81
Article 67	Additional national measures	82
Article 68	Clinical investigations in emergency situations	82
Article 69	Damage compensation	83
Article 70	Application for clinical investigations	83
Article 71	Assessment by Member States	84
Article 72	Conduct of a clinical investigation	85
Article 73	Electronic system on clinical investigations	86
Article 74	Clinical investigations regarding devices bearing the CE marking	87
Article 75	Substantial modifications to clinical investigations	87

Article 76 Corrective measures to be taken by Member States and information exchange between Member States.....	88
Article 77 Information from the sponsor at the end of a clinical investigation or in the event of a temporary halt or early termination	88
Article 78 Coordinated assessment procedure for clinical investigations.....	89
Article 79 Review of coordinated assessment procedure.....	91
Article 79a Clinical investigations in combined studies.....	91
Article 80 Recording and reporting of adverse events that occur during clinical investigations	91
Article 81 Implementing acts.....	92
Article 82 Requirements regarding other clinical investigations	93
Chapter VII Post-market surveillance, vigilance and market surveillance	93
SECTION 1 POST-MARKET SURVEILLANCE.....	93
Article 83 Post-market surveillance system of the manufacturer.....	93
Article 84 Post-market surveillance plan.....	94
Article 85 Post-market surveillance report.....	94
Article 86 Periodic safety update report.....	94
SECTION 2 VIGILANCE	95
Article 87 Reporting of serious incidents and field safety corrective actions.....	95
Article 87a Reporting of actively exploited vulnerabilities and severe incidents related to devices	96
Article 88 Trend reporting	97
Article 89 Analysis of serious incidents and field safety corrective actions	97
Article 90 Analysis of vigilance data	99
Article 91 Implementing acts.....	100
Article 92 Electronic system on vigilance and on post-market surveillance	100
SECTION 3 MARKET SURVEILLANCE.....	101
Article 93 Market surveillance activities	101
Article 94 Evaluation of devices suspected of presenting an unacceptable risk or other non-compliance	102
Article 95 Procedure for dealing with devices presenting an unacceptable risk to health and safety	103
Article 96 Procedure for evaluating national measures at Union level	104
Article 97 Other non-compliance.....	104
Article 98 Preventive health protection measures	105
Article 99 Good administrative practice	105
Article 100 Electronic system on market surveillance.....	106
Chapter VIII Cooperation between Member States, the MDCG [Medical Device Coordination Group], expert laboratories, expert panels and device registers, conflicts of interest and international cooperation	106
Article 101 Competent authorities.....	106
Article 102 Cooperation	107
Article 103 Medical Device Coordination Group	107
Article 104 Support by the Commission.....	108
Article 105 Tasks of the MDCG	108
Article 106 Expert panels [Provision of scientific, technical and clinical opinions and advice]	109
Article 106a Expert laboratories.....	112
Article 106b Support by the EMA.....	113
Article 107 Conflict of interests	114
Article 108 Device registers and data banks	114
SECTION 2 International cooperation.....	114
Article 108a International regulatory and administrative cooperation.....	114
Article 108b Reliance mechanisms.....	115
Article 108c Capacity building and technical assistance	115
Chapter IX Confidentiality, data protection, funding and penalties	116
Article 109 Confidentiality	116

Article 110 Data protection.....	116
Article 110a Submission of information or documents.....	116
Article 111 Levying of fees.....	116
Article 112 Funding of activities related to designation and monitoring of notified bodies	117
Article 113 Penalties	117
Chapter X Final provisions	117
Article 114 Committee procedure	117
Article 115 Exercise of the delegation	117
Article 116 Separate delegated acts for different delegated powers	118
Article 117 Amendment to Directive 2001/83/EC	118
Article 118 Amendment to Regulation (EC) No 178/2002	118
Article 119 Amendment to Regulation (EC) No 1223/2009	118
Article 120 Transitional provisions	119
Article 121 Evaluation	122
Article 122 Repeal.....	122
Article 123 Entry into force and date of application	123
ANNEX I GENERAL SAFETY AND PERFORMANCE REQUIREMENTS	126
Chapter I General requirements.....	126
Chapter II Requirements regarding design and manufacture	127
Chapter III Requirements regarding the information supplied with the device	136
ANNEX II TECHNICAL DOCUMENTATION.....	142
ANNEX III TECHNICAL DOCUMENTATION ON POST-MARKET SURVEILLANCE.....	146
ANNEX IV EU DECLARATION OF CONFORMITY	147
ANNEX V CE MARKING OF CONFORMITY.....	148
ANNEX VI INFORMATION TO BE SUBMITTED UPON THE REGISTRATION OF DEVICES AND ECONOMIC OPERATORS IN ACCORDANCE WITH ARTICLES 29(4) AND 31, CORE DATA ELEMENTS TO BE PROVIDED TO THE UDI DATABASE TOGETHER WITH THE UDI-DI IN ACCORDANCE WITH ARTICLES 28 AND 29, AND THE UDI SYSTEM	149
PART A INFORMATION TO BE SUBMITTED UPON THE REGISTRATION OF DEVICES AND ECONOMIC OPERATORS IN ACCORDANCE WITH ARTICLES 29(4) AND 31	149
PART B CORE DATA ELEMENTS TO BE PROVIDED TO THE UDI DATABASE TOGETHER WITH THE UDI-DI IN ACCORDANCE WITH ARTICLES 28 AND 29	149
PART C The UDI System	151
ANNEX VII REQUIREMENTS TO BE MET BY NOTIFIED BODIES	157
ANNEX VIII CLASSIFICATION RULES	176
Chapter I Definitions specific to classification rules	176
Chapter II Implementing rules	176
Chapter III Classification rules	177
ANNEX IX CONFORMITY ASSESSMENT BASED ON A QUALITY MANAGEMENT SYSTEM AND ON ASSESSMENT OF TECHNICAL DOCUMENTATION	183
Chapter I Quality Management System	183
Chapter II Assessment of the technical documentation	187
Chapter III Administrative provisions.....	192
ANNEX X CONFORMITY ASSESSMENT BASED ON TYPE EXAMINATION.....	193
ANNEX XI CONFORMITY ASSESSMENT BASED ON PRODUCT CONFORMITY VERIFICATION	195
PART A PRODUCTION QUALITY ASSURANCE	195
PART B PRODUCT VERIFICATION	198
ANNEX XII CERTIFICATES ISSUED BY A NOTIFIED BODY	201
Chapter I General Requirements	201
Chapter II Minimum content of the certificates	201
ANNEX XIII PROCEDURE FOR CUSTOM-MADE DEVICES	203
ANNEX XIV CLINICAL EVALUATION AND POST-MARKET CLINICAL FOLLOW-UP	204
PART A CLINICAL EVALUATION.....	204

PART B POST-MARKET CLINICAL FOLLOW-UP	205
ANNEX XV CLINICAL INVESTIGATIONS	207
Chapter I General requirements	207
Chapter II Documentation regarding the application for clinical investigation	207
Chapter III Other obligations of the sponsor	214
ANNEX XVI LIST OF GROUPS OF PRODUCTS WITHOUT AN INTENDED MEDICAL PURPOSE REFERRED TO IN ARTICLE 1(2).....	216
ANNEX XVII CORRELATION TABLE	217

Chapter I Scope and definitions

Article 1 *Subject matter and scope*

1. This Regulation lays down rules concerning the placing on the market, making available on the market or putting into service of medical devices for human use and accessories for such devices in the Union. This Regulation also applies to clinical investigations concerning such medical devices and accessories conducted in the Union.

2. This Regulation shall also apply, as from the date of application of common specifications adopted pursuant to Article 9, to the groups of products without an intended medical purpose that are listed in Annex XVI, **and their accessories**, taking into account the state of the art, and in particular existing harmonised standards for analogous devices with a medical purpose, based on similar technology. The common specifications for each of the groups of products listed in Annex XVI shall address, at least, application of risk management as set out in Annex I for the group of products in question and, where necessary, clinical evaluation regarding safety.

The necessary common specifications shall be adopted by 26 May 2021. They shall apply as from six months after the date of their entry into force or from 26 May 2021, whichever is the latest.

Notwithstanding Article 122, Member States' measures regarding the qualification of the products covered by Annex XVI as medical devices pursuant to Directive 93/42/EEC shall remain valid until the date of application, as referred to in the first subparagraph, of the relevant common specifications for that group of products.

This Regulation also applies to clinical investigations conducted in the Union concerning the products referred to in the first subparagraph.

3. Devices with both a medical and a non-medical intended purpose shall fulfil cumulatively the requirements applicable to devices with an intended medical purpose and those applicable to devices without an intended medical purpose.

4. For the purposes of this Regulation, medical devices, accessories for medical devices, and products listed in Annex XVI **and their accessories**, to which this Regulation applies pursuant to paragraph 2 shall hereinafter be referred to as 'devices'.

5. Where justified on account of the similarity between a device with an intended medical purpose placed on the market and a product without an intended medical purpose in respect of their characteristics and risks, the Commission is empowered to adopt delegated acts in accordance with Article 115 to amend the list in Annex XVI, by adding new groups of products, in order to protect the health and safety of users or other persons or other aspects of public health.

6. This Regulation does not apply to:

- (a) *in vitro* diagnostic medical devices covered by Regulation (EU) 2017/746;
- (b) medicinal products as defined in point 2 of Article 1 of Directive 2001/83/EC. In deciding whether a product falls under Directive 2001/83/EC or under this Regulation, particular account shall be taken of the principal mode of action of the product;
- (c) advanced therapy medicinal products covered by Regulation (EC) No 1394/2007;
- (d) human blood, blood products, plasma or blood cells of human origin or devices which incorporate, when placed on the market or put into service, such blood products, plasma or cells, except for devices referred to in paragraph 8 of this Article;

Proposal of Regulation (EU) 2017/745 Articles

(e) cosmetic products covered by Regulation (EC) No 1223/2009;

(f) transplants, tissues or cells of animal origin, or their derivatives, or products containing or consisting of them; however this Regulation does apply to devices manufactured utilising tissues or cells of animal origin, or their derivatives, which are non-viable or are rendered non-viable;

(g) organs intended for transplantation [transplants] falling within the scope of Directive 2010/53/EU of the European Parliament and of the Council* [covered by Directive 2004/23/EC] or substances [tissues or cells] of human origin falling within the scope of Regulation (EU) 2024/1938 of the European Parliament and of the Council**, or their derivatives, or products containing or consisting of them; however, this Regulation does apply to devices manufactured utilising derivatives of substances [tissues or cells] of human origin which are non-viable or are rendered non-viable;

* Regulation (EU) 2024/1938 of the European Parliament and of the Council of 13 June 2024 on standards of quality and safety for substances of human origin intended for human application and repealing Directives 2002/98/EC and 2004/23/EC (OJ L, 2024/1938, 17.7.2024, ELI: <http://data.europa.eu/eli/reg/2024/1938/oj>).

** Directive 2010/53/EU of the European Parliament and of the Council of 7 July 2010 on standards of quality and safety of human organs intended for transplantation (OJ L 207, 6.8.2010, p. 14, ELI: <http://data.europa.eu/eli/dir/2010/53/oj>);

(h) products, other than those referred to in points (d), (f) and (g), that contain or consist of viable biological material or viable organisms, including living micro-organisms, bacteria, fungi or viruses in order to achieve or support the intended purpose of the product;

(i) food covered by Regulation (EC) No 178/2002.

7. Any device which, when placed on the market or put into service, incorporates as an integral part an in vitro diagnostic medical device as defined in Article 2, point (2), of Regulation (EU) 2017/746 that has an action ancillary to that of the device in which it is incorporated, shall be governed by this Regulation. In that case, the relevant general safety and performance requirements set out in Annex I to of Regulation (EU) 2017/746, shall apply with regard to the safety and performance of the in vitro diagnostic medical device part of the device.

However, if the action of the in vitro diagnostic medical device is principal and not ancillary to that of the device in which it is incorporated, the integral product shall be governed by Regulation (EU) 2017/746. In that case, the relevant general safety and performance requirements set out in Annex I to this Regulation shall apply with regard to the safety and performance of the device part.

8. Any device which, when placed on the market or put into service, incorporates, as an integral part, a substance which, if used separately, would be considered to be a medicinal product as defined in point 2 of Article 1 of Directive 2001/83/EC, including a medicinal product derived from human blood or human plasma as defined in point 10 of Article 1 of that Directive, and that has an action ancillary to that of the device, shall be assessed and authorised in accordance with this Regulation.

However, if the action of that substance is principal and not ancillary to that of the device, the integral product shall be governed by Directive 2001/83/EC or Regulation (EC) No 726/2004 of the European Parliament and of the Council¹, as applicable. In that case, the relevant general safety and performance requirements set out in Annex I to this Regulation shall apply as far as the safety and performance of the device part are concerned.

¹ Regulation (EC) No 726/2004 of the European Parliament and of the Council of 31 March 2004 laying down Community procedures for the authorisation and supervision of medicinal products for human and veterinary use and establishing a European Medicines Agency (OJ L 136, 30.4.2004, p. 1).

Proposal of Regulation (EU) 2017/745 Articles

9. Any device which is intended to administer a medicinal product as defined in point 2 of Article 1 of Directive 2001/83/EC shall be governed by this Regulation, without prejudice to the provisions of that Directive and of Regulation (EC) No 726/2004 with regard to the medicinal product.

However, if the device intended to administer a medicinal product and the medicinal product are placed on the market in such a way that they form a single integral product which is intended exclusively for use in the given combination and which is not reusable, that single integral product shall be governed by Directive 2001/83/EC or Regulation (EC) No 726/2004, as applicable. In that case, the relevant general safety and performance requirements set out in Annex I to this Regulation shall apply as far as the safety and performance of the device part of the single integral product are concerned.

10. Any device which, when placed on the market or put into service, incorporates, as an integral part, non-viable substances [tissues or cells] of human origin or their derivatives that have an action ancillary to that of the device in which they are incorporated shall be assessed and authorised in accordance with this Regulation. ~~In that case, the provisions for donation, procurement and testing laid down in Directive 2004/23/EC shall apply.~~

However, if the action of those tissues or cells or their derivatives is principal and not ancillary to that of the device and the product is not governed by Regulation (EC) No 1394/2007, the product shall be governed by (EU) 2024/1938 [Directive 2004/23/EC]. In that case, the relevant general safety and performance requirements set out in Annex I to this Regulation shall apply as far as the safety and performance of the device part are concerned.

11. This Regulation is specific Union legislation within the meaning of Article 2(3) of Directive 2014/30/EU.

12. Devices which are also machinery within the meaning of point (a) of the second paragraph of Article 2 of Directive 2006/42/EC of the European Parliament and of the Council¹ shall, where a hazard relevant under that Directive exists, also meet the essential health and safety requirements set out in Annex I to that Directive to the extent to which those requirements are more specific than the general safety and performance requirements set out in Chapter II of Annex I to this Regulation.

13. This Regulation shall not affect the application of Directive 2013/59/Euratom.

14. This Regulation shall not affect the right of a Member State to restrict the use of any specific type of device in relation to aspects not covered by this Regulation.

15. This Regulation shall not affect national law concerning the organisation, delivery or financing of health services and medical care, such as the requirement that certain devices may only be supplied on a medical prescription, the requirement that only certain health professionals or healthcare institutions may dispense or use certain devices or that their use be accompanied by specific professional counselling.

16. Nothing in this Regulation shall restrict the freedom of the press or the freedom of expression in the media in so far as those freedoms are guaranteed in the Union and in the Member States, in particular under Article 11 of the Charter of Fundamental Rights of the European Union.

Article 2

Definitions

For the purposes of this Regulation, the following definitions apply:

(1) 'medical device' means any instrument, apparatus, appliance, software, implant, reagent, material or other article intended by the manufacturer to be used, alone or in combination, for human beings for one or more of the following specific medical purposes:

– diagnosis, prevention, monitoring, prediction, prognosis, treatment or alleviation of disease,

¹ Directive 2006/42/EC of the European Parliament and of the Council of 17 May 2006 on machinery, and amending Directive 95/16/EC (OJ L 157, 9.6.2006, p. 24).

Proposal of Regulation (EU) 2017/745 Articles

- diagnosis, monitoring, treatment, alleviation of, or compensation for, an injury or disability,
- investigation, replacement or modification of the anatomy or of a physiological or pathological process or state,
- providing information by means of *in vitro* examination of specimens derived from the human body, including organ, blood and tissue donations,

and which does not achieve its principal intended action by pharmacological, immunological or metabolic means, in or on the human body, but which may be assisted in its function by such means.

The following products shall also be deemed to be medical devices:

- devices for the control or support of conception;
- products specifically intended for the cleaning, disinfection or sterilisation of devices as referred to in Article 1(4) ~~and of those referred to in the first paragraph of this point.~~

(2) 'accessory for a medical device' means an article which, whilst not being itself a medical device, is intended by its manufacturer to be used together with one or several particular medical device(s) to specifically enable the medical device(s) to be used in accordance with its/their intended purpose(s) or to specifically and directly assist the medical functionality of the medical device(s) in terms of its/their intended purpose(s);

(3) 'custom-made device' means any device specifically made in accordance with a written prescription of any person authorised by national law by virtue of that person's professional qualifications which gives, under that person's responsibility, specific design characteristics, and is intended for the sole use of a particular patient exclusively to meet their individual conditions and needs.

However, mass-produced devices which need to be adapted to meet the specific requirements of any professional user and devices which are mass-produced by means of industrial manufacturing processes in accordance with the written prescriptions of any authorised person shall not be considered to be custom-made devices;

(4) 'active device' means any device, the operation of which depends on a source of energy other than that generated by the human body for that purpose, or by gravity, and which acts by changing the density of or converting that energy. Devices intended to transmit energy, substances or other elements between an active device and the patient, without any significant change, shall not be deemed to be active devices.

Software shall also be deemed to be an active device;

(5) 'implantable device' means any device, including those that are partially or wholly absorbed, which is intended:

- to be totally introduced into the human body, or
- to replace an epithelial surface or the surface of the eye,

by clinical intervention and which is intended to remain in place after the procedure.

Any device intended to be partially introduced into the human body by clinical intervention and intended to remain in place after the procedure for at least 30 days shall also be deemed to be an implantable device;

(6) 'invasive device' means any device which, in whole or in part, penetrates inside the body, either through a body orifice or through the surface of the body;

(7) 'generic device group' means a set of devices having the same or similar intended purposes and [or] a commonality of technology allowing them to be classified in a generic manner not reflecting specific characteristics;

(8) 'single-use device' means a device that is intended to be used on one individual during a single procedure;

Proposal of Regulation (EU) 2017/745 Articles

(9) 'falsified device' means any device with a false presentation of its identity and/or of its source and/or its CE marking certificates or documents relating to CE marking procedures. This definition does not include unintentional non-compliance and is without prejudice to infringements of intellectual property rights;

(10) 'procedure pack' means a combination of products packaged together and placed on the market with the purpose of being used for a specific medical purpose;

(11) 'system' means a combination of products, either packaged together or not, which are intended to be inter-connected or combined to achieve a specific medical purpose;

(12) 'intended purpose' means the use for which a device is intended according to the data supplied by the manufacturer on the label, in the instructions for use or in promotional or sales materials or statements and as specified by the manufacturer in the clinical evaluation;

(13) 'label' means the written, printed or graphic information appearing either on the device itself, or on the packaging of each unit or on the packaging of multiple devices;

(14) 'instructions for use' means the information provided by the manufacturer to inform the user of a device's intended purpose and proper use and of any precautions to be taken;

(15) 'Unique Device Identifier' ('UDI') means a series of numeric or alphanumeric characters that is created through internationally accepted device identification and coding standards and that allows unambiguous identification of specific devices on the market;

(16) 'non-viable' means having no potential for metabolism or multiplication;

(17) 'derivative' means a 'non-cellular substance' extracted from human or animal tissue or cells through a manufacturing process. The final substance used for manufacturing of the device in this case does not contain any cells or tissues;

~~(18) 'nanomaterial' means a natural, incidental or manufactured material containing particles in an unbound state or as an aggregate or as an particle and where, for 50 % or more of the particles in the number size distribution, one or more external dimensions is in the size range 1-100 nm;~~

~~Fullerenes, graphene flakes and single wall carbon nanotubes with one or more external dimensions below 1 nm shall also be deemed to be nanomaterials;~~

~~(19) 'particle', for the purposes of the definition of nanomaterial in point (18), means a minute piece of matter with defined physical boundaries;~~

~~(20) 'particle', for the purposes of the definition of nanomaterial in point (18), means a collection of weakly bound particles or aggregates where the resulting external surface area is similar to the sum of the surface areas of the individual components;~~

~~(21) 'aggregate', for the purposes of the definition of nanomaterial in point (18), means a particle comprising of strongly bound or fused particles;~~

(22) 'performance' means the ability of a device to achieve its intended purpose as stated by the manufacturer;

(23) 'risk' means the combination of the probability of occurrence of harm and the severity of that harm;

(24) 'benefit-risk determination' means the analysis of all assessments of benefit and risk of possible relevance for the use of the device for the intended purpose, when used in accordance with the intended purpose given by the manufacturer;

(25) 'compatibility' is the ability of a device, including software, when used together with one or more other devices in accordance with its intended purpose, to:

(a) perform without losing or compromising the ability to perform as intended, and/or

Proposal of Regulation (EU) 2017/745 Articles

(b) integrate and/or operate without the need for modification or adaption of any part of the combined devices, and/or

(c) be used together without conflict/interference or adverse reaction.

(26) 'interoperability' is the ability of two or more devices, including software, from the same manufacturer or from different manufacturers, to:

(a) exchange information and use the information that has been exchanged for the correct execution of a specified function without changing the content of the data, and/or

(b) communicate with each other, and/or

(c) work together as intended.

(27) 'making available on the market' means any supply of a device, other than an investigational device, for distribution, consumption or use on the Union market in the course of a commercial activity, whether in return for payment or free of charge;

(28) 'placing on the market' means the first making available of a device, other than an investigational device, on the Union market;

(29) 'putting into service' means the stage at which a device, other than an investigational device, has been made available to the final user as being ready for use on the Union market for the first time for its intended purpose;

(30) 'manufacturer' means a natural or legal person who manufactures or fully refurbishes a device or has a device designed, manufactured or fully refurbished, and markets that device under its name or trade mark;

(31) 'fully refurbishing', for the purposes of the definition of manufacturer, means the complete rebuilding of a device already placed on the market or put into service, or the making of a new device from used devices, to bring it into conformity with this Regulation, combined with the assignment of a new lifetime to the refurbished device;

(32) 'authorised representative' means any natural or legal person established within the Union who has received and accepted a written mandate from a manufacturer, located outside the Union, to act on the manufacturer's behalf in relation to specified tasks with regard to the latter's obligations under this Regulation;

(33) 'importer' means any natural or legal person established within the Union that places a device from a third country on the Union market;

(34) 'distributor' means any natural or legal person in the supply chain, other than the manufacturer or the importer, that makes a device available on the market, up until the point of putting into service;

(35) 'economic operator' means a manufacturer, an authorised representative, an importer, a distributor or the person referred to in Article 22(1) and 22(3);

(36) 'health institution' means an organisation the primary purpose of which is the care or treatment of patients or the promotion of public health;

(37) 'user' means any healthcare professional or lay person who uses a device;

(38) 'lay person' means an individual who does not have formal education in a relevant field of healthcare or medical discipline;

(39) 'reprocessing' means a process carried out on a used device in order to allow its safe reuse including cleaning, disinfection, sterilisation and related procedures, as well as testing and restoring the technical and functional safety of the used device;

Proposal of Regulation (EU) 2017/745 Articles

(40) 'conformity assessment' means the process demonstrating whether the requirements of this Regulation relating to a device have been fulfilled;

(41) 'conformity assessment body' means a body that performs third-party conformity assessment activities including calibration, testing, certification and inspection;

(42) 'notified body' means a conformity assessment body designated in accordance with this Regulation;

(43) 'CE marking of conformity' or 'CE marking' means a marking by which a manufacturer indicates that a device is in conformity with the applicable requirements set out in this Regulation and other applicable Union harmonisation legislation providing for its affixing;

(44) 'clinical evaluation' means a systematic and planned process to continuously generate, collect, analyse and assess the clinical data pertaining to a device in order to verify the safety and performance, including clinical benefits, of the device when used as intended by the manufacturer;

(45) 'clinical investigation' means any systematic investigation involving one or more human Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operators, undertaken to assess the safety or performance of a device;

(46) 'investigational device' means a device that is assessed in a clinical investigation;

(47) 'clinical investigation plan' means a document that describes the rationale, objectives, design, methodology, monitoring, statistical considerations, organisation and conduct of a clinical investigation;

(48) 'clinical data' means information concerning safety or performance that is generated from the use of a device and is sourced from any of the following:

- clinical investigations of the device concerned ~~or of a device for which equivalence to the device concerned can be demonstrated;~~

- ~~clinical investigation(s) or~~ other studies ~~published~~ [reported] in scientific literature ~~on the device concerned or of a device for which equivalence to the device concerned [in question] can be demonstrated;~~

- ~~other clinical experience~~ [reports] published in peer-reviewed scientific literature ~~on other clinical experience of either with the device concerned [in question] or a device for which equivalence to the device concerned [in question] can be demonstrated;~~

- clinically relevant information coming from post-market surveillance, in particular the post-market clinical follow-up;

(49) 'sponsor' means any individual, company, institution or organisation which takes responsibility for the initiation, for the management and setting up of the financing of the clinical investigation;

(50) 'subject' means an individual who participates in a clinical investigation;

(51) 'clinical evidence' means clinical data and clinical evaluation results pertaining to a device of a sufficient amount and quality to allow a qualified assessment of whether the device is safe and achieves the intended clinical benefit(s), when used as intended by the manufacturer;

(52) 'clinical performance' means the ability of a device, resulting from any direct or indirect medical effects which stem from its technical or functional characteristics, including diagnostic characteristics, to

Proposal of Regulation (EU) 2017/745 Articles

achieve its intended purpose as claimed by the manufacturer, thereby leading to a clinical benefit for patients, when used as intended by the manufacturer;

(53) 'clinical benefit' means the positive impact of a device on the health of an individual, expressed in terms of a meaningful, measurable, patient-relevant clinical outcome(s), including outcome(s) related to diagnosis, or a positive impact on patient management or public health;

(54) 'investigator' means an individual responsible for the conduct of a clinical investigation at a clinical investigation site;

(55) 'informed consent' means a Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operator's free and voluntary expression of his or her willingness to participate in a particular clinical investigation, after having been informed of all aspects of the clinical investigation that are relevant to the Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operator's decision to participate or, in the case of minors and of incapacitated Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operators, an authorisation or agreement from their legally designated representative to include them in the clinical investigation;

(56) 'ethics committee' means an independent body established in a Member State in accordance with the law of that Member State and empowered to give opinions for the purposes of this Regulation, taking into account the views of laypersons, in particular patients or patients' organisations;

(57) 'adverse event' means any untoward medical occurrence, unintended disease or injury or any untoward clinical signs, including an abnormal laboratory finding, in subjects, users or other persons, in the context of a clinical investigation, whether or not related to the investigational device;

(58) 'serious adverse event' means any adverse event that led to any of the following:

- (a) death,
- (b) serious deterioration in the health of the subject, that resulted in any of the following:
 - (i) life-threatening illness or injury,

Proposal of Regulation (EU) 2017/745 Articles

- (ii) permanent impairment of a body structure or a body function,
- (iii) hospitalisation or prolongation of patient hospitalisation,
- (iv) medical or surgical intervention to prevent life-threatening illness or injury or permanent impairment to a body structure or a body function,
- (v) chronic disease,
- (c) foetal distress, foetal death or a congenital physical or mental impairment or birth defect;
- (59) 'device deficiency' means any inadequacy in the identity, quality, durability, reliability, safety or performance of an investigational device, including malfunction, use errors or inadequacy in information supplied by the manufacturer;
- (60) 'post-market surveillance' means all activities carried out by manufacturers in cooperation with other economic operators to institute and keep up to date a systematic procedure to proactively collect and review experience gained from devices they place on the market, make available on the market or put into service for the purpose of identifying any need to immediately apply any necessary corrective or preventive actions;
- (61) 'market surveillance' means the activities carried out and measures taken by competent authorities to check and ensure that devices comply with the requirements set out in the relevant Union harmonisation legislation and do not endanger health, safety or any other aspect of public interest protection;
- (62) 'recall' means any measure aimed at achieving the return of a device that has already been made available to the end user;
- (63) 'withdrawal' means any measure aimed at preventing a device in the supply chain from being further made available on the market;
- (64) 'incident' means any malfunction or deterioration in the characteristics or performance of a device made available on the market, including use-error due to ergonomic features, as well as any inadequacy in the information supplied by the manufacturer and any undesirable side-effect;
- (65) 'serious incident' means any incident that directly or indirectly led, might have led or might lead to any of the following:
 - (a) the death of a patient, user or other person,
 - (b) the temporary or permanent serious deterioration of a patient's, user's or other person's state of health,
 - (c) a serious public health threat;
- (66) 'serious public health threat' means an event which could result in imminent risk of death, serious deterioration in a person's state of health, or serious illness, that may require prompt remedial action, and that may cause significant morbidity or mortality in humans, or that is unusual or unexpected for the given place and time;
- (67) 'corrective action' means action taken to eliminate the cause of a potential or actual non-conformity or other undesirable situation;
- (68) 'field safety corrective action' means corrective action taken by a manufacturer for technical or medical reasons to prevent or reduce the risk of a serious incident in relation to a device made available on the market;
- (69) 'field safety notice' means a communication sent by a manufacturer to users or customers in relation to a field safety corrective action;
- (70) 'harmonised standard' means a European standard as defined in point (1)(c) of Article 2 of Regulation (EU) No 1025/2012;

Proposal of Regulation (EU) 2017/745 Articles

(71) 'common specifications' (CS) means a set of technical and/or clinical requirements, other than a standard, that provides a means of complying with the legal obligations applicable to a device, process or system.

(72) 'well-established technology device' means a device that belongs to a generic device group, which fulfils the following criteria:

- (a) it has simple, common and stable design;
- (b) it has not been associated with safety issues in the past;
- (c) it has well-known clinical performance characteristics and comprises standard of care devices with little evolution in indications and the state of the art;
- (d) it has a long history on the Union market;

(73) 'combined study' means a clinical trial, as defined in Article 2(2), point (2), of Regulation (EU) No 536/2014, of one or more medicinal products, combined with a performance study of one or more in vitro diagnostic medical devices as defined in Article 2, point (42), of Regulation (EU) 2017/746, and/or with a clinical investigation of one or more devices;

(74) 'regulatory sandbox' means a controlled environment set up by a competent authority which offers manufacturers or prospective manufacturers the possibility to develop, test, validate and use, where appropriate in real-world conditions, an innovative product or technology potentially falling within the scope of this Regulation, pursuant to a sandbox plan for a limited time under regulatory supervision;

(75) 'sandbox plan' means a document agreed between the participating manufacturer(s) or prospective manufacturer(s) and the competent authority describing the objectives, conditions, timeframe, methodology and requirements for the activities carried out within the regulatory sandbox;

(76) 'Union regulatory sandbox' means a controlled environment set up by the Commission for testing alternative or new regulatory requirements or enforcement practices and appraising their validity in comparison with existing requirements and practices under this Regulation for a limited time.

Article 3

Amendment and implementation of certain definitions

1. The Commission is empowered to adopt delegated acts in accordance with Article 115 in order to amend the definition of ~~nanomaterial set out in point (18) and the related definitions in points (19), (20) and (21) of Article 2~~ well-established technology device set out in Article 2, point (72), in the light of technical and scientific progress and taking into account definitions agreed at Union and international level.

2. The Commission may, by means of implementing acts, draw up non-exhaustive lists of devices that fall under, or of devices that do not fall under the definition of well-established technology device in Article 2, point (72).

Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

Article 4

Regulatory status of products

1. The competent authorities of the Member States shall coordinate their activities when [Without prejudice to Article 2(2) of Directive 2001/83/EC, upon a duly substantiated request of a Member State, the Commission shall, after consulting the Medical Device Coordination Group established under Article 103 of this Regulation ('MDCG'), by means of implementing acts,] determining [determine] whether ~~or not~~ a specific product, or category, or group of products, falls within the definitions of 'medical device' set out in Article 2, point (1), or the definition of 'accessory for a medical device' set out in Article 2, point (2), or

Proposal of Regulation (EU) 2017/745 Articles

whether a product falls within the scope of Annex XVI or is an accessory for a product listed in that Annex. ~~Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3) of this Regulation.~~

2. [3], [4] The Member States [Commission] shall ensure an appropriate level of consultation of the relevant competent authorities of the Member States [that Member States share expertise] in the fields of ~~medical devices~~, in vitro diagnostic medical devices, medicinal products, substances of human origin (SoHO) [human tissues and cells], biocides, food products, cosmetics ~~and, if necessary~~, or other products subject to Union legislation, where the determination of whether a product has the regulatory status of a device involves aspects concerning the borderline with any of those types of products [in order to determine the appropriate regulatory status of a product, or category or group of products]. ~~When deliberating on the possible regulatory status as a device of products involving medicinal products, human tissues and cells, biocides or food products,~~ If that is the case, Member States [the Commission] shall also ensure an appropriate level of consultation of the relevant advisory or regulatory bodies established in the relevant Union legislation, such as the European Medicines Agency (EMA), the SoHO Coordination Board, the European Chemicals Agency (ECHA) and the European Food Safety Authority (EFSA), ~~as relevant.~~

3. Where a competent authority of a Member State, after having performed an evaluation in accordance with Article 94, considers that a product that is CE marked in accordance with Article 20, does not fall within the scope of this Regulation, it shall consult the competent authorities of the other Member States regarding its envisaged measure determining the regulatory status of the product in question.

4. Where a competent authority of a Member State raises a substantiated disagreement regarding the envisaged measure referred to in paragraph 3, the consulting authority shall refer the matter to an expert panel as referred to in Article 106 and give utmost consideration to the opinion of that expert panel.

5. The results of the coordination activities of the competent authorities in accordance with this Article and the opinions of the expert panel delivered in accordance with paragraph 4 of this Article and Article 4a(2) shall be made publicly available, without disclosing any confidential information as referred to in Article 109.

6. [2] The Commission may ~~also, on its own initiative, after consulting the MDCC, decide~~, by means of implementing acts, lay down the procedure, including timelines, for the application of paragraphs 1 to 4 of this Article and of Article 4a [on the issues referred to in paragraph 1 of this Article]. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

Article 4a

Opinion on and determination of the regulatory status of a product

1. A competent authority, a notified body, a manufacturer, a developer of a product or the Commission may submit a substantiated request for an opinion from an expert panel referred to in Article 106 on the question whether a specific product, or category or group of products, falls within the definitions of 'medical device' or 'accessory for a medical device', or whether a product falls within the scope of Annex XVI or is an accessory for a product listed in that Annex. Where, in such a request, the requester considers that the product in question is a device, the request shall also specify the proposed classification of the device in accordance with Article 51 and Annex VIII.

2. The expert panel shall provide its opinion without undue delay. The requester shall give utmost consideration to the opinion of the expert panel.

3. Having regard to the expert panel opinion referred to in paragraph 2 or in Article 4(4), a Member State may submit a substantiated request to the Commission to determine whether a specific product, or category or group of products, falls within the definitions of 'medical device' or 'accessory for a medical device', or whether a product falls within the scope of Annex XVI or is an accessory for a product listed in that Annex.

Proposal of Regulation (EU) 2017/745 Articles

The Commission shall decide on the substantiated request of the Member State or on its own initiative, by means of implementing acts, which shall be adopted in accordance with the examination procedure referred to in Article 114(3).

The Commission may ask the expert panel for clarifications or refer the opinion back to the expert panel for further consideration, including in cases where a Member State's substantiated request raises new questions of a scientific or technical nature.

4. This Article shall not apply where within the framework of another Union legislation the regulatory status of the product, or category or group of products concerned has been determined as falling within the scope of that other Union legislation, or where a procedure for the determination of the regulatory status is ongoing within the framework of another Union legislation.

Chapter II

Making available on the market and putting into service of devices, obligations of economic operators, reprocessing, CE marking, free movement

Article 5

Placing on the market and putting into service

1. A device may be placed on the market or put into service only if it complies with this Regulation when duly supplied and properly installed, maintained and used in accordance with its intended purpose.

2. A device shall meet the general safety and performance requirements set out in Annex I which apply to it, taking into account its intended purpose.

3. Demonstration of conformity with the general safety and performance requirements shall include a clinical evaluation in accordance with Article 61.

4. Devices that are manufactured and used within health institutions shall be considered as having been put into service.

5. With the exception of the relevant general safety and performance requirements set out in Annex I, the requirements of this Regulation shall not apply to devices, manufactured and used only within health institutions established in the Union, provided that all of the following conditions are met:

(a) the devices are not transferred to another legal entity, **except to another health institution in the duly justified interest of public health, patient safety or patient health, or to prepare or respond to a public health emergency;**

(b) manufacture and use of the devices occur under appropriate quality management systems,

(c) the health institution justifies in its documentation that the target patient group's specific needs cannot be met, or cannot be met at the appropriate level of performance by an equivalent device available on the market,

(d) **upon request by a [to its] competent authority,** the health institution provides information upon request on the use of such devices, which ~~shall~~ includes the [a] justification **referred to in points (a) and (c) [of their manufacturing, modification and use];**

(e) the health institution draws up a declaration, that it shall make publicly available, including:

(i) the name and address of the manufacturing health institution;

(ii) the details necessary to identify the devices;

(iii) a declaration that the devices meet the general safety and performance requirements set out in Annex I to this Regulation and, where applicable, information on which requirements are not fully met with a reasoned justification therefor,

Proposal of Regulation (EU) 2017/745 Articles

(f) the health institution draws up documentation that ~~makes it possible to have an understanding of the manufacturing facility, the manufacturing process, the design and performance data of the devices, including the intended purpose, and that~~ is sufficiently detailed to enable the competent authority to ascertain that the **relevant** general safety and performance requirements set out in Annex I ~~to this Regulation~~ are met;

~~(g) the health institution takes all necessary measures to ensure that all devices are manufactured in accordance with the documentation referred to in point (f), and~~

(h) the health institution reviews experience gained from clinical use of the devices and takes all necessary corrective actions.

Member States may require that such health institutions submit to the competent authority any further relevant information about such devices which have been manufactured and used on their territory. Member States shall retain the right to restrict the manufacture **or [and]** the use of any specific type of such devices and shall be permitted access to inspect the activities of the health institutions.

~~For the purposes of the first subparagraph, point (a), in the case of a transfer of the device to another health institution, the transferring and receiving health institutions shall ensure traceability of the device.~~

~~For the purposes of the first subparagraph, point (c), from the date that the health institution becomes aware that the target patient group's specific needs can be met by a device available on the market, it may continue to manufacture and use its device for a maximum period of 10 years.~~

~~For the purposes of the first subparagraph, point (h), where the device is transferred in accordance with point (a), the receiving health institution shall report any incident related to the device to the transferring health institution.~~

This paragraph shall not apply to devices that are manufactured on an industrial scale.

6. In order to ensure the uniform application of Annex I, the Commission may adopt implementing acts to the extent necessary to resolve issues of divergent interpretation and of practical application. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

~~7. The Commission is empowered to adopt delegated acts in accordance with Article 115 to amend the general safety and performance requirements set out in Annex I in order to adapt them to scientific or technical progress or to international developments or to add requirements in relation to emerging risks or technologies.~~

~~8. When adopting implementing acts pursuant to paragraph 6 of this Article, delegated acts pursuant to paragraph 7 of this Article or Common Specifications pursuant to Article 9 of this Regulation concerning devices that are high-risk AI systems as referred to in Article 6(1) of Regulation (EU) 2024/1689 of the European Parliament and of the Council***, or that use high-risk AI systems as safety components, the Commission shall take into account the requirements set out in Chapter III, Section 2, of that Regulation.~~

~~*** Regulation (EU) 2024/1689 of the European Parliament and of the Council of 13 June 2024 laying down harmonised rules on artificial intelligence and amending Regulations (EC) No 300/2008, (EU) No 167/2013, (EU) No 168/2013, (EU) 2018/858, (EU) 2018/1139 and (EU) 2019/2144 and Directives 2014/90/EU, (EU) 2016/797 and (EU) 2020/1828 (Artificial Intelligence Act) (OJ L, 2024/1689, 12.7.2024, ELI: <http://data.europa.eu/eli/reg/2024/1689/oj>).;~~

Article 6 *Distance sales*

1. A device offered by means of information society services, as defined in point (b) of Article 1(1) of Directive (EU) 2015/1535, to a natural or legal person established in the Union shall comply with this Regulation.

2. Without prejudice to national law regarding the exercise of the medical profession, a device that is not placed on the market but used in the context of a commercial activity, whether in return for payment or

Proposal of Regulation (EU) 2017/745 Articles

free of charge, for the provision of a diagnostic or therapeutic service offered by means of information society services as defined in point (b) of Article 1(1) of Directive (EU) 2015/1535 or by other means of communication, directly or through intermediaries, to a natural or legal person established in the Union shall comply with this Regulation.

(a) Any natural or legal person offering a device in accordance with paragraph 1 or a service in accordance with paragraph 2 shall provide in the offer at least the information referred to in Annex I, Section 23.2, points (a) to (d) and (m), and access to the instructions for use.

3. Upon request by a competent authority, any natural or legal person offering a device in accordance with paragraph 1 or ~~providing~~ a service in accordance with paragraph 2 shall make available a copy of the EU declaration of conformity drawn up in accordance with Article 19 or the statement drawn up in accordance with Article 21(2) for [of] the device concerned and cooperate with the competent authorities of the Member State where the device or the service is offered.

4. A Member State may, on grounds of protection of public health, require a provider of a service [information society services,] as defined in Article 1(1), point (b), of Directive (EU) 2015/1535, or a provider of a service in accordance with paragraph 2, to cease its activity.

Article 7

Claims

In the labelling, instructions for use, making available, putting into service and advertising of devices, it shall be prohibited to use text, names, trade marks, pictures and figurative or other signs that may mislead the user or the patient with regard to the device's intended purpose, safety and performance by:

- (a) ascribing functions and properties to the device which the device does not have;
- (b) creating a false impression regarding treatment or diagnosis, functions or properties which the device does not have;
- (c) failing to inform the user or the patient of a likely risk associated with the use of the device in line with its intended purpose;
- (d) suggesting uses for the device other than those stated to form part of the intended purpose for which the conformity assessment was carried out.

Without prejudice to national law regarding the exercise of the medical profession, the first subparagraph shall also apply to devices used for the provision of a service referred to in Article 6(2).

Article 8

Use of harmonised standards

1. Devices which are in conformity with the relevant harmonised standards, or the relevant parts of those standards, the references of which have been published in the *Official Journal of the European Union*, shall be presumed to be in conformity with the requirements of this Regulation covered by those standards or parts thereof.

The first subparagraph shall also apply to system or process requirements to be fulfilled in accordance with this Regulation by economic operators or sponsors, including those relating to quality management systems, risk management, post-market surveillance systems, clinical investigations, clinical evaluation or post-market clinical follow-up ('PMCF').

References in this Regulation to harmonised standards shall be understood as meaning harmonised standards the references of which have been published in the *Official Journal of the European Union*.

2. References in this Regulation to harmonised standards shall also include the monographs of the European Pharmacopoeia adopted in accordance with the Convention on the Elaboration of a European Pharmacopoeia, in particular on surgical sutures and on interaction between medicinal products and

Proposal of Regulation (EU) 2017/745 Articles

materials used in devices containing such medicinal products, provided that references to those monographs have been published in the *Official Journal of the European Union*.

Article 9

Common specifications

1. ~~Without prejudice to Article 1(2) and 17(5) and the deadline laid down in those provisions,~~ Where no harmonised standards exist or where relevant harmonised standards are not sufficient, or where there is a need to address public health concerns, the Commission, after having consulted the MDCG, may, by means of implementing acts, adopt common specifications (CS) in respect of the **requirements set out in this Regulation, in particular the reports and plans to be drawn up by manufacturers, the general safety and performance requirements set out in Annex I, the technical documentation set out in Annexes II and III, the conformity assessment procedures set out in Annexes IX to XI and the procedure for custom-made devices set out in Annex XIII, the clinical evaluation and post-market clinical follow-up set out in Annex XIV or the requirements regarding clinical investigation set out in Annex XV. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).**
2. Devices which are in conformity with the CS referred to in paragraph 1 shall be presumed to be in conformity with the requirements of this Regulation covered by those CS or the relevant parts of those CS.
3. Manufacturers shall comply with the CS referred to in paragraph 1 unless they can duly justify that they have adopted solutions that ensure a level of safety and performance that is at least equivalent thereto.
4. Notwithstanding paragraph 3, manufacturers of products listed in Annex XVI shall comply with the relevant CS for those products.

Article 10

General obligations of manufacturers

1. When placing their devices on the market or putting them into service, manufacturers shall ensure that they have been designed and manufactured in accordance with the requirements of this Regulation.
 2. Manufacturers shall establish, document, implement and maintain a system for risk management as described in Section 3 of Annex I.
 3. ~~Manufacturers shall conduct a clinical evaluation in accordance with the requirements set out in Article 61 and Annex XIV, including a PMCF.~~
 4. Manufacturers of devices other than custom-made devices shall draw up and keep up to date technical documentation for those devices. The technical documentation shall be such as to allow the conformity of the device with the requirements of this Regulation to be assessed. The technical documentation shall include the elements set out in Annexes II and III.
- The Commission is empowered to adopt delegated acts in accordance with Article 115 amending, in the light of technical progress, the Annexes II and III.
5. Manufacturers of custom-made devices shall draw up, keep up to date and keep available for competent authorities documentation in accordance with Section 2 of Annex XIII.
 6. Where compliance with the applicable requirements has been demonstrated following the applicable conformity assessment procedure, manufacturers of devices, other than custom-made or investigational devices, shall draw up an EU declaration of conformity in accordance with Article 19, and affix the CE marking of conformity in accordance with Article 20.
 7. ~~Manufacturers shall comply with the obligations relating to the UDI system referred to in Article 27 and with the registration obligations referred to in Articles 29 and 31.~~
 8. Manufacturers shall keep the technical documentation, the EU declaration of conformity and, if applicable, a copy of any relevant certificate, including any amendments and supplements, issued in

Proposal of Regulation (EU) 2017/745 Articles

accordance with Article 56, available for the competent authorities for a period of at least ten years after the last device covered by the EU declaration of conformity has been placed on the market. In the case of implantable devices, the period shall be at least 15 years after the last device has been placed on the market.

Upon request by a competent authority, the manufacturer shall, as indicated therein, provide that technical documentation in its entirety or a summary thereof.

A manufacturer with a registered place of business outside the Union shall, in order to allow its authorised representative to fulfil the tasks mentioned in Article 11(3), ensure that the authorised representative has the necessary documentation permanently available.

9. Manufacturers shall **put in place an appropriate quality management system** to ensure that procedures are in place to keep series production in conformity with the requirements of this Regulation. Changes in device design or characteristics and changes in the harmonised standards or CS by reference to which the conformity of a device is declared shall be adequately taken into account in a timely manner. Manufacturers of devices, other than investigational devices, shall establish, document, implement, maintain, keep up to date and continually improve a quality management system that shall ensure compliance with this Regulation in the most effective manner and in a manner that is proportionate to the risk class and the type of device.

The quality management system shall cover all parts and elements of a manufacturer's organisation dealing with the quality of processes, procedures and devices. It shall govern the structure, responsibilities, procedures, processes and management resources required to implement the principles and actions necessary to achieve compliance with the provisions of this Regulation.

~~The quality management system shall address at least the following aspects:~~

- ~~(a) a strategy for regulatory compliance, including compliance with conformity assessment procedures and procedures for management of modifications to the devices covered by the system;~~
- ~~(b) identification of applicable general safety and performance requirements and exploration of options to address those requirements;~~
- ~~(c) responsibility of the management;~~
- ~~(d) resource management, including selection and control of suppliers and sub-contractors;~~
- ~~(e) risk management as set out in in Section 3 of Annex I;~~
- ~~(f) clinical evaluation in accordance with Article 61 and Annex XIV, including PMCF;~~
- ~~(g) product realisation, including planning, design, development, production and service provision;~~
- ~~(h) verification of the UDI assignments made in accordance with Article 27(3) to all relevant devices and ensuring consistency and validity of information provided in accordance with Article 29;~~
- ~~(i) setting-up, implementation and maintenance of a post-market surveillance system, in accordance with Article 83;~~
- ~~(j) handling communication with competent authorities, notified bodies, other economic operators, customers and/or other stakeholders;~~
- ~~(k) processes for reporting of serious incidents and field safety corrective actions in the context of vigilance;~~
- ~~(l) management of corrective and preventive actions and verification of their effectiveness;~~
- ~~(m) processes for monitoring and measurement of output, data analysis and product improvement.~~

~~10. Manufacturers of devices shall implement and keep up to date the post-market surveillance system in accordance with Article 83.~~

Proposal of Regulation (EU) 2017/745 Articles

11. Manufacturers shall ensure that the device is accompanied by the information set out in Section 23 of Annex I in an official Union language(s) determined by the Member State in which the device is made available to the user or patient. The particulars on the label shall be indelible, easily legible, clearly comprehensible to the intended user or patient. ~~When determining the official language of the Union in which the information set out in Section 23 of Annex I or other information to be provided by the manufacturer shall be made available, Member States shall consider accepting another official language of the Union in which the information is made available, taking into consideration the technical knowledge, experience, education or training of the average intended user(s).~~

12. Manufacturers who consider or have reason to believe that a device which they have placed on the market or put into service is not in conformity with this Regulation shall immediately take the necessary corrective action to bring that device into conformity, to withdraw it or to recall it, as appropriate. They shall inform the distributors of the device in question and, where applicable, the authorised representative and importers accordingly.

Where the device presents a serious risk, manufacturers shall immediately inform the competent authorities of the Member States in which they made the device available and, where applicable, the notified body that issued a certificate for the device in accordance with Article 56, in particular, of the non-compliance and of any corrective action taken.

~~13. Manufacturers shall have a system for recording and reporting of incidents and field safety corrective actions as described in Articles 87 and 88.~~

14. Manufacturers shall, upon request by a competent authority, provide it with all the information and documentation necessary to demonstrate the conformity of the device, in an official Union language determined by the Member State concerned. The competent authority of the Member State in which the manufacturer has its registered place of business may require that the manufacturer provide samples of the device free of charge or, where that is impracticable, grant access to the device. Manufacturers shall cooperate with a competent authority, at its request, on any corrective action taken to eliminate or, if that is not possible, mitigate the risks posed by devices which they have placed on the market or put into service.

If the manufacturer fails to cooperate or the information and documentation provided is incomplete or incorrect, the competent authority may, in order to ensure the protection of public health and patient safety, take all appropriate measures to prohibit or restrict the device's being made available on its national market, to withdraw the device from that market or to recall it until the manufacturer cooperates or provides complete and correct information.

~~If a competent authority considers or has reason to believe that a device has caused damage, it shall, upon request, facilitate the provision of the information and documentation referred to in the first subparagraph to the potentially injured patient or user and, as appropriate, the patient's or user's successor in title, the patient's or user's health insurance company or other third parties affected by the damage caused to the patient or user, without prejudice to data protection rules and, unless there is an overriding public interest in disclosure, without prejudice to the protection of intellectual property rights.~~

~~The competent authority need not comply with the obligation laid down in the third subparagraph where disclosure of the information and documentation referred to in the first subparagraph is ordinarily dealt with in the context of legal proceedings.~~

15. Where manufacturers have their devices designed or manufactured by another legal or natural person the information on the identity of that person shall be part of the information to be submitted in accordance with Article 29(4). ~~In those cases, the manufacturer shall ensure that the relevant parts of the technical documentation are drawn up, kept up to date and, upon request, made available to the competent authorities in accordance with paragraphs 4 and 8 of this Article by the legal or natural person that has designed and manufactured the device. In addition, the manufacturer shall draw up, keep up to date and, upon request, make available to the competent authorities the remaining parts of the technical documentation, in particular those referred to in Section 2 of Annex II and in Annex III.~~

Proposal of Regulation (EU) 2017/745 Articles

~~16. Natural or legal persons may claim compensation for damage caused by a defective device in accordance with applicable Union and national law.~~

~~Manufacturers shall, in a manner that is proportionate to the risk class, type of device and the size of the enterprise, have measures in place to provide sufficient financial coverage in respect of their potential liability under Directive 85/374/EEC, without prejudice to more protective measures under national law.~~

Article 10a

Obligations in case of interruption or discontinuation of supply of certain devices

1. Where a manufacturer anticipates an interruption or a discontinuation of the supply of a device, other than a custom-made device, and where it is reasonably foreseeable that such interruption or discontinuation could result in serious harm or a risk of serious harm to patients or public health in one or more Member States, the manufacturer shall inform the competent authority of the Member State where it or its authorised representative is established, as well as the economic operators, health institutions and healthcare professionals to whom it directly supplies the device, of the anticipated interruption or discontinuation.

The information referred to in the first subparagraph shall, ~~other than in exceptional circumstances,~~ be provided at least six months before the anticipated interruption or discontinuation ~~or, if this is not possible, without undue delay after the manufacturer becomes aware of the anticipated interruption or discontinuation.~~ The manufacturer shall specify the reasons for the interruption or discontinuation in the information provided to the competent authority.

2. The competent authority that has received the information referred to in paragraph 1 shall, without undue delay, inform the competent authorities of the other Member States and the Commission of the anticipated interruption or discontinuation.

3. The economic operators who have received the information from the manufacturer in accordance with paragraph 1 or from another economic operator in the supply chain shall, without undue delay, inform any other economic operators, health institutions and healthcare professionals to whom they directly supply the device, of the anticipated interruption or discontinuation.

4. The Commission, where needed in cooperation with the EMA, shall set up, maintain, and manage an IT system to facilitate the reporting and information exchange regarding cases of interruption or discontinuation of the supply of devices in accordance with paragraphs 1, 2 and 3. That IT system shall be integrated in or interoperable with the European database on medical devices referred to in Article 33. It shall also enable health institutions and healthcare professionals to inform competent authorities about the unavailability or the immediate risk of unavailability of devices needed for the exercise of their professional activity.

5. The EMA, in collaboration with the Executive Steering Group on Shortages of Medical Devices (MDSSG) established by Article 21 of Regulation (EU) 2022/123, shall develop a methodology to identify the devices, or categories of devices, for which it is reasonably foreseeable that an interruption or discontinuation of supply could result in serious harm or a risk of serious harm to patients or public health as referred to in paragraph 1. Based on that methodology, the EMA, in collaboration with the MDSSG and in agreement with the Commission, shall draw up, publish and keep up to date a list of devices, or categories of devices, to which paragraphs 1, 2 and 3 shall apply. For the purpose of this paragraph, the MDCG, representatives of manufacturers, other relevant actors in the supply chain for the medical device sector and representatives of healthcare professionals, of patients and of consumers may be consulted as necessary.

6. The competent authorities of the Member States or the Commission may request the manufacturers of devices included in the list drawn up in EN 43 EN accordance with paragraph 5 to provide all necessary information regarding risks and weaknesses within the supply chain which may affect the supply of such devices, including production capacity and volume of sales.

Article 11

Authorised representative

1. Where the manufacturer of a device is not established in a Member State, the device may only be placed on the Union market if the manufacturer designates a sole authorised representative.

2. The designation shall constitute the authorised representative's mandate, it shall be valid only when accepted in writing by the authorised representative and shall be effective at least for all devices of the same generic device group.

3. The authorised representative shall perform the tasks specified in the mandate agreed between it and the manufacturer. The authorised representative shall provide a copy of the mandate to the competent authority, upon request.

The mandate shall require, and the manufacturer shall enable, the authorised representative to perform at least the following tasks in relation to the devices that it covers:

(a) verify that the EU declaration of conformity and technical documentation have been drawn up and, where applicable, that an appropriate conformity assessment procedure has been carried out by the manufacturer;

(b) keep available a copy of the technical documentation, the EU declaration of conformity and, if applicable, a copy of the relevant certificate, including any amendments and supplements, issued in accordance with Article 56, at the disposal of competent authorities for the period referred to in Article 10(8);

(c) comply with the registration obligations laid down in Article 31 and verify that the manufacturer has complied with the registration obligations laid down in Articles 27 and 29;

(d) in response to a request from a competent authority, provide that competent authority with all the information and documentation necessary to demonstrate the conformity of a device, in an official Union language determined by the Member State concerned;

(e) forward to the manufacturer any request by a competent authority of the Member State in which the authorised representative has its registered place of business for samples, or access to a device and verify that the competent authority receives the samples or is given access to the device;

(f) cooperate with the competent authorities on any preventive or corrective action taken to eliminate or, if that is not possible, mitigate the risks posed by devices;

(g) immediately inform the manufacturer about complaints and reports from healthcare professionals, patients and users about suspected incidents related to a device for which they have been designated;

(h) terminate the mandate if the manufacturer acts contrary to its obligations under this Regulation.

~~4. The mandate referred to in paragraph 3 of this Article shall not delegate the manufacturer's obligations laid down in Article 10(1), (2), (3), (4), (6), (7), (9), (10), (11) and (12).~~

~~5. Without prejudice to paragraph 4 of this Article, where the manufacturer is not established in a Member State and has not complied with the obligations laid down in Article 10, the authorised representative shall be legally liable for defective devices on the same basis as, and jointly and severally with, the manufacturer.~~

6. An authorised representative who terminates its mandate on the ground referred to in point (h) of paragraph 3 shall immediately inform the competent authority of the Member State in which it is established and, where applicable, the notified body that was involved in the conformity assessment for the device of the termination of the mandate and the reasons therefor.

7. Any reference in this Regulation to the competent authority of the Member State in which the manufacturer has its registered place of business shall be understood as a reference to the competent

Proposal of Regulation (EU) 2017/745 Articles

authority of the Member State in which the authorised representative, designated by a manufacturer referred to in paragraph 1, has its registered place of business.

Article 12

Change of authorised representative

The detailed arrangements for a change of authorised representative shall be clearly defined in an agreement between the manufacturer, where practicable the outgoing authorised representative, and the incoming authorised representative. That agreement shall address at least the following aspects:

- (a) the date of termination of the mandate of the outgoing authorised representative and date of beginning of the mandate of the incoming authorised representative;
- (b) the date until which the outgoing authorised representative may be indicated in the information supplied by the manufacturer, including any promotional material;
- (c) the transfer of documents, including confidentiality aspects and property rights;
- (d) the obligation of the outgoing authorised representative after the end of the mandate to forward to the manufacturer or incoming authorised representative any complaints or reports from healthcare professionals, patients or users about suspected incidents related to a device for which it had been designated as authorised representative.

Article 13

General obligations of importers

1. Importers shall place on the Union market only devices that are in conformity with this Regulation.
 2. In order to place a device on the market, importers shall verify that:
 - (a) the device has been CE marked and that the EU declaration of conformity of the device has been drawn up;
 - (b) a manufacturer is identified and that an authorised representative in accordance with Article 11 has been designated by the manufacturer;
 - (c) the device is labelled in accordance with this Regulation and accompanied by the required instructions for use;
 - (d) where applicable, a UDI has been assigned by the manufacturer in accordance with Article 27.
- Where an importer considers or has reason to believe that a device is not in conformity with the requirements of this Regulation, it shall not place the device on the market until it has been brought into conformity and shall inform the manufacturer and the manufacturer's authorised representative. Where the importer considers or has reason to believe that the device presents a serious risk or is a falsified device, it shall also inform the competent authority of the Member State in which the importer is established.
3. Importers shall indicate on the device or on its packaging or in a document accompanying the device their name, registered trade name or registered trade mark, their registered place of business and the address at which they can be contacted, so that their location can be established. They shall ensure that any additional label does not obscure any information on the label provided by the manufacturer.
 4. Importers shall verify that the device is registered in the electronic system in accordance with Article 29. Importers shall add their details to the registration in accordance with Article 31.
 5. Importers shall ensure that, while a device is under their responsibility, storage or transport conditions do not jeopardise its compliance with the general safety and performance requirements set out in Annex I and shall comply with the conditions set by the manufacturer, where available.

Proposal of Regulation (EU) 2017/745 Articles

6. Importers shall keep a register of complaints, of non-conforming devices and of recalls and withdrawals, and provide the manufacturer, authorised representative and distributors with any information requested by them, in order to allow them to investigate complaints.

7. Importers who consider or have reason to believe that a device which they have placed on the market is not in conformity with this Regulation shall immediately inform the manufacturer and its authorised representative. Importers shall co-operate with the manufacturer, the manufacturer's authorised representative and the competent authorities to ensure that the necessary corrective action to bring that device into conformity, to withdraw or recall it is taken. Where the device presents a serious risk, they shall also immediately inform the competent authorities of the Member States in which they made the device available and, if applicable, the notified body that issued a certificate in accordance with Article 56 for the device in question, giving details, in particular, of the non-compliance and of any corrective action taken.

8. Importers who have received complaints or reports from healthcare professionals, patients or users about suspected incidents related to a device which they have placed on the market shall immediately forward this information to the manufacturer and its authorised representative.

9. Importers shall, for the period referred to in Article 10(8), keep a copy of the EU declaration of conformity and, if applicable, a copy of any relevant certificate, including any amendments and supplements, issued in accordance with Article 56.

10. Importers shall cooperate with competent authorities, at the latter's request, on any action taken to eliminate or, if that is not possible, mitigate the risks posed by devices which they have placed on the market. Importers, upon request by a competent authority of the Member State in which the importer has its registered place of business, shall provide samples of the device free of charge or, where that is impracticable, grant access to the device.

Article 14

General obligations of distributors

1. When making a device available on the market, distributors shall, in the context of their activities, act with due care in relation to the requirements applicable.

2. Before making a device available on the market, distributors shall verify that all of the following requirements are met:

(a) the device has been CE marked and that the EU declaration of conformity of the device has been drawn up;

(b) the device is accompanied by the information to be supplied by the manufacturer in accordance with Article 10(11);

(c) for imported devices, the importer has complied with the requirements set out in Article 13(3);

(d) that, where applicable, a UDI has been assigned by the manufacturer **in accordance with Article 27(3)**.

In order to meet the requirements referred to in ~~points (a), (b) and (d) of~~ the first subparagraph the distributor may apply a sampling method that is representative of the devices supplied by that distributor.

Where a distributor considers or has reason to believe that a device is not in conformity with the requirements of this Regulation, it shall not make the device available on the market until it has been brought into conformity, and shall inform the manufacturer and, where applicable, the manufacturer's authorised representative, and the importer. Where the distributor considers or has reason to believe that the device presents a serious risk or is a falsified device, it shall also inform the competent authority of the Member State in which it is established.

3. Distributors shall ensure that, while the device is under their responsibility, storage or transport conditions comply with the conditions set by the manufacturer.

4. Distributors that consider or have reason to believe that a device which they have made available on the market is not in conformity with this Regulation shall immediately inform the manufacturer and, where applicable, the manufacturer's authorised representative and the importer. Distributors shall co-operate with the manufacturer and, where applicable, the manufacturer's authorised representative, and the importer, and with competent authorities to ensure that the necessary corrective action to bring that device into conformity, to withdraw or to recall it, as appropriate, is taken. Where the distributor considers or has reason to believe that the device presents a serious risk, it shall also immediately inform the competent authorities of the Member States in which it made the device available, giving details, in particular, of the non-compliance and of any corrective action taken.

5. Distributors that have received complaints or reports from healthcare professionals, patients or users about suspected incidents related to a device they have made available, shall immediately forward this information to the manufacturer and, where applicable, the manufacturer's authorised representative, and the importer. They shall keep a register of complaints, of non-conforming devices and of recalls and withdrawals, and keep the manufacturer and, where available, the authorised representative and the importer informed of such monitoring and provide them with any information upon their request.

6. Distributors shall, upon request by a competent authority, provide it with all the information and documentation that is at their disposal and is necessary to demonstrate the conformity of a device.

Distributors shall be considered to have fulfilled the obligation referred to in the first subparagraph when the manufacturer or, where applicable, the authorised representative for the device in question provides the required information. Distributors shall cooperate with competent authorities, at their request, on any action taken to eliminate the risks posed by devices which they have made available on the market.

Distributors, upon request by a competent authority, shall provide free samples of the device or, where that is impracticable, grant access to the device.

Article 15

Person responsible for regulatory compliance

1. Manufacturers shall have available within their organisation at least one person responsible for regulatory compliance who possesses the requisite expertise in the field of medical devices. ~~The requisite expertise shall be demonstrated by either of the following qualifications:~~

~~(a) a diploma, certificate or other evidence of formal qualification, awarded on completion of a university degree or of a course of study recognised as equivalent by the Member State concerned, in law, medicine, pharmacy, engineering or another relevant scientific discipline, and at least one year of professional experience in regulatory affairs or in quality management systems relating to medical devices;~~

~~(b) four years of professional experience in regulatory affairs or in quality management systems relating to medical devices.~~

Without prejudice to national provisions regarding professional qualifications, manufacturers of custom-made devices may demonstrate the requisite expertise referred to in the first subparagraph by having at least two years of professional experience within a relevant field of manufacturing.

2. Micro and small enterprises within the meaning of Commission Recommendation 2003/361/EC**** [(1)] shall not be required to have the person responsible for regulatory compliance within their organisation but shall have such person ~~permanently and continuously~~ at their disposal.

****Commission Recommendation of 6 May 2003 concerning the definition of micro, small and medium-sized enterprises (OJ L 124, 20.5.2003, p. 36, ELI: <http://data.europa.eu/eli/reco/2003/361/oj>).

~~(1) Commission Recommendation 2003/361/EC of 6 May 2003 concerning the definition of micro, small and medium-sized enterprises (OJ L 124, 20.5.2003, p. 36).~~

3. The person responsible for regulatory compliance shall at least be responsible for ensuring that:

Proposal of Regulation (EU) 2017/745 Articles

- (a) the conformity of the devices is appropriately checked, in accordance with the quality management system under which the devices are manufactured, before a device is released;
 - (b) the technical documentation and the EU declaration of conformity are drawn up and kept up-to-date;
 - (c) the post-market surveillance obligations are complied with in accordance with Article 83 [10(10)];
 - (d) the reporting obligations referred to in Articles 87 to 91 are fulfilled;
 - (e) in the case of investigational devices, the statement referred to in Section 4.1 of Chapter II of Annex XV is issued.
4. If a number of persons are jointly responsible for regulatory compliance in accordance with paragraphs 1, 2 and 3, their respective areas of responsibility shall be stipulated in writing.
5. The person responsible for regulatory compliance shall suffer no disadvantage within the manufacturer's organisation in relation to the proper fulfilment of his or her duties, regardless of whether or not they are employees of the organisation.
6. Authorised representatives shall have permanently and continuously at their disposal at least one person responsible for regulatory compliance who possesses the requisite expertise regarding the regulatory requirements for medical devices in the Union. ~~The requisite expertise shall be demonstrated by either of the following qualifications:~~
- ~~(a) a diploma, certificate or other evidence of formal qualification, awarded on completion of a university degree or of a course of study recognised as equivalent by the Member State concerned, in law, medicine, pharmacy, engineering or another relevant scientific discipline, and at least one year of professional experience in regulatory affairs or in quality management systems relating to medical devices;~~
 - ~~(b) four years of professional experience in regulatory affairs or in quality management systems relating to medical devices.~~

Article 16

Cases in which obligations of manufacturers apply to importers, distributors or other persons

1. A distributor, ~~an~~ importer or ~~another~~ [other] natural or legal person ~~who places a product on the market or puts it into service~~ shall assume the obligations incumbent on manufacturers if it does any of the following:
- (a) makes available on the market a device under its name, registered trade name or registered trade mark, except in cases where a distributor or importer enters into an agreement with a manufacturer whereby the manufacturer is identified as such on the label and is responsible for meeting the requirements placed on manufacturers in this Regulation;
 - (b) changes the intended purpose of a device already placed on the market or put into service;
 - (c) modifies a device already placed on the market or put into service in such a way that compliance with the applicable requirements may be affected.
- The first subparagraph shall not apply to any ~~healthcare professional or any other~~ person who, while not considered a manufacturer ~~as defined in point (30) of Article 2~~, assembles or adapts for an individual patient a device already on the market without changing its intended purpose.
2. For the purposes of point (c) of paragraph 1, the following shall not be considered to be a modification of a device that could affect its compliance with the applicable requirements:
- (a) provision, including translation, of the information supplied by the manufacturer, in accordance with Section 23 of Annex I, relating to a device already placed on the market and of further information which is necessary in order to market the device in the relevant Member State;

Proposal of Regulation (EU) 2017/745 Articles

(b) changes to the outer packaging of a device already placed on the market, including a change of pack size, if the repackaging is necessary in order to market the device in the relevant Member State and if it is carried out in such conditions that the original condition of the device cannot be affected by it. In the case of devices placed on the market in sterile condition, it shall be presumed that the original condition of the device is adversely affected if the packaging that is necessary for maintaining the sterile condition is opened, damaged or otherwise negatively affected by the repackaging.

3. A distributor or importer that carries out any of the activities mentioned in points (a) and (b) of paragraph 2 shall indicate on the device or, where that is impracticable, on its packaging or in a document accompanying the device, the activity carried out together with its name, registered trade name or registered trade mark, registered place of business and the address at which it can be contacted, so that its location can be established.

Distributors and importers shall ensure that they have in place a quality management system that includes procedures which ensure that the translation of information is accurate and up-to-date, and that the activities mentioned in points (a) and (b) of paragraph 2 are performed by a means and under conditions that preserve the original condition of the device and that the packaging of the repackaged device is not defective, of poor quality or untidy. The quality management system shall cover, *inter alia*, procedures ensuring that the distributor or importer is informed of any corrective action taken by the manufacturer in relation to the device in question in order to respond to safety issues or to bring it into conformity with this Regulation.

~~4. At least 28 days prior to making the relabelled or repackaged device available on the market, distributors or importers carrying out any of the activities mentioned in points (a) and (b) of paragraph 2 shall inform the manufacturer and the competent authority of the Member State in which they plan to make the device available of the intention to make the relabelled or repackaged device available and, upon request, shall provide the manufacturer and the competent authority with a sample or mock-up of the relabelled or repackaged device, including any translated label and instructions for use. Within the same period of 28 days, the distributor or importer shall submit to the competent authority a certificate, issued by a notified body designated for the type of devices that are subject to activities mentioned in points (a) and (b) of paragraph 2, attesting that the quality management system of the distributor or importer complies with the requirements laid down in paragraph 3.~~

Article 17

Single-use devices and ~~their~~ reprocessing of devices that are not for single use

1. A device shall only be intended for single-use where the manufacturer, in light of the design, construction, material, chemical, physical and biological properties of the device, cannot ensure that the device continues to meet the relevant safety and performance requirements when reused in accordance with its intended purpose after appropriate reprocessing. The manufacturer's justification of an indication of single use shall be part of the technical documentation referred to in Annex II.

2. [1] If the device is not intended for single-use, the manufacturer shall provide information about the appropriate reprocessing process for allowing reuse in the instructions for use in accordance with Annex I, Section 23.4, point (n). [Reprocessing and further use of single-use devices may only take place where permitted by national law and only in accordance with this Article.]

3. [2] Single-use devices and devices that cannot be further reprocessed may be subject to full refurbishing within the meaning of Article 2(31). The natural or legal person that carries out the full refurbishing shall be considered as the manufacturer of the fully refurbished device. [Any natural or legal person who reprocesses a single-use device to make it suitable for further use within the Union shall be considered to be the manufacturer of the reprocessed device and shall assume the obligations incumbent on manufacturers laid down in this Regulation, which include obligations relating to the traceability of the reprocessed device in accordance with Chapter III of this Regulation. The re-processor of the device shall be considered to be a producer for the purpose of Article 3(1) of Directive 85/374/EEC.]

Proposal of Regulation (EU) 2017/745 Articles

~~3. By way of derogation from paragraph 2, as regards single-use devices that are reprocessed and used within a health institution, Member States may decide not to apply all of the rules relating to manufacturers' obligations laid down in this Regulation provided that they ensure that:~~

~~(a) the safety and performance of the reprocessed device is equivalent to that of the original device and the requirements in points (a), (b), (d), (e), (f), (g) and (h) of Article 5(5) are complied with;~~

~~(b) the reprocessing is performed in accordance with CS detailing the requirements concerning:~~

~~—risk management, including the analysis of the construction and material, related properties of the device (reverse engineering) and procedures to detect changes in the design of the original device as well as of its planned application after reprocessing,~~

~~—the validation of procedures for the entire process, including cleaning steps,~~

~~—the product release and performance testing,~~

~~—the quality management system,~~

~~—the reporting of incidents involving devices that have been reprocessed, and~~

~~—the traceability of reprocessed devices.~~

~~Member States shall encourage, and may require, health institutions to provide information to patients on the use of reprocessed devices within the health institution and, where appropriate, any other relevant information on the reprocessed devices that patients are treated with.~~

~~Member States shall notify the Commission and the other Member States of the national provisions introduced pursuant to this paragraph and the grounds for introducing them. The Commission shall keep the information publicly available.~~

~~4. Member States may choose to apply the provisions referred to in paragraph 3 also as regards single-use devices that are reprocessed by an external reprocessor at the request of a health institution, provided that the reprocessed device in its entirety is returned to that health institution and the external reprocessor complies with the requirements referred to in points (a) and (b) of paragraph 3.~~

~~4. [5] The Commission may [shall] adopt, in accordance with Article 9(1), the necessary CS on general requirements regarding reprocessing of devices or fully refurbishing of single-use devices referred to in point (b) of paragraph 3 by 26 May 2021. Those CS shall be consistent with the latest scientific evidence and shall address the application of the general requirements on safety and performance laid down in this Regulation. In the event that those CS are not adopted by 26 May 2021, reprocessing shall be performed in accordance with any relevant harmonised standards and national provisions that cover the aspects outlined in point (b) of paragraph 3. Compliance with CS or, in the absence of CS, with any relevant harmonised standards and national provisions, shall be certified by a notified body.~~

~~6. Only single-use devices that have been placed on the market in accordance with this Regulation, or prior to 26 May 2021 in accordance with Directive 93/42/EEC, may be reprocessed.~~

~~7. Only reprocessing of single-use devices that is considered safe according to the latest scientific evidence may be carried out.~~

~~8. The name and address of the legal or natural person referred to in paragraph 2 and the other relevant information referred to in Section 23 of Annex I shall be indicated on the label and, where applicable, in the instructions for use of the reprocessed device.~~

~~The name and address of the manufacturer of the original single-use device shall no longer appear on the label, but shall be mentioned in the instructions for use of the reprocessed device.~~

~~9. A Member State that permits reprocessing of single-use devices may maintain or introduce national provisions that are stricter than those laid down in this Regulation and which restrict or prohibit, within its territory, the following:~~

Proposal of Regulation (EU) 2017/745 Articles

~~(a) the reprocessing of single-use devices and the transfer of single-use devices to another Member State or to a third country with a view to their reprocessing;~~

~~(b) the making available or further use of reprocessed single-use devices.~~

~~Member States shall notify the Commission and the other Member States of those national provisions. The Commission shall make such information publicly available.~~

~~10. The Commission shall by 27 May 2024 draw up a report on the operation of this Article and submit it to the European Parliament and to the Council. On the basis of that report, the Commission shall, if appropriate, make proposals for amendments to this Regulation.~~

Article 18

Implant card and information to be supplied to the patient with an implanted device

1. The manufacturer of an implantable device shall provide together with the device the following:

- (a) information allowing the identification of the device, including the device name, serial number, lot number, the UDI, the device model, as well as the name, address and the website of the manufacturer;
- (b) any warnings, precautions or measures to be taken by the patient or a healthcare professional with regard to reciprocal interference with reasonably foreseeable external influences, medical examinations or environmental conditions;
- (c) any information about the expected lifetime of the device and any necessary follow-up;
- (d) any other information to ensure safe use of the device by the patient, including the information in point (u) of Section 23.4 of Annex I.

The information referred to in the first subparagraph shall be provided, for the purpose of making it available to the ~~particular~~ patient who has been implanted with the device, by any means, **including in electronic or digital format**, that allow rapid access to that information and shall be stated in the language(s) determined by the concerned Member State. The information shall be written in a way that is readily understood by a lay person and shall be updated where appropriate. Updates of the information shall be made available to the patient via the website mentioned in point (a) of the first subparagraph.

In addition, the manufacturer shall provide the information referred to in point (a) of the first subparagraph on an implant card delivered with the device.

2. Member States shall require health institutions to make the information referred to in paragraph 1 available, by any means that allow rapid access to that information, to any patients who have been implanted with the device, together with the implant card, which shall bear their identity.

~~3. The following implants shall be exempted from the obligations laid down in this Article: sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors. The Commission is empowered to adopt delegated acts in accordance with Article 115 to amend this list by adding other types of implants to it or by removing implants therefrom.~~

~~Implants that are well-established technology devices shall be exempted from the obligations laid down in this Article.~~

Article 19

EU declaration of conformity

1. The EU declaration of conformity shall state that the requirements specified in this Regulation have been fulfilled in relation to the device that is covered. The manufacturer shall continuously update the EU declaration of conformity. The EU declaration of conformity shall, as a minimum, contain the information set out in Annex IV and shall be translated into an official Union language or languages required by the Member State(s) in which the device is made available.

Proposal of Regulation (EU) 2017/745 Articles

2. Where, concerning aspects not covered by this Regulation, devices are subject to other Union legislation which also requires an EU declaration of conformity by the manufacturer that fulfilment of the requirements of that legislation has been demonstrated, a single EU declaration of conformity shall be drawn up in respect of all Union acts applicable to the device. The declaration shall contain all the information required for identification of the Union legislation to which the declaration relates.

(a) Declarations of conformity in accordance with paragraphs 1 and 2 may be provided in electronic form.

3. By drawing up the EU declaration of conformity, the manufacturer shall assume responsibility for compliance with the requirements of this Regulation and all other Union legislation applicable to the device.

4. The Commission is empowered to adopt delegated acts in accordance with Article 115 amending the minimum content of the EU declaration of conformity set out in Annex IV in the light of technical progress.

Article 20 *CE marking of conformity*

1. Devices, other than custom-made or investigational devices, considered to be in conformity with the requirements of this Regulation shall bear the CE marking of conformity, as presented in Annex V.

2. The CE marking shall be subject to the general principles set out in Article 30 of Regulation (EC) No 765/2008.

3. The CE marking shall be affixed visibly, legibly and indelibly to the device or its sterile packaging. Where such affixing is not possible or not warranted on account of the nature of the device, the CE marking shall be affixed to the packaging. The CE marking shall also appear in any instructions for use and on any sales packaging.

4. The CE marking shall be affixed before the device is placed on the market. It may be followed by a pictogram or any other mark indicating a special risk or use.

5. Where applicable, the CE marking shall be followed by the identification number of the notified body responsible for the conformity assessment procedures set out in Article 52. The identification number shall also be indicated in any promotional material which mentions that a device fulfils the requirements for CE marking.

6. Where devices are subject to other Union legislation which also provides for the affixing of the CE marking, the CE marking shall indicate that the devices also fulfil the requirements of that other legislation.

Article 21 *Devices for special purposes*

1. Member States shall not create obstacles to:

(a) investigational devices being supplied to an investigator for the purpose of a clinical investigation if they meet the conditions laid down in Articles 62 to 80 and Article 82, in the implementing acts adopted pursuant to Article 81 and in Annex XV;

(b) custom-made devices being made available on the market if Article 52(8) and Annex XIII have been complied with.

The devices referred to in the first subparagraph shall not bear the CE marking, with the exception of the devices referred to in Article 74.

2. Custom-made devices shall be accompanied by the statement referred to in Section 1 of Annex XIII, which shall be made available to the particular patient or user identified by name, an acronym or a numerical code.

Proposal of Regulation (EU) 2017/745 Articles

Member States may require that the manufacturer of a custom-made device submit to the competent authority a list of such devices which have been made available in their territory.

3. At trade fairs, exhibitions, demonstrations or similar events, Member States shall not create obstacles to the showing of devices which do not comply with this Regulation, provided a visible sign clearly indicates that such devices are intended for presentation or demonstration purposes only and cannot be made available until they have been brought into compliance with this Regulation.

Article 22

Systems and procedure packs

1. Natural or legal persons shall draw up a statement if they combine devices bearing a CE marking with the following other devices or products, in a manner that is compatible with the intended purpose of the devices or other products and within the limits of use specified by their manufacturers, in order to place them on the market as a system or procedure pack:

(a) other devices bearing the CE marking;

(b) *in vitro* diagnostic medical devices bearing the CE marking in conformity with Regulation (EU) 2017/746;

(c) other products which are in conformity with legislation that applies to those products only where they are used within a medical procedure or their presence in the system or procedure pack is otherwise justified.

2. ~~the~~ the statement made pursuant to paragraph 1, shall contain at least the following information:

(a) an identification of the devices and, where applicable, other products included in the system or procedure pack, including where applicable their Basic UDI-DI;

(b) where applicable, an identification of the notified body involved in the sterilisation activities referred to in paragraph 3;

2 (c) [2] a declaration by the natural or legal person that: [the natural or legal person concerned shall declare that]

2 (c) (i) [2 (a)] they have verified the mutual compatibility of the devices and, if applicable, other products in accordance with the manufacturers' instructions and have carried out their activities in accordance with those instructions;

2 (c) (ii) [2 (b)] they have packaged the system or procedure pack and supplied relevant information to users incorporating the information to be supplied by the manufacturers of the devices or other products which have been put together;

2 (c) (iii) [2 (c)] the activity of combining devices and, if applicable, other products as a system or procedure pack was subject to appropriate methods of internal monitoring, verification and validation.

3. Any natural or legal person who sterilises systems or procedure packs referred to in paragraph 1 for the purpose of placing them on the market shall, at their choice, apply one of the procedures set out in Annex IX or the procedure set out in Part A of Annex XI. The application of those procedures and the involvement of the notified body shall be limited to the aspects of the procedure relating to ensuring sterility until the sterile packaging is opened or damaged. The natural or legal person shall draw up a statement declaring that sterilisation has been carried out in accordance with the manufacturer's instructions.

4. Where the system or procedure pack incorporate devices which do not bear the CE marking or where the chosen combination of devices is not compatible in view of their original intended purpose, or where the sterilisation has not been carried out in accordance with the manufacturer's instructions the system or procedure pack shall be treated as a device in its own right and shall be subject to the relevant conformity

Proposal of Regulation (EU) 2017/745 Articles

assessment procedure pursuant to Article 52. The natural or legal person shall assume the obligations incumbent on manufacturers.

5. The systems or procedure packs referred to in paragraph 1 of this Article shall not themselves bear an additional CE marking but they shall bear the name, registered trade name or registered trade mark of the person referred to in paragraphs 1 and 3 of this Article as well as the address at which that person can be contacted, so that the person's location can be established. Systems or procedure packs shall be accompanied by the information referred to in Section 23 of Annex I. The statement referred to in paragraph 2 of this Article shall be kept at the disposal of the competent authorities, after the system or procedure pack has been put together, for the period that is applicable under Article 10(8) to the devices that have been combined. Where those periods differ, the longest period shall apply.

Article 23 *Parts and components*

1. Any natural or legal person who makes available on the market an item specifically intended to replace an identical or similar integral part or component of a device that is defective or worn in order to maintain or restore the function of the device without changing its performance or safety characteristics or its intended purpose, shall ensure that the item does not adversely affect the safety and performance of the device. Supporting evidence shall be kept available for the competent authorities of the Member States.

2. An item that is intended specifically to replace a part or component of a device and that significantly changes the performance or safety characteristics or the intended purpose of the device shall be considered to be a device and shall meet the requirements laid down in this Regulation.

Article 24 *Free movement*

Except where otherwise provided for in this Regulation, Member States shall not refuse, prohibit or restrict the making available on the market or putting into service within their territory of devices which comply with the requirements of this Regulation.

Chapter III Identification and traceability of devices, registration of devices and of economic operators, summary of safety and clinical performance, European database on medical devices

Article 25 *Identification within the supply chain*

1. Distributors and importers shall co-operate with manufacturers or authorised representatives to achieve an appropriate level of traceability of devices.

2. Economic operators shall be able to identify the following to the competent authority, for the period referred to in Article 10(8):

- (a) any economic operator to whom they have directly supplied a device;
- (b) any economic operator who has directly supplied them with a device;
- (c) any health institution or healthcare professional to which they have directly supplied a device.

Article 26 *Medical devices nomenclature*

To facilitate the functioning of the European database on medical devices ('Eudamed') as referred to in Article 33, the Commission shall ensure that an internationally recognised medical devices nomenclature is available free of charge to manufacturers and other natural or legal persons required by this Regulation

Proposal of Regulation (EU) 2017/745 Articles

to use that nomenclature. The Commission shall also endeavour to ensure that that nomenclature is available to other stakeholders free of charge, where reasonably practicable.

Article 27

Unique Device Identification system

1. The Unique Device Identification system ('UDI system') described in Part C of Annex VI shall allow the identification and facilitate the traceability of devices, other than custom-made and investigational devices, and shall consist of the following:

(a) production of a UDI that comprises the following:

(i) a UDI device identifier ('UDI-DI') specific to a manufacturer and a device, providing access to the information laid down in Part B of Annex VI;

(ii) a UDI production identifier ('UDI-PI') that identifies the unit of device production and if applicable the packaged devices, as specified in Part C of Annex VI;

(b) placing of the UDI on the label of the device or on its packaging;

(ba) a Basic UDI-DI, as defined in Part C of Annex VI;

(c) storage of the UDI by economic operators, health institutions and healthcare professionals, in accordance with the conditions laid down in paragraphs 8 and 9 of this Article respectively;

(d) establishment of an electronic system for Unique Device Identification ('UDI database') in accordance with Article 28.

2. The Commission shall, by means of implementing acts, designate one or several entities to operate a system for assignment of UDIs pursuant to this Regulation ('issuing entity'). That entity or those entities shall satisfy all of the following criteria:

(a) the entity is an organisation with legal personality;

(b) its system for the assignment of UDIs is adequate to identify a device throughout its distribution and use in accordance with the requirements of this Regulation;

(c) its system for the assignment of UDIs conforms to the relevant international standards;

(d) the entity gives access to its system for the assignment of UDIs to all interested users in accordance with a set of predetermined and transparent terms and conditions **that take into account the interests of micro, small and medium-sized enterprises within the meaning of Recommendation 2003/361/EC;**

(e) the entity undertakes to do the following:

(i) operate its system for the assignment of UDIs for at least ten years after its designation;

(ii) make available to the Commission and to the Member States, upon request, information concerning its system for the assignment of UDIs;

(iii) remain in compliance with the criteria for designation and the terms of designation.

(iv) offer its system for the assignment of UDIs to manufacturers that are micro and small enterprises within the meaning of Recommendation 2003/361/EC under preferential conditions that take into account the specific needs of such enterprises and are proportionate to their size.

When designating issuing entities, the Commission shall endeavour to ensure that UDI carriers, as defined in Part C of Annex VI, are universally readable regardless of the system used by the issuing entity, with a view to minimising financial and administrative burdens for economic operators and health institutions.

3. Before placing a device, other than a custom-made device **or investigational device**, on the market, the manufacturer shall, **in accordance with the rules of the issuing entity designated by the Commission in**

Proposal of Regulation (EU) 2017/745 Articles

accordance with paragraph 2, assign to the device ~~and, if applicable, to all higher levels of packaging~~ a Basic UDI-DI and a UDI as defined in Part C of Annex VI [, a UDI created in compliance with the rules of the issuing entity designated by the Commission in accordance with paragraph 2]. If applicable, the manufacturer shall assign a UDI-DI to all higher levels of packaging.

~~Before a device, other than a custom-made or investigational device, is placed on the market the manufacturer shall ensure that the information referred to in Part B of Annex VI of the device in question are correctly submitted and transferred to the UDI database referred to in Article 28.~~

(a) Before placing on the market a system or procedure pack pursuant to Article 22(1) and (3), the natural or legal person responsible shall assign to the system or procedure pack, in accordance with the rules of the issuing entity designated by the Commission in accordance with paragraph 2, a Basic UDI-DI and UDI as defined in Part C of Annex VI.

(b) For devices that are the subject of a conformity assessment as referred to in Article 52(3) and in Article 52(4), second and third subparagraphs, a Basic UDI-DI referred to in paragraph 1 of this Article shall be assigned before the manufacturer applies to a notified body for that assessment.

4. UDI carriers shall be placed on the label of the device and on all higher levels of packaging. Higher levels of packaging shall not be understood to include shipping containers.

5. The UDI shall be used for reporting serious incidents and field safety corrective actions in accordance with Article 87.

6. The Basic UDI-DI, as defined in Part C of Annex VI, of the device shall appear on the EU declaration of conformity referred to in Article 19.

7. As part of the technical documentation referred to in Annex II, the manufacturer shall keep up-to-date a list of all UDIs that it has assigned.

8. Economic operators shall store and keep, preferably by electronic means, the UDI of the devices which they have supplied or with which they have been supplied, if those devices belong to:

- class III implantable devices;
- the devices, categories or groups of devices determined by a measure referred to in point (a) of paragraph 11.

9. Health institutions shall store and keep preferably by electronic means the UDI of the devices which they have supplied or with which they have been supplied, if those devices belong to class III implantable devices.

For devices other than class III implantable devices, Member States shall encourage, and may require, health institutions to store and keep, preferably by electronic means, the UDI of the devices with which they have been supplied.

Member States shall encourage, and may require, healthcare professionals to store and keep preferably by electronic means, the UDI of the devices with which they have been supplied with.

10. The Commission is empowered to adopt delegated acts in accordance with Article 115:

(a) amending the list of information set out in Parts A and B of Annex VI in the light of technical progress; ~~and~~

(b) amending Annex VI in the light of ~~experience obtained from the implementation of the UDI system, or in the light of~~ international developments and technical progress in the field of Unique Device Identification.

11. The Commission may, by means of implementing acts, specify the detailed arrangements and the procedural aspects for the UDI system with a view to ensuring its harmonised application in relation to any of the following:

Proposal of Regulation (EU) 2017/745 Articles

(a) determining the devices, categories or groups of devices to which the obligation laid down in paragraph 8 is to apply;

(b) specifying the data to be included in the UDI-PI of specific devices or device groups;

(c) determining the UDI related obligations laid down in this Article, Article 29 and Annex VI Part C that shall not apply to certain devices, categories or groups of devices in view of the highly individualised characterisation of those devices, or in view of their risk class, the number of devices placed on the market and the financial and administrative burden related to the assignment of UDI.

The implementing acts referred to in the first subparagraph shall be adopted in accordance with the examination procedure referred to in Article 114(3).

12. When adopting the measures referred to in paragraph 11, the Commission shall take into account all of the following:

(a) confidentiality and data protection as referred to in Articles 109 and 110 respectively;

(b) the risk-based approach;

(c) the cost-effectiveness of the measures;

(d) the convergence of UDI systems developed at international level;

(e) the need to avoid duplications in the UDI system;

(f) the needs of the healthcare systems of the Member States, and where possible, compatibility with other medical device identification systems that are used by stakeholders.

Article 28 UDI database

1. The Commission, after consulting the MDCG shall set up and manage a UDI database to validate, collate, process and make available to the public the information mentioned in Part B of Annex VI.

2. When designing the UDI database, the Commission shall take into account the general principles set out in Section 5 of Part C of Annex VI. The UDI database shall be designed in particular such that no UDI-PIs and no commercially confidential product information can be included therein.

3. The core data elements to be provided to the UDI database, referred to in Part B of Annex VI, shall be accessible to the public ~~free of charge, except the element referred to in point 13 of that Part.~~

4. The technical design of the UDI database shall ensure maximum accessibility to information stored therein, including multi-user access and automatic uploads and downloads of that information. The Commission shall provide for technical and administrative support to manufacturers and other users of the UDI database.

Article 29 Registration of devices *and systems or procedure packs*

1. Before placing a device, other than a custom-made device *or investigational device*, on the market, the manufacturer shall, ~~in accordance with the rules of the issuing entity referred to in Article 27(2), provide the [assign a]~~ Basic UDI-DI ~~as defined in Part C of Annex VI to the device and shall provide it to the UDI database together with the other core data elements referred to in Part B of Annex VI related to that device, as applicable.~~ *The manufacturer shall keep the information provided to the UDI database updated.*

2. Before placing on the market a system or procedure pack *as referred to in [pursuant to] Article 22(1) and (3), that is not a custom-made device,* the natural or legal person responsible shall *provide [assign] to the system or procedure pack, in compliance with the rules of the issuing entity, the [a] Basic UDI-DI and shall provide it* to the UDI database together with the other core data elements referred to in Part B of

Annex VI related to that system or procedure pack. The natural or legal person referred to in Article 22(1) shall keep up to date the information provided to the UDI database.

3. For devices that are the subject of a conformity assessment as referred to in Article 52(3) and in Article 52(4), second and third subparagraphs, ~~the assignment of a Basic UDI-DI referred to in paragraph 1 of this Article shall be done before the manufacturer applies to a notified body for that assessment.~~

~~For the devices referred to in the first subparagraph, the notified body shall include a reference to the Basic UDI-DI on the certificate issued in accordance with point (a) of Section 4 of Chapter I of Annex XII and confirm in Eudamed that the information referred to in Section 2.2 of Part B of Annex VI [Section 2.2 of Part A of Annex VI] is correct. After the issuing of the relevant certificate and before placing the device on the market, the manufacturer shall provide the Basic UDI-DI to the UDI database together with the other core data elements referred to in Part B of Annex VI related to that device.~~

~~4. Before placing a device on the market, other than a custom-made device, the manufacturer shall enter or if, already provided, verify in Eudamed the information referred to in Section 2 of Part A of Annex VI, with the exception of Section 2.2 thereof, and shall thereafter keep the information updated.~~

Article 30

Electronic system for registration of economic operators

1. The Commission, after consulting the MDCG, shall set up and manage an electronic system to create the single registration number referred to in Article 31(2) and to collate and process information that is necessary and proportionate to identify the manufacturer and, where applicable, the authorised representative, ~~and the importer and the person referred to in Article 22(1)~~. The details regarding the information to be provided to that electronic system by the economic operators are laid down in Part A, Section 1, of Annex VI.

2. Member States may maintain or introduce national provisions on registration of distributors of devices which have been made available on their territory. ~~Where national distributor databases require information on devices, such databases shall allow for the retrieval of the device information from the electronic systems referred to in Article 33(2), points (a) and (b).~~

3. Within two weeks of placing a device, other than a custom-made device, on the market, importers shall verify that the manufacturer or authorised representative has provided to the electronic system the information referred to in paragraph 1.

Where applicable, importers shall inform the relevant authorised representative or manufacturer if the information referred to in paragraph 1 is not included or is incorrect. Importers shall add their details to the relevant entry/entries.

Article 31

*Registration of **economic operators** [manufacturers, authorised representatives and importers]*

1. Before placing a device, other than a custom-made device, on the market, manufacturers, authorised representatives, ~~and importers and persons referred to in Article 22(1) of this Regulation~~ shall, in order to register, submit to the electronic system referred to in Article 30 the information referred to in ~~Section 1 of~~ Part A of Annex VI, provided that they have not already registered in accordance with this Article. In cases where the conformity assessment procedure requires the involvement of a notified body pursuant to Article 52, the information referred to in ~~Section 1 of~~ Part A of Annex VI shall be provided to that electronic system before applying to the notified body.

2. ~~Without undue delay, the competent authority shall verify the data entered pursuant to paragraph 1 [After having verified the data entered pursuant to paragraph 1, the competent authority shall], obtain a single registration number ('SRN') from the electronic system referred to in Article 30 and issue it to the manufacturer, the authorised representative, or the importer or the person referred to in Article 22(1).~~

Proposal of Regulation (EU) 2017/745 Articles

3. The manufacturer shall use the SRN when applying to a notified body for conformity assessment and for accessing Eudamed in order to fulfil its obligations under Article 29.

4. Within ~~two weeks~~ [one week] of any change occurring in relation to the information referred to in paragraph 1 of this Article, the economic operator shall update the data in the electronic system referred to in Article 30.

5. Not later than one year after submission of the information in accordance with paragraph 1, and every second year thereafter, the economic operator shall confirm the accuracy of the data. In the event of a failure to do so within six months of those deadlines, any Member State may take appropriate corrective measures within its territory until that economic operator complies with that obligation.

~~6. Without prejudice to the economic operator's responsibility for the data, the competent authority shall verify the confirmed data referred to in Section 1 of Part A of Annex VI.~~

7. The data entered pursuant to paragraph 1 of this Article in the electronic system referred to in Article 30 shall be accessible to the public, ~~except the information regarding the person responsible for regulatory compliance referred to in Part A, point 1.4, of Annex VI.~~

8. The competent authority may use the data to charge the manufacturer, the authorised representative, ~~or the importer or the person referred to in Article 22(1)~~ a fee pursuant to Article 111.

Article 32

Summary of safety and clinical performance

1. For ~~class IIb~~ implantable devices and for class III devices, other than custom-made or investigational devices ~~and well-established technology devices~~, the manufacturer shall draw up a summary of safety and clinical performance.

The summary of safety and clinical performance shall be written in a way that is clear to the intended user and, ~~if relevant, to the patient and~~ shall be made available to the public via Eudamed.

The draft of the summary of safety and clinical performance shall be part of the documentation to be submitted to the notified body involved in the conformity assessment pursuant to Article 52 ~~and shall be validated by that body. After its validation, the notified body shall upload the summary to Eudamed.~~ The manufacturer shall ensure that the summary of safety and clinical performance is available in Eudamed as ~~part of the information on the device to be provided pursuant to Article 29(1) and~~ mention on the label or instructions for use where that summary is available.

2. The summary of safety and clinical performance shall include at least the following aspects:

- (a) the identification of the device and the manufacturer, including the Basic UDI-DI and, if already issued, the SRN;
- (b) the intended purpose of the device and any indications, contraindications and target populations;
- (c) a description of the device, including a reference to previous generation(s) or variants if such exist, and a description of the differences, as well as, where relevant, a description of any accessories, other devices and products, which are intended to be used in combination with the device;
- (d) possible diagnostic or therapeutic alternatives;
- (e) reference to any harmonised standards and CS applied;
- (f) the summary of clinical evaluation as referred to in Annex XIV, and relevant information on post-market clinical follow-up;
- (g) suggested profile and training for users;
- (h) information on any residual risks ~~or~~ [and any] undesirable effects, ~~and any~~ warnings and precautions.

Proposal of Regulation (EU) 2017/745 Articles

3. The Commission may, by means of implementing acts, set out the form and the presentation of the data elements to be included in the summary of safety and clinical performance. Those implementing acts shall be adopted in accordance with the advisory procedure referred to in Article 114(2).

Article 33

European database on medical devices

1. The Commission, after consulting the MDCG, shall set up, maintain and manage the European database on medical devices ('Eudamed') for the following purposes:

- (a) to enable the public to be adequately informed about devices placed on the market, the corresponding certificates issued by notified bodies and about the relevant economic operators;
- (b) to enable unique identification of devices within the internal market and to facilitate their traceability;
- (c) to enable the public to be adequately informed about clinical investigations and to enable sponsors of clinical investigations to comply with obligations under Articles 62 to 80, Article 82, and any acts adopted pursuant to Article 81;
- (d) to enable manufacturers to comply with the information obligations laid down in Articles 87 to 90 or in any acts adopted pursuant to Article 91;
- (e) to enable the competent authorities of the Member States and the Commission to carry out their tasks relating to this Regulation on a well-informed basis and to enhance the cooperation between them.

2. Eudamed shall include the following electronic systems:

- (a) the electronic system for registration of devices referred to in Article 29(4);
- (b) the UDI-database referred to in Article 28;
- (c) the electronic system on registration of economic operators referred to in Article 30;
- (d) the electronic system on notified bodies and on certificates referred to in Article 57;
- (e) the electronic system on clinical investigations referred to in Article 73;
- (f) the electronic system on vigilance and post-market surveillance referred to in Article 92;
- (g) the electronic system on market surveillance referred to in Article 100.

By way of derogation from the first subparagraph, the Commission may decide that one or more of the electronic systems referred to in that subparagraph are not to be included in Eudamed. In that case, the Commission shall ensure that those electronic systems are interoperable with Eudamed.

3. When designing Eudamed and, where applicable, any of the electronic systems that are not included in Eudamed, the Commission shall give due consideration to compatibility with national databases and national web-interfaces to allow for import and export of data.

4. The data shall be entered into Eudamed and, where applicable, any of the electronic systems that are not included in Eudamed by the Member States, notified bodies, economic operators and sponsors as specified in the provisions on the electronic systems referred to in paragraph 2. The Commission shall provide for technical and administrative support to users of Eudamed and, where applicable, any of the electronic systems that are not included in Eudamed.

5. All the information collated and processed by Eudamed and, where applicable, any of the electronic systems that are not included in Eudamed shall be accessible to the Member States and to the Commission. The information shall be accessible to notified bodies, economic operators, sponsors and the public to the extent specified in the provisions on the electronic systems referred to in paragraph 2.

The Commission shall ensure that public parts of Eudamed and, where applicable, any of the electronic systems that are not included in Eudamed, are presented in a user-friendly and easily-searchable format.

Proposal of Regulation (EU) 2017/745 Articles

6. Eudamed ~~and, where applicable, any of the electronic systems that are not included in Eudamed~~, shall contain personal data only insofar as necessary for the electronic systems ~~referred to in paragraph 2 of this Article~~ to collate and process information in accordance with this Regulation. Personal data shall be kept in a form which permits identification of data subjects for periods no longer than those referred to in Article 10(8).

7. The Commission and the Member States shall ensure that data subjects may effectively exercise their rights to information, of access, to rectification and to object in accordance with Regulation (EC) No 45/2001 and Directive 95/46/EC, respectively. They shall also ensure that data subjects may effectively exercise the right of access to data relating to them, and the right to have inaccurate or incomplete data corrected and erased. Within their respective responsibilities, the Commission and the Member States shall ensure that inaccurate and unlawfully processed data are deleted, in accordance with the applicable legislation. Corrections and deletions shall be carried out as soon as possible, but no later than 60 days after a request is made by a data subject.

8. The Commission shall, by means of implementing acts, lay down the detailed arrangements necessary for the setting up and maintenance of Eudamed ~~and, where applicable, any of the electronic systems that are not included in Eudamed~~. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3). When adopting those implementing acts, the Commission shall ensure that, as far as possible, the systems are developed in such a way as to avoid having to enter the same information twice within the same **electronic system [module]** or in different **electronic systems [modules of the system]**.

9. In relation to its responsibilities under this Article and the processing of personal data involved therein, the Commission shall be considered to be the controller of Eudamed and its electronic systems **as well as, where applicable, any of the electronic systems that are not included in Eudamed**.

Article 34

Functionality of Eudamed

1. The Commission shall, in collaboration with the MDCG, draw up the functional specifications for Eudamed ~~and, where applicable, any of the electronic systems that are not included in Eudamed~~. ~~The Commission shall draw up a plan for the implementation of those specifications by 26 May 2018.~~

2. The Commission shall inform the MDCG when, on the basis of independent audit reports, it has verified that one or more of the electronic systems referred to in Article 33(2) are functional and meet the functional specifications drawn up pursuant to paragraph 1 of this Article.

3. The Commission shall, after consultation with the MDCG and when it is satisfied that the conditions referred to in paragraph 2 have been fulfilled, publish a notice to that effect in the *Official Journal of the European Union*.

Chapter IV

Notified bodies

Article 35

Authorities responsible for notified bodies

1. Any Member State that intends to designate a conformity assessment body as a notified body, or has designated a notified body, to carry out conformity assessment activities under this Regulation shall appoint an authority ('authority responsible for notified bodies'), which may consist of separate constituent entities under national law and shall be responsible for setting up and carrying out the necessary procedures for the assessment, designation and notification of conformity assessment bodies and for the monitoring of notified bodies, including subcontractors and subsidiaries of those bodies.

Proposal of Regulation (EU) 2017/745 Articles

2. The authority responsible for notified bodies shall be established, organised and operated so as to safeguard the objectivity and impartiality of its activities and to avoid any conflicts of interests with conformity assessment bodies.
3. The authority responsible for notified bodies shall be organised in a manner such that each decision relating to designation or notification is taken by personnel different from those who carried out the assessment.
4. The authority responsible for notified bodies shall not perform any activities that notified bodies perform on a commercial or competitive basis.
5. The authority responsible for notified bodies shall safeguard the confidential aspects of the information it obtains. However, it shall exchange information on notified bodies with other Member States, the Commission and, when required, with other regulatory authorities.
6. The authority responsible for notified bodies shall have a sufficient number of competent personnel permanently available for the proper performance of its tasks.

6a. Without prejudice to other administrative or judicial remedies, a manufacturer or notified body may raise with the authority responsible for notified bodies, in a duly substantiated manner, any unresolved dispute arising from the application of the requirements set out in Annex VII and the involvement of a notified body in the conformity assessment in accordance with Article 52 and Annexes IX, X and XI. The authority shall hear and decide within 90 days. Where the manufacturer is established in another Member State than the notified body, the authority responsible for the notified body shall consult the competent authority of the Member State where the manufacturer is established.

In duly justified cases, the authority responsible for notified bodies may seek guidance from the MDCG, which it shall take duly into account. Each authority responsible for notified bodies shall inform the Commission and the MDCG at least annually about the disputes raised with it pursuant to the first subparagraph, their outcome and the parties involved. That information shall be taken into consideration in the framework of the monitoring of notified bodies in accordance with Article 44.

By way of derogation from the first subparagraph and without prejudice to other administrative or judicial remedies, a Member State may choose to assign the tasks set out in this paragraph to another authority or to an out of court dispute resolution body.

Where the authority responsible for notified bodies is a different authority from the national competent authority for medical devices, it shall ensure that the national authority responsible for medical devices is consulted on relevant matters.

7. Member States shall make publicly available general information on their measures governing the assessment, designation and notification of conformity assessment bodies and for the monitoring of notified bodies, and on changes which have a significant impact on such tasks.

8. The **authorities** [authority] responsible for notified bodies shall ~~participate in the peer review activities provided for in Article 48~~ coordinate their activities to be carried out in accordance with this Chapter, cooperate with each other and with the Commission and resolve issues of diverging opinions between themselves to ensure a harmonised application of the requirements relating to notified bodies.

Article 36

Requirements relating to notified bodies

1. Notified bodies shall fulfil the tasks for which they are designated in accordance with this Regulation **in the public interest**. They shall satisfy the organisational and general requirements and the quality management, resource and process requirements **as set out in further detail in Annex VII** that are necessary to fulfil those tasks **in an effective, independent, diligent and expeditious manner**. ~~In particular, notified bodies shall comply with Annex VII.~~

~~In order to meet the requirements referred to in the first subparagraph, notified bodies shall have permanent availability of sufficient administrative, technical and scientific personnel in accordance with~~

Proposal of Regulation (EU) 2017/745 Articles

~~Section 3.1.1 of Annex VII and personnel with relevant clinical expertise in accordance with Section 3.2.4 of Annex VII, where possible employed by the notified body itself.~~

~~The personnel referred to in Sections 3.2.3 and 3.2.7 of Annex VII shall be employed by the notified body itself and shall not be external experts or subcontractors.~~

Notified bodies designated for conformity assessment activities relating to devices that are high-risk AI systems as referred to in Article 6(1) of Regulation (EU) 2024/1689, or that use high-risk AI systems as safety components, shall also meet the requirements set out in Article 31(4), (5), (10) and (11) of that Regulation.

2. Notified bodies shall make available and submit upon request all relevant documentation, including the manufacturer's documentation, to the authority responsible for notified bodies to allow it to conduct its assessment, designation, notification, monitoring and surveillance activities and to facilitate the assessment outlined in this Chapter.

3. In order to ensure the uniform application of the requirements set out in Annex VII, the Commission may adopt implementing acts, to the extent necessary to resolve issues of divergent interpretation and of practical application. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

4. The Commission is empowered to adopt delegated acts in accordance with Article 115 in order to amend Annex VII in order to adapt to technical or scientific progress regarding conformity assessment in the field of medical devices, including developments at international level.

Article 37

Subsidiaries and subcontracting

1. Where a notified body subcontracts specific tasks connected with conformity assessment or has recourse to a subsidiary for specific tasks connected with conformity assessment, it shall verify that the subcontractor or the subsidiary meets the applicable requirements set out in Annex VII and shall inform the authority responsible for notified bodies accordingly.

2. Notified bodies shall take full responsibility for the tasks performed on their behalf by subcontractors or subsidiaries.

3. Notified bodies shall make publicly available a list of their subsidiaries.

~~4. Conformity assessment activities may be subcontracted or carried out by a subsidiary provided that the legal or natural person that applied for conformity assessment has been informed accordingly.~~

5. Notified bodies shall keep at the disposal of the authority responsible for notified bodies all relevant documents concerning the verification of the qualifications of the subcontractor or the subsidiary and the work carried out by them under this Regulation.

Article 38

Application by conformity assessment bodies for designation

1. Conformity assessment bodies shall submit an application for designation to the authority responsible for notified bodies.

2. The application shall specify the conformity assessment activities as defined in this Regulation, and the types of devices for which the body is applying to be designated, and shall be supported by documentation demonstrating compliance with Annex VII.

In respect of the organisational and general requirements and the quality management requirements set out in Sections 1 and 2 of Annex VII, a valid accreditation certificate and the corresponding evaluation report delivered by a national accreditation body in accordance with Regulation (EC) No 765/2008 may be submitted and shall be taken into consideration during the assessment described in Article 39. However,

Proposal of Regulation (EU) 2017/745 Articles

the applicant shall make available all the documentation referred to in the first subparagraph to demonstrate compliance with those requirements upon request.

3. The notified body shall update the documentation referred to in paragraph 2 whenever relevant changes occur, in order to enable the authority responsible for notified bodies to monitor and verify continuous compliance with all the requirements set out in Annex VII.

Article 39

Assessment of the application

1 The authority responsible for notified bodies shall within 30 days check that the application referred to in Article 38 is complete and shall request the applicant to provide any missing information. Once the application is complete that authority shall send it to the Commission.

~~The authority responsible for notified bodies shall review the application and supporting documentation in accordance with its own procedures and shall draw up a preliminary assessment report.~~

~~2. The authority responsible for notified bodies shall submit the preliminary assessment report to the Commission which shall immediately transmit it to the MDCG.~~

2 [3] Within 14 days of the receipt of the application referred to in paragraph 1 [submission referred to in paragraph 2] of this Article, the Commission, in consultation [conjunction] with the MDCG, shall appoint a joint assessment team made up of three experts, chosen from the list referred to in Article 40(2). In view of the specific circumstances of the application, a different number of experts may be appointed [unless the specific circumstances require a different number of experts, chosen from the list referred to in Article 40(2)].

At least One of the experts shall be an [a] expert representing [representative of] the Commission who shall coordinate the activities of the joint assessment team. The other two experts shall be experts nominated by [come from] Member States other than the one in which the applicant conformity assessment body is established.

In order to be chosen in accordance with the first subparagraph, the experts shall be [The joint assessment team shall be comprised of experts who are] competent to assess the conformity assessment activities and the types of devices which are the subject of the application or, in particular when the assessment procedure is initiated in accordance with Article 47(3), to ensure that the specific concern can be appropriately assessed.

3 The experts appointed in accordance with paragraph 2 and the authority responsible for notified bodies shall form a joint assessment team that shall be coordinated by the expert representing the Commission.

4 Within 90 days of the appointment of the experts in accordance with paragraph 3 [its appointment], the joint assessment team shall review the application and supporting documentation [documentation submitted with the application] in accordance with Article 38 and shall draw up a preliminary assessment report. During that period, The joint assessment team may request any clarifications [provide feedback to, or require clarification] from, the applicant conformity assessment body [authority responsible for notified bodies on the application and on the planned on-site assessment]. The joint assessment team shall submit the preliminary assessment report to the applicant conformity assessment body.

~~The authority responsible for notified bodies together with the joint assessment team shall plan and conduct an on-site assessment of the applicant conformity assessment body and, where relevant, of any subsidiary or subcontractor, located inside or outside the Union, to be involved in the conformity assessment process.~~

~~The on-site assessment of the applicant body shall be led by the authority responsible for notified bodies.~~

5 Where, based on the review of the application and supporting documentation referred to in this paragraph, the joint assessment team considers that the applicant body does not and, within a reasonable period of time, will not be able to meet the requirements of this Regulation and that any further

Proposal of Regulation (EU) 2017/745 Articles

assessment in accordance with paragraphs 6 to 9 is not appropriate, it shall draw up a final assessment report and the authority responsible for the notified bodies shall reject the application.

6 Unless the application is rejected, the joint assessment team shall plan and conduct an on-site assessment of the applicant conformity assessment body and, where relevant, of any subsidiary or subcontractor, located inside or outside the Union, that is to be involved in the conformity assessment process.

Where an on-site assessment of the applicant body, a subsidiary or a subcontractor is temporarily impossible or impracticable due to exceptional circumstances, the joint assessment team may decide to carry out the assessment by other appropriate means.

At the end of the on-site assessment, the joint assessment team shall list for the applicant conformity assessment body any non-compliances resulting from the assessment and summarise the assessment by the joint assessment team.

7 [5] Where the joint assessment team has identified non-compliances, the applicant conformity assessment body shall submit to the joint assessment team a proposed corrective and preventive action plan to address the non-compliances effectively and in a timely manner.

That plan shall indicate the root cause of the identified non-compliances and shall include a timeframe for implementation of the actions therein.

[Findings regarding non-compliance of an applicant conformity assessment body with the requirements set out in Annex VII shall be raised during the assessment process and discussed between the authority responsible for notified bodies and the joint assessment team with a view to reaching consensus and resolving any diverging opinions, with respect to the assessment of the application.

At the end of the on-site assessment, the authority responsible for notified bodies shall list for the applicant conformity assessment body the non-compliances resulting from the assessment and summarise the assessment by the joint assessment team.

Within a specified timeframe, the applicant conformity assessment body shall submit to the national authority a corrective and preventive action plan to address the non-compliances.]

8 [6] Within 30 days after the finalisation of the on-site assessment, the joint assessment team shall submit the assessment report to the applicant conformity assessment body and, where applicable, inform that body about any non-compliances resulting from the assessment.

Where the joint assessment team has not identified any non-compliances, its assessment report shall be considered the final assessment report and paragraph 11 shall apply.

[The joint assessment team shall document any remaining diverging opinions with respect to the assessment within 30 days of completion of the on-site assessment and send them to the authority responsible for notified bodies.]

9 Within 30 days of receipt of the proposed corrective and preventive action plan referred to in paragraph 8, the joint assessment team shall assess whether the non-compliances identified during the assessment have been appropriately addressed and, where necessary, provide any comments on it to the applicant conformity assessment body, including requests for further clarification and modifications.

The applicant conformity assessment body and the joint assessment team shall endeavour to agree on a final corrective and preventive action plan in due course.

10 [7] Within 30 days of receipt of the final corrective and preventive action plan, or of the moment the joint assessment team concludes that no agreement on a final plan has been reached, the joint assessment team shall draw up its final assessment report. That report shall include the result of the assessment, conclusions regarding the corrective and preventive action plans and whether the non-compliances have been appropriately addressed and, where applicable, the recommended scope of designation.

[The authority responsible for notified bodies shall following receipt of a corrective and preventive action plan from the applicant body assess whether non-compliances identified during the assessment have

Proposal of Regulation (EU) 2017/745 Articles

been appropriately addressed. This plan shall indicate the root cause of the identified non-compliances and shall include a timeframe for implementation of the actions therein.

The authority responsible for notified bodies shall having confirmed the corrective and preventive action plan forward it and its opinion thereon to the joint assessment team. The joint assessment team may request of the authority responsible for notified bodies further clarification and modifications.

The authority responsible for notified bodies shall draw up its final assessment report which shall include:

- the result of the assessment,
- confirmation that the corrective and preventive actions have been appropriately addressed and, where required, implemented,
- any remaining diverging opinion with the joint assessment team, and, where applicable,
- the recommended scope of designation.]

11 [8] The joint assessment team shall submit its final assessment report to the MDCG without undue delay.

Based on the findings of the final assessment report, the authority responsible for notified bodies shall submit to the MDCG a draft decision on the designation of the notified body or reject the application.

[The authority responsible for notified bodies shall submit its final assessment report and, if applicable, the draft designation to the Commission, the MDCG and the joint assessment team.]

12 [9] sentence 1 Within 21 days of receipt of the draft decision on the designation referred to in paragraph 11, the MDCG shall issue a recommendation with regard to the envisaged designation, which the authority responsible for notified bodies shall duly take into consideration for its final decision on the designation of the notified body. That 21-day period may be extended once for a further 21 days on justified grounds.

[The joint assessment team shall provide a final opinion regarding the assessment report prepared by the authority responsible for notified bodies and, if applicable, the draft designation within 21 days of receipt of those documents to the Commission, which shall immediately submit that final opinion to the MDCG.]

13 [9] sentence 2 Where the authority responsible for notified bodies does not agree with the recommendation of the MDCG, it shall submit to the MDCG a duly justified request to reconsider its recommendation. Within 30 days of receipt of that request, the MDCG shall either confirm its recommendation or issue a new recommendation.

[Within 42 days of receipt of the opinion of the joint assessment team, the MDCG shall issue a recommendation with regard to the draft designation, which the authority responsible for notified bodies shall duly take into consideration for its decision on the designation of the notified body.]

14 Where no agreement can be reached between the MDCG and the authority responsible for notified bodies, either party may refer the matter to the Commission.

Within 180 days of receipt of the referral, the Commission shall, after consulting the MDCG, the authority responsible for notified bodies and, where necessary, the applicant conformity assessment body concerned, evaluate the draft decision on the designation and decide, by means of implementing act, whether or not the draft designation is justified.

15 If, at any stage of the process, consensus cannot be reached within the joint assessment team on any issue, any member of the joint assessment team may refer the issue to the MDCG, which shall provide its views without undue delay and at the latest within 60 days from the referral.

16 [10] The Commission may, by means of implementing acts, adopt measures setting out ~~the~~ detailed arrangements specifying ~~the~~ procedures and ~~reports~~ any relevant documentation for the following:

- (a) ~~for~~ the application for designation referred to in Article 38; ~~and~~
- (b) the assessment of the application set out in this Article;
- (c) the nomination and selection of experts referred to in Article 40;

(d) the monitoring of notified bodies pursuant to Article 44.

Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

Article 40

Nomination of experts for joint assessment ~~of applications for notification~~

1. All [The] Member States that have appointed an authority responsible for notified bodies and the Commission shall nominate experts who are qualified in the assessment, designation or monitoring of conformity assessment bodies in the field of medical devices or in the assessment of manufacturers' technical documentation and who will be available to participate in the activities referred to in Articles 39, 44 and 48. Member States that have not appointed an authority responsible for notified bodies may nominate experts who have those qualifications.

The nominated experts shall commit to participate in joint assessments.

2. The Commission shall maintain a list of the experts nominated pursuant to paragraph 1 of this Article, together with information on their specific field of competence and expertise. That list shall be made available to Member States competent authorities through the electronic system referred to in Article 57.

Article 40a

Funding of activities related to designation and monitoring of notified bodies

1. Conformity assessment bodies and notified bodies shall pay a fee for the assessment of their application for designation and their monitoring, including the costs for the involvement of experts nominated in accordance with Article 40 in those assessment and monitoring activities.

2. The structure and level of the fees and the scale and type of recoverable costs shall be established by the Commission by means of implementing acts. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

Article 41

Language requirements

All documents required pursuant to Articles 38 and 39 shall be drawn up in a language or languages which shall be determined by the Member State concerned.

Member States, in applying the first paragraph, shall consider accepting and using a commonly understood language in the medical field, for all or part of the documentation concerned.

The applicant [Commission] shall provide translations of the documentation pursuant to Articles 38 and 39, or parts thereof, into an official Union language, such as is necessary for that documentation to be readily understood by the joint assessment team referred to in Article 39(4) [appointed in accordance with Article 39(3)].

Article 42

Designation and notification procedure

1. Member States may only designate conformity assessment bodies in accordance with the recommendation of the MDCG issued in accordance with Article 39(11) or (12) or in accordance with an implementing act adopted by the Commission in accordance with Article 39(14) [for which the assessment pursuant to Article 39 was completed and which comply with Annex VII].

2. Member States shall notify the Commission and the other Member States of the conformity assessment bodies they have designated, using the electronic notification tool within the database of notified bodies developed and managed by the Commission (NANDO).

Proposal of Regulation (EU) 2017/745 Articles

3. The notification shall clearly specify, using the codes referred to in paragraph 13 of this Article, the scope of the designation indicating the conformity assessment activities as defined in this Regulation and the types of devices which the notified body is authorised to assess and, without prejudice to Article 44, any conditions associated with the designation.
4. The notification shall be accompanied by the ~~final assessment report of the authority responsible for notified bodies, the final opinion of the joint assessment team referred to in Article 39(9) and the recommendation of the MDCG. Where the notifying Member State does not follow the recommendation of the MDCG, it shall provide a duly substantiated justification.~~
5. The notifying Member State shall, without prejudice to Article 44, inform the Commission and the other Member States of any conditions associated with the designation and provide documentary evidence regarding the arrangements in place to ensure that the notified body will be monitored regularly and will continue to satisfy the requirements set out in Annex VII.
- ~~6. Within 28 days of the notification referred to in paragraph 2, a Member State or the Commission may raise written objections, setting out its arguments, with regard either to the notified body or to its monitoring by the authority responsible for notified bodies. Where no objection is raised, the Commission shall publish in NANDO the notification within 42 days of its having been notified as referred to in paragraph 2.~~
- ~~7. When a Member State or the Commission raises objections in accordance with paragraph 6, the Commission shall bring the matter before the MDCG within 10 days of the expiry of the period referred to in paragraph 6. After consulting the parties involved, the MDCG shall give its opinion at the latest within 40 days of the matter having been brought before it. Where the MDCG is of the opinion that the notification can be accepted, the Commission shall publish in NANDO the notification within 14 days.~~
- ~~8. Where the MDCG, after having been consulted in accordance with paragraph 7, confirms the existing objection or raises another objection, the notifying Member State shall provide a written response to the MDCG opinion within 40 days of its receipt. The response shall address the objections raised in the opinion, and set out the reasons for the notifying Member State's decision to designate or not designate the conformity assessment body.~~
- ~~9. Where the notifying Member State decides to uphold its decision to designate the conformity assessment body, having given its reasons in accordance with paragraph 8, the Commission shall publish in NANDO the notification within 14 days of being informed thereof.~~
10. When publishing the notification in NANDO, the Commission shall add to the electronic system referred to in Article 57 the information relating to the notification of the notified body along with the documents [referred to \[mentioned\]](#) in paragraph 4 of this Article ~~and the opinion and responses referred to in paragraphs 7 and 8 of this Article.~~
11. The designation shall become valid the day after the notification is published in NANDO. The published notification shall state the scope of lawful conformity assessment activity of the notified body.
12. The conformity assessment body concerned may perform the activities of a notified body only after the designation has become valid in accordance with paragraph 11.
13. The Commission shall by 26 November 2017, by means of implementing acts, draw up a list of codes and corresponding types of devices for the purpose of specifying the scope of the designation of notified bodies. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3). The Commission, after consulting the MDCG, may update this list based, *inter alia*, on information arising from the coordination activities described in Article 48.

Article 43

Identification number and list of notified bodies

1. The Commission shall assign an identification number to each notified body for which the notification becomes valid in accordance with Article 42(11). It shall assign a single identification number even when the body is notified under several Union acts. If they are successfully designated in accordance with this Regulation, bodies notified pursuant to Directives 90/385/EEC and 93/42/EEC shall retain the identification number assigned to them pursuant to those Directives.

2. The Commission shall make the list of the bodies notified under this Regulation, including the identification numbers that have been assigned to them and the conformity assessment activities as defined in this Regulation and the types of devices for which they have been notified, accessible to the public in NANDO. It shall also make this list available on the electronic system referred to in Article 57. The Commission shall ensure that the list is kept up to date.

Article 44

Monitoring ~~and re-assessment~~ of notified bodies

1. Notified bodies shall, without delay, and at the latest within 15 days, inform the authority responsible for notified bodies of relevant changes which may affect their compliance with the requirements set out in Annex VII or their ability to conduct the conformity assessment activities relating to the devices for which they have been designated.

2. The authorities responsible for notified bodies shall monitor the notified bodies established on their territory and their subsidiaries and subcontractors to ensure ongoing compliance with the requirements and the fulfilment of its obligations set out in this Regulation. Notified bodies shall, upon request by their authority responsible for notified bodies, supply all relevant information and documents, required to enable the authority ~~and the joint assessment team~~ [the Commission and other Member States] to verify compliance.

3. Where the Commission or the authority of a Member State submits a request to a notified body established on the territory of another Member State relating to a conformity assessment carried out by that notified body, it shall send a copy of that request to the authority responsible for notified bodies of that other Member State. The notified body concerned shall respond without delay and within 15 days at the latest to the request. The authority responsible for notified bodies of the Member State in which the body is established shall ensure that requests submitted by authorities of any other Member State or by the Commission are resolved by the notified body unless there is a legitimate reason for not doing so in which case the matter may be referred to the MDCG.

(a) The authority responsible for notified bodies shall conduct its monitoring and assessment activities in accordance with an assessment programme considering the full scope of the notified body's activities. That programme shall ensure that the authority can effectively monitor the continued compliance of the notified body with the requirements of this Regulation and shall provide a reasoned schedule for the frequency of assessment of the notified body and, where applicable, its subsidiaries and/or subcontractors for a period of at least two years. The authority responsible for notified bodies shall submit the programme for monitoring or assessment for each notified body for which it is responsible to the MDCG and to the Commission. The authority shall address any request for clarification or modification made by the Commission or the MDCG.

4. At least once a year, the authorities responsible for notified bodies shall ~~re-assess~~ whether ~~each~~ [the] notified ~~body~~ [bodies] established on their respective territory and, where appropriate, the subsidiaries and subcontractors under the responsibility of those notified bodies still satisfy the requirements and fulfil their obligations set out in ~~this Regulation, in particular~~ Annex VII.

Where necessary, ~~the assessment~~ [That review] shall include an on-site audit of ~~the~~ [each] notified body, ~~and of~~ its subsidiaries or subcontractors.

Proposal of Regulation (EU) 2017/745 Articles

~~The authority responsible for notified bodies shall conduct its monitoring and assessment activities according to an annual assessment plan to ensure that it can effectively monitor the continued compliance of the notified body with the requirements of this Regulation. That plan shall provide a reasoned schedule for the frequency of assessment of the notified body and, in particular, associated subsidiaries and subcontractors.~~ The authority, the MDCG or the Commission may request the participation of experts from the Commission and other Member States in the annual assessment of a notified body. [The authority shall submit its annual plan for monitoring or assessment for each notified body for which it is responsible to the MDCG and to the Commission.]

(a) At least every two years, the annual assessment of a notified body shall be carried out by a joint assessment team including the authority responsible for notified bodies and two experts from the list referred to in Article 40(2) appointed by the Commission in consultation with the MDCG. At least one of those experts shall be an expert representing the Commission. In light of the specific circumstances of the monitoring, the Commission may appoint a different number of experts. The joint assessment team shall be coordinated by the expert representing the Commission.

If, at any stage of the process, consensus cannot be reached within the joint assessment team on any issue, any member of the joint assessment team may refer the issue to the MDCG, which shall provide its views without undue delay and at the latest within 60 days from the referral.

(b) At the end of any assessment carried out pursuant to paragraph 4 or 4a, the authority responsible for notified bodies or the joint assessment team, as applicable, shall inform the notified body about any non-compliances resulting from the assessment and summarise their assessment.

Where non-compliances have been identified, the notified body shall submit a proposed corrective and preventive action plan to address the non-compliances. That plan shall indicate the root cause of the identified non-compliances and shall include a timeframe for implementation of the actions set out therein.

The authority responsible for notified bodies or the joint assessment team, as applicable, shall assess whether the non-compliances identified during the assessment have been appropriately addressed in the plan referred to in the second subparagraph and, where necessary, provide any comments on the plan to the notified body, including requests for further clarification and modifications. The notified body and the authority or the joint assessment team, as applicable, shall endeavour to agree on a final corrective and preventive action plan in due course.

(c) After receipt of the final corrective and preventive action plan, or where the authority responsible for notified bodies or the joint assessment team, as applicable, have not identified non-compliances or conclude that no agreement on a final plan has been reached, the authority or the joint assessment team, as applicable, shall draw up their final monitoring report which shall include the result of the assessment and, where applicable, conclusions regarding the corrective and preventive action plan and, where applicable, any recommendations regarding the notified body's designation.

The authority responsible for notified bodies or the joint assessment team, as applicable, shall submit their final monitoring report to the MDCG without undue delay.

The authority responsible for notified bodies shall monitor the implementation of the corrective and preventive action plan by the notified body, as appropriate.

Where the final monitoring report concludes that the notified body no longer meets the requirements set out in this Regulation, or where the notified body fails to implement the corrective and preventive action plan, the authority responsible for notified bodies shall follow the procedure set out in Article 46(4).

5. The monitoring of notified bodies ~~by the authority responsible for notified bodies~~ shall include observed audits of notified body personnel, including where necessary any personnel from subsidiaries and subcontractors, as that personnel is in the process of conducting quality management system assessments at a manufacturer's facility.

6. The monitoring of notified bodies ~~conducted by the authority responsible for notified bodies~~ shall consider data arising from market surveillance, vigilance and post-market surveillance ~~to help guide its activities.~~

Proposal of Regulation (EU) 2017/745 Articles

The authority responsible for notified bodies shall provide for a systematic follow-up of complaints and other information, including from other Member States, which may indicate non-fulfilment of the obligations by a notified body or its deviation from common or best practice.

7. The authority responsible for notified bodies may in addition to regular monitoring or on-site assessments conduct short-notice, unannounced or 'for-cause' reviews if needed to address a particular issue or to verify compliance.

~~8. The authority responsible for notified bodies shall review the assessments by notified bodies of manufacturers' technical documentation, in particular the clinical evaluation documentation as further outlined in Article 45.~~

During the monitoring of a notified body, an appropriate number of the notified body's assessments of manufacturers' quality management system and technical documentation, in particular the clinical evaluation documentation, shall be reviewed either off-site or on-site. The sampling of files shall be representative of the types and risk of devices certified by the notified body, in particular high-risk devices.

~~9. The authority responsible for notified bodies shall document and record any findings regarding non-compliance of the notified body with the requirements set out in Annex VII and shall monitor the timely implementation of corrective and preventive actions.~~

~~10. 5 years after notification of a notified body, and again every 5 years thereafter, a complete re-assessment to determine whether the notified body still satisfies the requirements set out in Annex VII shall be conducted by the authority responsible for notified bodies of the Member State in which the body is established and by a joint assessment team in accordance with the procedure described in Article 39.~~

~~The authority responsible for notified bodies of the Member State in which the notified body is established may conduct a complete re-assessment prior to the dates referred to in the first subparagraph, upon request by the notified body or where, based on the results of the annual assessments conducted in accordance with paragraph 4 of this Article, it has concerns regarding the continued fulfilment by the notified body of the requirements set out in Annex VII.~~

~~Complete re-assessments that have already started prior to 11 March 2023 shall continue to be conducted, unless the authority responsible for notified bodies of the Member State in which the notified body is established decides to suspend or terminate the ongoing complete re-assessment, taking into account its own resources and the resources of the notified body already spent on the re-assessment, as well as the results of the annual assessments conducted in accordance with paragraph 4 of this Article. Before suspending or terminating an ongoing complete re-assessment, the authority responsible for notified bodies shall hear the notified body concerned.~~

~~11. The Commission is empowered to adopt delegated acts in accordance with Article 115 in order to amend paragraph 10 to modify the frequency at which the complete re-assessment referred to in that paragraph is to be carried out.~~

12. The authority responsible for notified bodies shall draw up an annual summary of [The Member States shall report to the Commission and to the MDCG, at least once a year, on] their monitoring ~~and on-site assessment~~ activities regarding notified bodies and, where applicable, subsidiaries and subcontractors. ~~The report shall provide details of the outcome of those activities, including activities pursuant to paragraph 7, and shall be treated as confidential by the MDCG and the Commission;~~ That summary shall be made publicly available through the electronic system referred to in Article 57. [however it shall contain a summary which shall be made publicly available. The summary of the report shall be uploaded to the electronic system referred to in Article 57.]

Article 45

~~Review of notified body assessment of technical documentation and clinical evaluation documentation~~

~~1. The authority responsible for notified bodies, as part of its ongoing monitoring of notified bodies, shall review an appropriate number of notified body assessments of manufacturers' technical documentation, in~~

Proposal of Regulation (EU) 2017/745 Articles

~~particular the clinical evaluation documentation as referred to in points (c) and (d) of Section 6.1 of Annex II to verify the conclusions drawn by the notified body based on the information presented by the manufacturer. The reviews by the authority responsible for notified bodies shall be conducted both off-site and on-site.~~

~~2. The sampling of files to be reviewed in accordance with paragraph 1 shall be planned and representative of the types and risk of devices certified by the notified body, in particular high risk devices, and be appropriately justified and documented in a sampling plan, which shall be made available by the authority responsible for notified bodies to the MDCG upon request.~~

~~3. The authority responsible for notified bodies shall review whether the assessment by the notified body was conducted appropriately and shall check the procedures used, associated documentation and the conclusions drawn by the notified body. Such checking shall include the technical documentation and clinical evaluation documentation of the manufacturer upon which the notified body has based its assessment. Such reviews shall be conducted utilising GS.~~

~~4. Those reviews shall also form part of the re-assessment of notified bodies in accordance with Article 44(10) and the joint assessment activities referred to in Article 47(3). The reviews shall be conducted utilising appropriate expertise.~~

~~5. Based on the reports of the reviews and assessments by the authority responsible for notified bodies or joint assessment teams, on input from the market surveillance, vigilance and post-market surveillance activities described in Chapter VII, on the continuous monitoring of technical progress, or on the identification of concerns and emerging issues on the safety and performance of devices, the MDCG may recommend that the sampling, carried out under this Article, cover a greater or lesser proportion of the technical documentation and clinical evaluation documentation assessed by a notified body.~~

~~6. The Commission may, by means of implementing acts, adopt measures setting out the detailed arrangements, associated documents for, and coordination of, the review of assessments of technical documentation and clinical evaluation documentation, as referred to in this Article. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).~~

Article 46

Changes to designations and notifications

1. The authority responsible for notified bodies shall notify the Commission and the **MDCG** [other Member States] of any **application** for relevant changes to the designation of a notified body.

The procedures **set out** [described] in Articles 39 and 42 shall apply to **significant** extensions of the scope of the designation. **By way of derogation from Article 39, an on-site assessment shall not be conducted where the Joint Assessment Team considers such assessment not to be necessary for the assessment of the requested scope extension.**

For changes to the designation other than **significant** extensions of its scope, the procedures laid down in **paragraphs 2 to 9** [the following paragraphs] shall apply.

2. **After having assessed the application for changes, the authority responsible for notified bodies shall notify the Commission of the relevant changes to the designation.** The Commission shall **without undue delay** [immediately] publish the amended notification in NANDO. The Commission shall **also, without undue delay** [immediately], enter information on the changes to the designation of the notified body in the electronic system referred to in Article 57.

3. Where a notified body decides to cease its conformity assessment activities it shall inform the authority responsible for notified bodies and the manufacturers concerned as soon as possible and in the case of a planned cessation one year before ceasing its activities. The certificates may remain valid for a temporary period of nine months after cessation of the notified body's activities on condition that another notified body has confirmed in writing that it will assume responsibilities for the devices covered by those certificates. The new notified body shall complete a full assessment of the devices affected by the end of

Proposal of Regulation (EU) 2017/745 Articles

that time period before issuing new certificates for those devices. Where the notified body has ceased its activity, the authority responsible for notified bodies shall withdraw the designation.

4. Where an authority responsible for notified bodies has ascertained that a notified body no longer meets the requirements set out in [this Regulation \[Annex VII\]](#) or that it is failing to fulfil its obligations or has not implemented the necessary corrective measures, the authority shall suspend, restrict, or fully or partially withdraw the designation or impose conditions on the notified body, depending on the seriousness of the failure to meet those requirements or fulfil those obligations. A suspension shall not exceed a period of one year, renewable once for the same period.

The authority responsible for notified bodies shall immediately inform the Commission and the other Member States of any suspension, restriction or withdrawal of a designation.

5. Where its designation has been suspended, restricted, or fully or partially withdrawn, the notified body shall inform the manufacturers concerned at the latest within 10 days.

6. In the event of restriction, suspension or withdrawal of a designation, the authority responsible for notified bodies shall take appropriate steps to ensure that the files of the notified body concerned are kept and make them available to authorities in other Member States responsible for notified bodies and to authorities responsible for market surveillance at their request.

7. In the event of restriction, suspension or withdrawal of a designation, the authority responsible for notified bodies shall:

- (a) assess the impact on the certificates issued by the notified body;
- (b) submit a report on its findings to the Commission and the other Member States within three months of having notified the changes to the designation;
- (c) require the notified body to suspend or withdraw, within a reasonable period of time determined by the authority, any certificates which were unduly issued to ensure the safety of devices on the market;
- (d) enter into the electronic system referred to in Article 57 information in relation to certificates of which it has required their suspension or withdrawal;
- (e) inform the competent authority for medical devices of the Member State in which the manufacturer has its registered place of business through the electronic system referred to in Article 57 of the certificates for which it has required suspension or withdrawal. That competent authority shall take the appropriate measures, where necessary to avoid a potential risk to the health or safety of patients, users or others.

8. With the exception of certificates unduly issued, and where a designation has been suspended or restricted, the certificates shall remain valid in the following circumstances:

- (a) the authority responsible for notified bodies has confirmed, within one month of the suspension or restriction, that there is no safety issue in relation to certificates affected by the suspension or restriction, and the authority responsible for notified bodies has outlined a timeline and actions anticipated to remedy the suspension or restriction; or
- (b) the authority responsible for notified bodies has confirmed that no certificates relevant to the suspension will be issued, amended or re-issued during the course of the suspension or restriction and states whether the notified body has the capability of continuing to monitor and remain responsible for existing certificates issued for the period of the suspension or restriction. In the event that the authority responsible for notified bodies determines that the notified body does not have the capability to support existing certificates issued, the manufacturer shall provide, to the competent authority for medical devices of the Member State in which the manufacturer of the device covered by the certificate has its registered place of business, within three months of the suspension or restriction, a written confirmation that another qualified notified body is temporarily assuming the functions of the notified body to monitor and remain responsible for the certificates during the period of suspension or restriction.

Proposal of Regulation (EU) 2017/745 Articles

9. With the exception of certificates unduly issued, and where a designation has been withdrawn, the certificates shall remain valid for a period of nine months in the following circumstances:

- (a) where the competent authority for medical devices of the Member State in which the manufacturer of the device covered by the certificate has its registered place of business has confirmed that there is no safety issue associated with the devices in question; and
- (b) another notified body has confirmed in writing that it will assume immediate responsibilities for those devices and will have completed assessment of them within twelve months of the withdrawal of the designation.

In the circumstances referred to in the first subparagraph, the competent authority for medical devices of the Member State in which the manufacturer of the device covered by the certificate has its place of business may extend the provisional validity of the certificates for further periods of three months, which altogether shall not exceed twelve months.

The authority or the notified body assuming the functions of the notified body affected by the change of designation shall immediately inform the Commission, the other Member States and the other notified bodies thereof.

Article 47

Challenge to the competence of notified bodies

1. The Commission, in conjunction with the MDCG, shall investigate all cases where concerns have been brought to its attention regarding the continued fulfilment by a notified body, or of one or more of its subsidiaries or subcontractors, of the requirements set out in [this Regulation \[Annex VII or the obligations to which they are subject\]](#). It shall ensure that the relevant authority responsible for notified bodies is informed and is given an opportunity to investigate those concerns.
2. The notifying Member State shall provide the Commission, on request, with all information regarding the designation of the notified body concerned.
3. The Commission, in [consultation \[conjunction\]](#) with the MDCG, may initiate, ~~as applicable,~~ the assessment ~~procedure by a joint assessment team as referred to [described] in Article 39(2) to (10) [39(3) and (4)],~~ where there is reasonable concern about the ongoing compliance of a notified body, or a subsidiary or subcontractor of the notified body, with the requirements set out in [the Regulation \[Annex VII\]](#) and where the investigation by the authority responsible for notified bodies is not deemed to have fully addressed the concerns or upon request of the authority responsible for notified bodies. The reporting and outcome of that assessment shall follow the principles [set out in \[of\] Article 39](#). ~~Alternatively, depending on the severity of the issue, the Commission, in conjunction with the MDCG, may request that the authority responsible for notified bodies allow the participation of up to two experts from the list established pursuant to Article 40 in an on-site assessment as part of the planned monitoring and assessment activities in accordance with Article 44 and as outlined in the annual assessment plan described in Article 44(4).~~
4. Where the Commission ascertains that a notified body no longer meets the requirements for its designation, it shall inform the notifying Member State accordingly and request it to take the necessary corrective measures, including the suspension, restriction or withdrawal of the designation if necessary.

Where the Member State fails to take the necessary corrective measures, the Commission may, by means of implementing acts, suspend, restrict or withdraw the designation. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3). It shall notify the Member State concerned of its decision and update NANDO and the electronic system referred to in Article 57.
5. The Commission shall ensure that all confidential information obtained in the course of its investigations is treated accordingly.

Article 48

Peer review and exchange of experience between authorities responsible for notified bodies

1. The Commission shall provide for the organisation of exchange of experience and coordination of administrative practice between the authorities responsible for notified bodies. Such exchange shall cover elements including:

- (a) development of best practice documents relating to the activities of the authorities responsible for notified bodies;
- (b) development of guidance documents for notified bodies in relation to the implementation of this Regulation;
- (c) training and qualification of the experts referred to in Article 40;
- (d) monitoring of trends relating to changes to notified body designations and notifications and trends in certificate withdrawals and transfers between notified bodies;
- (e) monitoring of the application and applicability of scope codes referred to in Article 42(13);
- (f) development of a mechanism for peer reviews between authorities and the Commission;
- (g) methods of communication to the public on the monitoring and surveillance activities of authorities and the Commission on notified bodies.

~~2. The authorities responsible for notified bodies shall participate in a peer review every third year through the mechanism developed pursuant to paragraph 1 of this Article. Such reviews shall normally be conducted in parallel with the on-site joint assessments described in Article 39. Alternatively, an authority may make the choice of having such reviews take place as part of its monitoring activities referred to in Article 44.~~

~~3. The Commission shall participate in the organisation and provide support to the implementation of the peer review mechanism.~~

~~4. The Commission shall compile an annual summary report of the peer review activities which shall be made publicly available.~~

~~5. The Commission may, by means of implementing acts, adopt measures setting out the detailed arrangements and related documents for the peer review mechanism and training and qualification as referred to in paragraph 1 of this Article. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).~~

Article 49

Coordination of notified bodies

1. The Commission shall ensure that appropriate coordination and cooperation between notified bodies is put in place and operated in the form of a coordination group of notified bodies in the field of medical devices, including in vitro diagnostic medical devices. **The coordination group [This group]** shall meet on a regular basis, at least annually, **and report to the MDCG.**

~~The bodies notified under this Regulation shall participate in the work of that group.~~

The Commission may establish the specific arrangements for the functioning of the coordination group **of notified bodies.**

~~2. Notified bodies shall ensure that their specialised personnel participate in the relevant activities of the coordination group.~~

~~3. All bodies notified under this Regulation and under Regulation (EU) 2017/746 shall actively participate in the work of the coordination group to support the implementation of this Regulation and of Regulation (EU) 2017/746 by sharing experience and developing common templates and technical guidance to facilitate harmonisation and common approaches regarding conformity assessment activities. They shall~~

Proposal of Regulation (EU) 2017/745 Articles

cooperate with each other, with the Commission, with the relevant authorities of the Member States, with expert panels and, where applicable, expert laboratories or European Union reference laboratories to ensure a harmonised application of the requirements set out in this Regulation and in Regulation (EU) 2017/746.

4. All bodies notified under this Regulation and under Regulation (EU) 2017/746 shall adhere to a code of conduct developed by the coordination group and approved by the MDCG. The code of conduct shall set out the principles of public interest, highest professional competence and integrity, impartiality, independence, transparency, proportionality, predictability and accountability to which notified bodies commit when exercising the rights and obligations conferred on them by this Regulation or Regulation (EU) 2017/746.

Article 50

Access to notified bodies and fees [List of standard fees]

1. Notified bodies shall establish lists of their ~~standard~~ fees for the conformity assessment activities that they carry out and shall make those lists publicly available. They shall notify the lists to the Commission, which shall make references to them available to the public on a dedicated website.

2. Notified bodies shall apply at least a 50 % fee reduction for manufacturers that are micro enterprises within the meaning of Recommendation 2003/361/EC and at least a 25 % fee reduction for small enterprises within the meaning of that Recommendation. They shall apply at least a 50 % fee reduction for manufacturers that apply for conformity assessment of an orphan device referred to in Article 52a(3).

Notified bodies shall provide manufacturers that are micro or small enterprises within the meaning of Recommendation 2003/361/EC the possibility to defer the payment of fees until the relevant conformity assessment activity is finalised.

3. The Commission, in consultation with the MDCG, may adopt implementing acts to specify the structure and level of the fees referred to in paragraph 1, taking into account the need to:

1. establish and maintain high standards of quality and safety of devices;
2. ensure the availability of devices;
3. protect the interests of micro, small or medium-sized enterprises within the meaning of Recommendation 2003/361/EC;
4. support innovation and competitiveness.

4. Notified bodies shall ensure that manufacturers, which are micro, small or medium-sized enterprises within the meaning of Recommendation 2003/361/EC, have access to their conformity assessment activities in a manner that is not less favourable than the manner in which access is provided to other manufacturers.

5. Notified bodies shall deal with any request for conformity assessment activities from a manufacturer and, within 15 days of receipt of the request, inform the manufacturer accordingly.

6. When duly justified in the interest of public health or patient health or safety, the authority responsible for notified bodies may instruct a notified body to accept a manufacturer's request for conformity assessment activities falling within that notified body's scope of designation.

Chapter V
Classification and conformity assessment

SECTION 1
CLASSIFICATION

Article 51
Classification of devices

1. Devices shall be divided into classes I, IIa, IIb and III, taking into account the intended purpose of the devices and their inherent risks. Classification shall be carried out in accordance with Annex VIII.

2. The competent authorities shall coordinate their activities when determining the classification of a device, or a category or group of devices. The results of the coordination activities of the competent authorities, including the results of any decision or measure adopted by a competent authority in accordance with Articles 51a or 51b and any opinion issued by an expert panel in relation to classification, shall be made publicly available, without disclosing any confidential information as referred to in Article 109.

~~Any dispute between the manufacturer and the notified body concerned, arising from the application of Annex VIII, shall be referred for a decision to the competent authority of the Member State in which the manufacturer has its registered place of business. In cases where the manufacturer has no registered place of business in the Union and has not yet designated an authorised representative, the matter shall be referred to the competent authority of the Member State in which the authorised representative referred to in the last indent of point (b) of the second paragraph of Section 2.2 of Annex IX has its registered place of business. Where the notified body concerned is established in a Member State other than that of the manufacturer, the competent authority shall adopt its decision after consultation with the competent authority of the Member State that designated the notified body.~~

~~The competent authority of the Member State in which the manufacturer has its registered place of business shall notify the MDCG and the Commission of its decision. The decision shall be made available upon request.~~

3. At the request of a Member State the Commission shall after consulting the MDCG, decide, by means of implementing acts, on the following:

(a) application of Annex VIII to a given device, or category or group of devices, with a view to determining the classification of such devices;

(b) that a device, or category or group of devices, shall for reasons of public health based on new scientific evidence, or based on any information which becomes available in the course of the vigilance and market surveillance activities be reclassified, by way of derogation from Annex VIII, **taking into consideration the principle of proportionality and classification of devices at international level.**

4. The Commission may also, on its own initiative and after consulting the MDCG, decide, by means of implementing acts, on the issues referred to in points (a) and (b) of paragraph 3.

5. In order to ensure the uniform application of **the rules set out in** Annex VIII, and taking account of the relevant scientific opinions of the relevant scientific committees **or expert panels**, the Commission **is empowered to [may]** adopt implementing acts to the extent necessary to resolve issues of divergent interpretation and of practical application.

6. The implementing acts referred to in paragraphs 3, 4 and 5 of this Article shall be adopted in accordance with the examination procedure referred to in Article 114(3).

7. The Commission is empowered to adopt delegated acts in accordance with Article 115 to amend Annex VIII in order to adapt it to technical or scientific progress or to take into account developments regarding classification of devices at international level.

Article 51a

Classification in the event of a dispute between manufacturer and notified body

1. A manufacturer or a notified body may refer any dispute between them arising from the application of Annex VIII to the competent authority of the Member State in which the manufacturer has its registered place of business. In cases where the manufacturer has no registered place of business in the Union and has not yet designated an authorised representative, the matter shall be referred to the competent authority of the Member State in which the authorised representative referred to in Section 2.2, second paragraph, point (b), last indent, of Annex IX has its registered place of business.
2. Within 30 days of receipt of the referral referred to in paragraph 1, the competent authority shall consult the other Member States regarding its draft classification decision.
3. Where, within 30 days of receipt of the consultation referred to in the paragraph 2, no substantiated disagreement has been raised by a Member State, the competent authority shall adopt its decision within 90 days of receipt of the referral referred to in paragraph 1.
4. Where, within 30 days of receipt of the consultation referred to in paragraph 2, a substantiated disagreement has been raised by a Member State regarding the draft classification decision, the matter shall be referred to an expert panel as referred to in Article 106. That expert panel shall deliver an opinion on the classification of the device within 30 days. The competent authority may ask the expert panel for clarifications on its opinion.
5. Within 30 days of receipt of the expert panel opinion, or any requested clarification, referred to in paragraph 4, the competent authority shall adopt its decision, giving utmost consideration to the expert panel opinion. It shall notify the other Member States and the Commission of its decision without undue delay.
6. The Commission may, by means of implementing acts, lay down further details of the procedure for the application of this Article and of Article 51b. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

Article 51b

Challenges to the classification of CE marked devices

1. Where a competent authority, after having performed an evaluation in accordance with Article 94, considers that a device that is CE marked in accordance with Article 20, is not classified in accordance with Annex VIII, it shall consult the other Member States regarding its envisaged measure on the classification of the device.
2. Where, within 30 days of receipt of the consultation referred to in paragraph 1, no substantiated disagreement is raised by a Member State, the competent authority may adopt the measure on the classification of the device in question and shall notify the other Member States and the Commission of its decision giving the reasons for the decision.
3. Where, within 30 days of receipt of the consultation referred to in paragraph 1, a substantiated disagreement is raised by a Member State regarding the envisaged measure on the classification, the matter shall be referred to an expert panel referred to in Article 106, which shall deliver an opinion on the classification of the device within 30 days. The competent authority may ask the expert panel for clarifications on its opinion.
4. The competent authority shall give utmost consideration to the expert panel opinion. Where the competent authority adopts a measure on the classification, it shall notify the other Member States and the Commission of its measure without undue delay.

SECTION 2 CONFORMITY ASSESSMENT

Article 52

Conformity assessment procedures

1. Prior to placing a device on the market, manufacturers shall undertake an assessment of the conformity of that device, in accordance with the applicable conformity assessment procedures set out in Annexes IX to XI.

2. Prior to putting into service a device that is not placed on the market, manufacturers shall undertake an assessment of the conformity of that device, in accordance with the applicable conformity assessment procedures set out in Annexes IX to XI.

3. Manufacturers of class III devices, other than custom-made or investigational devices, shall be subject to a conformity assessment as specified in Annex IX. Alternatively, the manufacturer may choose to apply a conformity assessment as specified in Annex X coupled with a conformity assessment as specified in Annex XI.

By way of derogation from the first subparagraph, class III devices that are well-established technology devices shall be subject to a conformity assessment as specified in Chapters I and III of Annex IX, including an assessment of the technical documentation of one representative device per generic device group.

4. Manufacturers of class IIb devices, other than custom-made or investigational devices, shall be subject to a conformity assessment as specified in Chapters I and III of Annex IX, ~~and~~ including an assessment of the technical documentation ~~as specified in Section 4 of that Annex~~ of ~~at least~~ one representative device per generic device group ~~or, in the case of non-implantable class IIb devices that are well-established technology devices, one representative device per each category of devices.~~

~~However~~ By way of derogation from the first subparagraph, for class IIb implantable devices, except ~~well-established technology devices~~ [sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors], the assessment of the technical documentation as specified in Section 4 of Annex IX shall apply for every device. Alternatively, the manufacturer may choose to apply a conformity assessment based on type examination as specified in Annex X coupled with a conformity assessment based on product conformity verification as specified in Annex XI.

~~5. Where justified in view of well-established technologies, similar to those used in the exempted devices listed in the second subparagraph of paragraph 4 of this Article, being used in other class IIb implantable devices, or where justified in order to protect the health and safety of patients, users or other persons or other aspects of public health, the Commission is empowered to adopt delegated acts in accordance with Article 115 to amend that list by adding other types of class IIb implantable devices to that list or removing devices therefrom.~~

6. Manufacturers of class IIa devices, other than custom-made or investigational devices, shall be subject to a conformity assessment as specified in Chapters I and III of Annex IX, ~~and~~ including an assessment of the technical documentation ~~as specified in Section 4 of that Annex~~ of ~~at least~~ one representative device for each category of devices.

Alternatively, the manufacturer may choose to draw up the technical documentation set out in Annexes II and III coupled with a conformity assessment as specified in Section 10 or Section 18 of Annex XI. The assessment of the technical documentation shall apply for ~~at least~~ one representative device for each category of devices.

7. Manufacturers of class I devices, other than custom-made or investigational devices, shall declare the conformity of their products by issuing the EU declaration of conformity referred to in Article 19 after drawing up the technical documentation set out in Annexes II and III. If those devices are placed on the market in sterile condition, have a measuring function or are reusable surgical instruments, the

Proposal of Regulation (EU) 2017/745 Articles

manufacturer shall apply the procedures set out in Chapters I and III of Annex IX, or in [Section 10a](#) [Part A] of Annex XI. However, the involvement of the notified body in those procedures shall be limited:

- (a) in the case of devices placed on the market in sterile condition, to the aspects relating to establishing, securing and maintaining sterile conditions;
- (b) in the case of devices with a measuring function, to the aspects relating to the conformity of the devices with the metrological requirements;
- (c) in the case of reusable surgical instruments, to the aspects relating to the reuse of the device, in particular cleaning, disinfection, sterilization, maintenance and functional testing and the related instructions for use.

Where the manufacturer of class I reusable surgical instruments has applied harmonised standards or CS covering all relevant aspects referred to in the first subparagraph, point (c), the involvement of a notified body is not required.

8. Manufacturers of custom-made devices shall follow the procedure set out in Annex XIII and draw up the statement set out in Section 1 of that Annex before placing such devices on the market.

In addition to the procedure applicable pursuant to the first subparagraph, manufacturers of class III custom-made implantable devices shall be subject to the conformity assessment as specified in Chapter I of Annex IX. Alternatively, the manufacturer may choose to apply a conformity assessment as specified in Part A of Annex XI.

9. In addition to the procedures applicable pursuant to paragraph 3, 4, 6, or 7 of this Article, in the case of devices referred to in the first subparagraph of Article 1(8), the procedure specified in Section 5.2 of Annex IX or Section 6 of Annex X, as applicable, shall also apply.

10. In addition to the procedures applicable pursuant to paragraph 3, 4, 6, or 7 of this Article, in the case of devices that are covered by this Regulation in accordance with point (f) or (g) of Article 1(6) and with the first subparagraph of Article 1(10), the procedure specified in Section 5.3 of Annex IX or Section 6 of Annex X, as applicable, shall also apply.

11. In addition to the procedures applicable pursuant to paragraph 3, 4, 6, or 7, in the case of devices that are composed of substances or of combinations of substances that are intended to be introduced into the human body via a body orifice or applied to the skin and that are absorbed by or locally dispersed in the human body, the procedure specified in Section 5.4 of Annex IX or Section 6 of Annex X, as applicable, shall also apply.

~~12. The Member State in which the notified body is established may require that all or certain documents, including the technical documentation, audit, assessment and inspection reports, relating to the procedures referred to in paragraphs 1 to 7 and 9 to 11 be made available in an official Union language(s) determined by that Member State. In the absence of such requirement, those documents shall be available in any official Union language acceptable to the notified body.~~

~~The documents relating to the procedures referred to in paragraphs 1 to 7 shall be available in any official Union language acceptable to the notified body.~~

13. Investigational devices shall be subject to the requirements set out in Articles 62 to 81.

14. The Commission may, by means of implementing acts, specify detailed arrangements and procedural aspects ~~with a view to ensuring the harmonised application of the conformity assessment procedures by the notified bodies~~ for any of the following aspects:

- ~~1. the basis for the selection of the representative device for the assessment of the technical documentation as referred to in paragraphs 3, 4 and 6;~~
- ~~2. the modalities of unannounced on-site audits and sample tests to be conducted by notified bodies in accordance with Section 3.4 of Annex IX, taking into account the risk-class and the type of device,~~

Proposal of Regulation (EU) 2017/745 Articles

3. the physical, laboratory or other tests to be carried out by notified bodies in the context of sample tests, assessment of the technical documentation and type examination;

the modalities of the conformity assessment procedures regarding breakthrough devices and orphan devices set out in Article 52a;

~~(a) the frequency and the sampling basis of the assessment of the technical documentation on a representative basis as set out in the third paragraph of Section 2.3 and in Section 3.5 of Annex IX in the case of class IIa and class IIb devices, and in Section 10.2 of Annex XI in the case of class IIa devices;~~

~~(b) the minimum frequency of unannounced on-site audits and sample tests to be conducted by notified bodies in accordance with Section 3.4 of Annex IX, taking into account the risk class and the type of device;~~

~~(c) the physical, laboratory or other tests to be carried out by notified bodies in the context of sample tests, assessment of the technical documentation and type examination in accordance with Sections 3.4 and 4.3 of Annex IX, Section 3 of Annex X and Section 15 of Annex XI.~~

The implementing acts referred to in the first subparagraph shall be adopted in accordance with the examination procedure referred to in Article 114(3).

15. The Commission is empowered to adopt delegated acts in accordance with Article 115 to amend Annexes IX, X, XI and XIII in order to adapt to technical or scientific progress or to developments regarding conformity assessment of devices at international level and to take into account the needs of particular devices in view of their special characteristics.

Article 52a

Conformity assessment of breakthrough devices and of orphan devices

1. For the conformity assessment of breakthrough devices and orphan devices, for which a notified body is involved in the conformity assessment, the procedures laid down in Article 52 shall apply subject to the specific arrangements set out in this Article.

2. A device shall be considered a breakthrough device if it meets the following criteria:

(a) it is expected to introduce in the Union a high degree of novelty with respect to the device technology, related clinical procedure or the application of the device in clinical practice;

(b) it is expected to provide a significant positive clinical impact on patients or public health, for a life-threatening or irreversibly debilitating disease or condition, by either of the following:

(b) (i) offering a significant positive clinical or health impact compared to available alternatives and the state of the art;

(b) (ii) fulfilling an unmet medical need where there is an absence or insufficiency of available alternative options for that purpose.

3. A device shall be considered an orphan device if it meets the following criteria:

(a) it is intended for the treatment, diagnosis, or prevention of a disease or condition that presents in not more than 12 000 individuals in the Union per year;

(b) at least one of the following criteria is met:

(b) (i) there are insufficient available alternatives;

(b) (ii) the device is expected to provide a clinical benefit compared to available alternatives or the state of the art, taking into account both device-specific factors and patient population-specific factors.

4. Upon a duly substantiated request by a manufacturer or a notified body, an expert panel referred to in Article 106 shall provide an opinion as to whether the criteria set out in paragraph 2 or 3 of this Article, as applicable, are fulfilled. That opinion shall be published on a dedicated website without disclosing any

Proposal of Regulation (EU) 2017/745 Articles

confidential information as referred to in Article 109 and shall be duly taken into consideration by the manufacturer and the notified body.

5. Where the opinion of the expert panel confirms the fulfilment of the criteria set out in paragraph 2 or 3 of this Article, the manufacturer of a breakthrough device or of an orphan device, as applicable, may request advice from the expert panels referred to in Article 106 regarding its clinical development strategy and appropriate preclinical or clinical data for the clinical evaluation of the device.

6. For a confirmed breakthrough device or an orphan device, as applicable, the notified body involved in the conformity assessment procedure set out in Article 52 shall prioritise the conformity assessment of that device and apply, where appropriate, a rolling review with a view to reduce assessment timelines.

The notified body shall give due consideration to an opinion or advice provided by the expert panels in accordance with paragraph 4 or 5 and, where it does not follow such opinion or advice, it shall provide duly justified reasons. The notified body may ask the expert panel to clarify the opinion it has provided.

7. The notified body shall issue a certificate pursuant to Article 56 where the pre- market clinical evidence, even if based on limited clinical data, is deemed adequate, provided that either of the following conditions is fulfilled:

(a) the benefit of the immediate availability on the market of the device outweighs the risk associated with the fact that additional clinical data are still required;

(b) the benefit-risk-ratio of the device is favourable and the manufacturer commits to providing additional data from post-market clinical follow-up activities.

Where appropriate, the notified body shall limit the validity of the certificate and specify any conditions for or limitations to the certificate's validity in accordance with Article 56, such as a requirement for the manufacturer to conduct specific post-market clinical follow-up activities within a specified period of time.

8. The Commission is empowered to adopt delegated acts in accordance with Article 115 in order to amend this Article to adapt to technical or scientific progress or to take into account developments regarding conformity assessment of breakthrough devices or orphan devices at international level.

9. The Commission may, by means of implementing acts, lay down further details of the procedure for the conformity assessment of breakthrough devices or orphan devices set out in this Article. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

Article 52b

Digitalisation of technical documentation, conformity assessment and reports

1. The manufacturer may draw up and make available in a digital format the technical documentation and any reports or other documents required pursuant to this Regulation. That digital format may be a structured machine-readable format, provided that it is possible to transform it into human-readable format, and that there is version-control to enable the conduct of retrospective conformity checks. Where the technical documentation, reports or other documents are to be submitted to and assessed by a notified body, the manufacturer shall agree with the notified body on the digital format.

2. Where necessary to ensure that the digital format referred to in paragraph 1 is reliable, interoperable and standardised, the Commission may establish minimum requirements or functional specifications for the digital format by means of common specifications as referred to in Article 9.

Article 53

Involvement of notified bodies in conformity assessment procedures

1. Where the conformity assessment procedure requires the involvement of a notified body, the manufacturer may apply to a notified body of its choice, provided that the chosen notified body is designated for conformity assessment activities related to the types of devices concerned. The

Proposal of Regulation (EU) 2017/745 Articles

manufacturer may not lodge an application in parallel with another notified body for the same conformity assessment procedure.

2. The notified body concerned shall, by means of the electronic system referred to in Article 57, inform the other notified bodies of any manufacturer that withdraws its application prior to the notified body's decision regarding the conformity assessment.

3. When applying to a notified body under paragraph 1, manufacturers shall declare whether they have withdrawn an application with another notified body prior to the decision of that notified body and provide information about any previous application for the same conformity assessment that has been refused by another notified body.

4. The notified body may require any information or data from the manufacturer which is necessary in order to properly conduct the chosen conformity assessment procedure.

5. Notified bodies and the personnel of notified bodies shall carry out their conformity assessment activities **in the public interest and** with the highest degree of professional integrity and the requisite technical and scientific competence in the specific field and shall be free from all pressures and inducements, particularly financial, which might influence their judgement or the results of their conformity assessment activities, especially as regards persons or groups with an interest in the results of those activities.

Article 54

*Clinical evaluation consultation procedure for certain **high-risk** [class III and class IIb] devices*

1. In addition to the procedures applicable pursuant to Article 52, a notified body shall ~~also~~ follow the procedure regarding clinical evaluation consultation as specified in Section 5.1 of Annex IX or as referred to in Section 6 of Annex X, as applicable, when performing a conformity assessment of class III implantable devices, **other than custom-made devices** [the following devices].

Where justified in order to protect the health and safety of patients, users or other persons or other aspects of public health, the Commission is empowered to adopt delegated acts in accordance with Article 115 to amend paragraph 1 in order to add other types of devices to those devices that are subject to the clinical evaluation consultation procedure referred to in the first subparagraph.

~~(a)~~ class III implantable devices, and

~~(b) class IIb active devices intended to administer and/or remove a medicinal product, as referred to in Section 6.4 of Annex VIII (Rule 12).~~

2. **By way of derogation from paragraph 1, the procedure referred to in that paragraph is not required in the following cases:** [The procedure referred to in paragraph 1 shall not be required for the devices referred to therein:]

(a) **for a** [in the case of] in the case of renewal of a certificate issued under this Regulation;

(b) where the device has been designed by modifying a device already marketed ~~by the same manufacturer~~ for the same intended purpose, provided that the manufacturer has demonstrated to the satisfaction of the notified body that the modifications do not adversely affect the benefit-risk ratio of the device; ~~or~~

(c) where the principles of the clinical evaluation of the device type or category have been addressed in a **harmonised standard referred to in Article 8 or in CS** referred to in Article 9 and the notified body confirms that the clinical evaluation of the manufacturer for this device is in compliance with the relevant **harmonised standard or CS** for clinical evaluation of that kind of device.

3. The notified body shall notify the competent authorities, the authority responsible for notified bodies and the Commission through the electronic system referred to in Article 57 of whether or not the procedure referred to in paragraph 1 of this Article is to be applied. That notification shall be accompanied by the clinical evaluation assessment report.

4. The notified body shall give utmost consideration to the opinion issued by the expert panel in the framework of the clinical evaluation consultation procedure. Where the notified body has not followed the views and recommendations expressed in that opinion, it shall provide a substantiated justification of the reasons therefore and its final clinical evaluation assessment report to the authority responsible for notified bodies of the Member State in which it is established, to the expert panel that issued the opinion and to the Commission.

~~The Commission shall draw up an annual overview of devices which have been subject to the procedure specified in Section 5.1 of Annex IX and referred to in Section 6 of Annex X. The annual overview shall include the notifications in accordance with paragraph 3 of this Article and point (e) of Section 5.1 of Annex IX and a listing of the cases where the notified body did not follow the advice from the expert panel. The Commission shall submit this overview to the European Parliament, to the Council and to the MDCG.~~

~~5. The Commission shall by 27 May 2025 draw up a report on the operation of this Article and submit it to the European Parliament and to the Council. The report shall take into account the annual overviews and any available relevant recommendations from the MDCG. On the basis of that report the Commission shall, if appropriate, make proposals for amendments to this Regulation.~~

Article 55

Mechanism for scrutiny of conformity assessments ~~of certain class III and class IIb devices~~

1. The MDCG or the Commission may, based on reasonable concerns, request advice from an expert panel in relation to the safety and performance of any device. For that purpose, the MDCG or the Commission may request the notified body that issued the certificate for the device in question to submit to the expert panel its clinical evaluation assessment report and any subsequent surveillance assessment reports regarding that device. The expert panel may request the notified body or the manufacturer to submit additional information needed for its assessment.

~~A notified body shall notify the competent authorities of certificates it has granted to devices for which the conformity assessment has been performed pursuant to Article 54(1). Such notification shall take place through the electronic system referred to in Article 57 and shall include the summary of safety and clinical performance pursuant to Article 32, the assessment report by the notified body, the instructions for use referred to in Section 23.4 of Annex I, and, where applicable, the scientific opinion of the expert panels referred to in Section 5.1 of Annex IX or Section 6 of Annex X, as applicable. In the case of divergent views between the notified body and the expert panels, a full justification shall also be included.~~

2. The MDCG or the Commission may, based on reasonable concerns, request advice from one or more expert laboratories, based on laboratory testing, in relation to the safety and performance of any device, provided that the device falls within the scope of designation of those expert laboratories. For that purpose, the MDCG or the Commission may request the notified body that issued the certificate for the device in question to submit to the expert laboratories its clinical evaluation assessment report and any subsequent surveillance assessment reports regarding that device. The expert laboratories may request the notified body or the manufacturer to submit samples of the device or any additional information needed for their assessment.

~~A competent authority and, where applicable, the Commission may, based on reasonable concerns apply further procedures in accordance with Article 44, 45, 46, 47 or 94 and, where deemed necessary, take appropriate measures in accordance with Articles 95 and 97.~~

3. The notified body shall give utmost consideration to the advice provided by the expert panel or the expert laboratory and, where needed, take any appropriate measures, including those referred to in Article 56(3) and (4).

~~The MDCG and, where applicable, the Commission, may, based on reasonable concerns, request scientific advice from the expert panels in relation to the safety and performance of any device.~~

Article 56
Certificates of conformity

1. The notified bodies shall issue certificates [The certificates issued by the notified bodies] in accordance with Annexes IX, X and XI ~~shall be~~ in an official Union language ~~and immediately upload them in Eudamed determined by the Member State in which the notified body is established or otherwise in an official Union language acceptable to the notified body~~. The minimum content of the certificates shall be as set out in Annex XII..

2. The validity of certificates shall not be limited in time, unless in exceptional cases where the notified body considers it necessary to limit the period of validity based on duly justified grounds. In those cases, the notified body shall indicate the period of validity on the certificate. If the period of validity of the certificate is limited, on application by the manufacturer, the notified body may, following an assessment performed in accordance with Annex VII, Section 4.11, extend the validity of the certificate. [The certificates shall be valid for the period they indicate, which shall not exceed five years. On application by the manufacturer, the validity of the certificate may be extended for further periods, each not exceeding five years, based on a re-assessment in accordance with the applicable conformity assessment procedures.] Any supplement to a certificate shall remain valid as long as the certificate which it supplements is valid.

(a) During the validity of the certificate, the notified body shall carry out appropriate surveillance activities, including periodic reviews taking into consideration developments of the state of the art. Those reviews shall be proportionate to the risk class of the device.

3. Notified bodies may impose conditions [restrictions] on the validity of the certificate, such as limiting to the intended purpose of a device to certain groups of patients or requiring [require] the manufacturers to undertake specific PMCF studies pursuant to Part B of Annex XIV.

4. Where a notified body finds that the requirements of this Regulation are no longer met by the manufacturer, it shall, taking account of the principle of proportionality, suspend or withdraw the certificate issued or impose any conditions [restrictions] on it unless compliance with such requirements is ensured by appropriate corrective action taken by the manufacturer within an appropriate deadline set by the notified body. The notified body shall give the reasons for its decision.

5. The notified body shall enter in the electronic system referred to in Article 57 any information regarding certificates issued, including amendments and supplements thereto, and regarding suspended, reinstated, withdrawn or refused certificates and conditions [restrictions] imposed on certificates. Such information shall be accessible to the public.

6. In the light of technical progress, the Commission is empowered to adopt delegated acts in accordance with Article 115 amending the minimum content of the certificates set out in Annex XII.

Article 57
Electronic system on notified bodies and on certificates of conformity

1. The Commission, after consulting the MDCG, shall set up and manage an electronic system to collate and process the following information:

- (a) the list of subsidiaries referred to in Article 37(3);
- (b) the list of experts referred to in Article 40(2);
- (c) the information relating to the notification referred to in Article 42(10) and the amended notifications referred to in Article 46(2);
- (d) the list of notified bodies referred to in Article 43(2);
- (e) the summary of the report referred to in Article 44(12);

Proposal of Regulation (EU) 2017/745 Articles

(f) the notifications for conformity assessments and [information or documents](#) [\[certificates\]](#) referred to in Article 54(3) and [\(3a\)](#) [\[55\(1\)\]](#);

(g) withdrawal or refusals of applications for the certificates as referred to in Article 53(2) and Section 4.3 of Annex VII;

(h) the information regarding certificates referred to in Article 56(5);

~~(i) the summary of safety and clinical performance referred to in Article 32.~~

2. The information collated and processed by the electronic system shall be accessible to the competent authorities of the Member States, to the Commission, where appropriate to the notified bodies and where provided elsewhere in this regulation or in Regulation (EU) 2017/746 to the public.

Article 58

Voluntary change of notified body

1. In cases where a manufacturer terminates its contract with a notified body and enters into a contract with another notified body in respect of the conformity assessment of the same device, the detailed arrangements for the change of notified body shall be clearly defined in an agreement between the manufacturer, the incoming notified body and, where practicable the outgoing notified body. That agreement shall cover at least the following aspects:

(a) the date on which the certificates issued by the outgoing notified body become invalid;

(b) the date until which the identification number of the outgoing notified body may be indicated in the information supplied by the manufacturer, including any promotional material;

(c) the transfer of documents, including confidentiality aspects and property rights;

(d) the date after which the conformity assessment tasks of the outgoing notified body is assigned to the incoming notified body;

(e) the last serial number or lot number for which the outgoing notified body is responsible.

2. The outgoing notified body shall withdraw the certificates it has issued for the device concerned on the date on which they become invalid.

Article 59

Derogation from the conformity assessment procedures

1. ~~By way of derogation from Article 52 of this Regulation or, for the period from 24 April 2020 to 25 May 2021, by way of derogation from Article 9(1) and (2) of Directive 90/385/EEC or from Article 11(1) to (6) of Directive 93/42/EEC and~~ on a duly justified request, any competent authority may authorise ~~for a limited period of time~~ the placing on the market or putting into service within the territory of the Member State concerned of a specific device for which the applicable ~~conformity assessment~~ procedures ~~referred to in these Articles~~ have not been carried out, ~~provided that~~ [\[but\]](#) the use of ~~that device~~ [\[which\]](#) is in the interest of public health, patient safety or patient health.

~~(a) By way of derogation from Article 6(2) and on a duly justified request, any competent authority may authorise for a limited period of time the provision of a diagnostic or therapeutic service referred to in that Article to a natural or legal person established within the territory of the Member State concerned using a device for which the applicable conformity assessment procedures set out in this Regulation have not been carried out, provided that the provision of that service is in the interest of public health, patient safety or patient health.~~

2. The Member State shall inform the Commission, ~~and~~ the other Member States ~~and the relevant expert panels referred to in Article 106~~ of any decision to authorise the placing on the market or putting into service of a device, ~~or the provision of a service~~, in accordance with paragraph 1 ~~or 1a~~, where such authorisation is granted for use other than for a single patient. The Member State ~~shall also make~~

Proposal of Regulation (EU) 2017/745 Articles

information about such authorisations publicly available. [may inform the Commission and the other Member States of any authorisation granted in accordance with Article 9(9) of Directive 90/385/EEC or Article 11(13) of Directive 93/42/EEC before 24 April 2020.]

3. Where a request pursuant to paragraph 1 or paragraph 1a has been submitted to competent authorities in more than one Member State and based on an opinion of an expert panel referred to in Article 106 [Following a notification pursuant to paragraph 2 of this Article], the Commission, in exceptional cases relating to public health, patient safety or patient health, may, by means of implementing acts, extend for a limited period of time the validity of an authorisation granted by a Member State in accordance with paragraph 1 or paragraph 1a to the territory of the Union, or provide an authorisation referred to in paragraph 1 or paragraph 1a for [of this Article or, when granted before 24 April 2020, in accordance with Article 9(9) of Directive 90/385/EEC or Article 11(13) of Directive 93/42/EEC to] the territory of the Union. The Commission may set out the conditions under which the device may be placed on the market or put into service, or under which the diagnostic or therapeutic service may be provided. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

4. In the event of a public health emergency at Union level recognised in accordance with Article 23 of Regulation (EU) 2022/2371 of the European Parliament and of the Council****, the Commission may, by means of implementing acts, on its own initiative after consulting the MDCG, authorise the placing on the market or putting into service of a device in accordance with paragraph 3. The authorisation shall cease to apply at the latest when the recognition of the public health emergency is terminated pursuant to Article 23(2) of Regulation (EU) 2022/2371. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

On duly justified imperative grounds of urgency relating to the health and safety of humans, the Commission may adopt immediately applicable implementing acts in accordance with the procedure referred to in Article 114(4).

5. The Commission may, by means of implementing acts, lay down rules further specifying the procedure set out in this Article. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

**** Regulation (EU) 2022/2371 of the European Parliament and of the Council of 23 November 2022 on serious cross-border threats to health and repealing Decision No 1082/2013/EU (OJ L 314, 6.12.2022, p. 26, ELI: <http://data.europa.eu/eli/reg/2022/2371/oj>).

On duly justified imperative grounds of urgency relating to the health and safety of humans, the Commission shall adopt immediately applicable implementing acts in accordance with the procedure referred to in Article 114(4).

Article 59a

Derogations from certain requirements in the event of a serious cross-border threat to health, disaster or crisis

1. Upon a duly justified request by the manufacturer, a competent authority may authorise for a limited period of time, by way of derogation from the relevant provisions in Annexes II, III, IX, X and XI, an exemption from the requirements related to changes to the manufacturing, design or intended purpose of a CE marked device, where it is in the interest of public health, patient safety or patient health, in either of the following circumstances:

- (a) a serious cross-border threat to health as defined in Article 3, point (1), of Regulation (EU) 2022/2371;
- (b) a disaster or a crisis within the meaning of Regulation (EU) .../... of the European Parliament and of the Council*****+.

2. The manufacturer shall ensure that the manufactured devices remain in conformity with the relevant general safety and performance requirements set out in Annex I.

Proposal of Regulation (EU) 2017/745 Articles

3. The competent authority may request the notified body that issued a certificate for the device in question, to assist it in the assessment of a request referred to in paragraph 1.

4. Where applicable, the manufacturer shall keep the notified body that issued a certificate for the device in question, informed about any changes made regarding the manufacturing, design or intended purpose of a CE marked device in accordance with the authorisation referred to in paragraph 1.

5. Where a request pursuant to paragraph 1 has been submitted to competent authorities in more than one Member State, the Commission, in exceptional cases relating to public health, patient safety or patient health, may, by means of implementing acts, extend for a limited period of time the validity of an exemption granted by a Member State in accordance with paragraph 1 to the territory of the Union, or provide an exemption referred to in paragraph 1 for the territory of the Union. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

OJ: Please insert in the text the number of the Regulation contained in 2025/0223(COD) (Proposal for a Regulation on the Union Civil Protection Mechanism and Union support for health emergency preparedness and response, and repealing Decision No 1313/2013/EU) and insert the number, date, title and OJ reference of that Regulation in the footnote.

On duly justified imperative grounds of urgency relating to the health and safety of humans, the Commission may adopt immediately applicable implementing acts in accordance with the procedure referred to in Article 114(4).

***** Regulation (EU) .../... of the European Parliament and of the Council of ... on ... (OJ ..., ELI).

Article 59b

Regulatory sandboxes at national level

1. Member States, on their own initiative or upon a substantiated request by a manufacturer or a prospective manufacturer, may establish one or more regulatory sandboxes to which the application of certain requirements of Chapters V or VI or of Annexes I, VIII IX, X, XI, XIV or XV would not be appropriate. The Member States shall designate the competent authority that is responsible for the supervision of the regulatory sandbox.

Member States may also establish regulatory sandboxes jointly with other Member States.

2. The activities within a regulatory sandbox shall take place pursuant to a specific sandbox plan that shall clearly identify the requirements of this Regulation referred to in paragraph 1, which, by way of derogation from this Regulation, are temporarily adapted or waived in the regulatory sandbox, a justification that the application of those requirements is considered not appropriate and an explanation as to how potential risks related to the adaptation or waiver are controlled and mitigated. The plan shall also identify the reasonable duration of the regulatory sandbox necessary to achieve its objectives and the participants in the regulatory sandbox and their respective roles.

3. A regulatory sandbox shall be set up only if the following conditions are met:

- (a) the device is expected to address unmet medical needs or to provide a significant clinical benefit to patients or to the health system compared with similar existing alternatives or the state of the art;
- (b) the application of the requirements of this Regulation referred to in paragraph 1 would impede or significantly delay the development of the device and access by healthcare professionals or lay users to such device.

4. The Member State may request an expert panel referred to in Article 106 to provide scientific, technical or regulatory advice on the draft sandbox plan.

5. Any participant in the regulatory sandbox shall, without undue delay, inform the competent authority that is responsible for the supervision of the regulatory sandbox about any harm occurred in relation to the implementation of the regulatory sandbox. The competent authority shall take immediate and adequate corrective measures, including to suspend, revoke or restrict the scope of the regulatory sandbox.

Proposal of Regulation (EU) 2017/745 Articles

6. Manufacturers and prospective manufacturers participating in a regulatory sandbox shall remain liable under applicable Union and national law for any damage inflicted on third parties as a result of their activities taking place in the regulatory sandbox.

7. The Member State shall inform the Commission and the MDCG about the establishment of a regulatory sandbox and keep them informed about its implementation and outcome.

Article 59c

Union regulatory sandboxes

1. The Commission, on its own initiative or upon a substantiated request by a Member State, may establish, by means of implementing acts for a limited time and pursuant to a specific plan, Union regulatory sandboxes, which shall inform whether the existing requirements appropriately regulate a specific type of device with particular characteristics or emerging technologies, and there is a risk that the existing requirements:

(a) would impede or significantly delay the development of such devices and access by healthcare professionals or lay users to those devices; or

(b) would not adequately protect the health and safety of patients, users or other persons or other aspects of public health.

Union regulatory sandboxes shall not involve the placing on the market or putting into service of devices which do not comply with this Regulation.

2. The Commission shall request an expert panel as referred to in Article 106 to provide scientific, technical or regulatory advice on the design of a Union regulatory sandbox.

3. The Commission shall inform the MDCG about the establishment of a regulatory sandbox and keep it informed about its outcome.

4. The Commission may, by means of implementing acts, specify common principles or the detailed arrangements for the establishment, operation and supervision of regulatory sandboxes pursuant to Article 59b or of Union regulatory sandboxes pursuant to this Article. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

5. The Commission is empowered to adopt delegated acts in accordance with Article 115 to amend this Article or Article 59b in order to adapt it to scientific, technical or regulatory progress and to take into account developments regarding regulatory sandboxes, including in areas other than medical devices.

Article 60

Certificate of free sale

1. For the purpose of export and upon request by a manufacturer or an authorised representative, the Member State in which the manufacturer or the authorised representative has its registered place of business shall issue a certificate of free sale declaring that the manufacturer or the authorised representative, as applicable, has its registered place of business on its territory and that the device in question bearing the CE marking in accordance with this Regulation may be marketed in the Union. The certificate of free sale shall set out the Basic UDI-DI of the device as provided to the UDI database under Article 29. Where a notified body has issued a certificate pursuant to Article 56, the certificate of free sale shall set out the unique number identifying the certificate issued by the notified body, as referred to in Section 3 of Chapter II of Annex XII.

(a) The person referred to in Article 22(1) or (3) may request the competent authority of the Member State where it is established to issue a certificate of free sale for a system or procedure pack for which it has drawn up a statement in accordance with Article 22.

(b) The competent authority shall make the certificates of free sale issued in accordance with paragraphs 1 and 1a publicly available in Eudamed.

Proposal of Regulation (EU) 2017/745 Articles

2. The Commission may, by means of implementing acts, establish a model for certificates of free sale, taking into account international practice as regards the use of certificates of free sale. Those implementing acts shall be adopted in accordance with the advisory procedure referred to in Article 114(2).

Chapter VI Clinical evaluation and clinical investigations

Article 61 *Clinical evaluation*

1. ~~To that end~~ Manufacturers shall plan, conduct and document a clinical evaluation in accordance with this Article and with Part A of Annex XIV **to confirm the safety and performance of the device** [Confirmation of conformity with relevant general safety and performance requirements set out in Annex I] under ~~the~~ normal conditions of **use in accordance with the intended purpose of the device** [the intended use of the device], and **shall evaluate any** [the evaluation of the] undesirable side-effects and ~~of the~~ acceptability of the benefit-risk ratio referred to in Sections 1 and 8 of Annex I, ~~shall be based on clinical data providing sufficient clinical evidence, including where applicable relevant data as referred to in Annex III.~~

The manufacturer shall specify and justify the level of clinical evidence necessary to **confirm the safety and performance of the device** [demonstrate conformity with the relevant general safety and performance requirements]. That level of clinical evidence shall be appropriate in view of the characteristics of the device and its intended purpose, **taking into consideration paragraph 10.**

~~The clinical evaluation, its results and the clinical evidence derived from it shall be documented in a clinical evaluation report as referred to in Section 4 of Annex XIV, which, except for custom-made devices, shall be part of the technical documentation referred to in Annex II relating to the device concerned.~~

2. For **class IIb and class III devices** [all class III devices and for the class IIb devices] ~~referred to in point (b) of Article 54(1), a~~ [the] manufacturer may, prior to its clinical investigation **and/or** clinical evaluation, consult an expert panel as referred to in Article 106, with the aim of reviewing the manufacturer's intended clinical development strategy **or** [and] proposals for clinical investigation. The manufacturer **and the** notified body involved in any future conformity assessment procedure shall, in the clinical evaluation report ~~referred to in paragraph 12 of this Article and the clinical evaluation assessment report,~~ give due consideration to the **advice of** [views expressed by] the expert panel ~~Such consideration shall be documented and where they do not follow the advice, they shall provide duly justified reasons.~~

~~The manufacturer may not invoke any rights to the views expressed by the expert panel with regard to any future conformity assessment procedure.~~

3. A clinical evaluation shall follow a defined and methodologically sound procedure based on the following:

(a) a critical evaluation of the relevant scientific literature currently available relating to the safety, performance, design characteristics and intended purpose of the device, where the following conditions are satisfied:

– it is demonstrated that the device subject to clinical evaluation for the intended purpose is equivalent to the device to which the data relate, in accordance with Section 3 of Annex XIV, and

– the data adequately demonstrate compliance with the relevant general safety and performance requirements;

(b) a critical evaluation of the results of all available clinical investigations, taking duly into consideration whether the investigations were performed under Articles 62 to 80, any acts adopted pursuant to Article 81, and Annex XV; and

(c) a consideration of currently available alternative treatment options for that purpose, if any.

Proposal of Regulation (EU) 2017/745 Articles

4. In the case of implantable **class IIb** devices and class III devices, **other than custom-made devices**, clinical investigations shall be performed, except if:

- the device has been designed by modifications of a device already marketed by the same manufacturer,
- the modified device has been demonstrated by the manufacturer to be equivalent to the marketed device, in accordance with Section 3 of Annex XIV and this demonstration has been endorsed by the notified body, and
- the clinical evaluation of the marketed device is sufficient to demonstrate conformity of the modified device with the relevant safety and performance requirements.

In this case the notified body shall check that the PMCF plan is appropriate and includes post market studies to demonstrate the safety and performance of the device.

In addition, clinical investigations need not be performed in the cases referred to in paragraph 6.

5. A manufacturer of a device demonstrated to be equivalent to an already marketed device not manufactured by it [him], may also rely on paragraph 4 in order not to perform a clinical investigation provided that the ~~following conditions are fulfilled in addition to what is required in that paragraph: the two manufacturers have a contract in place that explicitly allows the manufacturer of the second device full access to the technical documentation on an ongoing basis, and~~ original clinical evaluation has been performed in compliance with the requirements of this Regulation and the manufacturer ~~of the second device~~ provides clear evidence thereof to the notified body.

6. The requirement to perform clinical investigations pursuant to paragraph 4 shall not apply to implantable **class IIb** devices and class III devices:

(a) which have been lawfully placed on the market or put into service in accordance with Directive 90/385/EEC or Directive 93/42/EEC and for which the clinical evaluation:

- is based on sufficient clinical data, and
- is in compliance with the relevant product-specific CS for the clinical evaluation of that kind of device, where such a CS is available; or

(b) that are **well-established technology devices** [sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips or connectors] for which the clinical evaluation is based on sufficient clinical **evidence** [data] and is in compliance with the relevant product-specific CS, where such ~~a~~ CS **are** [is] available.

7. Cases in which paragraph 4 is not applied by virtue of paragraph 6 shall be justified in the clinical evaluation report by the manufacturer and in the clinical evaluation assessment report by the notified body.

~~8. Where justified in view of well-established technologies, similar to those used in the exempted devices listed in point (b) of paragraph 6 of this Article, being used in other devices, or where justified in order to protect the health and safety of patients, users or other persons or other aspects of public health, the Commission is empowered to adopt delegated acts in accordance with Article 115 to amend the list of exempted devices referred to in the second subparagraph of Article 52(4) and in point (b) of paragraph 6 of this Article, by adding other types of implantable or class III devices to that list or removing devices therefrom.~~

9. In the case of the products without an intended medical purpose listed in Annex XVI, the requirement to demonstrate a clinical benefit in accordance with this Chapter and Annexes XIV and XV shall be understood as a requirement to demonstrate the performance of the device. Clinical evaluations of those products shall be based on relevant data concerning safety, including data from post-market surveillance, PMCF, and, where applicable, specific clinical investigation. Clinical investigations shall be performed for those products unless reliance on existing clinical data from an analogous medical device is duly justified.

Proposal of Regulation (EU) 2017/745 Articles

10. Without prejudice to paragraph 4, where the **confirmation of [demonstration of conformity with] general safety and performance requirements** based on clinical data is not deemed appropriate, adequate justification **for any such exception** shall be given based on the results of the manufacturer's risk management and on consideration of the specifics of the interaction between the device and the human body, the clinical performance intended and the claims of the manufacturer **and the data available for the generic device group**. In such a case, the manufacturer shall duly substantiate in the technical documentation referred to in Annex II why it considers a demonstration of conformity with general safety and performance requirements that is based on the results of non-clinical testing methods alone, including performance evaluation, bench testing, **in vitro, ex vivo, in silico testing, computational modeling or simulation** and pre-clinical evaluation, to be adequate.

11. The clinical evaluation, **and** its documentation **and, where applicable and needed, the summary of safety and performance referred to in Article 32** shall be updated throughout the life cycle of the device concerned with **clinical data and findings** obtained from the implementation of the manufacturer's PMCF plan in accordance with Part B of Annex XIV and the post-market surveillance plan referred to in Article 84, **whenever those data and findings obtained from PMCF provide information relevant for the confirmation of safety and performance of the device.**

~~For class III devices and implantable devices, the PMCF evaluation report and, if indicated, the summary of safety and clinical performance referred to in Article 32 shall be updated at least annually with such data.~~

~~12. The clinical evaluation, its results and the clinical evidence derived from it shall be documented in a clinical evaluation report as referred to in Section 4 of Annex XIV, which, except for custom-made devices, shall be part of the technical documentation referred to in Annex II relating to the device concerned.~~

13. Where necessary to ensure the uniform application of Annex XIV, the Commission may, having due regard to technical and scientific progress, adopt implementing acts to the extent necessary to resolve issues of divergent interpretation and of practical application. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

14. The Commission is empowered to adopt delegated acts in accordance with Article 115 to amend Annex XIV in order to adapt it to technical and scientific progress and developments at international level, having due regard to the protection of the health and safety of patients, users or other persons and other aspects of public health.

Article 62

General requirements regarding clinical investigations conducted to demonstrate conformity of devices

1. Clinical investigations shall be designed, authorised, conducted, recorded and reported in accordance with the provisions of this Article, **and of Articles 63 to 73 and of Articles 75 to 80 [63 to 80]**, the acts adopted pursuant to Article 81, and Annex XV, where **they are** carried out **to assess an investigational device that has not been placed on the market or put into service in accordance with this Regulation and where carried out [as part of the clinical evaluation for conformity assessment purposes]** for one or more of the following purposes:

(a) to establish and verify that, under normal conditions of use, a device is designed, manufactured and packaged in such a way that it is suitable for one or more of the specific purposes listed in point (1) of Article 2, and achieves the performance intended as specified by its manufacturer;

(b) to establish and verify the clinical benefits of a device as specified by its manufacturer;

(c) to establish and verify the clinical safety of the device and to determine any undesirable side-effects, under normal conditions of use of the device, and assess whether they constitute acceptable risks when weighed against the benefits to be achieved by the device.

2. Where the sponsor of a clinical investigation is not established in the Union, that sponsor shall ensure that a natural or legal person is established in the Union as its legal representative. Such legal

Proposal of Regulation (EU) 2017/745 Articles

representative shall be responsible for ensuring compliance with the sponsor's obligations pursuant to this Regulation, and shall be the addressee for all communications with the sponsor provided for in this Regulation. Any communication with that legal representative shall be deemed to be a communication with the sponsor.

Member States may choose not to apply the first subparagraph to clinical investigations to be conducted solely on their territory, or on their territory and the territory of a third country, provided that they ensure that the sponsor establishes at least a contact person on their territory in respect of that clinical investigation who shall be the addressee for all communications with the sponsor provided for in this Regulation.

3. Clinical investigations shall be designed and conducted in such a way that the rights, safety, dignity and well-being of the subjects participating in a clinical investigation are protected and prevail over all other interests and the clinical data generated are scientifically valid, reliable and robust.

Clinical investigations shall be subject to scientific and ethical review. The ethical review shall be performed by an ethics committee in accordance with national law. Member States shall ensure that the procedures for review by ethics committees are compatible with the procedures set out in this Regulation for the assessment of the application for authorisation of a clinical investigation. At least one lay person shall participate in the ethical review.

4. A clinical investigation as referred to in paragraph 1 may be conducted only where all of the following conditions are met:

(a) the clinical investigation is the subject of an authorisation by the Member State(s) in which the clinical investigation is to be conducted, in accordance with this Regulation, unless otherwise stated;

(b) an ethics committee, set up in accordance with national law, has not issued a negative opinion in relation to the clinical investigation, which is valid for that entire Member State under its national law;

(c) the sponsor, or its legal representative or a contact person pursuant to paragraph 2, is established in the Union;

(d) vulnerable populations and subjects are appropriately protected in accordance with Articles 64 to 68;

(e) the anticipated benefits to the subjects or to public health justify the foreseeable risks and inconveniences and compliance with this condition is constantly monitored;

(f) the subject or, where the subject is not able to give informed consent, his or her legally designated representative has given informed consent in accordance with Article 63;

(g) the subject or, where the subject is not able to give informed consent, his or her legally designated representative, has been provided with the contact details of an entity where further information can be received in case of need;

(h) the rights of the subject to physical and mental integrity, to privacy and to the protection of the data concerning him or her in accordance with Directive 95/46/EC are safeguarded;

(i) the clinical investigation has been designed to involve as little pain, discomfort, fear and any other foreseeable risk as possible for the subjects, and both the risk threshold and the degree of distress are specifically defined in the clinical investigation plan and constantly monitored;

(j) the medical care provided to the subjects is the responsibility of an appropriately qualified medical doctor or, where appropriate, a qualified dental practitioner or any other person entitled by national law to provide the relevant patient care under clinical investigation conditions;

(k) no undue influence, including that of a financial nature, is exerted on the subject, or, where applicable, on his or her legally designated representatives, to participate in the clinical investigation;

(l) the investigational device(s) in question conform(s) to the applicable general safety and performance requirements set out in Annex I apart from the aspects covered by the clinical investigation and that, with

Proposal of Regulation (EU) 2017/745 Articles

regard to those aspects, every precaution has been taken to protect the health and safety of the subjects. This includes, where appropriate, technical and biological safety testing and pre-clinical evaluation, as well as provisions in the field of occupational safety and accident prevention, taking into consideration the state of the art;

(m) the requirements of Annex XV are fulfilled.

5. Any subject, or, where the subject is not able to give informed consent, his or her legally designated representative, may, without any resulting detriment and without having to provide any justification, withdraw from the clinical investigation at any time by revoking his or her informed consent. Without prejudice to Directive 95/46/EC, the withdrawal of the informed consent shall not affect the activities already carried out and the use of data obtained based on informed consent before its withdrawal.

6. The investigator shall be a person exercising a profession which is recognised in the Member State concerned as qualifying for the role of investigator on account of having the necessary scientific knowledge and experience in patient care. Other personnel involved in conducting a clinical investigation shall be suitably qualified, by education, training or experience in the relevant medical field and in clinical research methodology, to perform their tasks.

7. The facilities where the clinical investigation is to be conducted shall be suitable for the clinical investigation and shall be similar to the facilities where the device is intended to be used.

Article 63 *Informed consent*

1. Informed consent shall be written, dated and signed by the person performing the interview referred to in point (c) of paragraph 2, and by the subject or, where the subject is not able to give informed consent, his or her legally designated representative after having been duly informed in accordance with paragraph 2. Where the subject is unable to write, consent may be given and recorded through appropriate alternative means in the presence of at least one impartial witness. In that case, the witness shall sign and date the informed consent document. The subject or, where the subject is not able to give informed consent, his or her legally designated representative shall be provided with a copy of the document or the record, as appropriate, by which informed consent has been given. The informed consent shall be documented. Adequate time shall be given for the subject or his or her legally designated representative to consider his or her decision to participate in the clinical investigation.

2. Information given to the subject or, where the subject is not able to give informed consent, his or her legally designated representative for the purposes of obtaining his or her informed consent shall:

(a) enable the subject or his or her legally designated representative to understand:

(i) the nature, objectives, benefits, implications, risks and inconveniences of the clinical investigations;

(ii) the subject's rights and guarantees regarding his or her protection, in particular his or her right to refuse to participate in and the right to withdraw from the clinical investigation at any time without any resulting detriment and without having to provide any justification;

(iii) the conditions under which the clinical investigations is to be conducted, including the expected duration of the subject's participation in the clinical investigation; and

(iv) the possible treatment alternatives, including the follow-up measures if the participation of the subject in the clinical investigation is discontinued;

(b) be kept comprehensive, concise, clear, relevant, and understandable to the subject or his or her legally designated representative;

(c) be provided in a prior interview with a member of the investigating team who is appropriately qualified under national law;

(d) include information about the applicable damage compensation system referred to in Article 69; and

Proposal of Regulation (EU) 2017/745 Articles

(e) include the Union-wide unique single identification number of the clinical investigation referred to in Article 70(1) and information about the availability of the clinical investigation results in accordance with paragraph 6 of this Article.

3. The information referred to in paragraph 2 shall be prepared in writing and be available to the subject or, where the subject is not able to give informed consent, his or her legally designated representative.

4. In the interview referred to in point (c) of paragraph 2, special attention shall be paid to the information needs of specific patient populations and of individual subjects, as well as to the methods used to give the information.

5. In the interview referred to in point (c) of paragraph 2, it shall be verified that the subject has understood the information.

6. The subject shall be informed that a clinical investigation report and a summary presented in terms understandable to the intended user will be made available pursuant to Article 77(5) in the electronic system on clinical investigations referred to in Article 73 irrespective of the outcome of the clinical investigation, and shall be informed, to the extent possible, when they have become available.

7. This Regulation is without prejudice to national law requiring that, in addition to the informed consent given by the legally designated representative, a minor who is capable of forming an opinion and assessing the information given to him or her, shall also assent in order to participate in a clinical investigation.

Article 64

Clinical investigations on incapacitated subjects

1. In the case of incapacitated subjects who have not given, or have not refused to give, informed consent before the onset of their incapacity, a clinical investigation may be conducted only where, in addition to the conditions set out in Article 62(4), all of the following conditions are met:

(a) the informed consent of their legally designated representative has been obtained;

(b) the incapacitated subjects have received the information referred to in Article 63(2) in a way that is adequate in view of their capacity to understand it;

(c) the explicit wish of an incapacitated subject who is capable of forming an opinion and assessing the information referred to in Article 63(2) to refuse participation in, or to withdraw from, the clinical investigation at any time, is respected by the investigator;

(d) no incentives or financial inducements are given to subjects or their legally designated representatives, except for compensation for expenses and loss of earnings directly related to the participation in the clinical investigation;

(e) the clinical investigation is essential with respect to incapacitated subjects and data of comparable validity cannot be obtained in clinical investigations on persons able to give informed consent, or by other research methods;

(f) the clinical investigation relates directly to a medical condition from which the subject suffers;

(g) there are scientific grounds for expecting that participation in the clinical investigation will produce **either of the following** [a direct benefit to the incapacitated subject outweighing the risks and burdens involved]:

(g) (i) a direct benefit to the incapacitated subject outweighing the risks and burdens involved;

(g) (ii) a benefit for the population represented by the incapacitated subject provided that the clinical investigation poses only minimal risk to, and imposes minimal burden on, the incapacitated subject in comparison with the standard treatment of the subject's condition.

2. The subject shall as far as possible take part in the informed consent procedure.

Article 65

Clinical investigations on minors

A clinical investigation on minors may be conducted only where, in addition to the conditions set out in Article 62(4), all of the following conditions are met:

- (a) the informed consent of their legally designated representative has been obtained;
- (b) the minors have received the information referred to in Article 63(2) in a way adapted to their age and mental maturity and from investigators or members of the investigating team who are trained or experienced in working with children;
- (c) the explicit wish of a minor who is capable of forming an opinion and assessing the information referred to in Article 63(2) to refuse participation in, or to withdraw from, the clinical investigation at any time, is respected by the investigator;
- (d) no incentives or financial inducements are given to the subject or his or her legally designated representative except for compensation for expenses and loss of earnings directly related to the participation in the clinical investigation;
- (e) the clinical investigation is intended to investigate treatments for a medical condition that only occurs in minors or the clinical investigation is essential with respect to minors to validate data obtained in clinical investigations on persons able to give informed consent or by other research methods;
- (f) the clinical investigation either relates directly to a medical condition from which the minor concerned suffers or is of such a nature that it can only be carried out on minors;
- (g) there are scientific grounds for expecting that participation in the clinical investigation will produce either of the following [a direct benefit to the minor subject outweighing the risks and burdens involved]:
 - (g) (i) a direct benefit to the minor subject outweighing the risks and burdens involved;
 - (g) (ii) a benefit for the population represented by the minor subject provided that the clinical investigation will pose only minimal risk to, and will impose minimal burden on, the minor subject in comparison with the standard treatment of the subject's condition.
- (h) the minor shall take part in the informed consent procedure in a way adapted to his or her age and mental maturity;
- (i) if during a clinical investigation the minor reaches the age of legal competence to give informed consent as defined in national law, his or her express informed consent shall be obtained before that subject can continue to participate in the clinical investigation.

Article 66

Clinical investigations on pregnant or breastfeeding women

A clinical investigation on pregnant or breastfeeding women may be conducted only where, in addition to the conditions set out in Article 62(4), all of the following conditions are met:

- (a) the clinical investigation has the potential to produce a direct benefit for the pregnant or breastfeeding woman concerned, or her embryo, foetus or child after birth, outweighing the risks and burdens involved, or if the clinical investigation has not the potential to produce such direct, the following conditions are met:
 - (i) a clinical investigation of comparable effectiveness cannot be carried out on women who are not pregnant or breastfeeding;
 - (ii) the clinical investigation contributes to the attainment of results capable of benefitting pregnant or breastfeeding women or other women in relation to reproduction, or other embryos, fetuses or children;

(iii) the clinical investigation poses a minimal risk to, and imposes a minimal burden on, the pregnant or breastfeeding woman concerned, her embryo, foetus or child after birth;

(b) where research is undertaken on breastfeeding women, particular care is taken to avoid any adverse impact on the health of the child;

(c) no incentives or financial inducements are given to the subject except for compensation for expenses and loss of earnings directly related to the participation in the clinical investigation.

Article 67

Additional national measures

Member States may maintain additional measures regarding persons performing mandatory military service, persons deprived of liberty, persons who, due to a judicial decision, cannot take part in clinical investigations, or persons in residential care institutions.

Article 68

Clinical investigations in emergency situations

1. By way of derogation from point (f) of Article 62(4), from points (a) and (b) of Article 64(1) and from points (a) and (b) of Article 65, informed consent to participate in a clinical investigation may be obtained, and information on the clinical investigation may be given, after the decision to include the subject in the clinical investigation, provided that that decision is taken at the time of the first intervention on the subject, in accordance with the clinical investigation plan for that clinical investigation and that all of the following conditions are fulfilled:

(a) due to the urgency of the situation, caused by a sudden life-threatening or other sudden serious medical condition, the subject is unable to provide prior informed consent and to receive prior information on the clinical investigation;

(b) there are scientific grounds to expect that participation of the subject in the clinical investigation will have the potential to produce ~~a direct clinically relevant benefit for the subject resulting in a measurable health-related improvement alleviating the suffering and/or improving the health of the subject, or in the diagnosis of its condition~~ either of the following:

(i) a direct clinically relevant benefit for the subject resulting in a measurable health-related improvement alleviating the suffering and/or improving the health of the subject, or in the diagnosis of its condition;

(ii) a benefit for the population represented by the subject provided that the clinical investigation will pose only minimal risk to, and will impose minimal burden on, the subject in comparison with the standard treatment of the subject's condition;

(c) it is not possible within the therapeutic window to supply all prior information to and obtain prior informed consent from his or her legally designated representative;

(d) the investigator certifies that he or she is not aware of any objections to participate in the clinical investigation previously expressed by the subject;

(e) the clinical investigation relates directly to the subject's medical condition because of which it is not possible within the therapeutic window to obtain prior informed consent from the subject or from his or her legally designated representative and to supply prior information, and the clinical investigation is of such a nature that it may be conducted exclusively in emergency situations;

(f) the clinical investigation poses a minimal risk to, and imposes a minimal burden on, the subject in comparison with the standard treatment of the subject's condition.

Proposal of Regulation (EU) 2017/745 Articles

2. Following an intervention pursuant to paragraph 1 of this Article, informed consent in accordance with Article 63 shall be sought to continue the participation of the subject in the clinical investigation, and information on the clinical investigation shall be given, in accordance with the following requirements:

(a) regarding incapacitated subjects and minors, the informed consent shall be sought by the investigator from his or her legally designated representative without undue delay and the information referred to in Article 63(2) shall be given as soon as possible to the subject and to his or her legally designated representative;

(b) regarding other subjects, the informed consent shall be sought by the investigator without undue delay from the subject or his or her legally designated representative, whichever can be done sooner, and the information referred to in Article 63(2) shall be given as soon as possible to the subject or his or her legally designated representative, as applicable.

For the purposes of point (b) where informed consent has been obtained from the legally designated representative, informed consent to continue the participation in the clinical investigation shall be obtained from the subject as soon as he or she is capable of giving informed consent.

3. If the subject or, where applicable, his or her legally designated representative does not give consent, he or she shall be informed of the right to object to the use of data obtained from the clinical investigation.

Article 69

Damage compensation

1. Member States shall ensure that systems for compensation for any damage suffered by a subject resulting from participation in a clinical investigation conducted on their territory are in place in the form of insurance, a guarantee, or a similar arrangement that is equivalent as regards its purpose and which is appropriate to the nature and the extent of the risk.

2. The sponsor and the investigator shall make use of the system referred to in paragraph 1 in the form appropriate for the Member State in which the clinical investigation is conducted.

Article 70

Application for clinical investigations

1. The sponsor of a clinical investigation shall submit an application to the Member State(s) in which the clinical investigation is to be conducted (referred to for the purposes of this Article as 'Member State concerned') accompanied by the documentation referred to in Chapter II of Annex XV.

The application shall be submitted by means of the electronic system referred to in Article 73, which shall generate a Union-wide unique single identification number for the clinical investigation which shall be used for all relevant communication in relation to that clinical investigation. Within ten days of it receiving the application, the Member State concerned shall notify the sponsor as to whether the clinical investigation falls within the scope of this Regulation and as to whether the application dossier is complete in accordance with Chapter II of Annex XV.

2. Within one week of any change occurring in relation to the documentation referred to in Chapter II of Annex XV, the sponsor shall update the relevant data in the electronic system referred to in Article 73 and make that change to the documentation clearly identifiable. The Member State concerned shall be notified of the update by means of that electronic system.

3. Where the Member State concerned finds that the clinical investigation applied for does not fall within the scope of this Regulation or that the application dossier is not complete, it shall inform the sponsor thereof and shall set a time limit of maximum ten days for the sponsor to comment or to complete the application by means of the electronic system referred to in Article 73. The Member State concerned may extend this period by a maximum of 20 days where appropriate.

Proposal of Regulation (EU) 2017/745 Articles

Where the sponsor has not provided comments nor completed the application within the time limit referred to in the first subparagraph, the application shall be deemed to have lapsed. Where the sponsor considers the application does fall under the scope of this Regulation and/or is complete but the Member State concerned does not, the application shall be considered to have been rejected. The Member State concerned shall provide for an appeal procedure in respect of such refusal.

The Member State concerned shall notify the sponsor within five days of receipt of the comments or of the requested additional information, whether the clinical investigation is considered as falling within the scope of this Regulation and the application is complete.

4. The Member State concerned may also extend the period referred to in paragraph 1 and 3 each by a further five days.

5. For the purposes of this Chapter, the date on which the sponsor is notified in accordance with paragraph 1 or 3 shall be the validation date of the application. Where the sponsor is not notified, the validation date shall be the last day of the time periods referred to in paragraphs 1, 3 and 4 respectively.

6. During the period when the application is being assessed, the Member State may request additional information from the sponsor. The expiry of the period laid down in point (b) of paragraph 7 shall be suspended from the date of the first request until such time as the additional information has been received.

7. The sponsor may start the clinical investigation in the following circumstances:

(a) in the case of investigational class I devices or in the case of non-invasive class IIa and class IIb devices, unless otherwise stated by national law, immediately after the validation date of the application pursuant to paragraph 5, and provided that a negative opinion which is valid for the entire Member State, under national law, has not been issued by an ethics committee in the Member State concerned in respect of the clinical investigation;

(b) in the case of investigational devices, other than those referred to in point (a), as soon as the Member State concerned has notified the sponsor of its authorisation, and provided that a negative opinion which is valid for the entire Member State, under national law, has not been issued by an ethics committee in the Member State concerned in respect of the clinical investigation. The Member State shall notify the sponsor of the authorisation within 45 days of the validation date referred to in paragraph 5. The Member State may extend this period by a further 20 days for the purpose of consulting with experts.

8. The Commission is empowered to adopt delegated acts in accordance with Article 115 amending, in the light of technical progress and global regulatory developments, the requirements laid down in Chapter II of Annex XV.

9. In order to ensure the uniform application of the requirements laid down in Chapter II of Annex XV, the Commission may adopt implementing acts to the extent necessary to resolve issues of divergent interpretation and of practical application. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

Article 71

Assessment by Member States

1. Member States shall ensure that the persons validating and assessing the application, or deciding on it, do not have conflicts of interest, are independent of the sponsor, the investigators involved and of natural or legal persons financing the clinical investigation, as well as free of any other undue influence.

2. Member States shall ensure that the assessment is done jointly by an appropriate number of persons who collectively have the necessary qualifications and experience.

3. Member States shall assess whether the clinical investigation is designed in such a way that potential remaining risks to subjects or third persons, after risk minimization, are justified, when weighed against

Proposal of Regulation (EU) 2017/745 Articles

the clinical benefits to be expected. They shall, while taking into account applicable CS or harmonised standards, examine in particular:

- (a) the demonstration of compliance of the investigational device(s) with the applicable general safety and performance requirements, apart from the aspects covered by the clinical investigation, and whether, with regard to those aspects, every precaution has been taken to protect the health and safety of the subjects. This includes, where appropriate, assurance of technical and biological safety testing and pre-clinical evaluation;
- (b) whether the risk-minimisation solutions employed by the sponsor are described in harmonised standards and, in those cases where the sponsor does not use harmonised standards, whether the risk-minimisation solutions provide a level of protection that is equivalent to that provided by harmonised standards;
- (c) whether the measures planned for the safe installation, putting into service and maintenance of the investigational device are adequate;
- (d) the reliability and robustness of the data generated in the clinical investigation, taking account of statistical approaches, design of the investigation and methodological aspects, including sample size, comparator and endpoints;
- (e) whether the requirements of Annex XV are met;
- (f) in the case of devices for sterile use, evidence of the validation of the manufacturer's sterilisation procedures or information on the reconditioning and sterilisation procedures which have to be conducted by the investigation site;
- (g) the demonstration of the safety, quality and usefulness of any components of animal or human origin or of substances, which may be considered medicinal products in accordance with Directive 2001/83/EC.

4. Member States shall refuse the authorisation of the clinical investigation if:

- (a) the application dossier submitted pursuant to Article 70(1) remains incomplete;
- (b) the device or the submitted documents, especially the investigation plan and the investigator's brochure, do not correspond to the state of scientific knowledge, and the clinical investigation, in particular, is not suitable for providing evidence for the safety, performance characteristics or benefit of the device on subjects or patients,
- (c) the requirements of Article 62 are not met, or
- (d) any assessment under paragraph 3 is negative.

Member States shall provide for an appeal procedure in respect of a refusal pursuant to the first subparagraph.

Article 72 *Conduct of a clinical investigation*

1. The sponsor and the investigator shall ensure that the clinical investigation is conducted in accordance with the approved clinical investigation plan.

2. In order to verify that the rights, safety and well-being of subjects are protected, that the reported data are reliable and robust, and that the conduct of the clinical investigation is in compliance with the requirements of this Regulation, the sponsor shall ensure adequate monitoring of the conduct of a clinical investigation. The extent and nature of the monitoring shall be determined by the sponsor on the basis of an assessment that takes into consideration all characteristics of the clinical investigation including the following:

- (a) the objective and methodology of the clinical investigation; and

(b) the degree of deviation of the intervention from normal clinical practice.

3. All clinical investigation information shall be recorded, processed, handled, and stored by the sponsor or investigator, as applicable, in such a way that it can be accurately reported, interpreted and verified while the confidentiality of records and the personal data of the subjects remain protected in accordance with the applicable law on personal data protection.

4. Appropriate technical and organisational measures shall be implemented to protect information and personal data processed against unauthorised or unlawful access, disclosure, dissemination, alteration, or destruction or accidental loss, in particular where the processing involves transmission over a network.

5. Member States shall inspect, at an appropriate level, investigation site(s) to check that clinical investigations are conducted in accordance with the requirements of this Regulation and with the approved investigation plan.

6. The sponsor shall establish a procedure for emergency situations which enables the immediate identification and, where necessary, an immediate recall of the devices used in the investigation.

7. The processing of personal data in the context of a clinical investigation, including the secondary use of personal data initially collected for other investigations, shall be deemed to be carried out for scientific research purposes as referred to in Article 9(2), point (j), of Regulation (EU) 2016/679.

Article 73

Electronic system on clinical investigations

1. The Commission shall, in collaboration with the Member States, set up, manage and maintain an electronic system:

(a) to create the single identification numbers for clinical investigations referred to in Article 70(1);

(b) to be used as an entry point for the submission of all applications or notifications for clinical investigations referred to in Articles 70, 74, 75 and 78 and for all other submission of data, or processing of data in this context;

(c) for the exchange of information relating to clinical investigations in accordance with this Regulation between the Member States and between them and the Commission including the exchange of information referred to in Articles 70 and 76;

(d) for information to be provided by the sponsor in accordance with Article 77, including the clinical investigation report and its summary as required in paragraph 5 of that Article;

(e) for reporting on serious adverse events and device deficiencies and related updates referred to in Article 80.

2. When setting up the electronic system referred in paragraph 1 of this Article, the Commission shall ensure that it is interoperable with the EU database for clinical trials on medicinal products for human use set up in accordance with Article 81 of Regulation (EU) No 536/2014 of the European Parliament and of the Council¹ as concerns combined clinical investigations of devices with a clinical trial under that Regulation.

3. The information referred to in point (c) of paragraph 1 shall only be accessible to the Member States and the Commission. The information referred to in the other points of that paragraph shall be accessible to the public, unless, for all or parts of that information, confidentiality of the information is justified on any of the following grounds:

(a) protection of personal data in accordance with Regulation (EC) No 45/2001;

¹ Regulation (EU) No 536/2014 of the European Parliament and of the Council of 16 April 2014 on clinical trials on medicinal products for human use, and repealing Directive 2001/20/EC (OJ L 158, 27.5.2014, p. 1).

Proposal of Regulation (EU) 2017/745 Articles

(b) protection of commercially confidential information, especially in the investigators brochure, in particular through taking into account the status of the conformity assessment for the device, unless there is an overriding public interest in disclosure;

(c) effective supervision of the conduct of the clinical investigation by the Member State(s) concerned.

4. No personal data of subjects shall be publicly available.

5. The user interface of the electronic system referred to in paragraph 1 shall be available in all official languages of the Union.

Article 74

Clinical investigations regarding devices bearing the CE marking

1. Where a clinical investigation is to be conducted to further assess, within the scope of its intended purpose **and in accordance with its PMCF plan**, a device which already bears the CE marking in accordance with Article 20(1), ('PMCF investigation'), and where the investigation would involve submitting subjects to procedures additional to those performed under the normal conditions of use of the device and those additional procedures are invasive or burdensome, the sponsor shall notify the Member States concerned at least 30 days prior to its commencement by means of the electronic system referred to in Article 73. The sponsor shall include the documentation referred to in Chapter II, **Sections 1, 3 and 4**, of Annex XV as part of the notification. **Article 62(4)**, points (b) to (k) and (m), **Article 75(1)**, **Articles 76 and 77**, **Article 80(5) and (6)**, and the relevant provisions of Annex XV shall apply to PMCF investigations **involving additional invasive or burdensome procedures**.

2. Where a clinical investigation is to be conducted to assess, outside the scope of its intended purpose, a device which already bears the CE marking in accordance with Article 20(1), Articles 62 to 81 shall apply.

Article 75

Substantial modifications to clinical investigations

1. If a sponsor intends to introduce modifications to a clinical investigation that are likely to have a substantial impact on the safety, health or rights of the subjects or on the robustness or reliability of the clinical data generated by the investigation, it shall notify, ~~within one week~~, by means of the electronic system referred to in Article 73, the Member State(s) in which the clinical investigation is being or is to be conducted of the reasons for and the nature of those modifications. The sponsor shall include an updated version of the relevant documentation referred to in Chapter II of Annex XV as part of the notification. Changes to the relevant documentation shall be clearly identifiable.

2. **Where the clinical investigation has been the subject of an authorisation in accordance with Article 62(4), point (a)**, the Member State shall assess any substantial modification to the clinical investigation in accordance with the procedure laid down in Article 71.

3. The sponsor may implement the modifications referred to in paragraph 1 **as soon as the Member State concerned has notified the sponsor of its authorisation or, where there is no authorisation**, at the earliest 38 days after the notification **from the sponsor** referred to in that paragraph, unless:

(a) the Member State in which the clinical investigation is being or is to be conducted has notified the sponsor of its refusal based on the grounds referred to in Article 71(4) or on considerations of public health, subject and user safety or health, of public policy, or

(b) an ethics committee in that Member State has issued a negative opinion in relation to the substantial modification to the clinical investigation, which, in accordance with national law, is valid for that entire Member State.

4. The Member State(s) concerned may extend the period referred to in paragraph 3 by a further seven days, for the purpose of consulting with experts.

Article 76

Corrective measures to be taken by Member States and information exchange between Member States

1. Where a Member State in which a clinical investigation is being or is to be conducted has grounds for considering that the requirements set out in this Regulation are not met, it may take at least any of the following measures on its territory:

- (a) revoke the authorisation for the clinical investigation;
- (b) suspend or terminate the clinical investigation;
- (c) require the sponsor to modify any aspect of the clinical investigation.

2. Before the Member State concerned takes any of the measures referred to in paragraph 1 it shall, except where immediate action is required, ask the sponsor or the investigator or both for their opinion. That opinion shall be delivered within seven days.

3. Where a Member State has taken a measure referred to in paragraph 1 of this Article or has refused a clinical investigation, or has been notified by the sponsor of the early termination of a clinical investigation on safety grounds, that Member State shall communicate the corresponding decision and the grounds therefor to all Member States and the Commission by means of the electronic system referred to in Article 73.

4. Where an application is withdrawn by the sponsor prior to a decision by a Member State, that information shall be made available through the electronic system referred to in Article 73 to all Member States and the Commission.

Article 77

Information from the sponsor at the end of a clinical investigation or in the event of a temporary halt or early termination

1. If the sponsor has temporarily halted a clinical investigation or has terminated a clinical investigation early, it shall inform within 15 days the Member State in which that clinical investigation has been temporarily halted or terminated early, through the electronic system referred to in Article 73, of the temporary halt or early termination, providing a justification. In the event that the sponsor has temporarily halted or terminated early the clinical investigation on safety grounds, it shall inform all Member States in which that clinical investigation is being conducted thereof within 24 hours.

2. The end of a clinical investigation shall be deemed to coincide with the last visit of the last subject unless another point in time for such end is set out in the clinical investigation plan.

3. The sponsor shall notify each Member State in which a clinical investigation was being conducted of the end of that clinical investigation in that Member State. That notification shall be made within 15 days of the end of the clinical investigation in relation to that Member State.

4. If an investigation is conducted in more than one Member State, the sponsor shall notify all Member States in which that clinical investigation was conducted of the end of the clinical investigation in all Member States. That notification shall be made within 15 days of that end of the clinical investigation.

5. Irrespective of the outcome of the clinical investigation, within one year of the end of the clinical investigation or within three months of the early termination or temporary halt, the sponsor shall submit to the Member States in which a clinical investigation was conducted a clinical investigation report as referred to in Section 2.8 of Chapter I and Section 7 of Chapter III of Annex XV.

The clinical investigation report shall be accompanied by a summary presented in terms that are easily understandable to the intended user. Both the report and summary shall be submitted by the sponsor by means of the electronic system referred to in Article 73.

Where, for scientific reasons, it is not possible to submit the clinical investigation report within one year of the end of the investigation, it shall be submitted as soon as it is available. In such case, the clinical

Proposal of Regulation (EU) 2017/745 Articles

investigation plan referred to in Section 3 of Chapter II of Annex XV shall specify when the results of the clinical investigation are going to be available, together with a justification.

6. The Commission shall issue guidelines regarding the content and structure of the summary of the clinical investigation report.

In addition, the Commission may issue guidelines for the formatting and sharing of raw data, for cases where the sponsor decides to share raw data on a voluntary basis. Those guidelines may take as a basis and adapt, where possible, existing guidelines for sharing of raw data in the field of clinical investigations.

7. The summary and the clinical investigation report referred to in paragraph 5 of this Article shall become publicly accessible through the electronic system referred to in Article 73, at the latest when the device is registered in accordance with Article 29 and before it is placed on the market. In cases of early termination or temporary halt the summary and the report shall become publicly accessible immediately after submission.

If the device is not registered in accordance with Article 29 within one year of the summary and the report having been entered into the electronic system pursuant to paragraph 5 of this Article, they shall become publicly accessible at that point in time.

Article 78

Coordinated assessment procedure for clinical investigations

1. By means of the electronic system referred to in Article 73, the sponsor of a clinical investigation to be conducted in more than one Member State may submit, for the purpose of Article 70, a single application that, upon receipt, is transmitted electronically to all Member States in which the clinical investigation is to be conducted.

2. The sponsor shall propose in the single application referred to in paragraph 1 that one of the Member States in which the clinical investigation is to be conducted acts as coordinating Member State. The Member States in which the clinical investigation is to be conducted shall, within six days of submission of the application, agree on one of them taking the role of the coordinating Member State. If they do not agree on a coordinating Member State, the coordinating Member State proposed by the sponsor shall assume that role.

3. Under the direction of the coordinating Member State referred to in paragraph 2, the Member States concerned shall coordinate their assessment of the application, in particular of the documentation referred to in Chapter II of Annex XV.

However, the completeness of the documentation referred to in Sections 1.13, 3.1.3, 4.2, 4.3 and 4.4 of Chapter II of Annex XV shall be assessed separately by each Member State concerned in accordance with Article 70(1) to (5).

4. With regard to documentation other than that referred to in the second subparagraph of paragraph 3, the coordinating Member State shall:

(a) within six days of receipt of the single application, notify the sponsor that it is the coordinating Member State ('notification date');

(b) for the purpose of the validation of the application, take into account any considerations submitted within seven days of the notification date by any Member State concerned;

(c) within 10 days of the notification date, assess whether the clinical investigation falls within the scope of this Regulation and whether the application is complete, and shall notify the sponsor accordingly. Article 70(1) and (3) to (5) shall apply to the coordinating Member State in relation to that assessment;

(d) establish the results of its assessment in a draft assessment report to be transmitted within 26 days of the validation date to the Member States concerned. By day 38 after the validation date the other Member States concerned shall transmit their comments and proposals on the draft assessment report and the

Proposal of Regulation (EU) 2017/745 Articles

underlying application to the coordinating Member State which shall take due account of those comments and proposals in its finalisation of the final assessment report, to be transmitted within 45 days of the validation date to the sponsor and the other Member States concerned.

The final assessment report shall be taken into account by all Member States concerned when deciding on the sponsor's application in accordance with [paragraph 11 \[Article 70\(7\)\]](#).

~~5. As regards the assessment of the documentation referred to in the second subparagraph of paragraph 3,~~ each Member State concerned may request, on a single occasion, additional information from the sponsor. The sponsor shall submit the requested additional information within ~~the period set by the Member State concerned, which shall not exceed~~ 12 days ~~of [from] the~~ receipt of the request. The expiry of the last deadline pursuant to point (d) of paragraph 4 shall be suspended from the date of the request until such time as the additional information has been received.

6. For class IIb and class III devices, the coordinating Member State may also extend the periods referred to in paragraph 4 by a further 20 [50] days, for the purpose of consulting with experts.

7. The Commission may, by means of implementing acts, further specify the procedures and timescales for coordinated assessments to be taken into account by Member States concerned when deciding on the sponsor's application. Such implementing acts may also set out the procedures and timescales for coordinated assessment in the case of substantial modifications pursuant to paragraph 12 of this Article, in the case of reporting of adverse events pursuant to Article 80(4) and in the case of clinical investigations of combination products between medical devices and medicinal products, where the latter are under a concurrent coordinated assessment of a clinical trial under Regulation (EU) No 536/2014. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

8. Where the conclusion of the coordinating Member State concerning the area of coordinated assessment is that the conduct of the clinical investigation is acceptable or acceptable subject to compliance with specific conditions, that conclusion shall be deemed to be the conclusion of all Member States concerned.

Notwithstanding the first subparagraph, a Member State concerned may only disagree with the conclusion of the coordinating Member State concerning the area of coordinated assessment on the following grounds:

- (a) when it considers that participation in the clinical investigation would lead to a subject receiving treatment inferior to that received in normal clinical practice in that Member State concerned;
- (b) infringement of national law; or
- (c) considerations as regards subject safety and data reliability and robustness submitted under point (d) of paragraph 4.

Where one of the Member States concerned disagrees with the conclusion on the basis of the second subparagraph of this paragraph, it shall communicate its disagreement, together with a detailed justification, through the electronic system referred to in Article 73, to the Commission, to all other Member States concerned and to the sponsor.

9. Where the conclusion of the coordinating Member State concerning the area of coordinated assessment is that the clinical investigation is not acceptable, that conclusion shall be deemed to be the conclusion of all Member States concerned.

10. A Member State concerned shall refuse to authorise a clinical investigation if it disagrees with the conclusion of the coordinating Member State as regards any of the grounds referred to in the second subparagraph of paragraph 8, or if it finds, on duly justified grounds, that the aspects addressed in Sections 1.13, 3.1.3, 4.2, 4.3 and 4.4 of Chapter II of Annex XV are not complied with, or where an ethics committee has issued a negative opinion in relation to that clinical investigation, which is valid, in

Proposal of Regulation (EU) 2017/745 Articles

accordance with national law, for that entire Member State. That Member State shall provide for an appeal procedure in respect of such refusal.

11. Each Member State concerned shall notify the sponsor through the electronic system referred to in Article 73 as to whether the clinical investigation is authorised, whether it is authorised subject to conditions, or whether authorisation has been refused. Notification shall be done by way of one single decision within five days of the transmission, pursuant to point (d) of paragraph 4, by the coordinating Member State of the final assessment report. Where an authorisation of a clinical investigation is subject to conditions, those conditions may only be such that, by their nature, they cannot be fulfilled at the time of that authorisation.

12. Any substantial modifications as referred to in Article 75 shall be notified to the Member States concerned by means of the electronic system referred to in Article 73. Any assessment as to whether there are grounds for disagreement as referred to in the second subparagraph of paragraph 8 of this Article shall be carried out under the direction of the coordinating Member State, except for substantial modifications concerning Sections 1.13, 3.1.3, 4.2, 4.3 and 4.4 of Chapter II of Annex XV, which shall be assessed separately by each Member State concerned.

13. The Commission shall provide administrative support to the coordinating Member State in the accomplishment of its tasks under this Chapter.

14. All Member States shall be required to apply the procedure set out in this Article from the date corresponding to 5 years from the date of publication of the notice referred to in Article 34(3), informing that the electronic system referred to in Article 33(2), point (e), is functional and meets the functional specifications drawn up pursuant to Article 34(1).

Before the date set out in the first subparagraph of this paragraph and at the earliest 6 months from the date of publication of the notice referred to in that subparagraph, the procedure set out in this Article shall be applied only by those Member States in which the clinical investigation is to be conducted which have agreed to apply it.

15. The Commission is empowered to adopt delegated acts in accordance with Article 115 to amend this Article in light of experience gained from the practical application of the coordinated assessment procedure, in particular as regards timelines and the authorisation of clinical investigations subject to a coordinated assessment.

Article 79

~~Review of coordinated assessment procedure~~

~~By 27 May 2026, the Commission shall submit to the European Parliament and to the Council a report on experience gained from the application of Article 78 and, if necessary, propose a review of Article 78(14) and point (h) of Article 123(3).~~

Article 79a

Clinical investigations in combined studies

Clinical investigations that are part of combined studies, and which are subject to authorisation in accordance with Article 62, may be carried out in accordance with Article 14c of Regulation (EU) No 536/2014.

If the sponsor chooses to apply Article 14c of Regulation (EU) No 536/2014, the requirements laid down therein and in any implementing or delegated acts adopted in accordance with that Article shall apply instead of the corresponding requirements laid down in this Regulation.

Article 80

Recording and reporting of adverse events that occur during clinical investigations

1. The sponsor shall fully record all of the following:

Proposal of Regulation (EU) 2017/745 Articles

(a) any adverse event of a type identified in the clinical investigation plan as being critical to the evaluation of the results of that clinical investigation;

(b) any serious adverse event;

(c) any device deficiency that might have led to a serious adverse event if appropriate action had not been taken, intervention had not occurred, or circumstances had been less fortunate;

(d) any new findings in relation to any event referred to in points (a) to (c).

2. The sponsor shall report, without delay to all Member States in which the clinical investigation is being conducted, all of the following by means of the electronic system referred to in Article 73:

(a) any serious adverse event that has a causal relationship with the investigational device, the comparator or the investigation procedure or where such causal relationship is reasonably possible;

(b) any device deficiency that might have led to a serious adverse event if appropriate action had not been taken, intervention had not occurred, or circumstances had been less fortunate;

(c) any new findings in relation to any event referred to in points (a) and (b).

The time period for reporting shall take account of the severity of the event. Where necessary to ensure timely reporting, the sponsor may submit an initial report that is incomplete followed up by a complete report.

Upon request by any Member State in which the clinical investigation is being conducted, the sponsor shall provide all information referred to in paragraph 1.

3. The sponsor shall also report to the Member States in which the clinical investigation is being conducted any event referred to in paragraph 2 of this Article that occurred in third countries in which a clinical investigation is performed under the same clinical investigation plan as the one applying to a clinical investigation covered by this Regulation by means of the electronic system referred to in Article 73.

4. In the case of a clinical investigation for which the sponsor has used the single application referred to in Article 78, the sponsor shall report any event as referred to in paragraph 2 of this Article by means of the electronic system referred to in Article 73. Upon receipt, this report shall be transmitted electronically to all Member States in which the clinical investigation is being conducted.

Under the direction of the coordinating Member State referred to in Article 78(2), the Member States shall coordinate their assessment of serious adverse events and device deficiencies to determine whether to modify, suspend or terminate the clinical investigation or whether to revoke the authorisation for that clinical investigation.

This paragraph shall not affect the rights of the other Member States to perform their own evaluation and to adopt measures in accordance with this Regulation in order to ensure the protection of public health and patient safety. The coordinating Member State and the Commission shall be kept informed of the outcome of any such evaluation and the adoption of any such measures.

5. In the case of PMCF investigations referred to in Article 74(1), the provisions on vigilance laid down in Articles 87 to 90 and in the acts adopted pursuant to Article 91 shall apply instead of this Article.

6. Notwithstanding paragraph 5, this Article shall apply where a causal relationship between the serious adverse event and the preceding investigational procedure has been established.

Article 81 Implementing acts

The Commission may, by means of implementing acts, establish the detailed arrangements and procedural aspects necessary for the implementation of this Chapter as regards the following:

Proposal of Regulation (EU) 2017/745 Articles

- (a) harmonised electronic forms for the application for clinical investigations and their assessment as referred to in Articles 70 and 78, taking into account specific categories or groups of devices;
- (b) the functioning of the electronic system referred to in Article 73;
- (c) harmonised electronic forms for the notification of PMCF investigations as referred to in Article 74(1), and of substantial modifications as referred to in Article 75;
- (d) the exchange of information between Member States as referred to in Article 76;
- (e) harmonised electronic forms for the reporting of serious adverse events and device deficiencies as referred to in Article 80;
- (f) the timelines for the reporting of serious adverse events and device deficiencies, taking into account the severity of the event to be reported as referred to in Article 80;
- (g) uniform application of the requirements regarding the clinical evidence or data needed to demonstrate compliance with the general safety and performance requirements set out in Annex I.

The implementing acts referred to in the first paragraph shall be adopted in accordance with the examination procedure referred to in Article 114(3).

~~Article 82~~

~~Requirements regarding other clinical investigations~~

~~1. Clinical investigations, not performed pursuant to any of the purposes listed in Article 62(1), shall comply with the provisions of Article 62 (2) and (3), points (b), (c), (d), (f), (h), and (i) of Article 62(4) and Article 62(6).~~

~~2. In order to protect the rights, safety, dignity and well-being of subjects and the scientific and ethical integrity of clinical investigations not performed for any of the purposes listed in Article 62(1), each Member State shall define any additional requirements for such investigations, as appropriate for each Member State concerned.~~

Chapter VII

Post-market surveillance, vigilance and market surveillance

SECTION 1

POST-MARKET SURVEILLANCE

Article 83

Post-market surveillance system of the manufacturer

1. For each device, manufacturers shall plan, establish, document, implement, maintain and update a post-market surveillance system in a manner that is proportionate to the risk class and appropriate for the type of device. That system shall be an integral part of the manufacturer's quality management system referred to in Article 10(9).
2. The post-market surveillance system shall be suited to actively and systematically gathering, recording and analysing relevant data on the quality, performance and safety of a device throughout its entire lifetime, and to drawing the necessary conclusions and to determining, implementing and monitoring any preventive and corrective actions.
3. Data gathered by the manufacturer's post-market surveillance system shall in particular be used:
 - (a) to update the benefit-risk determination and to improve the risk management as referred to in Chapter I of Annex I;
 - (b) to update the design and manufacturing information, the instructions for use and the labelling;

- (c) to update the clinical evaluation;
- (d) to update the summary of safety and clinical performance referred to in Article 32;
- (e) for the identification of needs for preventive, corrective or field safety corrective action;
- (f) for the identification of options to improve the usability, performance and safety of the device;
- (g) when relevant, to contribute to the post-market surveillance of other devices; and
- (h) to detect and report trends in accordance with Article 88.

The technical documentation shall be updated accordingly.

4. If, in the course of the post-market surveillance, a need for preventive or corrective action, or both, is identified, the manufacturer shall implement the appropriate measures. ~~and inform~~ The competent authorities concerned ~~and, where applicable, the notified body, may request the manufacturer to inform them when such preventive or corrective action is taken to reduce a risk that may compromise the safety or performance of the device.~~ Where a serious incident is identified or a field safety corrective action is implemented, it shall be reported in accordance with Article 87.

Article 84

Post-market surveillance plan

The post-market surveillance system referred to in Article 83 shall be based on a post-market surveillance plan, the requirements for which are set out in Section 1 of Annex III. ~~For devices other than custom-made devices, the post-market surveillance plan shall be part of the technical documentation specified in Annex II.~~

Article 85

Post-market surveillance report

Manufacturers of class I devices shall prepare a post-market surveillance report summarising the results and conclusions of the analyses of the post-market surveillance data gathered as a result of the post-market surveillance plan referred to in Article 84 together with a rationale and description of any preventive and corrective actions taken. The report shall be updated when necessary and made available to the competent authority upon request.

Article 86

Periodic safety update report

1. Manufacturers of class IIa, class IIb and class III devices, ~~other than custom-made devices~~, shall prepare a periodic safety update report ('PSUR') for each device, ~~or [and]~~ where relevant for each category or group of devices, summarising the results and conclusions of the analyses of the post-market surveillance data gathered as a result of the post-market surveillance plan referred to in Article 84, together with a description of any preventive and corrective actions taken, including their rationale. Throughout the lifetime of the device concerned, that PSUR shall set out:

- (a) the conclusions of the benefit-risk determination;
- (b) the main findings of the PMCF; and
- (c) the volume of sales of the device and an estimate evaluation of the size and other characteristics of the population using the device and, where practicable, the usage frequency of the device.

Manufacturers of class IIb and class III devices shall update the PSUR ~~in the first year [at least annually] after the certificate is issued and every two years thereafter or when there is a significant change in the benefit-risk determination or in the acceptability of undesirable side-effects. The [That] PSUR shall except in the case of custom-made devices, be part of the technical documentation as specified in Annexes II and III.~~

Proposal of Regulation (EU) 2017/745 Articles

Manufacturers of class IIa devices shall update the PSUR when necessary ~~and at least every two years~~. The [That] PSURs shall, ~~except in the case of custom-made devices~~, be part of the technical documentation specified in Annexes ~~II and III~~.

~~For custom-made devices, the PSUR shall be part of the documentation referred to in Section 2 of Annex XIII.~~

2. For class III devices or ~~class IIb~~ implantable devices, ~~manufacturers shall submit PSURs by means of the electronic system referred to in Article 92 to the notified body involved in the conformity assessment in accordance with Article 52 other than well-established technology devices, the notified body shall review the PSUR during the surveillance assessment the report and add its evaluation to that electronic system with details of any action taken. The manufacturer and the notified body shall make such PSURs and the evaluation by the notified body shall be made available to competent authorities through the electronic system referred to in Article 92.~~

3. For devices other than those referred to in paragraph 2, manufacturers shall make PSURs available to the notified body involved in the conformity assessment and, upon request, to competent authorities.

SECTION 2 VIGILANCE

Article 87

Reporting of serious incidents and field safety corrective actions

1. Manufacturers of devices made available on the Union market, other than investigational devices, shall report, to the relevant competent authorities, in accordance with Articles 92(5) and (7), the following:

(a) any serious incident involving devices made available on the Union market, except expected ~~undesirable~~ side-effects which are clearly documented in the product information and quantified in the technical documentation and are subject to trend reporting pursuant to Article 88;

(b) any field safety corrective action in respect of devices made available on the Union market, including any field safety corrective action undertaken in a third country in relation to a device which is also legally made available on the Union market, if the reason for the field safety corrective action is not limited to the device made available in the third country.

The reports referred to in the first subparagraph shall be submitted ~~without undue delay~~ through the electronic system referred to in Article 92.

2. As a general rule, the time period for the reporting referred to in paragraph 1 shall take account of the severity of the serious incident.

3. Manufacturers shall report any serious incident as referred to in ~~point (a) of~~ paragraph 1, ~~first subparagraph, point (a)~~, immediately after they have established ~~that there is a~~ [the] causal relationship between that incident and their device or that such causal relationship is reasonably possible and not later than ~~30~~ [15] days after they become aware of the incident.

4. Notwithstanding paragraph 3, in the event of a serious public health threat the report referred to in paragraph 1 shall be provided immediately, and not later than 2 days after the manufacturer becomes aware of that threat.

5. Notwithstanding paragraph 3, in the event of death or an unanticipated serious deterioration in a person's state of health the report shall be provided immediately after the manufacturer has established or as soon as it suspects a causal relationship between the device and the serious incident but not later than 10 days after the date on which the manufacturer becomes aware of the serious incident.

6. Where necessary to ensure timely reporting, the manufacturer may submit an initial report that is incomplete followed up by a complete report.

Proposal of Regulation (EU) 2017/745 Articles

7. If, after becoming aware of a potentially reportable incident, the manufacturer is uncertain about whether the incident is reportable, it shall nevertheless submit a report within the timeframe required in accordance with paragraphs 2 to 5.

8. Except in cases of urgency in which the manufacturer needs to undertake field safety corrective action immediately, the manufacturer shall, without undue delay, report the field safety corrective action referred to in point (b) of paragraph 1 in advance of the field safety corrective action being undertaken.

9. For similar serious incidents that occur with the same device or device type and for which the root cause has been identified or a field safety corrective action implemented or where the incidents are common and well documented, the manufacturer may provide periodic summary reports instead of individual serious incident reports, on condition that the coordinating competent authority referred to in Article 89(9), in consultation with the competent authorities referred to in point (a) of Article 92(8), has agreed with the manufacturer on the format, content and frequency of the periodic summary reporting. Where a single competent authority is referred to in points (a) and (b) of Article 92(8), the manufacturer may provide periodic summary reports following agreement with that competent authority.

10. The Member States shall take appropriate measures such as organising targeted information campaigns, to encourage and enable healthcare professionals, users and patients to report to the competent authorities suspected serious incidents referred to in point (a) of paragraph 1.

The competent authorities shall record centrally at national level reports they receive from healthcare professionals, users and patients.

11. Where a competent authority of a Member State obtains such reports on suspected serious incidents referred to in point (a) of paragraph 1 from healthcare professionals, users or patients, it shall take the necessary steps to ensure that the manufacturer of the device concerned is informed of the suspected serious incident without delay.

Where the manufacturer of the device concerned considers that the incident is a serious incident, it shall provide a report in accordance with paragraphs 1 to 5 of this Article on that serious incident to the competent authority of the Member State in which that serious incident occurred and shall take the appropriate follow-up action in accordance with Article 89.

Where the manufacturer of the device concerned considers that the incident is not a serious incident or is an expected undesirable side-effect, which will be covered by trend reporting in accordance with Article 88, it shall provide an explanatory statement. If the competent authority does not agree with the conclusion of the explanatory statement, it may require the manufacturer to provide a report in accordance with paragraphs 1 to 5 of this Article and require it to ensure that appropriate follow-up action is taken in accordance with Article 89.

Article 87a

Reporting of actively exploited vulnerabilities and severe incidents related to devices

1. Without prejudice to the reporting obligations regarding serious incidents and field safety corrective actions set out in Article 87, the manufacturer of a device shall report to the computer security incident response teams ('CSIRTs'), designated as coordinators of the Member States where a device has been made available, and to the European Union Agency for Cybersecurity (ENISA), either of the following:

- a) any actively exploited vulnerability as defined in Article 3, point (42), of Regulation (EU) 2024/2847 of the European Parliament and of the Council***** contained in the device;
- b) any severe incident as referred in Article 14(5) of Regulation (EU) 2024/2847 having an impact on the security of the device.

***** Regulation (EU) 2024/2847 of the European Parliament and of the Council of 23 October 2024 on horizontal cybersecurity requirements for products with digital elements and amending Regulations (EU) No 168/2013 and (EU) 2019/1020 and Directive (EU) 2020/1828 (Cyber Resilience Act) (OJ L, 024/2847, 20.11.2024, ELI: <http://data.europa.eu/eli/reg/2024/2847/oj>).

Proposal of Regulation (EU) 2017/745 Articles

2. The manufacturer shall submit the report referred to in paragraph 1 through the electronic system referred to in Article 92 not later than 30 days after it becomes aware of the actively exploited vulnerability or the severe incident.

3. The report referred to in paragraph 1, as well as any report submitted by a manufacturer in accordance with Article 87 that also qualifies as actively exploited vulnerability or severe incident, shall be made available simultaneously to the CSIRTs designated as coordinators of the Member States in which the device has been made available and to ENISA.

4. For the purposes of this Article, the CSIRTs designated as coordinators and ENISA shall have access to Eudamed.

Article 88

Trend reporting

1. Manufacturers shall report, by means of the electronic system referred to in Article 92, any statistically significant increase in the frequency or severity of incidents that are not serious incidents or that are expected undesirable side-effects that could have a significant impact on the benefit-risk analysis referred to in Sections 1 and 8 of Annex I ~~and which have led or may lead to risks to the health or safety of patients, users or other persons that are unacceptable when weighed against the intended benefits.~~ The significant increase shall be established in comparison to the foreseeable frequency or severity of such incidents in respect of the device, or category or group of devices, in question during a specific period as specified in the technical documentation and product information.

The manufacturer shall specify how to manage the incidents referred to in the first subparagraph and the methodology used for determining any statistically significant increase in the frequency or severity of such incidents, as well as the observation period, in the post-market surveillance plan referred to in Article 84.

2. The competent authorities may conduct their own assessments on the trend reports referred to in paragraph 1 and require the manufacturer to adopt appropriate measures in accordance with this Regulation in order to ensure the protection of public health and patient safety. Each competent authority shall inform the Commission, the other competent authorities and the notified body that issued the certificate, of the results of such assessment and of the adoption of such measures.

Article 89

Analysis of serious incidents and field safety corrective actions

1. Following the reporting of a serious incident pursuant to Article 87(1), the manufacturer shall, without delay, perform the necessary investigations in relation to the serious incident and the devices concerned. This shall include a risk assessment of the incident and field safety corrective action taking into account criteria as referred to in paragraph 3 of this Article as appropriate.

The manufacturer shall cooperate with the competent authorities ~~and where relevant with the notified body concerned~~ during the investigations referred to in the first subparagraph and shall not perform any investigation which involves altering the device or a sample of the batch concerned in a way which may affect any subsequent evaluation of the causes of the incident, prior to informing the competent authorities of such action.

2. Member States shall take the necessary steps to ensure that any information regarding a serious incident that has occurred within their territory, or a field safety corrective action that has been or is to be undertaken within their territory, and that is brought to their knowledge in accordance with Article 87 is evaluated centrally at national level by their competent authority, if possible together with the manufacturer, ~~and, where relevant, the notified body concerned.~~

3. In the context of the evaluation referred to in paragraph 2, the competent authority shall evaluate the risks arising from the reported serious incident and evaluate any related field safety corrective actions, taking into account the protection of public health and criteria such as causality, detectability and probability of recurrence of the problem, frequency of use of the device, probability of occurrence of direct

Proposal of Regulation (EU) 2017/745 Articles

or indirect harm, the severity of that harm, the clinical benefit of the device, intended and potential users, and population affected. The competent authority shall also evaluate the adequacy of the field safety corrective action envisaged or undertaken by the manufacturer and the need for, and kind of, any other corrective action, in particular taking into account the principle of inherent safety contained in Annex I.

Upon request by the national competent authority, manufacturers shall provide all documents necessary for the risk assessment.

(a) The competent authority may request the notified body that issued a certificate for the device in question in accordance with Article 56 to provide assistance in evaluating a corrective action related to a serious incident or a field safety corrective action.

4. The competent authority shall monitor the manufacturer's investigation of a serious incident. Where necessary, a competent authority may intervene in a manufacturer's investigation or initiate an independent investigation.

5. The manufacturer shall provide a final report to the competent authority setting out its findings from the investigation by means of the electronic system referred to in Article 92. The report shall set out conclusions and where relevant indicate corrective actions to be taken.

6. In the case of devices referred to in Article 1(8), first subparagraph, and where the serious incident or field safety corrective action is confirmed by the manufacturer to ~~may~~ be related to a substance which, if used separately, would be considered to be a medicinal product, the evaluating competent authority or the coordinating competent authority referred to in paragraph 9 ~~of this Article~~ shall inform the national competent authority or the EMA, ~~depending on~~ which issued the scientific opinion on that substance in accordance with [under] Article 52(9), of that serious incident or field safety corrective action.

In the case of devices falling within the scope of [covered by] this Regulation in accordance with Article 1(6), point (g), or in accordance with Article 1(10), the competent authority or the coordinating competent authority referred to in paragraph 9 of this Article shall inform the competent authority for substances of human origin that was consulted by the notified body in accordance with Article 52(10), provided that the serious incident or field safety corrective action is confirmed by the manufacturer to be related to the derivatives of substances of human origin utilised for the manufacture of the device or to the non-viable substances of human origin or their derivatives that have an action ancillary to that of the device [and where the serious incident or field safety corrective action may be related to the derivatives of tissues or cells of human origin utilised for the manufacture of the device, and in the case of devices falling under this Regulation pursuant to Article 1(10), the competent authority or the coordinating competent authority referred to in paragraph 9 of this Article shall inform the competent authority for human tissues and cells that was consulted by the notified body in accordance with Article 52(10)].

7. If, after carrying out the evaluation in accordance with paragraph 3 of this Article, the evaluating competent authority identifies the need for additional corrective actions from the manufacturer to minimise the risk of recurrence of the serious incident, it shall, through the electronic system referred to in Article 92, inform, without delay, the other competent authorities of the corrective action taken or envisaged by the manufacturer or required of it to minimise the risk of recurrence of the serious incident, including information on the underlying events and the outcome of its assessment.

8. The manufacturer shall ensure that information about the field safety corrective action taken is brought without delay to the attention of users of the device in question by means of a field safety notice. The field safety notice shall be edited in an official Union language or languages determined by the Member State in which the field safety corrective action is taken. Except in cases of urgency, the content of the draft field safety notice shall be submitted to the evaluating competent authority or, in the cases referred to in paragraph 9, to the coordinating competent authority to allow them to make comments. Unless duly justified by the situation of the individual Member State, the content of the field safety notice shall be consistent in all Member States.

The field safety notice shall allow the correct identification of the device or devices involved, in particular by including the relevant UDIs, and the correct identification, in particular, by including the SRN, if already

Proposal of Regulation (EU) 2017/745 Articles

issued, of the manufacturer that has undertaken the field safety corrective action. The field safety notice shall explain, in a clear manner, without understating the level of risk, the reasons for the field safety corrective action with reference to the device malfunction and associated risks for patients, users or other persons, and shall clearly indicate all the actions to be taken by users.

The manufacturer shall enter the field safety notice in the electronic system referred to in Article 92 through which that notice shall be accessible to the public.

9. The competent authorities shall actively participate in a procedure in order to coordinate their assessments referred to in paragraph 3 **whenever such coordination is needed to ensure a high level of protection of the health and safety of patients, users and other persons or the protection of public health throughout the Union, and in particular** in the following cases:

- (a) where there is concern regarding a particular serious incident or cluster of serious incidents relating to the same device or type of device of the same manufacturer in more than one Member State;
- (b) where the appropriateness of a field safety corrective action that is proposed by a manufacturer in more than one Member State is in question.

That coordinated procedure shall cover the following:

- designation of a coordinating competent authority on a case by case basis, when required;
- defining the coordinated assessment process, including the tasks and responsibilities of the coordinating competent authority and the involvement of other competent authorities.

Unless otherwise agreed between the competent authorities, the coordinating competent authority shall be the competent authority of the Member State in which the manufacturer has its registered place of business.

The coordinating competent authority shall, through the electronic system referred to in Article 92, inform the manufacturer, the other competent authorities and the Commission that it has assumed the role of coordinating authority.

10. The designation of a coordinating competent authority shall not affect the rights of the other competent authorities to perform their own assessment and to adopt measures in accordance with this Regulation in order to ensure the protection of public health and patient safety. The coordinating competent authority and the Commission shall be kept informed of the outcome of any such assessment and the adoption of any such measures.

11. The Commission shall provide administrative support to the coordinating competent authority in the accomplishment of its tasks under this Chapter.

Article 90 Analysis of vigilance data

The Commission shall, in collaboration with the Member States, put in place systems and processes to actively monitor the data available in the electronic system referred to in Article 92, in order to identify trends, patterns or signals in the data that may reveal new risks or safety concerns.

Where a previously unknown risk is identified or the frequency of an anticipated risk significantly and adversely changes the benefit-risk determination, the competent authority or, where appropriate, the coordinating competent authority shall inform the manufacturer, or where applicable the authorised representative, which shall then take the necessary corrective actions.

Article 91
Implementing acts

The Commission may, by means of implementing acts, and after consulting [consultation of] the MDCG, adopt the detailed arrangements and procedural aspects necessary for the implementation of Articles 84 [85] to 90 and Article 92 as regards the following:

- (a) the typology of serious incidents and field safety corrective actions in relation to specific devices, or categories or groups of devices;
- (b) the reporting of serious incidents and field safety corrective actions and field safety notices, and the provision and content of the post-market surveillance plan, periodic summary reports, post-market surveillance reports, PSURs and trend reports by manufacturers as referred to in Articles 84, 85, 86, 87, 88 and 89 respectively;
- (c) standard structured forms for electronic and non-electronic reporting, including a minimum data set for reporting of suspected serious incidents by healthcare professionals, users and patients;
- (d) timelines for the reporting of field safety corrective actions, and for the provision by manufacturers of periodic summary reports and trend reports, taking into account the severity of the incident to be reported as referred to in Article 87;
- (e) harmonised forms for the exchange of information between competent authorities as referred to in Article 89;
- (f) procedures for the designation of a coordinating competent authority; the coordinated evaluation process, including tasks and responsibilities of the coordinating competent authority and involvement of other competent authorities in this process.

The implementing acts referred to in the first paragraph shall be adopted in accordance with the examination procedure referred to in Article 114(3).

Article 92
Electronic system on vigilance and on post-market surveillance

1. The Commission shall, in collaboration with the Member States, set up and manage an electronic system to collate and process the following information:

- (a) the reports by manufacturers on serious incidents and field safety corrective actions referred to in Article 87(1) and Article 89(5);
- (b) the periodic summary reports by manufacturers referred to in Article 87(9);
- (c) the reports by manufacturers on trends referred to in Article 88;
- (d) the PSURs referred to in Article 86;
- (e) the field safety notices by manufacturers referred to in Article 89(8);
- (f) the information to be exchanged between the competent authorities of the Member States and between them and the Commission in accordance with Article 89(7) and (9).

That electronic system shall include relevant links to the UDI database.

2. The information referred to in paragraph 1 of this Article shall be made available through the electronic system to the competent authorities of the Member States and to the Commission. The notified bodies shall also have access to that information to the extent that it relates to devices for which they issued a certificate in accordance with Article 56 [53].

3. The Commission shall ensure that healthcare professionals and the public have appropriate levels of access to the electronic system referred to in paragraph 1.

Proposal of Regulation (EU) 2017/745 Articles

4. On the basis of arrangements between the Commission and competent authorities of third countries or international organisations, the Commission may grant those competent authorities or international organisations access to the electronic system referred to in paragraph 1 at the appropriate level. Those arrangements shall be based on reciprocity and make provision for confidentiality and data protection equivalent to those applicable in the Union.
5. The reports on serious incidents referred to in point (a) of Article 87(1) shall be automatically transmitted, upon receipt, via the electronic system referred to in paragraph 1 of this Article, to the competent authority of the Member State in which the incident occurred.
6. The trend reports referred to in Article 88(1) shall be automatically transmitted upon receipt via the electronic system referred to in paragraph 1 of this Article to the competent authorities of the Member State in which the incidents occurred.
7. The reports on field safety corrective actions referred to in point (b) of Article 87(1) shall be automatically transmitted upon receipt via the electronic system referred to in paragraph 1 of this Article to the competent authorities of the following Member States:
 - (a) the Member States in which the field safety corrective action is being or is to be undertaken;
 - (b) the Member State in which the manufacturer has its registered place of business.
8. The periodic summary reports referred to in Article 87(9) shall be automatically transmitted upon receipt via the electronic system referred to in paragraph 1 of this Article to the competent authority of:
 - (a) the Member State or Member States participating in the coordination procedure in accordance with Article 89(9) and which have agreed on the periodic summary report;
 - (b) the Member State in which the manufacturer has its registered place of business.
9. The information referred to in paragraphs 5 to 8 of this Article shall be automatically transmitted, upon receipt, through the electronic system referred to in paragraph 1 of this Article, to the notified body that issued the certificate for the device in question in accordance with Article 56.

SECTION 3 MARKET SURVEILLANCE

Article 93 *Market surveillance activities*

1. The competent authorities shall perform appropriate checks on the conformity characteristics and performance of devices **and on the compliance of economic operators with the obligations set out in this Regulation** including, where appropriate, a review of documentation and physical or laboratory checks on the basis of adequate samples. The competent authorities shall, in particular, take account of established principles regarding risk assessment and risk management, vigilance data and complaints.
 - (a) **Member States shall ensure that their national competent authorities are provided with adequate and sufficient technical, financial and human resources, and with infrastructure to fulfil their tasks effectively under this Regulation.**
2. The competent authorities shall draw up annual surveillance activity plans ~~and allocate a sufficient number of material and competent human resources in order to carry out those activities~~, taking into account the European market surveillance programme, **which shall be developed and maintained** by the MDCG ~~pursuant to Article 105~~, and local circumstances.
3. In order to fulfil the obligations laid down in paragraph 1, the competent authorities:
 - (a) may require economic operators to, *inter alia*, make available the documentation and information necessary for the purpose of carrying out the authorities' activities and, where justified, to provide the necessary samples of devices or access to devices free of charge; and

Proposal of Regulation (EU) 2017/745 Articles

(b) shall carry out both announced and, if necessary, unannounced inspections of the premises of economic operators, as well as suppliers and/or subcontractors, and, where necessary, at the facilities of professional users.

4. The competent authorities shall prepare an annual summary of the results of their surveillance activities and make it accessible to other competent authorities by means of the electronic system referred to in Article 100.

5. The competent authorities may confiscate, destroy or otherwise render inoperable devices that present an unacceptable risk or falsified devices where they deem it necessary to do so in the interests of the protection of public health.

6. Following each inspection carried out for the purposes referred to in paragraph 1, the competent authority shall draw up a report on the findings of the inspection that concern compliance with the legal and technical requirements applicable under this Regulation. The report shall set out any corrective actions needed.

7. The competent authority which carried out the inspection shall communicate the content of the report referred to in paragraph 6 of this Article to the economic operator that has been the subject of the inspection. Before adopting the final report, the competent authority shall give that economic operator the opportunity to submit comments. That final inspection report shall be entered in the electronic system provided for in Article 100.

8. The Member States shall review and assess the functioning of their market surveillance activities. Such reviews and assessments shall be carried out at least every four years and the results thereof shall be communicated to the other Member States and the Commission. Each Member State shall make a summary of the results accessible to the public by means of the electronic system referred to in Article 100.

9. The competent authorities of the Member States shall coordinate their market surveillance activities, cooperate with each other and share with each other and with the Commission the results thereof, to provide for a harmonised and high level of market surveillance in all Member States.

Where appropriate, the competent authorities of the Member States shall agree on work-sharing, joint market surveillance activities and specialisation.

10. Where more than one authority in a Member State is responsible for market surveillance and external border controls, those authorities shall cooperate with each other, by sharing information relevant to their role and functions.

11. Where appropriate, the competent authorities of the Member States shall cooperate with the competent authorities of third countries with a view to exchanging information and technical support and promoting activities relating to market surveillance.

12. In respect of devices that are high-risk AI systems as referred to in Article 6(1) of Regulation (EU) 2024/1689, the competent authorities shall cooperate with market surveillance authorities of their Member State designated in accordance with Article 70 of Regulation (EU) 2024/1689.

Article 94

Evaluation of devices suspected of presenting an unacceptable risk or other non-compliance

~~Where~~ The competent authorities of a Member State shall, alone or in cooperation with the competent authorities of other Member States, carry out an evaluation of a device of an economic operator covering all the relevant requirements laid down in this Regulation relating to the risk presented by the device, or to any other non-compliance of the device or of the economic operator, where they, based on data obtained by vigilance or market surveillance activities or on other information, have reason to believe ~~that a device~~ either of the following:

(a) ~~the device may present an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health; or~~

(b) ~~the device or the economic operator concerned otherwise does not comply with the requirements laid down in this Regulation.~~

~~they shall carry out an evaluation of the device concerned covering all requirements laid down in this Regulation relating to the risk presented by the device, or to any other non-compliance of the device.~~

The relevant economic operators ~~and, where applicable and requested, the notified body that issued a certificate for the device in question~~ shall cooperate with the competent authorities.

~~The competent authorities of the Member States may request any economic operator or notified body to submit documentation available to them, where access to such documentation is needed in the interest of public health or patient safety or health.~~

Article 95

Procedure for dealing with devices presenting an unacceptable risk to health and safety

1. Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operator.

2. The competent authorities shall, without delay, notify the Commission, the other Member States and, where a certificate has been issued in accordance with Article 56 for the device concerned, the notified body that issued that certificate, of the results of the evaluation and of the actions which they have required the economic operators to take, by means of the electronic system referred to in Article 100.

3. The economic operators as referred to in paragraph 1 shall without delay ensure that all appropriate corrective action is taken, ~~within the period referred to in paragraph 1~~, throughout the Union in respect of all the devices concerned that they have made available on the market.

4. Where the economic operator as referred to in paragraph 1 does not take adequate corrective action within the period referred to in paragraph 1, the competent authorities shall take all appropriate measures to prohibit or restrict the making available of the device on their national market, to withdraw the device from that market or to recall it.

The competent authorities shall notify the Commission, the other Member States and the notified body referred to in paragraph 2 of this Article, without delay, of those measures, by means of the electronic system referred to in Article 100.

5. The notification referred to in paragraph 4 shall include all available details, in particular the data necessary for the identification and tracking of the non-compliant device, the origin of the device, the nature of and the reasons for the non-compliance alleged and the risk involved, the nature and duration of the national measures taken and the arguments put forward by the relevant economic operator.

6. Member States other than the Member State initiating the procedure shall, without delay, inform the Commission and the other Member States, by means of the electronic system referred to in Article 100, of any additional relevant information at their disposal relating to the non-compliance of the device concerned and of any measures adopted by them in relation to the device concerned.

Proposal of Regulation (EU) 2017/745 Articles

In the event of disagreement with the notified national measure, they shall without delay inform the Commission and the other Member States of their objections, by means of the electronic system referred to in Article 100.

7. Where, within two months of receipt of the notification referred to in paragraph 4, no objection has been raised by either a Member State or the Commission in respect of the notified [any] measures taken by a Member State, those measures shall be deemed to be justified.

In that case, all Member States shall ensure that corresponding appropriate restrictive or prohibitive measures, including withdrawing, recalling or limiting the availability of the device on their national market, are taken without delay in respect of the device concerned. **Paragraph 4 shall not apply to such measures adopted by the Member States.**

Article 96

Procedure for evaluating national measures at Union level

1. Where, within two months of receipt of the notification referred to in Article 95(4), objections are raised by a Member State against a measure taken by another Member State, or where the Commission considers the measure to be contrary to Union law, the Commission shall, after consulting the competent authorities concerned and, where necessary, the economic operators concerned, evaluate that national measure. On the basis of the results of that evaluation, the Commission may decide, by means of implementing acts, whether or not the national measure is justified. ~~Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).~~

2. Where the Commission considers the national measure to be justified as referred to in paragraph 1 of this Article, the second sub-paragraph of Article 95(7) shall apply. If the Commission considers that the national measure is [to be] unjustified, the Member State concerned, **as well as any Member State that has taken corresponding restrictive or prohibitive measures**, shall withdraw the measure.

Where the Commission does not adopt a decision pursuant to paragraph 1 of this Article within eight months of receipt of the notification referred to in Article 95(4), the national measure shall be considered to be justified.

3. Where a Member State or the Commission consider that the risk to health and safety emanating from a device cannot be mitigated satisfactorily by means of measures taken by the Member State or Member States concerned, the Commission, at the request of a Member State or on its own initiative, may take, by means of implementing acts, the necessary and duly justified measures to ensure the protection of health and safety, including measures restricting or prohibiting the placing on the market and putting into service of the device concerned. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

Article 97

Other non-compliance

1. Where, having performed an evaluation pursuant to Article 94, the competent authorities of a Member State find that a device **or an economic operator** does not comply with the requirements laid down in this Regulation but does not present an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall require the relevant economic operator to bring the non-compliance concerned to an end within a reasonable period that is clearly defined and communicated to the economic operator and that is proportionate to the non-compliance.

2. Where the economic operator does not bring the non-compliance to an end within the period referred to in paragraph 1 ~~of this Article~~, the Member State concerned shall, without delay, take all appropriate measures to restrict or prohibit the product being made available on the **national** market or to ensure that it is recalled or withdrawn from the **national** market. That Member State shall inform the Commission and the other Member States, without delay, of those measures, by means of the electronic system referred to in Article 100.

(a) The economic operator shall take any appropriate corrective action pursuant to paragraph 1 or 2 throughout the Union in respect of all the devices concerned that they have made available on the market, unless a competent authority takes other appropriate measures.

3. In order to ensure the uniform application of this Article, the Commission may, by means of implementing acts, specify appropriate measures to be taken by competent authorities to address given types of non-compliance. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

Article 98

Preventive health protection measures

1. Where a Member State, after having performed an evaluation which indicates a potential risk related to a device or a specific category or group of devices, considers that, in order to protect the health and safety of patients, users or other persons or other aspects of public health, the making available on the market or putting into service of a device or a specific category or group of devices should be prohibited, restricted or made subject to particular requirements or that such device or category or group of devices should be withdrawn from the market or recalled, it may take any necessary and justified measures.

2. The Member State referred to in paragraph 1 shall immediately notify the Commission and all other Member States, giving the reasons for its decision, by means of the electronic system referred to in Article 100.

3. The Commission, in consultation with the MDCG and, where necessary, the economic operators concerned, shall assess the national measures taken. The Commission may decide, by means of implementing acts, whether the national measures are justified or not. In the absence of a Commission decision within six months of their notification, the national measures shall be considered to be justified. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

4. Where the assessment referred to in paragraph 3 of this Article demonstrates that the making available on the market or putting into service of a device, specific category or group of devices should be prohibited, restricted or made subject to particular requirements or that such device or category or group of devices should be withdrawn from the market or recalled in all Member States in order to protect the health and safety of patients, users or other persons or other aspects of public health, the Commission may adopt implementing acts to take the necessary and duly justified measures. ~~Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).~~

Article 99

Good administrative practice

1. Any measure adopted by the competent authorities of the Member States pursuant to Articles 95 to 98 shall state the exact grounds on which it is based. Where such a measure is addressed to a specific economic operator, the competent authority shall notify without delay the economic operator concerned of that measure, and shall at the same time inform that economic operator of the remedies available under the law or the administrative practice of the Member State concerned and of the time limits to which such remedies are subject. Where the measure is of general applicability, it shall be appropriately published.

2. Except in cases where immediate action is necessary for reasons of unacceptable risk to human health or safety, the economic operator concerned shall be given the opportunity to make submissions to the competent authority within an appropriate period of time that is clearly defined before any measure is adopted.

Where action has been taken without the economic operator having had the opportunity to make submissions as referred to in the first subparagraph, it shall be given the opportunity to make submissions as soon as possible and the action taken shall be reviewed promptly thereafter.

Proposal of Regulation (EU) 2017/745 Articles

3. Any measure adopted shall be immediately withdrawn or amended upon the economic operator's demonstrating that it has taken effective corrective action and that the device is in compliance with the requirements of this Regulation.

4. Where a measure adopted pursuant to Articles 95 to 98 concerns a device for which a notified body has been involved in the conformity assessment, the competent authorities shall by means of the electronic system referred to in Article 100 inform the relevant notified body and the authority responsible for the notified body of the measure taken.

Article 100

Electronic system on market surveillance

1. The Commission, in collaboration with the Member States, shall set up and manage an electronic system to collate and process the following information:

- (a) summaries of the results of the surveillance activities referred to in Article 93(4);
- (b) the final inspection report referred to in Article 93(7);
- (c) information in relation to devices presenting an unacceptable risk to health and safety as referred to in Article 95(2), (4) and (6);
- (d) information in relation to non-compliance of products as referred to in Article 97(2);
- (e) information in relation to the preventive health protection measures referred to in Article 98(2);
- (f) summaries of the results of the reviews and assessments of the market surveillance activities of the Member States referred to in 93(8).

2. The information referred to in paragraph 1 of this Article shall be immediately transmitted through the electronic system to all competent authorities concerned and, where applicable, to the notified body that issued a certificate in accordance with Article 56 for the device concerned and be accessible to the Member States and to the Commission.

3. Information exchanged between Member States shall not be made public where to do so might impair market surveillance activities and co-operation between Member States.

Chapter VIII

Cooperation between Member States, the MDCG [Medical Device Coordination Group], expert laboratories, expert panels and device registers, conflicts of interest and international cooperation

SECTION 1

Cooperation between Member States, the MDCG, expert panels, expert laboratories and device registers and conflict of interests

Article 101

Competent authorities

The Member States shall designate the competent authority or authorities responsible for the implementation and practical application of this Regulation. They shall ensure that those authorities are entrusted [entrust their authorities] with sufficient powers, resources, equipment and knowledge to effectively and efficiently perform [necessary for the proper performance of] their tasks pursuant to this Regulation. The Member States shall communicate the names and contact details of the competent authorities to the Commission which shall publish a list of competent authorities.

Article 102
Cooperation

1. The competent authorities of the Member States shall cooperate with each other and with the Commission. The Commission shall provide for the organisation of exchanges of information necessary to enable this Regulation to be applied uniformly.

~~2. Member States shall with the support of the Commission participate, where appropriate, in initiatives developed at international level with the aim of ensuring cooperation between regulatory authorities in the field of medical devices.~~

Article 103
Medical Device Coordination Group

1. A Medical Device Coordination Group ('MDCG') is hereby established.

2. Each Member State shall appoint to the MDCG, for a three-year term which may be renewed, one member and one alternate each with expertise in the field of medical devices, and one member and one alternate with expertise in the field of *in vitro* diagnostic medical devices. A Member State may choose to appoint only one member and one alternate, each with expertise in both fields.

The members of the MDCG shall be chosen for their competence and experience in the field of medical devices and *in vitro* diagnostic medical devices. They shall represent the competent authorities of the Member States ~~and the authorities responsible for notified bodies of the Member States, as applicable.~~
~~The names and affiliation of members shall be made public by the Commission.~~

The alternates shall represent and vote for the members in their absence.

3. The MDCG shall meet at regular intervals and, where the situation requires, upon request by the Commission or a Member State. The meetings shall be attended either by the members appointed for their role and expertise in the field of medical devices, or by the members appointed for their expertise in the field of *in vitro* diagnostic medical devices, or by the members appointed for their expertise in both fields, or their alternates, as appropriate.

4. The MDCG shall use its best endeavours to reach consensus. If such consensus cannot be reached, the MDCG shall decide by a majority of its members. Members with diverging positions may request that their positions and the grounds on which they are based are recorded in the MDCG's position.

5. The MDCG shall be chaired by a representative of the Commission. The chair shall not take part in votes of the MDCG.

6. The MDCG may invite, on a case-by-case basis, experts and other third parties to attend meetings or provide written contributions.

7. The MDCG ~~shall establish a sub-group with members representing the authorities responsible for notified bodies and~~ may establish other standing or temporary sub-groups. Where appropriate, ~~representatives of the coordination group referred to in Article 49 and~~ organisations representing the interests of the medical device industry, ~~in particular of micro, small and medium-sized enterprises within the meaning of Recommendation 2003/361/EC,~~ healthcare professionals, laboratories, patients and consumers at Union level shall be invited to ~~the MDCG and its [such]~~ sub-groups in the capacity of observers.

8. The MDCG shall establish its rules of procedure which shall, in particular, lay down procedures for the following:

- the adoption of opinions or recommendations or other positions, including in cases of urgency;
- the delegation of tasks to reporting and co-reporting members;
- the implementation of Article 107 regarding conflict of interests;

Proposal of Regulation (EU) 2017/745 Articles

– the functioning of sub-groups.

9. The MDCG shall have the tasks laid down in Article 105 of this Regulation ~~and Article 99 of Regulation (EU) 2017/746.~~

Article 104

Support by the Commission

The Commission shall support the functioning of the cooperation between national competent authorities. It shall, in particular, provide for the organisation of exchanges of experience between the competent authorities and provide technical, scientific and logistic support to the MDCG and its sub-groups. ~~It shall organise the meetings of the MDCG and its sub-groups, participate in those meetings and ensure the appropriate follow-up.~~

Article 105

Tasks of the MDCG

~~Under this Regulation, the MDCG shall have the following tasks:~~

- ~~(a) to contribute to the assessment of applicant conformity assessment bodies and notified bodies pursuant to the provisions set out in Chapter IV;~~
- ~~(b) to advise the Commission, at its request, in matters concerning the coordination group of notified bodies as established pursuant to Article 49;~~
- ~~(c) to contribute to the development of guidance aimed at ensuring effective and harmonised implementation of this Regulation, in particular regarding the designation and monitoring of notified bodies, application of the general safety and performance requirements and conduct of clinical evaluations and investigations by manufacturers, assessment by notified bodies and vigilance activities;~~
- ~~(d) to contribute to the continuous monitoring of technical progress and assessment of whether the general safety and performance requirements laid down in this Regulation and Regulation (EU) 2017/746 are adequate to ensure safety and performance of devices, and thereby contribute to identifying whether there is a need to amend Annex I to this Regulation;~~
- ~~(e) to contribute to the development of device standards, of CS and of scientific guidelines, including product specific guidelines, on clinical investigation of certain devices in particular implantable devices and class III devices;~~
- ~~(f) to assist the competent authorities of the Member States in their coordination activities in particular in the fields of classification and the determination of the regulatory status of devices, clinical investigations, vigilance and market surveillance including the development and maintenance of a framework for a European market surveillance program with the objective of achieving efficiency and harmonisation of market surveillance in the Union, in accordance with Article 93;~~
- ~~(g) to provide advice, either on its own initiative or at request of the Commission, in the assessment of any issue related to the implementation of this Regulation;~~
- ~~(h) to contribute to harmonised administrative practice with regard to devices in the Member States.~~

In addition to the tasks assigned to it in other provisions of this Regulation and Regulation (EU) 2017/746, the MDCG shall in particular assist the Commission in the assessment of any issue related to the implementation of those Regulations and contribute to the development of guidance aimed at ensuring their effective and harmonised implementation.

Article 106

Expert panels [Provision of scientific, technical and clinical opinions and advice]

~~1. The Commission shall, by means of implementing acts and in consultation with the MDCG, make provision for expert panels to be designated for the assessment of the clinical evaluation in relevant medical fields as referred to in paragraph 9 of this Article and to provide views in accordance with Article 48(6) of Regulation (EU) 2017/746 on the performance evaluation of certain *in vitro* diagnostic medical devices and, where necessary, for categories or groups of devices, or for specific hazards relating to categories or groups of devices, observing the principles of highest scientific competence, impartiality, independence and transparency. The same principles shall apply where the Commission decides to appoint expert laboratories in accordance with paragraph 7 of this Article.~~

~~2. Expert panels and expert laboratories may be designated in areas where the Commission, in consultation with the MDCG, has identified a need for the provision of consistent scientific, technical and/or clinical advice or laboratory expertise in relation to the implementation of this Regulation. Expert panels and expert laboratories may be appointed on a standing or temporary basis.~~

~~3. Expert panels shall consist of advisors appointed by the Commission on the basis of up-to-date clinical, scientific or technical expertise in the field and with a geographical distribution that reflects the diversity of scientific and clinical approaches in the Union. The Commission shall determine the number of members of each panel in accordance with the requisite needs.~~

~~The members of expert panels shall perform their tasks with impartiality and objectivity. They shall neither seek nor take instructions from notified bodies or manufacturers. Each member shall draw up a declaration of interests which shall be made publicly available.~~

~~The Commission shall establish systems and procedures to actively manage and prevent potential conflicts of interest.~~

~~4. Expert panels shall take into account relevant information provided by stakeholders including patients' organisations and healthcare professionals when preparing their scientific opinions.~~

~~5. The Commission, following consultation with the MDCG, may appoint advisors to expert panels following publication in the *Official Journal of the European Union* and on the Commission website following a call for expressions of interest. Depending on the type of task and the need for specific expertise, advisors may be appointed to the expert panels for a maximum period of three years and their appointment may be renewed.~~

~~6. The Commission, following consultation with the MDCG, may include advisors on a central list of available experts who, whilst not being formally appointed to a panel, are available to provide advice and to support the work of the expert panel as needed. That list shall be published on the Commission website.~~

~~7. The Commission may, by means of implementing acts and following consultation with the MDCG, designate expert laboratories, on the basis of their expertise in:~~

- ~~— physico-chemical characterisation, or~~
- ~~— microbiological, biocompatibility, mechanical, electrical, electronic or non-clinical biological and toxicological testing~~
- ~~of specific devices, categories or groups of devices.~~

~~The Commission shall only designate expert laboratories for which a Member State or the Joint Research Centre has submitted an application for designation.~~

~~8. Expert laboratories shall satisfy the following criteria:~~

- ~~(a) have adequate and appropriately qualified staff with adequate knowledge and experience in the field of the devices for which they are designated;~~

Proposal of Regulation (EU) 2017/745 Articles

~~(b) possess the necessary equipment to carry out the tasks assigned to them;~~

~~(c) have the necessary knowledge of international standards and best practices;~~

~~(d) have an appropriate administrative organisation and structure;~~

~~(e) ensure that their staff observe the confidentiality of information and data obtained in carrying out their tasks.~~

~~9. Expert panels appointed for clinical evaluation in relevant medical fields shall fulfil the tasks provided for in Article 54(1) and Article 61(2) and Section 5.1 of Annex IX or Section 6 of Annex X, as applicable.~~

~~10. Expert panels and expert laboratories may have the following tasks, depending on the requisite needs:~~

~~(a) to provide scientific, technical and clinical assistance to the Commission and the MDCG in relation to the implementation of this Regulation;~~

~~(b) to contribute to the development and maintenance of appropriate guidance and CS for:~~

~~—clinical investigations,~~

~~—clinical evaluation and PMCF,~~

~~—performance studies,~~

~~—performance evaluation and post-market performance follow-up,~~

~~—physico-chemical characterisation, and~~

~~—microbiological, biocompatibility, mechanical, electrical, electronic or non-clinical toxicological testing for specific devices, or a category or group of devices, or for specific hazards related to a category or group of devices;~~

~~(c) to develop and review clinical evaluation guidance and performance evaluation guidance for performance of conformity assessment in line with the state of the art with regard to clinical evaluation, performance evaluation, physico-chemical characterisation, and microbiological, biocompatibility, mechanical, electrical, electronic or non-clinical toxicological testing;~~

~~(d) to contribute to the development of standards at international level, ensuring that such standards reflect the state of the art;~~

~~(e) to provide opinions in response to consultations by manufacturers in accordance with Article 61(2), notified bodies and Member States in accordance with paragraphs 11 to 13 of this Article.~~

~~(f) to contribute to identification of concerns and emerging issues on the safety and performance of medical devices;~~

~~(g) to provide views in accordance with Article 48(4) of Regulation (EU) 2017/746 on the performance evaluation of certain *in vitro* diagnostic medical devices.~~

~~11. The Commission, shall facilitate the access of Member States and notified bodies and manufacturers to advice provided by expert panels and expert laboratories concerning, *inter alia*, the criteria for an appropriate data set for assessment of the conformity of a device, in particular with regard to the clinical data required for clinical evaluation, with regard to physico-chemical characterisation, and with regard to microbiological, biocompatibility, mechanical, electrical, electronic and non-clinical toxicological testing.~~

~~12. When adopting its scientific opinion in accordance with paragraph 9, the members of the expert panels shall use their best endeavours to reach consensus. If consensus cannot be reached, the expert panels shall decide by a majority of their members, and the scientific opinion shall mention the divergent positions and the grounds on which they are based.~~

~~The Commission shall publish the scientific opinion and advice delivered in accordance with paragraphs 9 and 11 of this Article, ensuring consideration of aspects of confidentiality as set out in Article 109. The~~

Proposal of Regulation (EU) 2017/745 Articles

~~clinical evaluation guidance referred to in point (c) of paragraph 10 shall be published following consultation with the MDCG.~~

~~13. The Commission may require manufacturers and notified bodies to pay fees for the advice provided by expert panels and expert laboratories. The structure and the level of fees as well as the scale and structure of recoverable costs shall be adopted by the Commission by means of implementing acts, taking into account the objectives of the adequate implementation of this Regulation, protection of health and safety, support of innovation and cost effectiveness and the necessity to achieve active participation in the expert panels. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).~~

~~14. The fees payable to the Commission in accordance with the procedure under paragraph 13 of this Article shall be set in a transparent manner and on the basis of the costs for the services provided. The fees payable shall be reduced in the case of a clinical evaluation consultation procedure initiated in accordance with point (c) of Section 5.1 of Annex IX involving a manufacturer who is a micro, small or medium sized enterprise within the meaning of Recommendation 2003/361/EC.~~

~~15. The Commission is empowered to adopt delegated acts in accordance with Article 115 to amend the tasks of expert panels and expert laboratories referred to in paragraph 10 of this Article.~~

1. The Commission shall, by means of implementing acts and following consultation with the MDCG, make provision for expert panels to be designated to provide scientific, clinical, technical or regulatory opinions and advice in relation to the implementation of this Regulation and of Regulation (EU) 2017/746 to the Commission, the MDCG, Member States, notified bodies or manufacturers.

Expert panels may be designated on a standing or temporary basis.

2. Expert panels shall consist of experts with proven up-to-date clinical, scientific, technical or regulatory expertise in the field of medical devices or in vitro diagnostic medical devices reflecting the diversity of scientific and clinical approaches in the Union.

Experts shall be appointed following publication of a call for expressions of interest. Depending on the type of task and the need for specific expertise, experts may be appointed to the panels for a maximum period of three years and their appointment may be renewed.

When expert panels are requested to provide an opinion on the regulatory status of a product which involves aspects concerning the borderline with other types of products, experts with expertise in the field of the relevant other products shall be involved.

3. The experts shall observe the principles of highest scientific competence and perform their tasks with impartiality, objectivity and transparency. They shall neither seek nor take instructions from notified bodies or manufacturers. Each expert shall draw up a declaration of interests, which shall be made publicly available.

4. Expert panels shall take into account relevant information provided by stakeholders including patients' organisations and healthcare professionals' associations.

5. Experts may be included on a list of available experts who, whilst not being formally appointed to a panel, are available to provide advice and to support the work of the expert panels as needed.

6. Experts from notified bodies shall not be involved in the clinical evaluation consultation procedure provided for in Article 54(1).

7. In addition to the tasks assigned to them in other provisions of this Regulation and of Regulation (EU) 2017/746, the expert panels may have the following tasks:

(a) to provide scientific, clinical, technical and regulatory advice to the Commission, the MDCG, Member States or notified bodies in relation to the implementation of this Regulation or Regulation (EU) 2017/746;

(b) to contribute to the development and maintenance of appropriate guidance and CS supporting the implementation of this Regulation or Regulation (EU) 2017/746;

Proposal of Regulation (EU) 2017/745 Articles

(c) to contribute to the development of standards at Union or international level, ensuring that such standards reflect the state of the art;

(d) to contribute to the identification of concerns and emerging issues concerning the safety and performance of medical devices, including in vitro diagnostic medical devices.

8. When adopting their opinions or advice, the members of the expert panels shall use their best endeavours to reach consensus. If consensus cannot be reached, the expert panels shall decide by a majority of their members, and the opinion or advice shall mention the divergent positions and the grounds on which they are based.

9. The Commission shall require manufacturers and notified bodies to pay fees for opinions and advice provided by expert panels. The structure and the level of the fees, as well as the scale and type of recoverable costs, shall be established by the Commission by means of implementing acts, taking into account the objectives of the adequate implementation of this Regulation, protection of health and safety, support of innovation and cost-effectiveness and the necessity to achieve active participation in the expert panels. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

The fees referred to in the first subparagraph shall be established in a transparent manner and on the basis of the costs for the services provided. The fees shall be reduced for manufacturers which are micro, small or medium- sized enterprises within the meaning of Recommendation 2003/361/EC, including in the case of a clinical evaluation consultation procedure initiated in accordance with Section 5.1, point (c), of Annex IX involving a manufacturer who is a micro, small or medium-sized enterprise within the meaning of Recommendation 2003/361/EC. The fees related to the opinions and advice provided by expert panels are payable to EMA pursuant to Article 30, point (f), of Regulation (EU) 2022/123.

10. The Commission is empowered to adopt delegated acts in accordance with Article 115 to amend paragraph 7 of this Article by adding, adapting or removing tasks from the list of tasks of the expert panels.

Article 106a Expert laboratories

1. The Commission may, by means of implementing acts and following consultation with the MDCG, designate expert laboratories, on the basis of their expertise in the field of testing of medical devices, such as physico-chemical characterisation, microbiological, biocompatibility, mechanical, electrical, electronic or non-clinical toxicological testing of specific devices, categories or groups of devices.

The Commission shall only designate expert laboratories for which a Member State or the Joint Research Centre has submitted an application for designation.

2. Expert laboratories shall satisfy the following criteria:

- (a) have adequate and appropriately qualified staff with adequate knowledge and experience in the field of the devices for which they are designated;
- (b) have at their disposal the necessary equipment to carry out the tasks assigned to them;
- (c) have the necessary knowledge of international standards and best practices;
- (d) have an appropriate administrative organisation and structure;
- (e) ensure that their staff observe the confidentiality of information and data obtained in carrying out their tasks;
- (f) act in the public interest and in an independent manner.

3. Expert laboratories may have the following tasks:

Proposal of Regulation (EU) 2017/745 Articles

- (a) to provide scientific and technical assistance to the Commission, the MDCG, the EMA, Member States and notified bodies in relation to the implementation of this Regulation;
 - (b) to contribute to the development and maintenance of appropriate guidance and CS supporting the implementation of this Regulation;
 - (c) to contribute to the development of standards at Union or international level, ensuring that such standards reflect the state of the art;
 - (d) to contribute to identification of concerns and emerging issues on the safety and performance of medical devices;
 - (e) to provide scientific and technical assistance to Member States and the Commission in vigilance and market surveillance activities.
4. The Commission is empowered to adopt delegated acts in accordance with Article 115 to amend paragraph 3 of this Article by adding, adapting or removing tasks from the list of tasks of the expert laboratories.
5. The expert laboratories may charge fees in accordance with predetermined and transparent terms and conditions for scientific and technical assistance provided to the EMA, Member States or notified bodies for carrying out the requested task. For the provision of services within the public interest on request of the Commission or the MDCG, the expert laboratories may be granted a Union financial contribution.

Article 106b Support by the EMA

1. The EMA shall, on behalf of the Commission, provide scientific, technical and administrative support to the national competent authorities designated under this Regulation and under Regulation (EU) 2017/746 to facilitate the exchange of experience, cooperation and coordination with a view to ensuring a uniform application of such Regulations, in particular in the following areas:
- (a) regulatory status of products and classification of devices in accordance with Articles 4, 4a, 51, 51a and 51b of this Regulation and Articles 3, 3a, 47, 47a and 47b of Regulation (EU) 2017/746;
 - (b) derogations from the applicable conformity assessment procedures in accordance with Articles 59 and 59a of this Regulation and Articles 54 and 54a of Regulation (EU) 2017/746;
 - (c) clinical evaluation, clinical investigations, performance evaluation and performance studies in accordance with Chapter VI of this Regulation and Chapter VI of Regulation (EU) 2017/746, including support to the coordinating Member State for the coordinated assessment procedure for clinical investigations and performance studies referred to in Article 78 of this Regulation and Article 74 of Regulation (EU) 2017/746;
 - (d) vigilance and market surveillance in accordance with Chapter VII of this Regulation and Chapter VII of Regulation (EU) 2017/746, including support to the coordinating competent authority for the coordinated procedure referred to in Article 89(9) of this Regulation and Article 84(9) of Regulation (EU) 2017/746.
2. The EMA shall provide scientific, technical and administrative support to the Commission for the establishment of Union regulatory sandboxes in accordance with Article 59c of this Regulation and Article 54c of Regulation (EU) 2017/746.
3. The EMA shall set up a support scheme for manufacturers of medical devices and in vitro diagnostic medical devices, which are micro, small and medium- sized within the meaning of Recommendation 2003/361/EC, regarding the requirements of this Regulation and of Regulation (EU) 2017/746.
4. The EMA shall have access to Eudamed and any electronic system referred to in Article 33(2) of Regulation (EU) 2017/745 or in Article 30(2) of Regulation (EU) 2017/746 that is not included in Eudamed.

Article 107

Conflict of interests

1. Members of the MDCG, its sub-groups, and members of expert panels and expert laboratories shall not have financial or other interests in the medical device industry which could affect their impartiality. They shall undertake to act in the public interest and in an independent manner. They shall declare any direct or indirect interests they may have in the medical device industry and update that declaration whenever a relevant change occurs. The declaration of interests shall be made publicly available ~~on the Commission website~~. This Article shall not apply to the representatives of stakeholder organisations participating in the MDCG or its sub-groups.

2. Experts and other third parties invited by the MDCG on a case-by-case basis shall declare any interests they may have in the issue in question.

Article 108

Device registers and data banks

The Commission, the Member States ~~and the EMA~~ shall take all appropriate measures to encourage the establishment of registers and databanks for specific types of devices ~~setting common principles to collect comparable information~~, where appropriate including in vitro diagnostic medical devices, ~~setting common principles to collect comparable information~~. Such registers and databanks shall contribute to the independent evaluation of the long-term safety and performance of devices, or the traceability of implantable devices, or all of such characteristics.

SECTION 2

International cooperation

Article 108a

International regulatory and administrative cooperation

1. The Commission shall pursue international regulatory cooperation in the field of medical devices and in vitro diagnostic medical devices with a view to promoting a high level of protection of public health and patient safety, fostering innovation and enhancing efficiency of regulatory compliance through global convergence. For that purpose, the Commission and the Member States shall contribute to the development and adoption of global principles, standards, and guidance which provide a high level of international convergence in the field of medical devices and in vitro diagnostic medical devices, including in relation to safety, performance, quality management systems, conformity assessment, and post-market surveillance.

2. The Commission shall participate in relevant international fora in the field of medical devices and in vitro diagnostic medical devices, including the International Medical Device Regulators Forum (IMDRF), the Medical Device Single Audit Programme (MDSAP) and International Standardisation Organisations.

3. In its activities in accordance with paragraphs 1 and 2, the Commission shall be supported by experts nominated by the Member States in view of their competence in medical devices or in vitro diagnostic medical devices.

4. The Commission may sign administrative arrangements with authorities of third countries and with international organisations for the purpose of regulatory cooperation in the field of medical devices and in vitro diagnostic medical devices, including:

- (a) exchange of information and best practices;
- (b) joint or coordinated inspections and assessments;
- (c) coordinated actions relating to safety issues, including recalls or safety communications.

Proposal of Regulation (EU) 2017/745 Articles

Where the Commission signs such administrative arrangements which include the exchange of information or data, the administrative arrangement shall provide for protection of such information or data in accordance with Article 109 of this Regulation or in Article 102 of Regulation (EU) 2017/746, as applicable.

5. The Union shall finance the activities of the Commission and the Member States relating to international cooperation as referred to in this Article and the reliance mechanisms referred to in Article 108b.

Article 108b

Reliance mechanisms

1. The Commission, may participate in bilateral or multilateral reliance mechanisms or reliance programmes in the field of medical devices and in vitro diagnostic medical devices, which shall enable the use of assessments, inspections, and other regulatory decisions carried out or taken by regulatory authorities of third countries or international organisations or international bodies, provided that the following conditions are fulfilled:

(a) the reliance mechanism or reliance programme ensures a level of health and safety protection equivalent to that required under this Regulation or Regulation (EU) 2017/746, as applicable;

(b) effective arrangements for mutual exchange of information, transparency, and oversight are in place, providing for the confidentiality of information and data referred to in Article 109 of this Regulation or Article 102 of Regulation (EU) 2017/746, as applicable.

2. The Commission may invite Member States to nominate experts, in view of their competence in medical devices or in vitro diagnostic medical devices, who participate in reliance mechanisms or reliance programmes referred to in paragraph 1.

3. Reliance mechanisms or reliance programmes referred to in paragraph 1 shall be taken into consideration by competent authorities, economic operators or notified bodies in the framework of the implementation of this Regulation or of Regulation (EU) 2017/746, as applicable.

4. The Commission may adopt implementing acts establishing detailed rules for the recognition of reliance mechanisms or reliance programmes, which may include conditions for participation of notified bodies to reliance mechanisms or reliance programmes, requirements relating to the scope of assessments, inspections or other regulatory decisions carried out or taken in the course of reliance mechanisms or reliance programmes, and procedural safeguards for manufacturers. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).

Article 108c

Capacity building and technical assistance

1. The Commission may support the strengthening of regulatory capacity in third countries, including through the provision of technical assistance, training, exchange of experts, and dissemination of best practices.

2. The Commission may invite Member States to nominate experts, in view of their competence in medical devices or in vitro diagnostic medical devices, to participate in its activities referred to in paragraph 1.3.

3. The activities referred to in paragraph 1 may be financed through relevant Union programmes or external action instruments.

Chapter IX
Confidentiality, data protection, funding and penalties

Article 109
Confidentiality

1. Unless otherwise provided for in this Regulation and without prejudice to existing national provisions and practices in the Member States on confidentiality, all parties involved in the application of this Regulation shall respect the confidentiality of information and data obtained in carrying out their tasks in order to protect the following:

- (a) personal data, in accordance with Article 110;
 - (b) commercially confidential information and trade secrets of a natural or legal person, including intellectual property rights; unless disclosure is in the public interest;
 - (c) the effective implementation of this Regulation, in particular for the purpose of inspections, investigations or audits.
2. Without prejudice to paragraph 1, information exchanged on a confidential basis between competent authorities and between competent authorities and the Commission shall not be disclosed without the prior agreement of the originating authority.
3. Paragraphs 1 and 2 shall not affect the rights and obligations of the Commission, Member States and notified bodies with regard to exchange of information and the dissemination of warnings, nor the obligations of the persons concerned to provide information under criminal law.
4. The Commission and Member States may exchange confidential information with regulatory authorities of third countries with which they have concluded bilateral or multilateral confidentiality arrangements.

Article 110
Data protection

- 1. Member States shall apply Directive 95/46/EC to the processing of personal data carried out in the Member States pursuant to this Regulation.
- 2. Regulation (EC) No 45/2001 shall apply to the processing of personal data carried out by the Commission pursuant to this Regulation.

Article 110a
Submission of information or documents

The submission of information or documents in accordance with this Regulation shall take place electronically.

Article 111
Levying of fees

- 1. This Regulation shall be without prejudice to the possibility for Member States and the Commission to levy fees for the activities set out in this Regulation, provided that the level of the fees is set in a transparent manner and on the basis of cost-recovery principles.
- 2. Member States shall inform the Commission and the other Member States at least three months before the structure and level of fees is to be adopted. The structure and level of fees shall be made publicly available on request.

Article 112

~~Funding of activities related to designation and monitoring of notified bodies~~

~~The costs associated with joint assessment activities shall be covered by the Commission. The Commission shall, by means of implementing acts, lay down the scale and structure of recoverable costs and other necessary implementing rules. Those implementing acts shall be adopted in accordance with the examination procedure referred to in Article 114(3).~~

Article 113

Penalties

The Member States shall lay down the rules on penalties applicable for infringement of the provisions of this Regulation and shall take all measures necessary to ensure that they are implemented. The penalties provided for shall be effective, proportionate, and dissuasive. The Member States shall notify the Commission of those rules and of those measures by 25 February 2021 and shall notify it without delay of any subsequent amendment affecting them.

Chapter X
Final provisions

Article 114

Committee procedure

1. The Commission shall be assisted by a Committee on Medical Devices. That Committee shall be a committee within the meaning of Regulation (EU) No 182/2011.

2. Where reference is made to this paragraph, Article 4 of Regulation (EU) No 182/2011 shall apply.

3. Where reference is made to this paragraph, Article 5 of Regulation (EU) No 182/2011 shall apply.

Where the committee delivers no opinion, the Commission shall not adopt the draft implementing act and the third subparagraph of Article 5(4) of Regulation (EU) No 182/2011 shall apply.

4. Where reference is made to this paragraph, Article 8 of Regulation (EU) No 182/2011, in conjunction with Article 4 or Article 5 thereof, as appropriate, shall apply.

Article 115

Exercise of the delegation

1. The power to adopt delegated acts is conferred on the Commission subject to the conditions laid down in this Article.

2. The power to adopt delegated acts referred to in Articles 1(5), 3, 10(4), 18(3), 19(4), 27(10), 44(11), 52(5), 56(6), 61(8), 70(8) and 106(15) shall be conferred on the Commission for a period of five years from 25 May 2017. The Commission shall draw up a report in respect of the delegation of power not later than nine months before the end of the five year period. The delegation of power shall be tacitly extended for periods of an identical duration, unless the European Parliament or the Council opposes such extension not later than three months before the end of each period.

3. The delegation of power referred to in Articles 1(5), 3, 10(4), 18(3), 19(4), 27(10), 44(11), 52(5), 56(6), 61(8), 70(8) and 106(15) may be revoked at any time by the European Parliament or by the Council. A decision to revoke shall put an end to the delegation of the power specified in that decision. It shall take effect the day following the publication of the decision in the *Official Journal of the European Union* or at a later date specified therein. It shall not affect the validity of any delegated acts already in force.

4. Before adopting a delegated act, the Commission shall consult experts designated by each Member State in accordance with the principles laid down in the Interinstitutional Agreement of 13 April 2016 on Better Law-Making.

Proposal of Regulation (EU) 2017/745 Articles

5. As soon as it adopts a delegated act, the Commission shall notify it simultaneously to the European Parliament and to the Council.

6. A delegated act adopted pursuant to Articles 1(5), 3, 10(4), 18(3), 19(4), 27(10), 44(11), 52(5), 56(6), 61(8), 70(8) and 106(15) shall enter into force only if no objection has been expressed either by the European Parliament or by the Council within a period of three months of notification of that act to the European Parliament and the Council or if, before the expiry of that period, the European Parliament and the Council have both informed the Commission that they will not object. That period shall be extended by three months at the initiative of the European Parliament or of the Council.

Article 116

~~Separate delegated acts for different delegated powers~~

~~The Commission shall adopt a separate delegated act in respect of each power delegated to it pursuant to this Regulation.~~

Article 117

Amendment to Directive 2001/83/EC

In Annex I to Directive 2001/83/EC, point 12 of Section 3.2. is replaced by the following:

'(12) Where, in accordance with the second subparagraph of Article 1(8) or the second subparagraph of Article 1(9) of Regulation (EU) 2017/745 of the European Parliament and of the Council*, a product is governed by this Directive, the marketing authorisation dossier shall include, where available, the results of the assessment of the conformity of the device part with the relevant general safety and performance requirements set out in Annex I to that Regulation contained in the manufacturer's EU declaration of conformity or the relevant certificate issued by a notified body allowing the manufacturer to affix a CE marking to the medical device.

If the dossier does not include the results of the conformity assessment referred to in the first subparagraph and where for the conformity assessment of the device, if used separately, the involvement of a notified body is required in accordance with Regulation (EU) 2017/745, the authority shall require the applicant to provide an opinion on the conformity of the device part with the relevant general safety and performance requirements set out in Annex I to that Regulation issued by a notified body designated in accordance with that Regulation for the type of device in question.

Article 118

Amendment to Regulation (EC) No 178/2002

In the third subparagraph of Article 2 of Regulation (EC) No 178/2002, the following point is added:

'(i) medical devices within the meaning of Regulation (EU) 2017/745 of the European Parliament and of the Council*.

Article 119

Amendment to Regulation (EC) No 1223/2009

In Article 2 of Regulation (EC) No 1223/2009, the following paragraph is added:

* Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices, amending Directive 2001/83/EC, Regulation (EC) No 178/2002 and Regulation (EC) No 1223/2009 and repealing Council Directives 90/385/EEC and 93/42/EEC (OJ L 117, 5.5.2017, p. 1).¹

¹ Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices, amending Directive 2001/83/EC, Regulation (EC) No 178/2002 and Regulation (EC) No 1223/2009 and repealing Council Directives 90/385/EEC and 93/42/EEC (OJ L 117, 5.5.2017, p. 1).¹

Proposal of Regulation (EU) 2017/745 Articles

'4. The Commission may, at the request of a Member State or on its own initiative, adopt the necessary measures to determine whether or not a specific product or group of products falls within the definition 'cosmetic product'. Those measures shall be adopted in accordance with the regulatory procedure referred to in Article 32(2).'

Article 120

Transitional provisions

1. From 26 May 2021, any publication of a notification in respect of a notified body in accordance with Directives 90/385/EEC and 93/42/EEC shall become void.

2. Certificates issued by notified bodies in accordance with Directives 90/385/EEC and 93/42/EEC prior to 25 May 2017 shall remain valid until the end of the period indicated on the certificate, except for certificates issued in accordance with Annex 4 to Directive 90/385/EEC or Annex IV to Directive 93/42/EEC which shall become void at the latest on 27 May 2022.

Certificates issued by notified bodies in accordance with Directives 90/385/EEC and 93/42/EEC from 25 May 2017 that were still valid on 26 May 2021 and that have not been withdrawn afterwards shall remain valid after the end of the period indicated on the certificate until the date set out in paragraph 3a of this Article applicable for the relevant risk class of the devices. Certificates issued by notified bodies in accordance with those Directives from 25 May 2017 that were still valid on 26 May 2021 and that have expired before 20 March 2023 shall be considered to be valid until the dates set out in paragraph 3a of this Article only if one of the following conditions is fulfilled:

- (a) before the date of expiry of the certificate, the manufacturer and a notified body have signed a written agreement in accordance with Section 4.3, second subparagraph, of Annex VII to this Regulation for the conformity assessment in respect of the device covered by the expired certificate or in respect of a device intended to substitute that device;
- (b) a competent authority of a Member State has granted a derogation from the applicable conformity assessment procedure in accordance with Article 59(1) of this Regulation or has required the manufacturer, in accordance with Article 97(1) of this Regulation, to carry out the applicable conformity assessment procedure.

3. By way of derogation from Article 5 and provided the conditions set out in paragraph 3c of this Article are met, devices referred to in paragraphs 3a and 3b of this Article may be placed on the market or put into service until the dates set out in those paragraphs.

3a. Devices which have a certificate that was issued in accordance with Directive 90/385/EEC or Directive 93/42/EEC and that is valid by virtue of paragraph 2 of this Article may be placed on the market or put into service until the following dates:

- (a) 31 December 2027, for all class III devices, and for class IIb implantable devices except sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors;
- (b) 31 December 2028, for class IIb devices other than those covered by point (a) of this paragraph, for class IIa devices, and for class I devices placed on the market in sterile condition or having a measuring function.

3b. Devices for which the conformity assessment procedure pursuant to Directive 93/42/EEC did not require the involvement of a notified body, for which the declaration of conformity was drawn up prior to 26 May 2021 and for which the conformity assessment procedure pursuant to this Regulation requires the involvement of a notified body, may be placed on the market or put into service until 31 December 2028.

3c. Devices referred to in paragraphs 3a and 3b of this Article may be placed on the market or put into service until the dates referred to in those paragraphs only if the following conditions are met:

- (a) those devices continue to comply with Directive 90/385/EEC or Directive 93/42/EEC, as applicable;

Proposal of Regulation (EU) 2017/745 Articles

(b) there are no significant changes in the design and intended purpose;

(c) the devices do not present an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health;

(d) no later than 26 May 2024, the manufacturer has put in place a quality management system in accordance with Article 10(9);

(e) no later than 26 May 2024, the manufacturer or the authorised representatives has lodged a formal application with a notified body in accordance with Section 4.3, first subparagraph, of Annex VII for conformity assessment in respect of a device referred to in paragraph 3a or 3b of this Article or in respect of a device intended to substitute that device, and, no later than 26 September 2024, the notified body and the manufacturer have signed a written agreement in accordance with Section 4.3, second subparagraph, of Annex VII.

3d. By way of derogation from paragraph 3 of this Article, the requirements of this Regulation relating to post-market surveillance, market surveillance, vigilance, registration of economic operators and of devices shall apply to devices referred to in paragraphs 3a and 3b of this Article in place of the corresponding requirements in Directives 90/385/EEC and 93/42/EEC.

3e. Without prejudice to Chapter IV and paragraph 1 of this Article, the notified body that issued the certificate referred to in paragraph 3a of this Article shall continue to be responsible for the appropriate surveillance in respect of the applicable requirements relating to the devices it has certified, unless the manufacturer has agreed with a notified body designated in accordance with Article 42 that the latter shall carry out such surveillance.

No later than 26 September 2024, the notified body that has signed the written agreement referred to in paragraph 3c, point (e), of this Article shall be responsible for the surveillance in respect of the devices covered by the written agreement. Where the written agreement covers a device intended to substitute a device which has a certificate that was issued in accordance with Directive 90/385/EEC or Directive 93/42/EEC, the surveillance shall be conducted in respect of the device that is being substituted.

The arrangements for the transfer of the surveillance from the notified body that issued the certificate to the notified body designated in accordance with Article 42 shall be clearly defined in an agreement between the manufacturer and the notified body designated in accordance with Article 42 and, where practicable, the notified body that issued the certificate. The notified body designated in accordance with Article 42 shall not be responsible for conformity assessment activities carried out by the notified body that issued the certificate.

3f. By way of derogation from Article 5, class III custom-made implantable devices may be placed on the market or put into service until 26 May 2026 without a certificate issued by a notified body in accordance with the conformity assessment procedure referred to in Article 52(8), second subparagraph, provided that no later than 26 May 2024, the manufacturer or the authorised representatives has lodged a formal application with a notified body in accordance with Section 4.3, first subparagraph, of Annex VII for conformity assessment, and, no later than 26 September 2024, the notified body and the manufacturer have signed a written agreement in accordance with Section 4.3, second subparagraph, of Annex VII.

4. Devices lawfully placed on the market pursuant to Directives 90/385/EEC and 93/42/EEC prior to 26 May 2021, and devices lawfully placed on the market from 26 May 2021 pursuant to paragraphs 3, 3a, 3b and 3f of this Article, may continue to be made available on the market or put into service.

5. By way of derogation from Directives 90/385/EEC and 93/42/EEC, devices which comply with this Regulation may be placed on the market prior to 26 May 2021.

6. By way of derogation from Directives 90/385/EEC and 93/42/EEC, conformity assessment bodies which comply with this Regulation may be designated and notified prior to 26 May 2021. Notified bodies which are designated and notified in accordance with this Regulation may carry out the conformity assessment procedures laid down in this Regulation and issue certificates in accordance with this Regulation prior to 26 May 2021.

Proposal of Regulation (EU) 2017/745 Articles

7. As regards devices subject to the consultation procedure laid down in Article 54, paragraph 5 of this Article shall apply provided that the necessary appointments to the MDCG and expert panels have been made.

9. Authorisations granted by the competent authorities of the Member States in accordance with Article 9(9) of Directive 90/385/EEC or Article 11(13) of Directive 93/42/EEC shall keep the validity indicated in the authorisation.

10. Devices falling within the scope of this Regulation in accordance with point (g) of Article 1(6) which have been legally placed on the market or put into service in accordance with the rules in force in the Member States prior to 26 May 2021 may continue to be placed on the market and put into service in the Member States concerned

11. Clinical investigations which have started to be conducted in accordance with Article 10 of Directive 90/385/EEC or Article 15 of Directive 93/42/EEC prior to 26 May 2021 may continue to be conducted. As of 26 May 2021, however, the reporting of serious adverse events and device deficiencies shall be carried out in accordance with this Regulation.

12. Until the Commission has designated, pursuant to Article 27(2), issuing entities, GS1, HIBCC and ICCBBA shall be considered to be designated issuing entities.

13. Article 10a shall also apply to devices referred to in paragraphs 3a and 3b of this Article.

14. By way of derogation from Article 5 and from paragraphs 3 to 3e of this Article, a device as referred to in paragraph 3a or 3b of this Article that meets the criteria for an orphan device as referred to in Article 52a(3) may be placed on the market or put into service after the dates referred to in paragraphs 3a and 3b of this Article if the following conditions are met:

(a) an expert panel referred to in Article 106 has issued an opinion confirming the fulfilment of the criteria for an orphan device as referred to in Article 52a(3);

(b) there are no significant changes in the design and intended purpose of the device;

(c) the device does not present an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health.

The requirements of this Regulation, with the exception of Chapters IV, V and VI, shall apply to a device referred to in the first subparagraph.

By way of derogation from Article 86(1), manufacturers of class IIa devices, class IIb devices and class III devices placed on the market or put into service in accordance with this paragraph shall update the PSUR at least annually. On an annual basis, the manufacturer shall submit the PSUR and, where applicable, an update of the summary of safety and clinical performance to the competent authority of the Member State in which it is established.

The competent authority of the Member State may require the manufacturer to conduct defined post-market surveillance or PMCF activities within a specified period of time to generate additional clinical data to confirm the safety and performance of the device and to evaluate any undesirable side-effects and the acceptability of the benefit-risk ratio.

Devices placed on the market or put into service in accordance with this paragraph, which do not have a valid certificate in accordance with paragraph 2, shall not bear the CE marking. In its EU declaration of conformity, the manufacturer shall make reference to the fact that the device is an orphan device placed on the market or put into service in accordance with this provision.

The manufacturer shall inform the intended users about the fact that the device is an orphan device placed on the market or put into service in accordance with this provision, where applicable, in the summary of safety and clinical performance and in the instructions for use or any other accompanying documentation. At least every 10 years, the manufacturer shall request an opinion from an expert panel referred to in Article 106 confirming the fulfilment of the criteria for an orphan device as referred to in Article 52a(3).

15. As regards devices for which a conformity assessment procedure is pending on ...[OP please insert the date = six months after the date of entry into force of this Regulation], or for which a certificate is issued by a notified body before that date, the manufacturer and the notified body may agree to continue applying the provisions of this Regulation in the form applicable before ...[OP please insert the date = six months after the date of entry into force of this Regulation] until the conformity assessment procedure is finalised or until the certificate is renewed.

Article 121

Evaluation

~~By 27 May 2027, the Commission shall assess the application of this Regulation and produce an evaluation report on the progress towards achievement of the objectives contained herein including an assessment of the resources required to implement this Regulation. Special attention shall be given to the traceability of medical devices through the storage, pursuant to Article 27, of the UDI by economic operators, health institutions and health professionals.~~

No sooner than ...[Publications Office, please insert the date five years after the date of application of this Regulation], the Commission shall carry out an evaluation of this Regulation and present a report on the main findings to the European Parliament and the Council.

Member States and notified bodies shall provide the Commission with the information necessary for the preparation of that report.

Article 122

Repeal

Without prejudice to Article 120(3) to (3e) and (4) of this Regulation, and without prejudice to the obligations of the Member States and manufacturers as regards vigilance and to the obligations of manufacturers as regards the making available of documentation, under Directives 90/385/EEC and 93/42/EEC, those Directives are repealed with effect from 26 May 2021, with the exception of:

– Articles 8 and 10, Article 10b(1), points (b) and (c), and Article 10b(2) and (3) of Directive 90/385/EEC, Article 10, Article 14a(1), points (c) and (d), Article 14a(2) and (3) and Article 15 of Directive 93/42/EEC, and the obligations relating to vigilance and clinical investigations provided for in the corresponding Annexes to those Directives, which are repealed, as applicable, with effect from the date referred to in Article 123(3), point (d), of this Regulation in respect of the application of the obligations and requirements that relate to the electronic systems referred to in Article 33(2), points (e) and (f), respectively, of this Regulation;

– Article 10a, Article 10b(1), point (a), and Article 11(5) of Directive 90/385/EEC, Article 14(1) and (2), Article 14a(1), points (a) and (b), and Article 16(5) of Directive 93/42/EEC, and the obligations relating to registration of devices and economic operators, and to certificate notifications, provided for in the corresponding Annexes to those Directives, which are repealed, as applicable, with effect from the date referred to in Article 123(3), point (d), of this Regulation in respect of the application of the obligations and requirements that relate to the electronic systems referred to in Article 33(2), points (a) to (d), respectively, of this Regulation;

– Article 9(9) of Directive 90/385/EEC and Article 11(13) of Directive 93/42/EEC, which are repealed with effect from 24 April 2020.

As regards the devices referred to in Article 120(3) to (3e) and (4) of this Regulation, the Directives referred to in the first paragraph of this Article shall continue to apply to the extent necessary for the application of those paragraphs.

Notwithstanding the first paragraph, Regulations (EU) No 207/2012 and (EU) No 722/2012 shall remain in force and continue to apply unless and until repealed by implementing acts adopted by the Commission pursuant to this Regulation.

Proposal of Regulation (EU) 2017/745 Articles

References to the repealed Directives shall be understood as references to this Regulation and shall be read in accordance with the correlation table laid down in Annex XVII to this Regulation.

Article 123

Entry into force and date of application

1. This Regulation shall enter into force on the twentieth day following that of its publication in the *Official Journal of the European Union*.

2. It shall apply from 26 May 2021.

3. By way of derogation from paragraph 2:

(a) Articles 35 to 50 shall apply from 26 November 2017. However, from that date until 26 May 2021, the obligations on notified bodies pursuant to Articles 35 to 50 shall apply only to those bodies which submit an application for designation in accordance with Article 38;

(b) Articles 101 and 103 shall apply from 26 November 2017;

(c) Article 102 shall apply from 26 May 2018;

(d) without prejudice to the obligations of the Commission pursuant to Article 34, the obligations and requirements that relate to any of the electronic systems referred to in Article 33(2) shall apply from the date corresponding to 6 months from the date of publication of the notice referred to in Article 34(3), informing that the relevant electronic system is functional and meets the functional specifications drawn up pursuant to Article 34(1). The provisions referred to in the preceding sentence are:

- Article 29,
- Article 31,
- Article 32,
- Article 33(4),
- the second sentence of Article 40(2),
- Article 42(10),
- Article 43(2),
- the second subparagraph of Article 44(12),
- points (d) and (e) of Article 46(7),
- Article 53(2),
- Article 54(3),
- Article 55(1),
- Article 56(5),
- Articles 70 to 77,
- Article 78(1) to (13), without prejudice to Article 78(14),
- Articles 79 to 82,
- Article 86(2),
- Articles 87 and 88,
- Article 89(5) and (7), and the third subparagraph of Article 89(8),
- Article 90,

Proposal of Regulation (EU) 2017/745 Articles

- Article 93(4), (7) and (8),
- Article 95(2) and (4),
- the last sentence of Article 97(2),
- Article 99(4),
- Article 120(3d).

Until the date of application of the provisions referred to in the first subparagraph of this point, the corresponding provisions of Directives 90/385/EEC and 93/42/EEC regarding information on vigilance reporting, clinical investigations, registration of devices and economic operators, and certificate notifications shall continue to apply.

After the date of application of the provisions referred to in the first subparagraph of this point, where Member States maintain national databases, the relevant information available in Eudamed for those national databases shall be retrieved from Eudamed.

(e) no later than 12 months from the date of publication of the notice referred to in Article 34(3) in respect of the electronic system referred to in Article 33(2), points (a) and (b), manufacturers shall ensure that the information to be entered in Eudamed in accordance with Article 29 is entered in that electronic system, including regarding the following devices, provided that those devices are also placed on the market from 6 months from the date of publication of that notice:

(i) devices, other than custom-made devices, for which the manufacturer has undertaken a conformity assessment in accordance with Article 52;

(ii) devices, other than custom-made devices, placed on the market pursuant to Article 120(3), (3a) or (3b), unless the device, for which the manufacturer has undertaken a conformity assessment in accordance with Article 52, is already registered in Eudamed;

(ea) no later than 18 months from the date of publication of the notice referred to in Article 34(3) in respect of the electronic system referred to in Article 33(2), point (d), notified bodies shall ensure that the information to be entered in Eudamed in accordance with Article 56(5) is entered in that electronic system, including regarding devices referred to in point (e)(i) of this paragraph; for those devices, only the latest relevant certificate and, where applicable, any subsequent decision taken by the notified body related to such certificate shall be entered;

(eb) by way of derogation from point (d), first subparagraph, of this paragraph, the obligations to upload the summary of safety and clinical performance in accordance with Article 32(1) and to notify competent authorities in accordance with Article 55(1), through the electronic system referred in Article 33(2), point (d), shall apply to devices referred to in point (e) of this paragraph when the certificate is entered in Eudamed in accordance with point (ea) of this paragraph;

(ec) without prejudice to point (d), first subparagraph, of this paragraph, when a manufacturer is to submit a PSUR in accordance with Article 86(2) of this Regulation, to report a serious incident or a field safety corrective action in accordance with Article 87 of this Regulation, or to submit a trend report in accordance with Article 88 of this Regulation through the electronic system referred to in Article 33(2), point (f), of this Regulation, it shall also register the device, which is the subject of the PSUR or the vigilance reporting, in the electronic system referred to in Article 33(2), points (a) and (b), of this Regulation, except if such device was placed on the market in accordance with Directive 90/385/EEC or 93/42/EEC;

(f) for implantable devices and for class III devices Article 27(4) shall apply from 26 May 2021. For class IIa and class IIb devices Article 27(4) shall apply from 26 May 2023. For class I devices Article 27(4) shall apply from 26 May 2025;

(g) with regard to reusable devices that are required to bear the UDI carrier on the device itself, Article 27(4) shall apply to:

Proposal of Regulation (EU) 2017/745 Articles

- (i) implantable devices and class III devices from 26 May 2023;
- (II) class IIa and class IIb devices from 26 May 2025;
- (iii) class I devices from 26 May 2027;

(i) Article 120(12) shall apply from 26 May 2019.

(j) Article 59 shall apply from 24 April 2020.

This Regulation shall be binding in its entirety and directly applicable in all Member States.

Done at Strasbourg, 5 April 2017.

For the European Parliament

For the Council

The President

The President

A. TAJANI

I. BORG

ANNEX I GENERAL SAFETY AND PERFORMANCE REQUIREMENTS

Chapter I General requirements

1. Devices shall achieve the performance intended by their manufacturer and shall be designed and manufactured in such a way that, during normal conditions of use, they are suitable for their intended purpose. They shall be safe and effective and shall not compromise the clinical condition or the safety of patients, or the safety and health of users or, where applicable, other persons, provided that any risks which may be associated with their use constitute acceptable risks when weighed against the benefits to the patient and are compatible with a high level of protection of health and safety, taking into account the generally acknowledged state of the art.

2. The requirement in this Annex to reduce risks as far as possible means the reduction of risks as far as possible without adversely affecting the benefit-risk ratio.

3. Manufacturers shall establish, implement, document and maintain a risk management system.

Risk management shall be understood as a continuous iterative process throughout the entire lifecycle of a device, requiring regular systematic updating. In carrying out risk management manufacturers shall:

- (a) establish and document a risk management plan for each device;
- (b) identify and analyse the known and foreseeable hazards associated with each device;
- (c) estimate and evaluate the risks associated with, and occurring during, the intended use and during reasonably foreseeable misuse;
- (d) eliminate or control the risks referred to in point (c) in accordance with the requirements of Section 4;
- (e) evaluate the impact of information from the production phase and, in particular, from the post-market surveillance system, on hazards and the frequency of occurrence thereof, on estimates of their associated risks, as well as on the overall risk, benefit-risk ratio and risk acceptability; and
- (f) based on the evaluation of the impact of the information referred to in point (e), if necessary amend control measures in line with the requirements of Section 4.

4. Risk control measures adopted by manufacturers for the design and manufacture of the devices shall conform to safety principles, taking account of the generally acknowledged state of the art. To reduce risks, Manufacturers shall manage risks so that the residual risk associated with each hazard as well as the overall residual risk is judged acceptable. In selecting the most appropriate solutions, manufacturers shall, in the following order of priority:

- (a) eliminate or reduce risks as far as possible through safe design and manufacture;
- (b) where appropriate, take adequate protection measures, including alarms if necessary, in relation to risks that cannot be eliminated; and
- (c) provide information for safety (warnings/precautions/contra-indications) and, where appropriate, training to users.

Manufacturers shall inform users of any residual risks.

5. In eliminating or reducing risks related to use error, the manufacturer shall:

- (a) reduce as far as possible the risks related to the ergonomic features of the device and the environment in which the device is intended to be used (design for patient safety), and

Proposal for Regulation (EU) 2017/745 – Annex I

(b) give consideration to the technical knowledge, experience, education, training and use environment, where applicable, and the medical and physical conditions of intended users (design for lay, professional, disabled or other users).

6. The characteristics and performance of a device shall not be adversely affected to such a degree that the health or safety of the patient or the user and, where applicable, of other persons are compromised during the lifetime of the device, as indicated by the manufacturer, when the device is subjected to the stresses which can occur during normal conditions of use and has been properly maintained in accordance with the manufacturer's instructions.

7. Devices shall be designed, manufactured and packaged in such a way that their characteristics and performance during their intended use are not adversely affected during transport and storage, for example, through fluctuations of temperature and humidity, taking account of the instructions and information provided by the manufacturer.

8. All known and foreseeable risks, and any undesirable side-effects, shall be minimised and be acceptable when weighed against the evaluated benefits to the patient and/or user arising from the achieved performance of the device during normal conditions of use.

9. For the devices referred to in Annex XVI, the general safety requirements set out in Sections 1 and 8 shall be understood to mean that the device, when used under the conditions and for the purposes intended, does not present a risk at all or presents a risk that is no more than the maximum acceptable risk related to the product's use which is consistent with a high level of protection for the safety and health of persons.

Chapter II

Requirements regarding design and manufacture

10. Chemical, physical and biological properties

10.1. Devices shall be designed and manufactured in such a way as to ensure that the characteristics and performance requirements referred to in Chapter I are fulfilled. Particular attention shall be paid to:

(a) the choice of materials and substances used, particularly as regards toxicity and, where relevant, flammability;

(b) the compatibility between the materials and substances used and biological tissues, cells and body fluids, taking account of the intended purpose of the device and, where relevant, absorption, distribution, metabolism and excretion;

(c) the compatibility between the different parts of a device which consists of more than one implantable part;

(d) the impact of processes on material properties;

(e) where appropriate, the results of biophysical or modelling research the validity of which has been demonstrated beforehand;

(f) the mechanical properties of the materials used, reflecting, where appropriate, matters such as strength, ductility, fracture resistance, wear resistance and fatigue resistance;

(g) surface properties; and

(h) the confirmation that the device meets any defined chemical and/or physical specifications.

10.2. Devices shall be designed, manufactured and packaged in such a way as to minimise the risk posed by contaminants and residues to patients, taking account of the intended purpose of the device, and to the persons involved in the transport, storage and use of the devices. Particular attention shall be paid to tissues exposed to those contaminants and residues and to the duration and frequency of exposure.

Proposal for Regulation (EU) 2017/745 – Annex I

10.3. Devices shall be designed and manufactured in such a way that they can be used safely with the materials and substances, including gases, with which they enter into contact during their intended use; if the devices are intended to administer medicinal products they shall be designed and manufactured in such a way as to be compatible with the medicinal products concerned in accordance with the provisions and restrictions governing those medicinal products and that the performance of both the medicinal products and of the devices is maintained in accordance with their respective indications and intended use.

10.4. Substances

10.4.1. Design and manufacture of devices

Devices shall be designed and manufactured in such a way as to reduce as far as possible the risks posed by substances or particles, including wear debris, degradation products and processing residues, that may be released from the device.

Devices, or those parts thereof or those materials used therein that:

- are invasive and come into direct contact with the human body,
- (re)administer medicines, body liquids or other substances, including gases, to/from the body, or
- transport or store such medicines, body fluids or substances, including gases, to be (re)administered to the body,

shall only contain the following substances in a concentration that is above 0,1 % weight by weight (w/w) where justified pursuant to Section 10.4.2:

- (a) substances which are carcinogenic, mutagenic or toxic to reproduction ('CMR'), of category 1A or 1B, in accordance with Part 3 of Annex VI to Regulation (EC) No 1272/2008 of the European Parliament and of the Council¹, or
- (b) substances having endocrine-disrupting properties for which there is scientific evidence of probable serious effects to human health and which are identified either in accordance with the procedure set out in Article 59 of Regulation (EC) No 1907/2006 of the European Parliament and of the Council² or, once a delegated act has been adopted by the Commission pursuant to the first subparagraph of Article 5(3) of Regulation (EU) No 528/2012 of the European Parliament and the Council³, in accordance with the criteria that are relevant to human health amongst the criteria established therein.

10.4.2. Justification regarding the presence of CMR and/or endocrine-disrupting substances

The justification for the presence of such substances shall be based upon:

- (a) an analysis and estimation of potential patient or user exposure to the substance;
- (b) an analysis of possible alternative substances, materials or designs, including, where available, information about independent research, peer-reviewed studies, scientific opinions from relevant scientific committees and an analysis of the availability of such alternatives;
- (c) argumentation as to why possible substance and/ or material substitutes, if available, or design changes, if feasible, are inappropriate in relation to maintaining the functionality, performance and the benefit-risk ratios of the product; including taking into account if the intended use of such devices includes

¹ Regulation (EC) No 1272/2008 of the European Parliament and of the Council of 16 December 2008 on classification, labelling and packaging of substances and mixtures, amending and repealing Directives 67/548/EEC and 1999/45/EC, and amending Regulation (EC) No 1907/2006 (OJ L 353, 31.12.2008, p. 1).

² Regulation (EC) No 1907/2006 of the European Parliament and of the Council of 18 December 2006 concerning the Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) (OJ L 396, 30.12.2006, p. 1).

³ Regulation (EU) No 528/2012 of the European Parliament and the Council of 22 May 2012 concerning the making available on the market of and use of biocidal products (OJ L 167, 27.06.2012, p. 1).

Proposal for Regulation (EU) 2017/745 – Annex I

treatment of children or treatment of pregnant or breastfeeding women or treatment of other patient groups considered particularly vulnerable to such substances and/or materials; and

(d) where applicable and available, the latest relevant scientific committee guidelines in accordance with Sections 10.4.3. and 10.4.4.

10.4.3. Guidelines on phthalates

For the purposes of Section 10.4., the Commission shall, as soon as possible and by 26 May 2018, provide the relevant scientific committee with a mandate to prepare guidelines that shall be ready before 26 May 2020. The mandate for the committee shall encompass at least a benefit-risk assessment of the presence of phthalates which belong to either of the groups of substances referred to in points (a) and (b) of Section 10.4.1. The benefit-risk assessment shall take into account the intended purpose and context of the use of the device, as well as any available alternative substances and alternative materials, designs or medical treatments. When deemed appropriate on the basis of the latest scientific evidence, but at least every five years, the guidelines shall be updated.

10.4.4. Guidelines on other CMR and endocrine-disrupting substances

Subsequently, the Commission shall mandate the relevant scientific committee to prepare guidelines as referred to in Section 10.4.3. also for other substances referred to in points (a) and (b) of Section 10.4.1., where appropriate.

10.4.5. Labelling

Where devices, parts thereof or materials used therein as referred to in Section 10.4.1. contain substances referred to in points (a) or (b) of Section 10.4.1. in a concentration above 0,1 % weight by weight (w/w), the presence of those substances shall be labelled on the device itself and/or on the packaging for each unit or, where appropriate, on the sales packaging, with the list of such substances. If the intended use of such devices includes treatment of children or treatment of pregnant or breastfeeding women or treatment of other patient groups considered particularly vulnerable to such substances and/or materials, information on residual risks for those patient groups and, if applicable, on appropriate precautionary measures shall be given in the instructions for use.

10.5. Devices shall be designed and manufactured in such a way as to reduce as far as possible the risks posed by the unintentional ingress of substances into the device taking into account the device and the nature of the environment in which it is intended to be used.

10.6. Devices shall be designed and manufactured in such a way as to reduce as far as possible the risks linked to the size and the properties of particles which are or can be released into the patient's or user's body, unless they come into contact with intact skin only. Special attention shall be given to nanomaterials as defined in Commission Recommendation C/2022/3689*.

* Commission Recommendation of 10 June 2022 on the definition of nanomaterial (OJ C 229, 14.6.2022, pp. 1).

11. Infection and microbial contamination

11.1. Devices and their manufacturing processes shall be designed in such a way as to eliminate or to reduce as far as possible the risk of infection to patients, users and, where applicable, other persons. The design shall:

- (a) reduce as far as possible and appropriate the risks from unintended cuts and pricks, such as needle stick injuries,
- (b) allow easy and safe handling,
- (c) reduce as far as possible any microbial leakage from the device and/or microbial exposure during use, and
- (d) prevent microbial contamination of the device or its content such as specimens or fluids.

Proposal for Regulation (EU) 2017/745 – Annex I

11.2. Where necessary devices shall be designed to facilitate their safe cleaning, disinfection, and/or re-sterilisation.

11.3. Devices labelled as having a specific microbial state shall be designed, manufactured and packaged to ensure that they remain in that state when placed on the market and remain so under the transport and storage conditions specified by the manufacturer.

11.4. Devices delivered in a sterile state shall be designed, manufactured and packaged in accordance with appropriate procedures, to ensure that they are sterile when placed on the market and that, unless the packaging which is intended to maintain their sterile condition is damaged, they remain sterile, under the transport and storage conditions specified by the manufacturer, until that packaging is opened at the point of use. It shall be ensured that the integrity of that packaging is clearly evident to the final user.

11.5. Devices labelled as sterile shall be processed, manufactured, packaged and, sterilised by means of appropriate, validated methods.

11.6. Devices intended to be sterilised shall be manufactured and packaged in appropriate and controlled conditions and facilities.

11.7. Packaging systems for non-sterile devices shall maintain the integrity and cleanliness of the product and, where the devices are to be sterilised prior to use, minimise the risk of microbial contamination; the packaging system shall be suitable taking account of the method of sterilisation indicated by the manufacturer.

11.8. The labelling of the device shall distinguish between identical or similar devices placed on the market in both a sterile and a non-sterile condition additional to the symbol used to indicate that devices are sterile.

12. Devices incorporating a substance considered to be a medicinal product and devices that are composed of substances or of combinations of substances that are absorbed by or locally dispersed in the human body.

12.1. In the case of devices referred to in the first subparagraph of Article 1(8), the quality, safety and usefulness of the substance which, if used separately, would be considered to be a medicinal product within the meaning of point (2) of Article 1 of Directive 2001/83/EC, shall be verified by analogy with the methods specified in Annex I to Directive 2001/83/EC, as required by the applicable conformity assessment procedure under this Regulation.

12.2. Devices that are composed of substances or of combinations of substances that are intended to be introduced into the human body, and that are absorbed by or locally dispersed in the human body shall comply, where applicable and in a manner limited to the aspects not covered by this Regulation, with the relevant requirements laid down in Annex I to Directive 2001/83/EC for the evaluation of absorption, distribution, metabolism, excretion, local tolerance, toxicity, interaction with other devices, medicinal products or other substances and potential for adverse reactions, as required by the applicable conformity assessment procedure under this Regulation.

13. Devices incorporating materials of biological origin

13.1. For devices manufactured utilising derivatives of substances [tissues or cells] of human origin which are non-viable or are rendered non-viable covered by this Regulation in accordance with Article 1(6), point (g), the following shall apply:

(a) donor registration, donor review [donation], collection [procurement] and testing of the substances [tissues or cells] of human origin shall be done in accordance with Regulation (EU) 2024/1938 [Directive 2004/23/EC];

(b) processing, preservation and any other handling of those substances [tissues or cells] of human origin or their derivatives shall be carried out so as to provide safety for patients, users and, where applicable, other persons. In particular, safety with regard to viruses and other transmissible agents shall be

Proposal for Regulation (EU) 2017/745 – Annex I

addressed by appropriate methods of sourcing and by implementation of validated methods of elimination or inactivation in the course of the manufacturing process;

(c) the traceability system for those devices shall be complementary to and compatible with the traceability and data protection requirements laid down in Regulation (EU) 2024/1938 [Directive 2004/23/EC and in Directive 2002/98/EC].

13.2. For devices manufactured utilising tissues or cells of animal origin, or their derivatives, which are non-viable or rendered non-viable the following shall apply:

(a) where feasible taking into account the animal species, tissues and cells of animal origin, or their derivatives, shall originate from animals that have been subjected to veterinary controls that are adapted to the intended use of the tissues. Information on the geographical origin of the animals shall be retained by manufacturers;

(b) sourcing, processing, preservation, testing and handling of tissues, cells and substances of animal origin, or their derivatives, shall be carried out so as to provide safety for patients, users and, where applicable, other persons. In particular safety with regard to viruses and other transmissible agents shall be addressed by implementation of validated methods of elimination or viral inactivation in the course of the manufacturing process, except when the use of such methods would lead to unacceptable degradation compromising the clinical benefit of the device;

(c) in the case of devices manufactured utilising tissues or cells of animal origin or their derivatives, as referred to in Regulation (EU) No 722/2012, the particular requirements laid down in that Regulation, or in any subsequent implementing rules adopted pursuant to this Regulation, shall apply

13.3. For devices manufactured utilising non-viable biological substances other than those referred to in Sections 13.1 and 13.2, the processing, preservation, testing and handling of those substances shall be carried out so as to provide safety for patients, users and, where applicable, other persons, including in the waste disposal chain. In particular, safety with regard to viruses and other transmissible agents shall be addressed by appropriate methods of sourcing and by implementation of validated methods of elimination or inactivation in the course of the manufacturing process.

14. Construction of devices and interaction with their environment

14.1. If the device is intended for use in combination with other devices or equipment the whole combination, including the connection system shall be safe and shall not impair the specified performance of the devices. Any restrictions on use applying to such combinations shall be indicated on the label and/or in the instructions for use. Connections which the user has to handle, such as fluid, gas transfer, electrical or mechanical coupling, shall be designed and constructed in such a way as to minimise all possible risks, such as misconnection.

14.2. Devices shall be designed and manufactured in such a way as to remove or reduce as far as possible:

(a) the risk of injury, in connection with their physical features, including the volume/pressure ratio, dimensional and where appropriate ergonomic features;

(b) risks connected with reasonably foreseeable external influences or environmental conditions, such as magnetic fields, external electrical and electromagnetic effects, electrostatic discharge, radiation associated with diagnostic or therapeutic procedures, pressure, humidity, temperature, variations in pressure and acceleration or radio signal interferences;

(c) the risks associated with the use of the device when it comes into contact with materials, liquids, and substances, including gases, to which it is exposed during normal conditions of use;

(d) the risks associated with the possible negative interaction between software and the IT environment within which it operates and interacts;

(e) the risks of accidental ingress of substances into the device;

Proposal for Regulation (EU) 2017/745 – Annex I

(f) the risks of reciprocal interference with other devices normally used in the investigations or for the treatment given; and

(g) risks arising where maintenance or calibration are not possible (as with implants), from ageing of materials used or loss of accuracy of any measuring or control mechanism.

14.3. Devices shall be designed and manufactured in such a way as to minimise the risks of fire or explosion during normal use and in single fault condition. Particular attention shall be paid to devices the intended use of which includes exposure to or use in association with flammable or explosive substances or substances which could cause combustion.

14.4. Devices shall be designed and manufactured in such a way that adjustment, calibration, and maintenance can be done safely and effectively.

14.5. Devices that are intended to be operated together with other devices or products shall be designed and manufactured in such a way that the interoperability and compatibility are reliable and safe.

14.6 Any measurement, monitoring or display scale shall be designed and manufactured in line with ergonomic principles, taking account of the intended purpose, users and the environmental conditions in which the devices are intended to be used.

14.7. Devices shall be designed and manufactured in such a way as to facilitate their safe disposal and the safe disposal of related waste substances by the user, patient or other person. To that end, manufacturers shall identify and test procedures and measures as a result of which their devices can be safely disposed after use. Such procedures shall be described in the instructions for use.

15. Devices with a diagnostic or measuring function

15.1. Diagnostic devices and devices with a measuring function, shall be designed and manufactured in such a way as to provide sufficient accuracy, precision and stability for their intended purpose, based on appropriate scientific and technical methods. The limits of accuracy shall be indicated by the manufacturer.

15.2. The measurements made by devices with a measuring function shall be expressed in legal units conforming to the provisions of Council Directive 80/181/EEC¹.

16. Protection against radiation

16.1. General

(a) Devices shall be designed, manufactured and packaged in such a way that exposure of patients, users and other persons to radiation is reduced as far as possible, and in a manner that is compatible with the intended purpose, whilst not restricting the application of appropriate specified levels for therapeutic and diagnostic purposes.

(b) The operating instructions for devices emitting hazardous or potentially hazardous radiation shall contain detailed information as to the nature of the emitted radiation, the means of protecting the patient and the user, and on ways of avoiding misuse and of reducing the risks inherent to installation as far as possible and appropriate. Information regarding the acceptance and performance testing, the acceptance criteria, and the maintenance procedure shall also be specified.

16.2. Intended radiation

(a) Where devices are designed to emit hazardous, or potentially hazardous, levels of ionizing and/or non-ionizing radiation necessary for a specific medical purpose the benefit of which is considered to outweigh the risks inherent to the emission, it shall be possible for the user to control the emissions. Such devices

¹ Council Directive 80/181/EEC of 20 December 1979 on the approximation of the laws of the Member States relating to units of measurement and on the repeal of Directive 71/354/EEC (OJ L 039, 15.2.1980, p. 40).

Proposal for Regulation (EU) 2017/745 – Annex I

shall be designed and manufactured to ensure reproducibility of relevant variable parameters within an acceptable tolerance.

(b) Where devices are intended to emit hazardous, or potentially hazardous, ionizing and/or non-ionizing radiation, they shall be fitted, where possible, with visual displays and/or audible warnings of such emissions.

16.3. Devices shall be designed and manufactured in such a way that exposure of patients, users and other persons to the emission of unintended, stray or scattered radiation is reduced as far as possible. Where possible and appropriate, methods shall be selected which reduce the exposure to radiation of patients, users and other persons who may be affected.

16.4. Ionising radiation

(a) Devices intended to emit ionizing radiation shall be designed and manufactured taking into account the requirements of the Directive 2013/59/Euratom laying down basic safety standards for protection against the dangers arising from exposure to ionising radiation.

(b) Devices intended to emit ionising radiation shall be designed and manufactured in such a way as to ensure that, where possible, taking into account the intended use, the quantity, geometry and quality of the radiation emitted can be varied and controlled, and, if possible, monitored during treatment.

(c) Devices emitting ionising radiation intended for diagnostic radiology shall be designed and manufactured in such a way as to achieve an image and/or output quality that are appropriate to the intended medical purpose whilst minimising radiation exposure of the patient and user.

(d) Devices that emit ionising radiation and are intended for therapeutic radiology shall be designed and manufactured in such a way as to enable reliable monitoring and control of the delivered dose, the beam type, energy and, where appropriate, the quality of radiation.

17. Electronic programmable systems – devices that incorporate electronic programmable systems and software that are devices in themselves

17.1. Devices that incorporate electronic programmable systems, including software, or software that are devices in themselves, shall be designed to ensure repeatability, reliability and performance in line with their intended use. In the event of a single fault condition, appropriate means shall be adopted to eliminate or reduce as far as possible consequent risks or impairment of performance.

17.2. For devices that incorporate software or for software that are devices in themselves, the software shall be developed and manufactured in accordance with the state of the art taking into account the principles of development life cycle, risk management, including information security, verification and validation.

17.3. Software referred to in this Section that is intended to be used in combination with mobile computing platforms shall be designed and manufactured taking into account the specific features of the mobile platform (e.g. size and contrast ratio of the screen) and the external factors related to their use (varying environment as regards level of light or noise).

17.4. Manufacturers shall set out minimum requirements concerning hardware, IT networks characteristics, IT security measures and cybersecurity, including protection against unauthorised access, necessary to run the software as intended.

18. Active devices and devices connected to them

18.1. For non-implantable active devices, in the event of a single fault condition, appropriate means shall be adopted to eliminate or reduce as far as possible consequent risks.

18.2. Devices where the safety of the patient depends on an internal power supply shall be equipped with a means of determining the state of the power supply and an appropriate warning or indication for when

Proposal for Regulation (EU) 2017/745 – Annex I

the capacity of the power supply becomes critical. If necessary, such warning or indication shall be given prior to the power supply becoming critical.

18.3. Devices where the safety of the patient depends on an external power supply shall include an alarm system to signal any power failure.

18.4. Devices intended to monitor one or more clinical parameters of a patient shall be equipped with appropriate alarm systems to alert the user of situations which could lead to death or severe deterioration of the patient's state of health.

18.5. Devices shall be designed and manufactured in such a way as to reduce as far as possible the risks of creating electromagnetic interference which could impair the operation of the device in question or other devices or equipment in the intended environment.

18.6. Devices shall be designed and manufactured in such a way as to provide a level of intrinsic immunity to electromagnetic interference such that is adequate to enable them to operate as intended.

18.7. Devices shall be designed and manufactured in such a way as to avoid, as far as possible, the risk of accidental electric shocks to the patient, user or any other person, both during normal use of the device and in the event of a single fault condition in the device, provided the device is installed and maintained as indicated by the manufacturer.

18.8. Devices shall be designed and manufactured in such a way as to protect, as far as possible, against unauthorised access that could hamper the device from functioning as intended.

19. Particular requirements for active implantable devices

19.1. Active implantable devices shall be designed and manufactured in such a way as to remove or minimize as far as possible:

(a) risks connected with the use of energy sources with particular reference, where electricity is used, to insulation, leakage currents and overheating of the devices,

(b) risks connected with medical treatment, in particular those resulting from the use of defibrillators or high-frequency surgical equipment, and

(c) risks which may arise where maintenance and calibration are impossible, including:

- excessive increase of leakage currents,
- ageing of the materials used,
- excess heat generated by the device,
- decreased accuracy of any measuring or control mechanism.

19.2. Active implantable devices shall be designed and manufactured in such a way as to ensure

- if applicable, the compatibility of the devices with the substances they are intended to administer, and
- the reliability of the source of energy.

19.3. Active implantable devices and, if appropriate, their component parts shall be identifiable to allow any necessary measure to be taken following the discovery of a potential risk in connection with the devices or their component parts.

19.4. Active implantable devices shall bear a code by which they and their manufacturer can be unequivocally identified (particularly with regard to the type of device and its year of manufacture); it shall be possible to read this code, if necessary, without the need for a surgical operation.

20. Protection against mechanical and thermal risks

20.1. Devices shall be designed and manufactured in such a way as to protect patients and users against mechanical risks connected with, for example, resistance to movement, instability and moving parts.

Proposal for Regulation (EU) 2017/745 – Annex I

20.2. Devices shall be designed and manufactured in such a way as to reduce to the lowest possible level the risks arising from vibration generated by the devices, taking account of technical progress and of the means available for limiting vibrations, particularly at source, unless the vibrations are part of the specified performance.

20.3. Devices shall be designed and manufactured in such a way as to reduce to the lowest possible level the risks arising from the noise emitted, taking account of technical progress and of the means available to reduce noise, particularly at source, unless the noise emitted is part of the specified performance.

20.4. Terminals and connectors to the electricity, gas or hydraulic and pneumatic energy supplies which the user or other person has to handle, shall be designed and constructed in such a way as to minimise all possible risks.

20.5. Errors likely to be made when fitting or refitting certain parts which could be a source of risk shall be made impossible by the design and construction of such parts or, failing this, by information given on the parts themselves and/or their housings.

The same information shall be given on moving parts and/or their housings where the direction of movement needs to be known in order to avoid a risk.

20.6. Accessible parts of devices (excluding the parts or areas intended to supply heat or reach given temperatures) and their surroundings shall not attain potentially dangerous temperatures under normal conditions of use.

21. Protection against the risks posed to the patient or user by devices supplying energy or substances

21.1. Devices for supplying the patient with energy or substances shall be designed and constructed in such a way that the amount to be delivered can be set and maintained accurately enough to ensure the safety of the patient and of the user.

21.2. Devices shall be fitted with the means of preventing and/or indicating any inadequacies in the amount of energy delivered or substances delivered which could pose a danger. Devices shall incorporate suitable means to prevent, as far as possible, the accidental release of dangerous levels of energy or substances from an energy and/or substance source.

21.3. The function of the controls and indicators shall be clearly specified on the devices. Where a device bears instructions required for its operation or indicates operating or adjustment parameters by means of a visual system, such information shall be understandable to the user and, as appropriate, the patient.

22. Protection against the risks posed by medical devices intended by the manufacturer for use by lay persons

22.1. Devices for use by lay persons shall be designed and manufactured in such a way that they perform appropriately for their intended purpose taking into account the skills and the means available to lay persons and the influence resulting from variation that can be reasonably anticipated in the lay person's technique and environment. The information and instructions provided by the manufacturer shall be easy for the lay person to understand and apply.

22.2. Devices for use by lay persons shall be designed and manufactured in such a way as to:

- ensure that the device can be used safely and accurately by the intended user at all stages of the procedure, if necessary after appropriate training and/or information,
- reduce, as far as possible and appropriate, the risk from unintended cuts and pricks such as needle stick injuries, and
- reduce as far as possible the risk of error by the intended user in the handling of the device and, if applicable, in the interpretation of the results.

22.3. Devices for use by lay persons shall, where appropriate, include a procedure by which the lay person:

- can verify that, at the time of use, the device will perform as intended by the manufacturer, and
- if applicable, is warned if the device has failed to provide a valid result.

Chapter III

Requirements regarding the information supplied with the device

23. Label and instructions for use

23.1. General requirements regarding the information supplied by the manufacturer

Each device shall be accompanied by the information needed to identify the device and its manufacturer, and by any safety and performance information relevant to the user, or any other person, as appropriate. Such information may appear on the device itself, on the packaging or in the instructions for use, ~~and shall, if the manufacturer has a website, be made available and kept up to date on the website,~~ taking into account the following:

- (a) The medium, format, content, legibility, and location of the label and instructions for use shall be appropriate to the particular device, its intended purpose and the technical knowledge, experience, education or training of the intended user(s). In particular, instructions for use shall be written in terms readily understood by the intended user and, where appropriate, supplemented with drawings and diagrams.
- (b) The information required on the label shall be provided on the device itself. If this is not practicable or appropriate, some or all of the information may appear on the packaging for each unit, and/or on the packaging of multiple devices.
- (c) Labels shall be provided in a human-readable format and may be supplemented by machine-readable information, such as radiofrequency identification ('RFID') or bar codes. ~~Labels may be provided in digital form to the extent, and only under the conditions, set out in any implementing rules adopted pursuant to this Regulation.~~
- (d) Instructions for use shall be provided together with devices. By way of exception, instructions for use shall not be required for class I and class IIa devices if such devices can be used safely without any such instructions and unless otherwise provided for elsewhere in this Section.
- (e) Where multiple devices are supplied to a single user and/or location, a single copy of the instructions for use may be provided if so agreed by the purchaser who in any case may request further copies to be provided free of charge.
- (f) Instructions for use may be provided to the user in non-paper format (e.g. electronic) to the extent, and only under the conditions, set out in [Commission Implementing Regulation \(EU\) 2021/2226 \[Regulation \(EU\) No 207/2012\]](#) or in any subsequent implementing rules adopted pursuant to this Regulation.

**** Commission Implementing Regulation (EU) 2021/2226 of 14 December 2021 laying down rules for the application of Regulation (EU) 2017/745 of the European Parliament and of the Council as regards electronic instructions for use of medical devices (OJ L 448, 15.12.2021, p. 32, ELI: http://data.europa.eu/eli/reg_impl/2021/2226/oj).**

- (g) Residual risks which are required to be communicated to the user and/or other person shall be included as limitations, contra-indications, precautions or warnings in the information supplied by the manufacturer.
- (h) Where appropriate, the information supplied by the manufacturer shall take the form of internationally recognised symbols. Any symbol or identification colour used shall conform to the harmonised standards or CS. In areas for which no harmonised standards or CS exist, the symbols and colours shall be described in the documentation supplied with the device.
- (i) ~~For a device that is exclusively in use with a medicinal product in accordance with Article 19 of [Proposal for a Directive on the Union code relating to medicinal products for human use, and repealing~~

Proposal for Regulation (EU) 2017/745 – Annex I

Directive 2011/83/EC and Directive 2009/35/EC] and packaged together with a medicinal product, the instructions for use may be included, where needed, as part of the co-packaging of the medicinal product with the device. Moreover, the information on the label of the device may be limited to the particulars referred to in Section 23.2, points (a) and (c), where, following agreement of the competent authority responsible for the authorisation of the medicinal product, the following conditions are met:

- the information necessary for safe use and correct functioning of the device is provided to the user with the summary of product characteristics and/or package leaflet of the medicinal product under the responsibility of the marketing authorisation holder set out in [Proposal for a Directive on the Union code relating to medicinal products for human use, and repealing Directive 2011/83/EC and Directive 2009/35/EC];

the traceability and identification of the device is ensured by the marketing authorisation holder.

23.2. Information on the label

The label shall bear all of the following particulars:

- (a) the name or trade name of the device;
- (b) the details strictly necessary for a user to identify the device, the contents of the packaging and, where it is not obvious for the user, the intended purpose of the device;
- (c) the name, registered trade name or registered trade mark of the manufacturer and the address of its registered place of business;
- (d) if the manufacturer has its registered place of business outside the Union, the name of the authorised representative and address of the registered place of business of the authorised representative;
- (e) where applicable, an indication that the device contains or incorporates:
 - a medicinal substance, including a human blood or plasma derivative, or
 - substances of human origin or their derivatives [tissues or cells, or their derivatives, of human origin], or
 - tissues or cells of animal origin, or their derivatives, as referred to in Regulation (EU) No 722/2012 or in any subsequent implementing rules adopted pursuant to this Regulation;
- (f) where applicable, information labelled in accordance with Section 10.4.5.;
- (g) the lot number or the serial number of the device preceded by the words LOT NUMBER or SERIAL NUMBER or an equivalent symbol, as appropriate;
- (h) the UDI carrier referred to in Article 27(4) and Part C of Annex VI;
- (i) an unambiguous indication of the time limit for using or implanting the device safely, expressed at least in terms of year and month, where this is relevant;
- (j) where there is no indication of the date until when it may be used safely, the date of manufacture. This date of manufacture may be included as part of the lot number or serial number, provided the date is clearly identifiable;
- (k) an indication of any special storage and/or handling condition that applies;
- (l) if the device is supplied sterile, an indication of its sterile state and the sterilisation method;
- (m) warnings or precautions to be taken that need to be brought to the immediate attention of the user of the device, and to any other person. This information may be kept to a minimum in which case more detailed information shall appear in the instructions for use, taking into account the intended users;
- (n) if the device is intended for single use, an indication of that fact. A manufacturer's indication of single use shall be consistent across the Union;

~~(o) if the device is a single-use device that has been reprocessed, an indication of that fact, the number of reprocessing cycles already performed, and any limitation as regards the number of reprocessing cycles;~~

Proposal for Regulation (EU) 2017/745 – Annex I

- (p) if the device is custom-made, the words 'custom-made device';
- (q) an indication that the device is a medical device **or an accessory for a medical device**. If the device is intended for clinical investigation only, the words 'exclusively for clinical investigation';
- (r) in the case of devices that are composed of substances or of combinations of substances that are intended to be introduced into the human body via a body orifice or applied to the skin and that are absorbed by or locally dispersed in the human body, the overall qualitative composition of the device and quantitative information on the main constituent or constituents responsible for achieving the principal intended action;
- (s) for active implantable devices, the serial number, and for other implantable devices, the serial number or the lot number.

23.3. Information on the packaging which maintains the sterile condition of a device ('sterile packaging')

The following particulars shall appear on the sterile packaging:

- (a) an indication permitting the sterile packaging to be recognised as such,
- (b) a declaration that the device is in a sterile condition,
- (c) the method of sterilisation,
- (d) the name and address of the manufacturer,
- (e) a description of the device,
- (f) if the device is intended for clinical investigations, the words 'exclusively for clinical investigations',
- (g) if the device is custom-made, the words 'custom-made device',
- (h) the month and year of manufacture,
- (i) an unambiguous indication of the time limit for using or implanting the device safely expressed at least in terms of year and month, and
- (j) an instruction to check the instructions for use for what to do if the sterile packaging is damaged or unintentionally opened before use.

23.4. Information in the instructions for use

The instructions for use shall contain all of the following particulars:

- (a) the particulars referred to in points (a), (c), (e), (f), (k), (l), (n) and (r) of Section 23.2;
- (b) the device's intended purpose with a clear specification of indications, contra-indications, the patient target group or groups, and of the intended users, as appropriate;
- (c) where applicable, a specification of the clinical benefits to be expected.
- (d) where applicable, links to the summary of safety and clinical performance referred to in Article 32;
- (e) the performance characteristics of the device;
- (f) where applicable, information allowing the healthcare professional to verify if the device is suitable and select the corresponding software and accessories;
- (g) any residual risks, contra-indications and any undesirable side-effects, including information to be conveyed to the patient in this regard;
- (h) specifications the user requires to use the device appropriately, e.g. if the device has a measuring function, the degree of accuracy claimed for it;

Proposal for Regulation (EU) 2017/745 – Annex I

- (i) details of any preparatory treatment or handling of the device before it is ready for use or during its use, such as sterilisation, final assembly, calibration, etc., including the levels of disinfection required to ensure patient safety and all available methods for achieving those levels of disinfection;
- (j) any requirements for special facilities, or special training, or particular qualifications of the device user and/or other persons;
- (k) the information needed to verify whether the device is properly installed and is ready to perform safely and as intended by the manufacturer, together with, where relevant:
 - details of the nature, and frequency, of preventive and regular maintenance, and of any preparatory cleaning or disinfection,
 - identification of any consumable components and how to replace them,
 - information on any necessary calibration to ensure that the device operates properly and safely during its intended lifetime, and
 - methods for eliminating the risks encountered by persons involved in installing, calibrating or servicing devices;
- (l) if the device is supplied sterile, instructions in the event of the sterile packaging being damaged or unintentionally opened before use;
- (m) if the device is supplied non-sterile with the intention that it is sterilised before use, the appropriate instructions for sterilisation;
- (n) if the device is reusable, information on the appropriate processes for allowing reuse, including cleaning, disinfection, packaging and, where appropriate, the validated method of re-sterilisation appropriate to the Member State or Member States in which the device has been placed on the market. Information shall be provided to identify when the device should no longer be reused, e.g. signs of material degradation or the maximum number of allowable reuses;
- (o) an indication, if appropriate, that a device can be reused only if it is reconditioned under the responsibility of the manufacturer to comply with the general safety and performance requirements;
- (p) if the device bears an indication that it is for single use, information on known characteristics and technical factors known to the manufacturer that could pose a risk if the device were to be re-used. This information shall be based on a specific section of the manufacturer's risk management documentation, where such characteristics and technical factors shall be addressed in detail. If in accordance with point (d) of Section 23.1. no instructions for use are required, this information shall be made available to the user upon request;
- (q) for devices intended for use together with other devices and/or general purpose equipment:
 - information to identify such devices or equipment, in order to obtain a safe combination, and/or
 - information on any known restrictions to combinations of devices and equipment;
- (r) if the device emits radiation for medical purposes:
 - detailed information as to the nature, type and where appropriate, the intensity and distribution of the emitted radiation,
 - the means of protecting the patient, user, or other person from unintended radiation during use of the device;
- (s) information that allows the user and/or patient to be informed of any warnings, precautions, contra-indications, measures to be taken and limitations of use regarding the device. That information shall, where relevant, allow the user to brief the patient about any warnings, precautions, contra-indications, measures to be taken and limitations of use regarding the device. The information shall cover, where appropriate:

Proposal for Regulation (EU) 2017/745 – Annex I

- warnings, precautions and/or measures to be taken in the event of malfunction of the device or changes in its performance that may affect safety,
 - warnings, precautions and/or measures to be taken as regards the exposure to reasonably foreseeable external influences or environmental conditions, such as magnetic fields, external electrical and electromagnetic effects, electrostatic discharge, radiation associated with diagnostic or therapeutic procedures, pressure, humidity, or temperature,
 - warnings, precautions and/or measures to be taken as regards the risks of interference posed by the reasonably foreseeable presence of the device during specific diagnostic investigations, evaluations, or therapeutic treatment or other procedures such as electromagnetic interference emitted by the device affecting other equipment,
 - if the device is intended to administer medicinal products, substances [tissues or cells] of human origin or tissues or cells of animal origin, or their derivatives, or biological substances, any limitations or incompatibility in the choice of substances to be delivered,
 - warnings, precautions and/or limitations related to the medicinal substance or biological material that is incorporated into the device as an integral part of the device; and
 - precautions related to materials incorporated into the device that contain or consist of CMR substances or endocrine-disrupting substances, or that could result in sensitisation or an allergic reaction by the patient or user;
- (t) in the case of devices that are composed of substances or of combinations of substances that are intended to be introduced into the human body and that are absorbed by or locally dispersed in the human body, warnings and precautions, where appropriate, related to the general profile of interaction of the device and its products of metabolism with other devices, medicinal products and other substances as well as contra-indications, undesirable side-effects and risks relating to overdose;
- (u) in the case of implantable devices, the overall qualitative and quantitative information on the materials and substances to which patients can be exposed;
- (v) warnings or precautions to be taken in order to facilitate the safe disposal of the device, its accessories and the consumables used with it, if any. This information shall cover, where appropriate:
- infection or microbial hazards such as explants, needles or surgical equipment contaminated with potentially infectious substances of human origin, and
 - physical hazards such as from sharps.
- If in accordance with the point (d) of Section 23.1 no instructions for use are required, this information shall be made available to the user upon request;
- (w) for devices intended for use by lay persons, the circumstances in which the user should consult a healthcare professional;
- (x) for the devices covered by this Regulation pursuant to Article 1(2), information regarding the absence of a clinical benefit and the risks related to use of the device;
- (y) date of issue of the instructions for use or, if they have been revised, date of issue and identifier of the latest revision of the instructions for use;
- (z) a notice to the user and/or patient that any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established;
- (aa) information to be supplied to the patient with an implanted device in accordance with Article 18;
- (ab) for devices that incorporate electronic programmable systems, including software, or software that are devices in themselves, minimum requirements concerning hardware, IT networks characteristics and

Proposal for Regulation (EU) 2017/745 – Annex I

IT security measures, including protection against unauthorised access, necessary to run the software as intended.

ANNEX II TECHNICAL DOCUMENTATION

The technical documentation and, if applicable, the summary thereof to be drawn up by the manufacturer shall be presented in a clear, organised, readily searchable and unambiguous manner and shall include in particular the elements listed in this Annex.

1. DEVICE DESCRIPTION AND SPECIFICATION, INCLUDING VARIANTS AND ACCESSORIES

1.1. Device description and specification

- (a) product or trade name and a general description of the device including its intended purpose and intended users;
- (b) the Basic UDI-DI as referred to in Part C of Annex VI assigned by the manufacturer to the device in question, as soon as identification of this device becomes based on a UDI system, or otherwise a clear identification by means of product code, catalogue number or other unambiguous reference allowing traceability;
- (c) the intended patient population and medical conditions to be diagnosed, treated and/or monitored and other considerations such as patient selection criteria, indications, contra-indications, warnings;
- (d) principles of operation of the device and its mode of action, scientifically demonstrated if necessary;
- (e) the rationale for the qualification of the product as a device;
- (f) the risk class of the device and the justification for the classification rule(s) applied in accordance with Annex VIII;
- (g) an explanation of any novel features;
- (h) a description of the accessories for a device, other devices and other products that are not devices, which are intended to be used in combination with it;
- (i) a description or complete list of the various configurations/variants of the device that are intended to be made available on the market;
- (j) a general description of the key functional elements, e.g. its parts/components (including software if appropriate), its formulation, its composition, its functionality and, where relevant, its qualitative and quantitative composition. Where appropriate, this shall include labelled pictorial representations (e.g. diagrams, photographs, and drawings), clearly indicating key parts/components, including sufficient explanation to understand the drawings and diagrams;
- (k) a description of the raw materials incorporated into key functional elements and those making either direct contact with the human body or indirect contact with the body, e.g., during extracorporeal circulation of body fluids;
- (l) technical specifications, such as features, dimensions and performance attributes, of the device and any variants/configurations and accessories that would typically appear in the product specification made available to the user, for example in brochures, catalogues and similar publications.

1.2. Reference to previous and similar generations of the device

- (a) an overview of the previous generation or generations of the device produced by the manufacturer, where such devices exist;
- (b) an overview of identified similar devices available on the Union or international markets, where such devices exist.

Proposal for Regulation (EU) 2017/745 – Annex II

2. INFORMATION TO BE SUPPLIED BY THE MANUFACTURER

A complete set of:

- the label or labels on the device and on its packaging, such as single unit packaging, sales packaging, transport packaging in case of specific management conditions, in the languages accepted in the Member States where the device is envisaged to be sold; and
- the instructions for use in the languages accepted in the Member States where the device is envisaged to be sold.

3. DESIGN AND MANUFACTURING INFORMATION

- (a) information to allow the design stages applied to the device to be understood;
- (b) complete information and specifications, including the manufacturing processes and their validation, their adjuvants, the continuous monitoring and the final product testing. Data shall be fully included in the technical documentation;
- (c) identification of all sites, including suppliers and sub-contractors, where design and manufacturing activities are performed.

4. GENERAL SAFETY AND PERFORMANCE REQUIREMENTS

The documentation shall contain information for the demonstration of conformity with the general safety and performance requirements set out in Annex I that are applicable to the device taking into account its intended purpose, and shall include a justification, validation and verification of the solutions adopted to meet those requirements. The demonstration of conformity shall include:

- (a) the general safety and performance requirements that apply to the device and an explanation as to why others do not apply;
- (b) the method or methods used to demonstrate conformity with each applicable general safety and performance requirement;
- (c) the harmonised standards, CS or other solutions applied; and
- (d) the precise identity of the controlled documents offering evidence of conformity with each harmonised standard, CS or other method applied to demonstrate conformity with the general safety and performance requirements. The information referred to under this point shall incorporate a cross-reference to the location of such evidence within the full technical documentation and, if applicable, the summary technical documentation.

5. BENEFIT-RISK ANALYSIS AND RISK MANAGEMENT

The documentation shall contain information on:

- (a) the benefit-risk analysis referred to in Sections 1 and 8 of Annex I, and
- (b) the solutions adopted and the results of the risk management referred to in Section 3 of Annex I.

6. PRODUCT VERIFICATION AND VALIDATION

The documentation shall contain the results and critical analyses of all verifications and validation tests and/or studies undertaken to demonstrate conformity of the device with the requirements of this Regulation and in particular the applicable general safety and performance requirements.

6.1. **Non-clinical**, Pre-clinical and clinical data

- (a) results of tests, such as engineering, laboratory, **in vitro**, **ex vivo**, **in silico testing**, **computational modeling**, simulated use and animal tests, and evaluation of published literature applicable to the device, taking into account its intended purpose, or to similar devices, regarding the pre-clinical safety of the device and its conformity with the specifications;

Proposal for Regulation (EU) 2017/745 – Annex II

(b) detailed information regarding test design, complete test or study protocols, methods of data analysis, in addition to data summaries and test conclusions regarding in particular:

- the biocompatibility of the device including the identification of all materials in direct or indirect contact with the patient or user;
- physical, chemical and microbiological characterisation;
- electrical safety and electromagnetic compatibility;
- software verification and validation (describing the software design and development process and evidence of the validation of the software, as used in the finished device. This information shall typically include the summary results of all verification, validation and testing performed both in-house and in a simulated or actual user environment prior to final release. It shall also address all of the different hardware configurations and, where applicable, operating systems identified in the information supplied by the manufacturer);
- stability, including shelf life; and
- performance and safety.

Where applicable, conformity with the provisions of Directive 2004/10/EC of the European Parliament and of the Council¹ shall be demonstrated.

Where no new testing has been undertaken, the documentation shall incorporate a rationale for that decision. An example of such a rationale would be that, biocompatibility testing on identical materials was conducted when those materials were incorporated in a previous version of the device that has been legally placed on the market or put into service;

(c) the clinical evaluation plan, the clinical evaluation report and its updates referred to in Article 61(1) [61(12)] and Part A of Annex XIV.

~~(d) the PMCF plan and PMCF evaluation report referred to in Part B of Annex XIV or a justification why a PMCF is not applicable.~~

6.2. Additional information required in specific cases

(a) Where a device incorporates, as an integral part, a substance which, if used separately, may be considered to be a medicinal product within the meaning of point 2 of Article 1 of Directive 2001/83/EC, including a medicinal product derived from human blood or human plasma, as referred to in the first subparagraph of Article 1(8), a statement indicating this fact. In this case, the documentation shall identify the source of that substance and contain the data of the tests conducted to assess its safety, quality and usefulness, taking account of the intended purpose of the device.

(b) Where a device is manufactured utilising substances [tissues or cells] of human origin or tissues or cells of animal origin, or their derivatives, and is covered by this Regulation in accordance with Article 1(6), points (f) and (g), and where a device incorporates, as an integral part, substances [tissues or cells] of human origin or their derivatives that have an action ancillary to that of the device and is covered by this Regulation in accordance with Article 1(10), first subparagraph, a statement indicating this fact. In such a case, the documentation shall identify all materials of human or animal origin used and provide detailed information concerning the conformity with Sections 13.1. or 13.2., respectively, of Annex I.

(c) In the case of devices that are composed of substances or combinations of substances that are intended to be introduced into the human body and that are absorbed by or locally dispersed in the human

¹ Directive 2004/10/EC of the European Parliament and of the Council of 11 February 2004 on the harmonisation of laws, regulations and administrative provisions relating to the application of the principles of good laboratory practice and the verification of their applications for tests on chemical substances (OJ L 50, 20.2.2004, p. 44).

Proposal for Regulation (EU) 2017/745 – Annex II

body, detailed information, including test design, complete test or study protocols, methods of data analysis, and data summaries and test conclusions, regarding studies in relation to:

- absorption, distribution, metabolism and excretion;
- possible interactions of those substances, or of their products of metabolism in the human body, with other devices, medicinal products or other substances, considering the target population, and its associated medical conditions;
- local tolerance; and
- toxicity, including single-dose toxicity, repeat-dose toxicity, genotoxicity, carcinogenicity and reproductive and developmental toxicity, as applicable depending on the level and nature of exposure to the device.

In the absence of such studies, a justification shall be provided.

(d) In the case of devices containing CMR or endocrine-disrupting substances referred to in Section 10.4.1 of Annex I, the justification referred to in Section 10.4.2 of that Annex.

(e) In the case of devices placed on the market in a sterile or defined microbiological condition, a description of the environmental conditions for the relevant manufacturing steps. In the case of devices placed on the market in a sterile condition, a description of the methods used, including the validation reports, with respect to packaging, sterilisation and maintenance of sterility. The validation report shall address bioburden testing, pyrogen testing and, if applicable, testing for sterilant residues.

(f) In the case of devices placed on the market with a measuring function, a description of the methods used in order to ensure the accuracy as given in the specifications.

(g) If the device is to be connected to other device(s) in order to operate as intended, a description of this combination/configuration including proof that it conforms to the general safety and performance requirements when connected to any such device(s) having regard to the characteristics specified by the manufacturer.

(h) Where the device incorporates as an integral part an in vitro diagnostic medical device that has an action ancillary to that of the device, as referred to in Article 1(7) of this Regulation, the documentation shall include the results of the assessment of the conformity of the in vitro diagnostic medical device part with the relevant general safety and performance requirements set out in Annex I to Regulation (EU) 2017/746 contained in the manufacturer's EU declaration of conformity or the relevant certificate issued by a notified body allowing the manufacturer to affix a CE marking to the in vitro diagnostic medical device. Where those results of the conformity assessment are not available and where for the conformity assessment of the in vitro diagnostic medical device, if used separately, the involvement of a notified body is required in accordance with Regulation (EU) 2017/746, an opinion on the conformity of the in vitro diagnostic medical device part with the relevant general safety and performance requirements set out in Annex I to Regulation 2017/746 issued by a notified body designated in accordance with that Regulation for the type of device in question shall be included in the documentation.

ANNEX III TECHNICAL DOCUMENTATION ON POST-MARKET SURVEILLANCE

The technical documentation on post-market surveillance to be drawn up by the manufacturer in accordance with Articles 83 to 86 shall be presented in a clear, organised, readily searchable and unambiguous manner and shall include in particular the elements described in this Annex.

1. The post-market surveillance plan drawn up in accordance with Article 84.

The manufacturer shall prove in a post-market surveillance plan that it complies with the obligation referred to in Article 83.

(a) The post-market surveillance plan shall address the collection and utilization of available information, in particular:

- information concerning serious incidents, including information from PSURs, and field safety corrective actions;
- records referring to non-serious incidents and data on any undesirable side-effects;
- information from trend reporting;
- relevant specialist or technical literature, databases and/or registers;
- information, including feedbacks and complaints, provided by users, distributors and importers; and
- publicly available information about similar medical devices.

(b) The post-market surveillance plan shall cover at least:

- a proactive and systematic process to collect any information referred to in point (a). The process shall allow a correct characterisation of the performance of the devices and shall also allow a comparison to be made between the device and similar products available on the market;
- effective and appropriate methods and processes to assess the collected data;
- suitable indicators and threshold values that shall be used in the continuous reassessment of the benefit-risk analysis and of the risk management as referred to in Section 3 of Annex I;
- effective and appropriate methods and tools to investigate complaints and analyse market-related experience collected in the field;
- methods and protocols to manage the incidents subject to the trend report as provided for in Article 88, including the methods and protocols to be used to establish any statistically significant increase in the frequency or severity of incidents as well as the observation period;
- methods and protocols to communicate effectively with competent authorities, notified bodies, economic operators and users;
- reference to procedures to fulfil the manufacturers obligations laid down in Articles 83, 84 and 86;
- systematic procedures to identify and initiate appropriate measures including corrective actions;
- effective tools to trace and identify devices for which corrective actions might be necessary; and
- a PMCF plan as referred to in Part B of Annex XIV, or a justification as to why a PMCF is not applicable.

2. The PSUR referred to in Article 86 or [and] the post-market surveillance report referred to in Article 85.


ANNEX IV
EU DECLARATION OF CONFORMITY

The EU declaration of conformity shall contain all of the following information:

1. Name, registered trade name or registered trade mark and, if already issued, SRN as referred to in Article 31 of the manufacturer, and, if applicable, its authorised representative, and the address of their registered place of business where they can be contacted and their location be established;
2. A statement that the EU declaration of conformity is issued under the sole responsibility of the manufacturer;
3. The Basic UDI-DI as referred to in Part C of Annex VI;
4. Product and trade name, product code, catalogue number or other unambiguous reference allowing identification and traceability of the device covered by the EU declaration of conformity, such as a photograph, where appropriate, as well as its intended purpose. Except for the product or trade name, the information allowing identification and traceability may be provided by the Basic UDI-DI referred to in point 3;
5. Risk class of the device in accordance with the rules set out in Annex VIII;
6. A statement that the device that is covered by the present declaration is in conformity with this Regulation and, if applicable, with any other relevant Union legislation that provides for the issuing of an EU declaration of conformity;
7. References to any CS used and in relation to which conformity is declared;
8. Where applicable, the name and identification number of the notified body, a description of the conformity assessment procedure performed and identification of the certificate or certificates issued;
9. Where applicable, additional information;
10. Place and date of issue of the declaration, name and function of the person who signed it as well as an indication for, and on behalf of whom, that person signed, signature.

ANNEX V
CE MARKING OF CONFORMITY

1. The CE marking shall consist of the initials 'CE' taking the following form:



2. If the CE marking is reduced or enlarged, the proportions given in the above graduated drawing shall be respected.

3. The various components of the CE marking shall have substantially the same vertical dimension, which may not be less than 5 mm. This minimum dimension may be waived for small-scale devices.

ANNEX VI

INFORMATION TO BE SUBMITTED UPON THE REGISTRATION OF DEVICES AND ECONOMIC OPERATORS IN ACCORDANCE WITH ARTICLES 29(4) AND 31, CORE DATA ELEMENTS TO BE PROVIDED TO THE UDI DATABASE TOGETHER WITH THE UDI-DI IN ACCORDANCE WITH ARTICLES 28 AND 29, AND THE UDI SYSTEM

PART A

INFORMATION TO BE SUBMITTED UPON THE REGISTRATION OF ~~DEVICES AND~~ ECONOMIC OPERATORS IN ACCORDANCE WITH ARTICLES ~~S 29(4) AND~~ 31

Manufacturers ~~and [or] where [when]~~ applicable, authorised representatives, and, ~~where [when]~~ applicable, importers ~~and, where applicable the persons referred to in Article 22(1)~~ shall submit the following information ~~referred to in Section 1 and shall ensure that the information on their devices referred to in Section 2 is complete, correct and updated by the relevant party relating to the economic operator:~~

~~1. Information relating to the economic operator~~

1. [1.1] type of economic operator (manufacturer, authorised representative, importer ~~or the person referred to in Article 22(1)~~),
2. [1.2] name, address and contact details, ~~including the digital contact~~, of the economic operator,
3. [1.3] where submission of information is carried out by another person on behalf of any of the economic operators mentioned under Section 1.1, the name, address and contact details, ~~including the digital contact~~, of that person,
4. [1.4] name, address and contact details, ~~including the digital contact~~, of the person or persons responsible for regulatory compliance referred to in Article 15.

~~2. Information relating to the device~~

~~2.1. Basic UDI-DI,~~

~~2.6. reprocessed single-use device (y/n),~~

PART B

CORE DATA ELEMENTS TO BE PROVIDED TO THE UDI DATABASE ~~TOGETHER WITH THE UDI-DI~~ IN ACCORDANCE WITH ARTICLES 28 AND 29

The manufacturer shall provide to the UDI database ~~the UDI-DI and all of~~ the following information relating to the manufacturer and the device:

1. [2] ~~the~~ Basic UDI-DI as referred to in Article 27 [29] and any additional UDI-DIs,

~~Part B 2 [Part A 2.2] For devices referred to in Article 120(3),~~ type, number and expiry date of the certificate issued by the notified body and the name or identification number of that notified body and the link to the information that appears on the certificate and was entered by the notified body in the electronic system on notified bodies and certificates,

~~Part B 3 [Part A 2.3] Member State in which the device is to~~ ~~be~~ or has been placed on the market in the Union,

~~Part B 4 [Part A 2.4] in the case of class IIa, class IIb or class III devices:~~ Member States where the device is or is to be made available,

~~Part B 5 [Part A 2.5] risk class of the device,~~

~~Part B 6 [Part A 2.7] presence of a substance which, if used separately, may be considered to be a medicinal product and name of that substance,~~

Proposal for Regulation (EU) 2017/745 – Annex VI

Part B 7 [Part A 2.8] presence of a substance which, if used separately, may be considered to be a medicinal product derived from human blood or human plasma and name of this substance,

Part B 8 [Part A 2.9] presence of tissues or cells of human origin, or their derivatives (y/n),

Part B 9 [Part A 2.10] presence of tissues or cells of animal origin, or their derivatives, as referred to in Regulation (EU) No 722/2012 **or any subsequent implementing rules adopted pursuant to that Regulation** (y/n),

Part B 10 [Part A 2.11] where applicable, the single identification number of the clinical investigation or investigations conducted in relation to the device or a link to the clinical investigation registration in the electronic system on clinical investigations,

Part B 11 [Part A 2.12] in the case of devices listed in Annex XVI, specification as to whether the intended purpose of the device is other than a medical purpose,

Part B 12 [Part A 2.13] in the case of devices designed and manufactured by another legal or natural person as referred in Article 10(15), the name, address and contact details, **including the digital contact**, of that legal or natural person,

Part B 13 [Part A 2.14] ~~in the case of class III or implantable devices, where applicable,~~ the summary of safety and clinical performance,

Part B 14 [Part A 2.15] status of the device (**placed** on the market, no longer placed on the market, recalled, field safety corrective action initiated).

15 [1] quantity per package configuration,

16 [3] the manner in which production of the device is controlled (expiry date or manufacturing date, lot number, serial number),

17 [4] if applicable, the unit of use UDI-DI (where a UDI is not labelled on the device at the level of its unit of use, a 'unit of use' DI shall be assigned so as to associate the use of a device with a patient),

18 [5] name and address of the manufacturer (as indicated on the label),

19 [6] the SRN issued in accordance with Article 31(2),

20 [7] if applicable, name and address of the authorised representative (as indicated on the label),

21 [8] the medical device nomenclature code as provided for in Article 26,

~~9, risk class of the device,~~

22 [10, 15] if applicable, name or trade name, and if applicable additional trade names,

23 [11] if applicable, device model, reference, or catalogue number,

24 [12] if applicable, clinical size (including volume, length, gauge, diameter),

25 [13] additional product description (optional),

26 [14] if applicable, storage and/or handling conditions (as indicated on the label or in the instructions for use),

27 [16] labelled as a single-use device (y/n),

28 [17] if applicable, the maximum number of reuses,

29 [18] device labelled sterile (y/n),

30 [19] need for sterilisation before use (y/n),

31 [20] containing latex (y/n),

32 [21] where applicable, information labelled in accordance with Section 10.4.5 of Annex I,

Proposal for Regulation (EU) 2017/745 – Annex VI

33 [22] if applicable, the instructions for use, or where available, the URL of the website where the instructions for use are made available, [URL for additional information, such as electronic instructions for use (optional),]

34 [23] if applicable, critical warnings or contra-indications,

~~24. status of the device (on the market, no longer placed on the market, recalled, field safety corrective action initiated).~~

PART C The UDI System

1. Definitions

Automatic identification and data capture ('AIDC')

AIDC is a technology used to automatically capture data. AIDC technologies include bar codes, smart cards, biometrics and RFID.

Basic UDI-DI

The Basic UDI-DI is the primary identifier of a device model. ~~The Basic UDI-DI is the main key in the database and relevant documentation (e.g. certificates, declaration of conformity, technical documentation and summary of safety and clinical performance) to connect devices with the same intended purpose, risk class and essential design and manufacturing characteristics. It is the DI assigned at the level of the device unit of use.~~ It is the main key for records in the UDI database and is referenced in relevant certificates and EU declarations of conformity.

Unit of Use DI

The Unit of Use DI serves to associate the use of a device with a patient in instances in which a UDI is not labelled on the individual device at the level of its unit of use, for example in the event of several units of the same device being packaged together.

Configurable device

A configurable device is a device that consists of several components which can be assembled by the manufacturer in multiple configurations. Those individual components may be devices in themselves.

Configurable devices include computed tomography (CT) systems, ultrasound systems, anaesthesia systems, physiological Monitoring systems, radiology information systems (RIS).

Configuration

Configuration is a combination of items of equipment, as specified by the manufacturer, that operate together as a device to achieve an intended purpose. The combination of items may be modified, adjusted or customized to meet specific needs.

Configurations include *inter alia*:

- gantries, tubes, tables, consoles and other items of equipment that can be configured/combined to deliver an intended function in computed tomography.
- ventilators, breathing circuits, vaporizers combined to deliver an intended function in anaesthesia.

UDI-DI

The UDI-DI is a unique numeric or alphanumeric code specific to a model of device and that is also used as the 'access key' to information stored in a UDI database.

Human Readable Interpretation ('HRI')

HRI is a legible interpretation of the data characters encoded in the UDI carrier.

Proposal for Regulation (EU) 2017/745 – Annex VI

Packaging levels

Packaging levels means the various levels of device packaging that contain a defined quantity of devices, such as a carton or case.

UDI-PI

The UDI-PI is a numeric or alphanumeric code that identifies the unit of device production.

The different types of UDI-PIs include serial number, lot number, software identification and manufacturing or expiry date or both types of date.

Radio Frequency Identification RFID

RFID is a technology that uses communication through the use of radio waves to exchange data between a reader and an electronic tag attached to an object, for the purpose of identification.

Shipping containers

A shipping container is a container in relation to which traceability is controlled by a process specific to logistics systems.

Unique Device Identifier ('UDI')

The UDI is a series of numeric or alphanumeric characters that is created through a globally accepted device identification and coding standard. It allows the unambiguous identification of a specific device on the market. The UDI is comprised of the UDI-DI and the UDI-PI.

The word 'Unique' does not imply serialisation of individual production units.

UDI carrier

The UDI carrier is the means of conveying the UDI by using AIDC and, if applicable, its HRI.

UDI carriers include, *inter alia*, ID/linear bar code, 2D/Matrix bar code, RFID.

2. General requirements

2.1. The affixing of the UDI is an additional requirement – it does not replace any other marking or labelling requirements laid down in Annex I to this Regulation.

2.2. The manufacturer shall assign and maintain unique UDIs for its devices.

2.3. Only the manufacturer may place the UDI on the device or its packaging.

2.4. Only coding standards provided by issuing entities designated by the Commission pursuant to Article 27(2) may be used.

3. The UDI

3.1. A UDI shall be assigned to the device itself or its packaging. Higher levels of packaging shall have their own UDI.

3.2. Shipping containers shall be exempted from the requirement in Section 3.1. By way of example, a UDI shall not be required on a logistics unit; where a healthcare provider orders multiple devices using the UDI or model number of individual devices and the manufacturer places those devices in a container for shipping or to protect the individually packaged devices, the container (logistics unit) shall not be subject to UDI requirements.

3.3. The UDI shall contain two parts: a UDI-DI and a UDI-PI.

3.4. The UDI-DI shall be unique at each level of device packaging.

Proposal for Regulation (EU) 2017/745 – Annex VI

3.5. If a lot number, serial number, software identification or expiry date appears on the label, it shall be part of the UDI-PI. If there is also a manufacturing date on the label, it does not need to be included in the UDI-PI. If there is only a manufacturing date on the label, this shall be used as the UDI-PI.

3.6. Each component that is considered to be a device and is commercially available on its own shall be assigned a separate UDI unless the components are part of a configurable device that is marked with its own UDI.

3.7. Systems and procedure packs as referred to in Article 22 shall be assigned and bear their own UDI.

3.8. The manufacturer shall assign the UDI to a device following the relevant coding standard.

3.9. A new UDI-DI shall be required whenever there is a change that could lead to misidentification of the device and/or ambiguity in its traceability; in particular any change of one of the following UDI database data elements shall require a new UDI-DI:

- (a) name or trade name,
- (b) device version or model,
- (c) labelled as single use,
- (d) packaged sterile,
- (e) need for sterilization before use,
- (f) quantity of devices provided in a package,
- (g) critical warnings or contra-indications: e.g. containing latex or DEHP.

3.10. Manufacturers that repackage and/or relabel devices, with their own label shall retain a record of the original device manufacturer's UDI.

4. UDI carrier

4.1. The UDI carrier (AIDC and HRI representation of the UDI) shall be placed on the label or on the device itself and on all higher levels of device packaging. Higher levels do not include shipping containers.

4.2. In the event of there being significant space constraints on the unit of use packaging the UDI carrier may be placed on the next higher packaging level.

4.3. For single-use devices of classes I and IIa packaged and labelled individually, the UDI carrier shall not be required to appear on the packaging but it shall appear on a higher level of packaging, e.g. a carton containing several individually packaged devices. However, when the healthcare provider is not expected to have access, in cases such as in home healthcare settings, to the higher level of device packaging, the UDI shall be placed on the packaging of the individual device.

4.4. For devices exclusively intended for retail point of sale the UDI-PIs in AIDC shall not be required to appear on the point of sale packaging.

4.5. When AIDC carriers other than the UDI carrier are part of the product labelling, the UDI carrier shall be readily identifiable.

4.6. If linear bar codes are used, the UDI-DI and UDI-PI may be concatenated or non-concatenated in two or more bar codes. All parts and elements of the linear bar code shall be distinguishable and identifiable.

4.7. If there are significant constraints limiting the use of both AIDC and HRI on the label, only the AIDC format shall be required to appear on the label. For devices intended to be used outside healthcare facilities such as devices for home care, the HRI shall however appear on the label even if this results in there being no space for the AIDC.

4.8. The HRI format shall follow the rules of the UDI code-issuing entity.

Proposal for Regulation (EU) 2017/745 – Annex VI

4.9. If the manufacturer is using RFID technology, a linear or 2D bar code in line with the standard provided by the issuing entities shall also be provided on the label.

4.10. Devices that are reusable shall bear a UDI carrier on the device itself. The UDI carrier for reusable devices that require cleaning, disinfection, sterilisation or refurbishing between patient uses shall be permanent and readable after each process performed to make the device ready for the subsequent use throughout the intended lifetime of the device, **unless the device is intended to be reused exclusively by or for the same patient**. The requirement of this Section shall not apply to devices in the following circumstances:

- (a) any type of direct marking would interfere with the safety or performance of the device;
- (b) the device cannot be directly marked because it is not technologically feasible.

4.11. The UDI carrier shall be readable during normal use and throughout the intended lifetime of the device.

4.12. If the UDI carrier is readily readable, or in the case of AIDC scannable, through the device's packaging, the placing of the UDI carrier on the packaging shall not be required.

4.13. In the case of single finished devices made up of multiple parts that must be assembled before their first use it shall be sufficient to place the UDI carrier on only one part of each device.

4.14. The UDI carrier shall be placed in a manner such that the AIDC can be accessed during normal operation or storage.

4.15. Bar code carriers that include both a UDI-DI and a UDI-PI may also include essential data for the device to operate or other data.

5. General principles of the UDI database

5.1. The UDI database shall support the use of all core UDI database data elements referred to in Part B of this Annex.

5.2. Manufacturers shall be responsible for the initial submission and updates of the identifying information and other device data elements in the UDI database.

5.3. Appropriate methods/procedures for validation of the data provided shall be implemented.

5.4. Manufacturers shall periodically verify the correctness of all of the data relevant to devices they have placed on the market, except for devices that are no longer available on the market.

5.5. The presence of the device UDI-DI in the UDI database shall not be assumed to mean that the device is in conformity with this Regulation.

5.6. The database shall allow for the linking of all the packaging levels of the device.

5.7. The data for new UDI-DIs shall be available at the time the device is placed on the market.

5.8. Manufacturers shall update the relevant UDI database record within 30 days of a change being made to an element, which does not require a new UDI-DI.

5.9. Internationally-accepted standards for data submission and updates shall, wherever possible, be used by the UDI database.

5.10. The user interface of the UDI database shall be available in all official languages of the Union. The use of free-text fields shall, however, be minimized in order to reduce translations.

5.11. Data relating to devices that are no longer available on the market shall be retained in the UDI database.

6. Rules for specific device types

6.1. Implantable devices:

Proposal for Regulation (EU) 2017/745 – Annex VI

6.1.1. Implantable devices shall, at their lowest level of packaging ('unit packs'), be identified, or marked using AIDC, with a UDI (UDI-DI + UDI-PI);

6.1.2. The UDI-PI shall have at least the following characteristics:

- (a) the serial number for active implantable devices,
- (b) the serial number or lot number for other implantable devices.

6.1.3. The UDI of the implantable device shall be identifiable prior to implantation.

6.2. Reusable devices requiring cleaning, disinfection, sterilisation or refurbishing between uses

6.2.1. The UDI of such devices shall be placed on the device and be readable after each procedure to make the device ready for the next use.

6.2.2. The UDI-PI characteristics such as the lot or serial number shall be defined by the manufacturer.

6.3. Systems and procedure packs as referred to in Article 22

6.3.1. The natural or legal person referred to in Article 22 shall be responsible for identifying the system or procedure pack with a UDI including both UDI-DI and UDI-PI.

6.3.2. Device contents of system or procedure packs shall bear a UDI carrier on their packaging or on the device itself.

Exemptions:

(a) individual single-use disposable devices, the uses of which are generally known to the persons by whom they are intended to be used, which are contained within a system or procedure pack, and which are not intended for individual use outside the context of the system or procedure pack, shall not be required to bear their own UDI carrier;

(b) devices that are exempted from bearing a UDI carrier on the relevant level of packaging shall not be required to bear a UDI carrier when included within a system or procedure pack.

6.3.3. Placement of the UDI carrier on systems or procedure packs

(a) The system or procedure pack UDI carrier shall as a general rule be affixed to the outside of the packaging.

(b) The UDI carrier shall be readable, or, in the case of AIDC, scannable, whether placed on the outside of the packaging of the system or procedure pack or inside transparent packaging.

6.4. Configurable devices:

6.4.1. A UDI shall be assigned to the configurable device in its entirety and shall be called the configurable device UDI.

6.4.2. The configurable device UDI-DI shall be assigned to groups of configurations, not per configuration within the group. A group of configurations is defined as the collection of possible configurations for a given device as described in the technical documentation.

6.4.3. A configurable device UDI-PI shall be assigned to each individual configurable device.

6.4.4. The carrier of the configurable device UDI shall be placed on the assembly that is most unlikely to be exchanged during the lifetime of the system and shall be identified as the configurable device UDI.

6.4.5. Each component that is considered a device and is commercially available on its own shall be assigned a separate UDI.

6.5. Device Software

6.5.1. UDI assignment Criteria

Proposal for Regulation (EU) 2017/745 – Annex VI

The UDI shall be assigned at the system level of the software. Only software which is commercially available on its own and software which constitutes a device in itself, shall be subject to that requirement.

The software identification shall be considered to be the manufacturing control mechanism and shall be displayed in the UDI-PI.

6.5.2. A new UDI-DI shall be required whenever there is a modification that changes:

- (a) the original performance;
- (b) the safety or the intended use of the software;
- (c) interpretation of data.

Such modifications include new or modified algorithms, database structures, operating platform, architecture or new user interfaces or new channels for interoperability.

6.5.3. Minor software revisions shall require a new UDI-PI and not a new UDI-DI.

Minor software revisions are generally associated with bug fixes, usability enhancements that are not for safety purposes, security patches or operating efficiency.

Minor software revisions shall be identified by a manufacturer-specific form of identification.

6.5.4. UDI placement criteria for software

(a) where the software is delivered on a physical medium, e.g. CD or DVD, each packaging level shall bear the human readable and AIDC representation of the complete UDI. The UDI that is applied to the physical medium containing the software and its packaging shall be identical to the UDI assigned to the system level software;

(b) the UDI shall be provided on a readily accessible screen for the user in an easily-readable plain-text format, such as an 'about' file, or included on the start-up screen;

(c) software lacking a user interface such as middleware for image conversion, shall be capable of transmitting the UDI through an application programming interface (API);

(d) only the human readable portion of the UDI shall be required in electronic displays of the software. The marking of UDI using AIDC shall not be required in the electronic displays, such as 'about' menu, splash screen etc.;

(e) the human readable format of the UDI for the software shall include the Application Identifiers (AI) for the standard used by the issuing entities, so as to assist the user in identifying the UDI and determining which standard is being used to create the UDI.

ANNEX VII REQUIREMENTS TO BE MET BY NOTIFIED BODIES

1. ORGANISATIONAL AND GENERAL REQUIREMENTS

1.1. Legal status and organisational structure

1.1.1. Each notified body shall be established under the national law of a Member State, or under the law of a third country with which the Union has concluded an agreement in this respect. Its legal personality and status shall be fully documented. Such documentation shall include information about ownership and the legal or natural persons exercising control over the notified body. **It shall also include information about the larger organisation to which the notified body belongs.**

1.1.2. If the notified body is a legal entity that is part of a larger organisation, the activities of that organisation as well as its organisational structure and governance, and the relationship with the notified body shall be clearly documented. In such cases, the requirements of Section 1.2 are applicable to both the notified body and the **larger organisation to which it belongs. The larger organisation and any of the legal entities belonging to it shall not be involved in the design, manufacture, marketing, installation or maintenance of the devices for which the notified body is designated or offer consultancy services for such activities, neither shall they represent the parties engaged in such activities.**

1.1.3. If a notified body wholly or partly owns legal entities established in a Member State or in a third country or is owned by another legal entity, the activities and responsibilities of those entities, as well as their legal and operational relationships with the notified body, shall be clearly defined and documented. Personnel of those entities performing conformity assessment activities under this Regulation shall be subject to the applicable requirements of this Regulation.

1.1.4. The organisational structure, allocation of responsibilities, reporting lines and operation of the notified body shall be such that they ensure that there is confidence in the performance by the notified body and in the results of the conformity assessment activities it conducts.

1.1.5. The notified body shall clearly document its organisational structure and the functions, responsibilities and authority of its top-level management and of other personnel who may have an influence upon the performance by the notified body and upon the results of its conformity assessment activities.

1.1.6. The notified body shall identify the persons in top-level management that have overall authority and responsibility for each of the following:

- the provision of adequate resources for conformity assessment activities;
- the development of procedures and policies for the operation of the notified body;
- the supervision of implementation of the procedures, policies and quality management systems of the notified body;
- the supervision of the notified body's finances;
- the activities and decisions taken by the notified body, including contractual agreements;
- the delegation of authority to personnel and/or committees, where necessary, for the performance of defined activities;
- the interaction with the authority responsible for notified bodies and the obligations regarding communications with other competent authorities, the Commission and other notified bodies.

1.2. Independence and impartiality

1.2.1. The notified body shall be a third-party body that is independent of the manufacturer of the device in relation to which it performs conformity assessment activities. The notified body shall also be independent of any other economic operator having an interest in the device as well as of any competitors of the

Proposal for Regulation (EU) 2017/745 – Annex VII

manufacturer. This does not preclude the notified body from carrying out conformity assessment activities for competing manufacturers.

1.2.2. The notified body shall be organised and operated so as to safeguard the independence, objectivity and impartiality of its activities. The notified body shall document and implement a structure and procedures for safeguarding impartiality and for promoting and applying the principles of impartiality throughout its organisation, personnel and assessment activities. Such procedures shall provide for the identification, investigation and resolution of any case in which a conflict of interest may arise, including involvement in consultancy services in the field of devices prior to taking up employment with the notified body. The investigation, outcome and its resolution shall be documented.

1.2.3. The notified body, its top-level management and the personnel responsible for carrying out the conformity assessment tasks shall not:

(a) be the designer, manufacturer, supplier, installer, purchaser, owner or maintainer of devices which they assess, nor the authorised representative of any of those parties. Such restriction shall not preclude the purchase and use of assessed devices that are necessary for the operations of the notified body and the conduct of the conformity assessment, or the use of such devices for personal purposes;

(b) be involved in the design, manufacture or construction, marketing, installation and use, or maintenance of the devices for which they are designated, nor represent the parties engaged in those activities;

(c) engage in any activity that may conflict with their independence of judgement or integrity in relation to conformity assessment activities for which they are designated;

(d) offer or provide any service which may jeopardise the confidence in their independence, impartiality or objectivity. In particular, they shall not offer or provide consultancy services to the manufacturer, its authorised representative, a supplier or a commercial competitor as regards the design, construction, marketing or maintenance of devices or processes under assessment, and

(e) be linked to any organisation which itself provides consultancy services as referred to in point (d). Such restriction does not preclude general training activities that are not client specific and that relate to regulation of devices or to related standards.

1.2.4. Involvement in consultancy services in the field of devices prior to taking up employment with a notified body shall be fully documented at the time of employment and potential conflicts of interest shall be monitored and resolved in accordance with this Annex. Personnel who were formerly employed by a specific client or provided consultancy services in the field of devices to that specific client prior to taking up employment with a notified body, shall not be assigned for conformity assessment activities for that specific client or companies belonging to the same group for a period of three years.

1.2.5. The impartiality of notified bodies, of their top level management and of the assessment personnel shall be guaranteed. The level of the remuneration of the top level management and assessment personnel of a notified body and subcontractors, involved in assessment activities shall not depend on the results of the assessments. Notified bodies shall make publicly available the declarations of interest of their top-level management.

1.2.6. If a notified body is owned by a public entity or institution, independence and absence of any conflict of interest shall be ensured and documented between, on the one hand, the authority responsible for notified bodies and/or the competent authority and, on the other hand, the notified body.

1.2.7. The notified body shall ensure and document that the activities of its subsidiaries or subcontractors, or of any associated body, including the activities of its owners do not affect its independence, impartiality or the objectivity of its conformity assessment activities.

1.2.8. The notified body shall operate in accordance with a set of consistent, fair and reasonable terms and conditions, taking into account the interests of small and medium-sized enterprises as defined in Recommendation 2003/361/EC in relation to fees.

1.2.9. The requirements laid down in this Section in no way preclude exchanges of technical information and regulatory guidance between a notified body and a manufacturer applying for conformity assessment. **The notified body shall have documented procedures in place to offer and carry out dialogues with the manufacturer before and after an application for conformity assessment is lodged.**

1.3. Confidentiality

1.3.1. The notified body shall have documented procedures in place ensuring that its personnel, committees, **branch offices**, subsidiaries, subcontractors, and any associated body or personnel of external bodies respect the confidentiality of the information which comes into its possession during the performance of conformity assessment activities, except when disclosure is required by law.

1.3.2. The personnel of a notified body shall observe professional secrecy in carrying out their tasks under this Regulation or any provision of national law giving effect to it, except in relation to the authorities responsible for notified bodies, competent authorities for medical devices in the Member States or the Commission. Proprietary rights shall be protected. The notified body shall have documented procedures in place in respect of the requirements of this Section.

1.4. Liability

1.4.1. The notified body shall take out appropriate liability insurance for its conformity assessment activities, unless liability is assumed by the Member State in question in accordance with national law or that Member State is directly responsible for the conformity assessment.

1.4.2. The scope and overall financial value of the liability insurance shall correspond to the level and geographic scope of activities of the notified body and be commensurate with the risk profile of the devices certified by the notified body. ~~The liability insurance shall cover cases where the notified body may be obliged to withdraw, restrict or suspend certificates.~~

1.4.3 By way of derogation from Section 1.4.1, the notified body may demonstrate coverage for liability through adherence to a guarantee fund that provides effective protection and is recognised by the Member State(s) concerned.

1.5. Financial requirements

The notified body shall have at its disposal the financial resources required to conduct its conformity assessment activities within its scope of designation and related business operations. It shall document and provide evidence of its financial capacity and its long-term economic viability, taking into account, where relevant, any specific circumstances during an initial start-up phase.

1.6. Participation in coordination activities

1.6.1. The notified body shall participate in ~~or ensure that its assessment personnel is informed of, any relevant standardisation activities and in~~ the activities of the notified body coordination group referred to in Article 49 and **ensure** that its assessment and decision-making personnel are informed of all relevant legislation, **standards**, guidance and best practice documents adopted in the framework of this Regulation.

1.6.2. The notified body shall take into consideration guidance and best practice documents.

2. QUALITY MANAGEMENT REQUIREMENTS

2.1. The notified body shall establish, document, implement, maintain and operate a quality management system that is appropriate to the nature, area and scale of its conformity assessment activities and is capable of supporting and demonstrating **in the most efficient manner** the consistent fulfilment of the requirements of this Regulation.

2.2. The quality management system of the notified body shall address at least the following:

- management system structure and documentation, including policies and objectives for its activities;
- policies for assignment of activities and responsibilities to personnel;

Proposal for Regulation (EU) 2017/745 – Annex VII

- assessment and decision-making processes in accordance with the tasks, responsibilities and role of the notified body's personnel and top-level management;
- the planning, conduct, evaluation and, if necessary, adaptation of its conformity assessment procedures;
- control of documents;
- control of records;
- management reviews;
- internal audits;
- corrective and preventive actions;
- complaints and appeals; and
- continuous training.

Where documents are used in various languages, the notified body shall ensure and control that they have the same content.

2.3. The top-level management of the notified body shall ensure that the quality management system is fully understood, implemented and maintained throughout the notified body organisation, including **branch offices**, subsidiaries and subcontractors involved in conformity assessment activities pursuant to this Regulation.

2.4. The notified body shall require all personnel to formally commit themselves by a signature or equivalent to comply with the procedures defined by the notified body. That commitment shall cover aspects relating to confidentiality and to independence from commercial and other interests, and any existing or prior association with clients. The personnel shall be required to complete written statements indicating their compliance with confidentiality, independence and impartiality principles.

3. RESOURCE REQUIREMENTS

3.1. General

3.1.1. Notified bodies shall be capable of carrying out all the tasks falling to them under this Regulation with the highest degree of professional integrity and the requisite competence in the specific field, whether those tasks are carried out by notified bodies themselves or on their behalf and under their responsibility.

In particular, notified bodies shall have the necessary personnel and possess or have access to all equipment, facilities and competence needed to perform properly the technical, scientific and administrative tasks entailed in the conformity assessment activities in relation to which they have been designated.

Such requirement presupposes at all times and for each conformity assessment procedure and each type of devices in relation to which they have been designated, that the notified body has permanent availability of sufficient administrative, technical and scientific personnel who possess experience and knowledge relating to the relevant devices and the corresponding technologies. Such personnel shall be in sufficient numbers to ensure that the notified body in question can perform the conformity assessment tasks, including the assessment of the medical functionality, clinical evaluations and the performance and safety of devices, for which it has been designated, having regard to the requirements of this Regulation, in particular, those set out in Annex I.

A notified body's cumulative competences shall be such as to enable it to assess the types of devices for which it is designated. The notified body shall have sufficient internal competence to critically evaluate assessments conducted by external expertise. Tasks which a notified body is precluded from subcontracting are set out in Section 4.1.

Personnel involved in the management of the operation of a notified body's conformity assessment activities for devices shall have appropriate knowledge to set up and operate a system for the selection of

Proposal for Regulation (EU) 2017/745 – Annex VII

assessment and verification staff, for verification of their competence, for authorisation and allocation of their tasks, for organisation of their initial and ongoing training and for the assignment of their duties and the monitoring of those staff, in order to ensure that personnel who carry out and perform assessment and verification operations are competent to fulfil the tasks required of them.

The notified body shall identify at least one individual within its top-level management as having overall responsibility for all conformity assessment activities in relation to devices.

3.1.2. The notified body shall ensure that personnel involved in conformity assessment activities maintain their qualification and expertise by implementing a system for exchange of experience and a continuous training and education programme.

3.1.3. The notified body shall clearly document the extent and limits of duties and responsibilities and the level of authorisation of the personnel, including any subcontractors and external experts, involved in conformity assessment activities and inform those personnel accordingly.

3.2. Qualification criteria in relation to personnel

3.2.1. The notified body shall establish and document qualification criteria and procedures for selection and authorisation of persons involved in conformity assessment activities, including as regards knowledge, experience and other competence required, and the required initial and ongoing training. The qualification criteria shall address the various functions within the conformity assessment process, such as auditing, product evaluation or testing, technical documentation review and decision-making, as well as the devices, technologies and areas, such as biocompatibility, sterilisation, tissues and cells of human and animal origin and clinical evaluation, covered by the scope of designation.

3.2.2. The qualification criteria referred to in Section 3.2.1 shall refer to the scope of a notified body's designation in accordance with the scope description used by the Member State for the notification referred to in Article 42(3), providing a sufficient level of detail for the required qualification within the subdivisions of the scope description.

Specific qualification criteria shall be defined at least for the assessment of:

- the pre-clinical evaluation,
- clinical evaluation,
- tissues and cells of human and animal origin,
- functional safety,
- software,
- packaging,
- devices that incorporate as an integral part a medicinal product,
- devices that are composed of substances or of combinations of substances that are absorbed by or locally dispersed in the human body and
- the different types of sterilisation processes.

3.2.3. The personnel responsible for establishing qualification criteria and for authorising other personnel to perform specific conformity assessment activities shall ~~be employed by the notified body itself and shall~~ not be external experts or be subcontracted. They shall have proven knowledge and experience in all of the following:

- Union devices legislation and relevant guidance documents;
- the conformity assessment procedures provided for in this Regulation;
- a broad base of knowledge of device technologies and the design and manufacture of devices;

Proposal for Regulation (EU) 2017/745 – Annex VII

- the notified body's quality management system, related procedures and the required qualification criteria;
- training relevant to personnel involved in conformity assessment activities in relation to devices;
- adequate experience in conformity assessments under this Regulation, **Regulation (EU) 2017/746** or previously applicable law within a notified body.

3.2.4. The notified body shall have permanent availability of personnel with relevant clinical expertise ~~and where possible such personnel shall be employed by the notified body itself~~. Such personnel shall be integrated throughout the notified body's assessment and decision-making process in order to:

- identify when specialist input is required for the assessment of the clinical evaluation conducted by the manufacturer and identify appropriately qualified experts;
- appropriately train external clinical experts in the relevant requirements of this Regulation, CS, guidance and harmonised standards and ensure that the external clinical experts are fully aware of the context and implications of their assessment and the advice they provide;
- be able to review and scientifically challenge the clinical data contained within the clinical evaluation, and any associated clinical investigations, and appropriately guide external clinical experts in the assessment of the clinical evaluation presented by the manufacturer;
- be able to scientifically evaluate and, if necessary, challenge the clinical evaluation presented, and the results of the external clinical experts' assessment of the manufacturer's clinical evaluation;
- be able to ascertain the comparability and consistency of the assessments of clinical evaluations conducted by clinical experts;
- be able to make an assessment of the manufacturer's clinical evaluation and a clinical judgement of the opinion provided by any external expert and make a recommendation to the notified body's decision maker; and
- be able to draw up records and reports demonstrating that the relevant conformity assessment activities have been appropriately carried out.

3.2.5. The personnel responsible for carrying out product-related reviews (product reviewers), such as technical documentation reviews or type examination, including aspects such as clinical evaluation, biological safety, sterilisation and software validation, shall have all of the following proven qualifications:

- successful completion of a university or a technical college degree or equivalent qualification in relevant studies, e.g. medicine, pharmacy, engineering or other relevant sciences;
- four years' professional experience in the field of healthcare products or related activities, such as in manufacturing, auditing or research, of which two years shall be in the design, manufacture, testing or use of the device or technology to be assessed or related to the scientific aspects to be assessed;
- knowledge of device legislation, including the general safety and performance requirements set out in Annex I;
- appropriate knowledge and experience of relevant harmonised standards, CS and guidance documents;
- appropriate knowledge and experience of risk management and related device standards and guidance documents;
- appropriate knowledge and experience of clinical evaluation;
- appropriate knowledge of the devices which they are assessing;
- appropriate knowledge and experience of the conformity assessment procedures laid down in Annexes IX to XI, in particular of the aspects of those procedures for which they are responsible, and adequate authorisation for carrying out those assessments;

Proposal for Regulation (EU) 2017/745 – Annex VII

– the ability to draw up records and reports demonstrating that the relevant conformity assessment activities have been appropriately carried out.

3.2.6. The personnel responsible for carrying out audits of the manufacturer's quality management system (site auditors) shall have all of the following proven qualifications:

- successful completion of a university or a technical college degree or equivalent qualification in relevant studies, such as medicine, pharmacy, engineering or other relevant sciences;
- four years' professional experience in the field of healthcare products or related activities, such as in manufacturing, auditing or research, of which two years shall be in the area of quality management;
- appropriate knowledge of devices legislation as well as related harmonised standards, CS and guidance documents;
- appropriate knowledge and experience of risk management and related device standards and guidance documents;
- appropriate knowledge of quality management systems and related standards and guidance documents;
- appropriate knowledge and experience of the conformity assessment procedures laid down in Annexes IX to XI, in particular of the aspects of those procedures for which they are responsible, and adequate authorisation for carrying out those audits;
- training in auditing techniques enabling them to challenge quality management systems;
- the ability to draw up records and reports demonstrating that the relevant conformity assessment activities have been appropriately carried out.

3.2.7. The personnel with overall responsibility for final reviews and decision-making on certification shall ~~be employed by the notified body itself and shall~~ not be external experts or be subcontracted. Those personnel shall, as a group, have proven knowledge and comprehensive experience of all of the following:

- devices legislation and relevant guidance documents;
- the device conformity assessments relevant to this Regulation;
- the types of qualifications, experience and expertise relevant to device conformity assessment;
- a broad base of knowledge of device technologies, including sufficient experience of conformity assessment of devices being reviewed for certification, the device industry and the design and manufacture of devices;
- the notified body's quality management system, related procedures and the required qualifications for personnel involved;
- the ability to draw up records and reports demonstrating that the conformity assessment activities have been appropriately carried out.

3.3. Documentation of qualification, training and authorisation of personnel

3.3.1. The notified body shall have a procedure in place to fully document the qualification of each member of personnel involved in conformity assessment activities and the satisfaction of the qualification criteria referred to in Section 3.2. Where in exceptional circumstances the fulfilment of the qualification criteria set out in Section 3.2. cannot be fully demonstrated, the notified body shall justify to the authority responsible for notified bodies the authorisation of those members of personnel to carry out specific conformity assessment activities.

3.3.2. For all of its personnel referred to in Sections 3.2.3 to 3.2.7, the notified body shall establish and maintain up to date:

- a matrix detailing the authorisations and responsibilities of the personnel in respect of conformity assessment activities; and

Proposal for Regulation (EU) 2017/745 – Annex VII

– records attesting to the required knowledge and experience for the conformity assessment activity for which they are authorised. The records shall contain a rationale for defining the scope of the responsibilities for each of the assessment personnel and records of the conformity assessment activities carried out by each of them.

3.4. Subcontractors and external experts

3.4.1. Notified bodies may, without prejudice to Section 3.2, subcontract certain clearly defined component parts of a conformity assessment activity. **They shall inform the manufacturer accordingly.**

The subcontracting of the auditing of quality management systems or of product related reviews as a whole shall not be permitted; nevertheless parts of those activities may be conducted by subcontractors and external auditors and experts working on behalf of the notified body. The notified body in question shall retain full responsibility for being able to produce appropriate evidence of the competence of subcontractors and experts to fulfil their specific tasks, for making a decision based on a subcontractor's assessment and for the work conducted by subcontractors and experts on its behalf.

The following activities may not be subcontracted by notified bodies:

- review of the qualifications and monitoring of the performance of external experts;
- auditing and certification activities where the subcontracting in question is to auditing or certification organisations;
- allocation of work to external experts for specific conformity assessment activities; and
- final review and decision making functions.

3.4.2. Where a notified body subcontracts certain conformity assessment activities either to an organisation or an individual, it shall have a policy describing the conditions under which subcontracting may take place, and shall ensure that:

- the subcontractor meets the relevant requirements of this Annex;
- subcontractors and external experts do not further subcontract work to organisations or personnel; and
- the natural or legal person that applied for conformity assessment has been informed of the requirements referred to in the first and second indent.

Any subcontracting or consultation of external personnel shall be properly documented, shall not involve any intermediaries and shall be subject to a written agreement covering, among other things, confidentiality and conflicts of interest. The notified body in question shall take full responsibility for the tasks performed by subcontractors.

3.4.3. Where subcontractors or external experts are used in the context of a conformity assessment, in particular regarding novel, invasive and implantable devices or technologies, the notified body in question shall have internal competence in each product area for which it is designated that is adequate for the purpose of leading the overall conformity assessment, verifying the appropriateness and validity of expert opinions and making decisions on certification.

3.5. Monitoring of competences, training and exchange of experience

3.5.1. The notified body shall establish procedures for the initial evaluation and on-going monitoring of the competence, conformity assessment activities and performance of all internal and external personnel, and subcontractors, involved in conformity assessment activities.

3.5.2. Notified bodies shall review at regular intervals, the competence of their personnel, identify training needs and draw up a training plan to maintain the required level of qualification and knowledge of individual personnel. That review shall at a minimum, verify that personnel:

- are aware of Union and national law in force on devices, relevant harmonised standards, CS, guidance documents and the results of the coordination activities referred to in Section 1.6; and

Proposal for Regulation (EU) 2017/745 – Annex VII

– take part in the internal exchange of experience and the continuous training and education programme referred to in Section 3.1.2.

4. PROCESS REQUIREMENTS

4.1. General

The notified body shall have in place documented processes and sufficiently detailed procedures for the conduct of each conformity assessment activity for which it is designated, comprising the individual steps from pre-application activities up to decision making and surveillance and taking into account, when necessary, the respective specificities of the devices.

The requirements laid down in Sections 4.3, 4.4, 4.7 and 4.8 shall ~~be fulfilled as part of the internal activities of notified bodies and shall~~ not be subcontracted or fulfilled by external experts.

4.2. notified body quotations and pre-application activities

The notified body shall:

- (a) publish a publicly available description of the application procedure by which manufacturers can obtain certification from it. That description shall include which languages are acceptable for submission of documentation and for any related correspondence **and the modalities for carrying out the dialogue referred to in Section 1.2.9 before an application is lodged;**
- (b) have documented procedures relating to, and documented details about, fees charged for specific conformity assessment activities and any other financial conditions relating to notified bodies' assessment activities for devices;
- (c) have documented procedures in relation to advertising of their conformity assessment services. Those procedures shall ensure that advertising or promotional activities in no way imply or are capable of leading to an inference that their conformity assessment will offer manufacturers earlier market access or be quicker, easier or less stringent than that of other notified bodies;
- (d) have documented procedures requiring the review of pre-application information, including the preliminary verification that the product is covered by this Regulation and its classification, prior to issuing any quotation to the manufacturer relating to a specific conformity assessment; and
- (e) ensure that all contracts relating to the conformity assessment activities covered by this Regulation are concluded directly between the manufacturer and the notified body and not with any other organisation.

4.3. Application review and contract

The notified body shall require a formal application signed by a manufacturer or an authorised representative containing all of the information and the manufacturer's declarations required by the relevant conformity assessment as referred to in Annexes IX to XI.

The contract between a notified body and a manufacturer shall take the form of a written agreement signed by both parties. It shall be kept by the notified body. This contract shall have clear terms and conditions and contain obligations that enable the notified body to act as required under this Regulation, including an obligation on the manufacturer to inform the notified body of vigilance reports, the right of the notified body to suspend, restrict or withdraw certificates issued and the duty of the notified body to fulfil its information obligations.

The notified body shall have documented procedures to review applications, addressing:

- (a) the completeness of those applications with respect to the requirements of the relevant conformity assessment procedure, as referred to in the corresponding Annex, under which approval has been sought,
- (b) the verification of the qualification of products covered by those applications as devices and their respective classifications,

Proposal for Regulation (EU) 2017/745 – Annex VII

(c) whether the conformity assessment procedures chosen by the applicant are applicable to the device in question under this Regulation,

(d) the ability of the notified body to assess the application based on its designation, and

(e) the availability of sufficient and appropriate resources.

The outcome of each review of an application shall be documented. Refusals or withdrawals of applications shall be notified to the electronic system referred to in Article 57 and shall be accessible to other notified bodies.

4.4. Allocation of resources

The notified body shall have documented procedures to ensure that all conformity assessment activities are conducted by appropriately authorised and qualified personnel who are sufficiently experienced in the evaluation of the devices, systems and processes and related documentation that are subject to conformity assessment.

For each application, the notified body shall determine the resources needed and identify one individual responsible for ensuring that the assessment of that application is conducted in accordance with the relevant procedures and for ensuring that the appropriate resources including personnel are utilised for each of the tasks of the assessment. The allocation of tasks required to be carried out as part of the conformity assessment and any changes subsequently made to this allocation shall be documented.

4.5. Conformity assessment activities

4.5.1. General

The notified body and its personnel shall carry out the conformity assessment activities with the highest degree of professional integrity and the requisite technical and scientific competence in the specific fields.

The notified body shall have expertise, facilities and documented procedures that are sufficient to effectively conduct the conformity assessment activities for which the notified body in question is designated, taking account of the relevant requirements set out in Annexes IX to XI, and in particular all of the following requirements:

- appropriately plan the conduct of each individual project,
- ensure that the composition of the assessment teams is such that there is sufficient experience in relation to the technology concerned, and that there is continuous objectivity and independence, and to provide for rotation of the members of the assessment team at appropriate intervals,
- specify the rationale for fixing time limits for completion of conformity assessment activities,
- assess the manufacturer's technical documentation and the solutions adopted to meet the requirements laid down in Annex I,
- review the manufacturer's procedures and documentation relating to the evaluation of pre-clinical aspects,
- review the manufacturer's procedures and documentation relating to clinical evaluation,
- address the interface between the manufacturer's risk management process and its appraisal and analysis of the pre-clinical and clinical evaluation and to evaluate their relevance for the demonstration of conformity with the relevant requirements in Annex I,
- carry out the specific procedures referred to in Sections 5.2 to 5.4 of Annex IX,
- in the case of class IIa or class IIb devices, assess the technical documentation of ~~devices selected on a~~ the representative device(s) basis;

Proposal for Regulation (EU) 2017/745 – Annex VII

- plan and periodically carry out appropriate surveillance audits and assessments, carry out or request certain tests to verify the proper functioning of the quality management system and to perform unannounced on site audits,

~~– relating to the sampling of devices, verify that the manufactured device is in conformity with the technical documentation; such requirements shall define the relevant sampling criteria and testing procedure prior to sampling;~~

- evaluate and verify a manufacturer's compliance with relevant Annexes.

- where appropriate, perform a rolling review of the manufacturer's data or documentation as they become available;

- in case of class I devices that are placed on the market in sterile condition, have measuring function or are reusable surgical instrument, assess the quality management system only in relation to the relevant special aspects of these devices;

- leverage evidence from previous assessments performed.

The notified body shall, where relevant, take into consideration available CS, guidance and best practice documents and harmonised standards, even if the manufacturer does not claim to be in compliance.

4.5.2. Quality management system auditing

(a) As part of the assessment of the quality management system, a notified body shall prior to ~~or in relation with~~ an audit and in accordance with its documented procedures:

- assess the documentation submitted in accordance with the relevant conformity assessment Annex, and draw up an audit programme which clearly identifies the number and sequence of activities required to demonstrate complete coverage of a manufacturer's quality management system and to determine whether it meets the requirements of this Regulation,

- identify links between, and allocation of responsibilities among, the various manufacturing sites, and identify relevant suppliers and/or subcontractors of the manufacturer, and consider the need to specifically audit any of those suppliers or subcontractors or both,

- clearly define, for each audit identified in the audit programme, the objectives, criteria and scope of the audit, and draw up an audit plan that adequately addresses and takes account of the specific requirements for the devices, technologies and processes involved,

- ~~clearly identify [draw up and keep up to date], for class IIa and class IIb devices, a sampling plan the representative devices selected for the assessment of technical documentation as referred to in Annexes II and III, covering the range of such devices covered by the manufacturer's application. That plan shall ensure that the entire range of devices covered by the certificate is sampled over the period of validity of the certificate, and~~

- select and assign appropriately qualified and authorised personnel for conducting the individual audits. The respective roles, responsibilities and authorities of the team members shall be clearly defined and documented.

(b) Based on the audit programme it has drawn up, the notified body shall, in accordance with its documented procedures:

- audit the manufacturer's quality management system, in order to verify that the quality management system ensures that the devices covered conform to the relevant provisions of this Regulation which apply to devices at every stage, from design through final quality control to ongoing surveillance, and shall determine whether the requirements of this Regulation are met,

- based on relevant technical documentation and in order to determine whether the manufacturer meets the requirements referred to in the relevant conformity assessment Annex, review and audit the manufacturer's processes and subsystems, in particular for:

Proposal for Regulation (EU) 2017/745 – Annex VII

- design and development,
- production and process controls,
- product documentation,
- purchasing controls including verification of purchased devices,
- corrective and preventive actions, including for post-market surveillance, and
- PMCF,

and review and audit requirements and provisions adopted by the manufacturer, including those in relation to fulfilling the general safety and performance requirements set out in Annex I.

The documentation shall be sampled in such a manner as to reflect the risks associated with the intended use of the device, the complexity of the manufacturing technologies, the range and classes of devices produced **or under certification** and any available post-market surveillance information,

– if not already covered by the audit programme, audit the control of processes on the premises of the manufacturer's suppliers, when the conformity of finished devices is significantly influenced by the activity of suppliers and, in particular when the manufacturer cannot demonstrate sufficient control over its suppliers,

– conduct assessments of the technical documentation based on its sampling plan and taking account of Sections 4.5.4. and 4.5.5. for pre-clinical and clinical evaluations, and

– the notified body shall ensure that audit findings are appropriately and consistently classified in accordance with the requirements of this Regulation and with relevant standards, or with best practice documents developed or adopted by the MDCG.

4.5.3. Product verification

Assessment of the technical documentation

For assessment of the technical documentation conducted in accordance with Chapter II of Annex IX, notified bodies shall have sufficient expertise, facilities and documented procedures for:

- the allocation of appropriately qualified and authorised personnel for the examination of individual aspects such as use of the device, biocompatibility, clinical evaluation, risk management, and sterilisation, and
- the assessment of conformity of the design with this Regulation. and for taking account of Sections 4.5.4. to 4.5.6. That assessment shall include examination of the implementation by manufacturers of incoming, in-process and final checks and the results thereof. If further tests or other evidence is required for the assessment of conformity with the requirements of this Regulation, the notified body in question shall carry out adequate physical or laboratory tests in relation to the device or request the manufacturer to carry out such tests.

Type-examinations

The notified body shall have documented procedures, sufficient expertise and facilities for the type-examination of devices in accordance with Annex X including the capacity to:

- examine and assess the technical documentation taking account of Sections 4.5.4. to 4.5.6., and verify that the type has been manufactured in conformity with that documentation;
- establish a test plan identifying all relevant and critical parameters which need to be tested by the notified body or under its responsibility;
- document its rationale for the selection of those parameters;

Proposal for Regulation (EU) 2017/745 – Annex VII

- carry out the appropriate examinations and tests in order to verify that the solutions adopted by the manufacturer meet the general safety and performance requirements set out in Annex I. Such examinations and tests shall include all tests necessary to verify that the manufacturer has in fact applied the relevant standards it has opted to use;
- agree with the applicant as to where the necessary tests will be performed if they are not to be carried out directly by the notified body; and
- assume full responsibility for test results. Test reports submitted by the manufacturer shall only be taken into account if they have been issued by conformity assessment bodies which are competent and independent of the manufacturer.

Verification by examination and testing of every product

The notified body shall:

- (a) have documented procedures, sufficient expertise and facilities for the verification by examination and testing of every product in accordance with Part B of Annex XI;
- (b) establish a test plan identifying all relevant and critical parameters which need to be tested by the notified body or under its responsibility in order to:
 - verify, for class IIb devices, the conformity of the device with the type described in the EU type-examination certificate and with the requirements of this Regulation which apply to those devices,
 - confirm, for class IIa devices, the conformity with the technical documentation referred to in Annexes II and III and with the requirements of this Regulation which apply to those devices;
- (c) document its rationale for the selection of the parameters referred to in point (b);
- (d) have documented procedures to carry out the appropriate assessments and tests in order to verify the conformity of the device with the requirements of this Regulation by examining and testing every product as specified in Section 15 of Annex XI;
- (e) have documented procedures providing for the reaching of an agreement with the applicant concerning when and where necessary tests that are not to be carried out by the notified body itself are to be performed; and
- (f) assume full responsibility for test results in accordance with documented procedures; test reports submitted by the manufacturer shall only be taken into account if they have been issued by conformity assessment bodies which are competent and independent of the manufacturer.

4.5.4. Pre-clinical evaluation assessment

The notified body shall have documented procedures in place for the review of the manufacturer's procedures and documentation relating to the evaluation of pre-clinical aspects. The notified body shall examine, validate and verify that the manufacturer's procedures and documentation adequately address:

- (a) the planning, conduct, assessment, reporting and, where appropriate, updating of the pre-clinical evaluation, in particular of
 - the scientific pre-clinical literature search, and
 - the pre-clinical testing, for example laboratory ~~testing~~, in vitro, ex vivo, in silico testing, simulated use testing, computer modelling, the use of animal models,
- (b) the nature and duration of body contact and the specific associated biological risks,
- (c) the interface with the risk management process, and
- (d) the appraisal and analysis of the available pre-clinical data and its relevance with regard to demonstrating conformity with the relevant requirements in Annex I.

Proposal for Regulation (EU) 2017/745 – Annex VII

The notified body's assessment of pre-clinical evaluation procedures and documentation shall address the results of literature searches and all validation, verification and testing performed and conclusions drawn, and shall typically include considering the use of alternative materials and substances and take account of the packaging, stability, including shelf life, of the finished device. Where no new testing has been undertaken by a manufacturer or where there are deviations from procedures, the notified body in question shall critically examine the justification presented by the manufacturer.

4.5.5. clinical evaluation assessment

The notified body shall have documented procedures in place relating to the assessment of a manufacturer's procedures and documentation relating to clinical evaluation both for initial conformity assessment and on an ongoing basis. The notified body shall examine, validate and verify that manufacturers' procedures and documentation adequately address:

- the planning, conduct, assessment, reporting and updating of the clinical evaluation as referred to in Annex XIV,
- post-market surveillance and PMCF,
- the interface with the risk management process,
- the appraisal and analysis of the available data and its relevance with regard to demonstrating conformity with the relevant requirements in Annex I, and
- the conclusions drawn with regard to the clinical evidence and drawing up of the clinical evaluation report.

These procedures referred to in the first paragraph shall take into consideration available CS, guidance and best practice documents.

The notified body's assessment of clinical evaluations as referred to in Annex XIV shall cover:

- the intended use specified by the manufacturer and claims for the device defined by it,
- the planning of the clinical evaluation,
- the methodology for the literature search,
- relevant documentation from the literature search,
- the clinical investigation,
- validity of equivalence claimed in relation to other devices, the demonstration of equivalence, the suitability and conclusions data from equivalent and similar devices,
- post-market surveillance and PMCF,
- the clinical evaluation report, and
- justifications in relation to non-performance of clinical investigations or PMCF.
- where applicable, the justification for the confirmation of the safety and performance that is based on the results of non-clinical testing methods alone.

In relation to clinical data from clinical investigations included within the clinical evaluation, the notified body in question shall ensure that the conclusions drawn by the manufacturer are valid in the light of the approved clinical investigation plan.

The notified body shall ensure that the clinical evaluation adequately addresses the relevant safety and performance requirements provided for in Annex I, that it is appropriately aligned with the risk management requirements, that it is conducted in accordance with Annex XIV and that it is appropriately reflected in the information provided relating to the device.

4.5.6. Specific Procedures

Proposal for Regulation (EU) 2017/745 – Annex VII

The notified body shall have documented procedures, sufficient expertise and facilities for the procedures referred to in Sections 5 and 6 of Annex IX, Section 6 of Annex X and Section 16 of Annex XI, for which they are designated.

In the case of devices manufactured utilising tissues or cells of animal origin or their derivatives, such as from **transmissible spongiform encephalopathy (TSE)** susceptible species, as referred to in Regulation (EU) No 722/2012, **or in any subsequent implementing rules adopted pursuant to this Regulation**, the notified body shall have documented procedures in place that fulfil the requirements laid down in **that Regulation (EU) No 722/2012, including for the preparation of a summary evaluation report for the relevant competent authority or those subsequent implementing rules.**

4.6. Reporting

The notified body shall:

- ensure that all steps of the conformity assessment are documented so that the conclusions of the assessment are clear and demonstrate compliance with the requirements of this Regulation and can represent objective evidence of such compliance to persons that are not themselves involved in the assessment, for example personnel in designating authorities,
- ensure that records that are sufficient to provide a discernible audit trail are available for quality management system audits,
- clearly document the conclusions of its assessment of clinical evaluation in a clinical evaluation assessment report, and
- for each specific project, provide a detailed report which shall be based on a standard format containing a minimum set of elements determined by the MDCG.

The report(s) of the notified body shall:

- clearly document the outcome of its assessment and draw clear conclusions from the verification of the manufacturer's conformity with the requirements of this Regulation,
- make a recommendation for a final review and for a final decision to be taken by the notified body; this recommendation shall be signed off by the member of personnel responsible in the notified body, and
- be provided to the manufacturer in question.

4.7. Final review

The notified body shall prior to making a final decision:

- ensure that the personnel assigned for the final review and decision-making on specific projects are appropriately authorised and are different from the personnel who have conducted the assessments,
- verify that the report or reports and supporting documentation needed for decision making, including concerning resolution of non-conformities noted during assessment, are complete and sufficient with respect to the scope of the application, and
- verify whether there are any unresolved non-conformities preventing issuance of a certificate.

4.8. Decisions and Certifications

The notified body shall have documented procedures for decision-making including as regards the allocation of responsibilities for the issuance, suspension, **limitation [restriction]** and withdrawal of certificates. Those procedures shall include the notification requirements laid down in Chapter V of this Regulation. The procedures shall allow the notified body in question to:

- decide, based on the assessment documentation and additional information available, whether the requirements of this Regulation are fulfilled,

Proposal for Regulation (EU) 2017/745 – Annex VII

- decide, based on the results of its assessment of the clinical evaluation and risk management, whether the post-market surveillance plan, including the PMCF plan, is adequate,
- decide on specific milestones for further review by the notified body of the up to date clinical evaluation,
- decide whether ~~specific~~ conditions or ~~limitations~~ [provisions] need to be defined for the certification,
- ~~where appropriate~~, decide, based on the novelty, risk classification, clinical evaluation and conclusions from the risk analysis of the device, on a period of certification ~~not exceeding five years~~,
- clearly document decision making and approval steps including approval by signature of the members of personnel responsible,
- clearly document responsibilities and mechanisms for communication of decisions, in particular, where the final signatory of a certificate differs from the decision maker or decision makers or does not fulfil the requirements laid down in Section 3.2.7,
- issue a certificate or certificates in accordance with the minimum requirements laid down in Annex XII ~~for a period of validity not exceeding five years~~ and ~~shall~~ indicate whether there are ~~specific~~ conditions or limitations associated with the certification,
- issue a certificate or certificates for the applicant alone and shall not issue certificates covering multiple entities, and
- ensure that the manufacturer is notified of the outcome of the assessment and the resultant decision and that they are entered into the electronic system referred to in Article 57.

4.9. Changes and modifications

The notified body shall have documented procedures and contractual arrangements with manufacturers in place relating to the manufacturers' information obligations and the assessment of changes to:

- the approved quality management system or systems or to the product-range covered,
- the approved design of a device,
- the intended ~~purpose~~ [use] of or claims made for the device,
- the approved type of a device, and
- any substance incorporated in or utilised for the manufacturing of a device and being subject to the specific procedures in accordance with Section 4.5.6.

The procedures and contractual arrangements referred to in the first paragraph shall ~~include measures for checking the significance of the changes referred to in the first paragraph~~ clearly distinguish between changes that do not need to be reported, that need to be reported without requiring prior approval and that require prior approval.';

In accordance with its documented procedures, the notified body in question shall, ~~where changes require prior approval~~:

~~Where appropriate, the notified body and the manufacturer shall agree on a predetermined change control plan enabling the manufacturer to implement changes in accordance with such a plan without prior information.~~

- ensure that manufacturers submit for prior approval plans for changes as referred to in the first paragraph and relevant information relating to such changes,
- assess the changes proposed and verify whether, after these changes, the quality management system, or the design of a device or type of a device, still meets the requirements of this Regulation, and
- notify the manufacturer of its decision and provide a report or as applicable a supplementary report, which shall contain the justified conclusions of its assessment.

4.10. Surveillance activities and post-certification monitoring

The notified body shall have documented procedures:

- defining how and when surveillance activities of manufacturers are to be conducted. Those procedures shall include arrangements for unannounced on-site audits of manufacturers and, where applicable, subcontractors and suppliers carrying out product tests and the monitoring of compliance with any conditions binding manufacturers and associated with certification decisions, such as updates to clinical data at defined intervals,
- for screening relevant sources of scientific and clinical data and post-market information relating to the scope of their designation. ~~That screening shall be conducted in the framework of the coordination group established in Article 49 to avoid unnecessary duplication and to enhance efficiency and work-sharing.~~ ~~The findings from the screening~~ [Such information] shall be taken into account in the planning and conduct of surveillance activities, ~~and~~
- ~~to assess whether any reported serious incident related to a serious public health threat or any field safety corrective action taken or envisaged by the manufacturer or required of it by a competent authority has an~~ [to review vigilance data to which they have access under Article 92(2) in order to estimate its] impact, ~~if any,~~ on the validity of existing certificates. The results of the evaluation and any decisions taken shall be thoroughly documented.

The notified body in question shall, ~~with regard to signals arising from vigilance data to which they have access under Article 92(2),~~ [upon receipt of information about vigilance cases from a manufacturer or competent authorities] decide which of the following options to apply, ~~where appropriate:~~

- ~~– not to take action on the basis that the vigilance case is clearly not related to the certification granted,~~
- observe the manufacturer's and competent authority's activities and the results of the manufacturer's investigation so as to determine whether the certification granted is at risk or whether adequate corrective action has been taken,
- perform extraordinary surveillance measures, such as document reviews, short-notice or unannounced audits and product testing, where it is likely that the certification granted is at risk,
- increase the frequency of surveillance audits,
- review specific products or processes on the occasion of the next audit of the manufacturer, or
- take any other relevant measure.

In relation to surveillance audits of manufacturers, the notified body shall have documented procedures to:

- conduct surveillance audits of the manufacturer ~~on at least an annual basis~~ which shall be planned and conducted in line with the relevant requirements in Section 4.5,
- ensure adequate assessment of the manufacturer's documentation on, and application of the provisions on, vigilance, the post-market surveillance, and PMCF,
- sample and test devices and technical documentation, during audits, according to pre-defined sampling criteria and testing procedures to ensure that the manufacturer continuously applies the approved quality management system,
- ensure that the manufacturer complies with the documentation and information obligations laid down in the relevant Annexes and that its procedures take into account best practices in the implementation of quality management systems,
- ensure that the manufacturer does not use quality management system or device approvals in a misleading manner,
- gather sufficient information to determine if the quality management system continues to comply with the requirements of this Regulation,

Proposal for Regulation (EU) 2017/745 – Annex VII

– ask the manufacturer, if non-conformities are detected, for corrections, corrective actions and, where applicable, preventive actions, and

– where necessary, impose **conditions or limitations** [specific restrictions] on the relevant certificate or suspend or withdraw it.

The notified body shall, if listed as part of the conditions for certification:

– conduct an in-depth review of the clinical evaluation as most recently updated by the manufacturer based on the manufacturer's post-market surveillance, on its PMCF and on clinical literature relevant to the condition being treated with the device or on clinical literature relevant to similar devices,

– clearly document the outcome of the in-depth review and address any specific concerns to the manufacturer or impose any specific conditions on it, and

– ensure that the clinical evaluation as most recently updated, is appropriately reflected in the instructions for use and, where applicable, the summary of safety and **clinical** performance.

4.11. Periodic reviews and extension of a certificate's period of validity [Re-certification]

The notified body shall have documented procedures in place relating to **periodic reviews** [the re-certification reviews and the renewal of certificates]. ~~Re-certification~~ of approved quality management systems or EU technical documentation assessment certificates or EU type-examination certificates ~~shall occur at least every five years.~~

~~The notified body shall have documented procedures relating to renewals of EU technical documentation assessment certificates and EU type-examination certificates and those procedures shall require the manufacturer in question to submit a summary of changes and scientific findings for the device, including:~~

Those procedures shall require the manufacturer in question to submit at predefined intervals a summary of changes and of relevant data gathered by the manufacturer's post-market surveillance system. **The notified body shall assess such information** [The notified body shall have documented procedures to assess the information referred to in the second paragraph] and shall pay particular attention to clinical data from post-market surveillance and PMCF activities undertaken since the previous certification or **periodic review** [re-certification], including appropriate updates to manufacturers' clinical evaluation reports, ~~without repeating assessments already conducted.~~

The notified body shall have documented procedures in place relating to the extension of the period of validity of a certificate in cases where it has exceptionally limited the period of validity. Those procedures shall require the manufacturer to submit prior to the expiry of the certificate the data or documentation specified by the notified body to enable it to decide about the extension of the period of validity of the certificate.

~~For the decision on re-certification, the notified body in question shall use the same methods and principles as for the initial certification decision. If necessary, separate forms shall be established for re-certification taking into account the steps taken for certification such as application and application review.~~

~~(a) all changes to the originally approved device, including changes not yet notified,~~

~~(b) experience gained from post-market surveillance,~~

~~(c) experience from risk management,~~

~~(d) experience from updating the proof of compliance with the general safety and performance requirements set out in Annex I,~~

~~(e) experience from reviews of the clinical evaluation, including the results of any clinical investigations and PMCF,~~

~~(f) changes to the requirements, to components of the device or to the scientific or regulatory environment,~~

~~(g) changes to applied or new harmonised standards, CS or equivalent documents, and~~

~~(h) changes in medical, scientific and technical knowledge, such as:~~

Proposal for Regulation (EU) 2017/745 – Annex VII

- ~~— new treatments,~~
- ~~— changes in test methods,~~
- ~~— new scientific findings on materials and components, including findings on their biocompatibility,~~
- ~~— experience from studies on comparable devices,~~
- ~~— data from registers and registries,~~
- ~~— experience from clinical investigations with comparable devices.~~

~~The notified body shall have documented procedures to assess the information referred to in the second paragraph and shall pay particular attention to clinical data from post-market surveillance and PMCF activities undertaken since the previous certification or re-certification, including appropriate updates to manufacturers' clinical evaluation reports.~~

~~For decision on re-certification the notified body in question shall use the same methods and principles as for the initial certification decision. If necessary, separate forms shall be established for re-certification taking into account the steps taken for certification such as application and application review.~~

~~The notified body shall have documented procedures in place relating to the extension of the period of validity of a certificate in cases where it has exceptionally limited the period of validity. Those procedures shall require the manufacturer to submit prior to the expiry of the certificate the data or documentation specified by the notified body to enable it to decide about the extension of the period of validity of the certificate.~~

ANNEX VIII CLASSIFICATION RULES

Chapter I

Definitions specific to classification rules

1. DURATION OF USE

- 1.1. 'Transient' means normally intended for continuous use for less than 60 minutes.
- 1.2. 'Short term' means normally intended for continuous use for between 60 minutes and 30 days.
- 1.3. 'Long term' means normally intended for continuous use for more than 30 days.

2. INVASIVE AND ACTIVE DEVICES

- 2.1. 'Body orifice' means any natural opening in the body, as well as the external surface of the eyeball, or any permanent artificial opening, such as a stoma.
- 2.2. 'Surgically invasive device' means:
 - (a) an invasive device which penetrates inside the body through the surface of the body, including through mucous membranes of body orifices with the aid or in the context of a surgical operation; and
 - (b) a device which produces penetration other than through a body orifice.
- 2.3. 'Reusable surgical instrument' means an instrument intended for surgical use in cutting, drilling, sawing, scratching, scraping, clamping, retracting, clipping or similar procedures, without a connection to an active device and which is intended by the manufacturer to be reused after appropriate procedures such as cleaning, disinfection and sterilisation have been carried out.
- 2.4. 'Active therapeutic device' means any active device used, whether alone or in combination with other devices, to support, modify, replace or restore biological functions or structures with a view to treatment or alleviation of an illness, injury or disability.
- 2.5. 'Active device intended for diagnosis and monitoring' means any active device used, whether alone or in combination with other devices, to supply information for detecting, diagnosing, monitoring or treating physiological conditions, states of health, illnesses or congenital deformities.
- 2.6. 'Central circulatory system' means the following blood vessels: *arteriae pulmonales, aorta ascendens, arcus aortae, aorta descendens to the bifurcatio aortae, arteriae coronariae, arteria carotis communis, arteria carotis externa, arteria carotis interna, arteriae cerebrales, truncus brachiocephalicus, venae cordis, venae pulmonales, vena cava superior and vena cava inferior.*
- 2.7. 'Central nervous system' means the brain, meninges and spinal cord.
- 2.8. 'Injured skin or mucus membrane' means an area of skin or a mucus membrane presenting a pathological change or change following disease or a wound.

Chapter II

Implementing rules

- 3.1. Application of the classification rules shall be governed by the intended purpose of the devices.
- 3.2. If the device in question is intended to be used in combination with another device, the classification rules shall apply separately to each of the devices. Accessories for a medical device **and accessories for a product listed in Annex XVI** shall be classified in their own right separately from the device with which they are used.
- 3.3. Software, which drives a device or influences the use of a device, shall fall within the same class as the device.

Proposal for Regulation (EU) 2017/745 – Annex VIII

If the software is independent of any other device, it shall be classified in its own right.

3.4. If the device is not intended to be used solely or principally in a specific part of the body, it shall be considered and classified on the basis of the most critical specified use.

3.5. If several rules, or if, within the same rule, several sub-rules, apply to the same device based on the device's intended purpose, the strictest rule and sub-rule resulting in the higher classification shall apply.

3.6. In calculating the duration referred to in Section 1, continuous use shall mean:

(a) the entire duration of use of the same device without regard to temporary interruption of use during a procedure or temporary removal for purposes such as cleaning or disinfection of the device. Whether the interruption of use or the removal is temporary shall be established in relation to the duration of the use prior to and after the period when the use is interrupted or the device removed; and

(b) the accumulated use of a device that is intended by the manufacturer to be replaced immediately with another of the same type.

3.7. A device is considered to allow direct diagnosis when it provides the diagnosis of the disease or condition in question by itself or when it provides decisive information for the diagnosis.

Chapter III Classification rules

4. NON-INVASIVE DEVICES

4.1. Rule 1

All non-invasive devices are classified as class I, unless one of the rules set out hereinafter applies.

4.2. Rule 2

All non-invasive devices intended for channelling or storing blood, body liquids, cells or tissues, liquids or gases for the purpose of eventual infusion, administration or introduction into the body are classified as class IIa:

- if they may be connected to a class IIa, class IIb or class III active device; or
- if they are intended for use for channelling or storing blood or other body liquids or for storing organs, parts of organs or body cells and tissues, except for blood bags; **by derogation from any other classification rule**, blood bags are classified as class IIb.

In all other cases such devices are classified as class I.

4.3. Rule 3

All non-invasive devices intended for modifying the biological or chemical composition of human tissues or cells, blood, other body liquids or other liquids intended for implantation or administration into the body are classified as class IIb, unless the treatment for which the device is used consists of filtration, centrifugation or exchanges of gas, heat, in which case they are classified as class IIa.

All non-invasive devices consisting of a substance or a mixture of substances intended to be used *in vitro* in direct contact with human cells, tissues or organs taken from the human body or used *in vitro* with human embryos before their implantation or administration into the body are classified as class III.

4.4. Rule 4

All non-invasive devices which come into contact with injured skin or mucous membrane are classified as:

- class I if they are intended to be used as a mechanical barrier, for compression or for absorption of exudates;

Proposal for Regulation (EU) 2017/745 – Annex VIII

- class IIb if they are intended to be used principally for injuries to skin which have breached the dermis or mucous membrane and can only heal by secondary intent;
- class IIa if they are principally intended to manage the micro-environment of injured skin or mucous membrane; and
- class IIa in all other cases.

This rule applies also to the invasive devices that come into contact with injured mucous membrane.

5. INVASIVE DEVICES

5.1. Rule 5

All invasive devices with respect to body orifices, other than surgically invasive devices, which are not intended for connection to an active device or which are intended for connection to a class I active device are classified as:

- class I if they are intended for transient use;
- class IIa if they are intended for short-term use, except if they are used in the oral cavity as far as the pharynx, in an ear canal up to the ear drum or in the nasal cavity, in which case they are classified as class I; and
- class IIb if they are intended for long-term use, except if they are used in the oral cavity as far as the pharynx, in an ear canal up to the ear drum or in the nasal cavity and are not liable to be absorbed by the mucous membrane, in which case they are classified as class IIa.

All invasive devices with respect to body orifices, other than surgically invasive devices, intended for connection to a class IIa, class IIb or class III active device, are classified as class IIa.

5.2. Rule 6

All surgically invasive devices intended for transient use are classified as class IIa unless they:

- are intended specifically to control, diagnose, monitor or correct a defect of the heart or of the central circulatory system through direct contact with those parts of the body, in which case they are classified as class III;
- are reusable surgical instruments **regardless of the body part with which they come into contact**, in which case they are classified as class I;
- are intended specifically for use in direct contact with the heart or central circulatory system or the central nervous system, in which case they are classified as class III;
- are intended to supply energy in the form of ionising radiation in which case they are classified as class IIb;
- have a biological effect or are wholly or mainly absorbed in which case they are classified as class IIb; or
- are intended to administer medicinal products by means of a delivery system, if such administration of a medicinal product is done in a manner that is potentially hazardous taking account of the mode of application, in which case they are classified as class IIb.

5.3. Rule 7

All surgically invasive devices intended for short-term use are classified as class IIa unless they:

- are intended specifically to control, diagnose, monitor or correct a defect of the heart or of the central circulatory system through direct contact with those parts of the body, in which case they are classified as class III;

Proposal for Regulation (EU) 2017/745 – Annex VIII

- are intended specifically for use in direct contact with the heart or central circulatory system or the central nervous system, in which case they are classified as class III;
- are intended to supply energy in the form of ionizing radiation in which case they are classified as class IIb;
- have a biological effect or are wholly or mainly absorbed in which case they are classified as class III;
- are intended to undergo chemical change in the body in which case they are classified as class IIb, except if the devices are placed in the teeth;
- are intended to administer medicines, in which case they are classified as class IIb; or
- are reusable surgical instruments regardless of the body part with which they come into contact, in which case they are classified as class I;

5.4. Rule 8

All implantable devices and long-term surgically invasive devices are classified as class IIb unless they:

- are intended to be placed in the teeth, in which case they are classified as class IIa;
- are intended to be used in direct contact with the heart, the central circulatory system or the central nervous system, in which case they are classified as class III;
- have a biological effect or are wholly or mainly absorbed, in which case they are classified as class III;
- are intended to undergo chemical change in the body in which case they are classified as class III, except if the devices are placed in the teeth;
- are intended to administer medicinal products, in which case they are classified as class III;
- are active implantable devices ~~or their accessories~~, in which cases they are classified as class III;
- are breast implants or surgical meshes, in which cases they are classified as class III;
- are total or partial joint replacements, in which case they are classified as class III, with the exception of ~~ancillary~~ components such as screws, wedges, plates and instruments ~~and other devices that are well-established technology devices~~; or
- are spinal disc replacement implants or are implantable devices that come into contact with the spinal column, in which case they are classified as class III with the exception of components such as screws, wedges, plates and instruments ~~and other devices that are well-established technology devices~~.

6. ACTIVE DEVICES

6.1. Rule 9

All active therapeutic devices ~~and all active products listed in Annex XVI~~ intended to administer or exchange energy are classified as class IIa unless their characteristics are such that they may administer energy to or exchange energy with the human body in a potentially hazardous way, taking account of the nature, the density and site of application of the energy, in which case they are classified as class IIb.

All active devices intended to control or monitor the performance of active therapeutic class IIb devices, ~~and of any active product listed in Annex XVI falling in class IIb~~, or intended directly to influence the performance of such devices are classified as class IIb.

All active devices intended to emit ionizing radiation for therapeutic purposes, including devices which control or monitor such devices, or which directly influence their performance, are classified as class IIb.

All active devices that are intended for controlling, monitoring or directly influencing the performance of active implantable devices are classified as class III.

6.2. Rule 10

Active devices intended for diagnosis and monitoring are classified as class IIa:

Proposal for Regulation (EU) 2017/745 – Annex VIII

– if they are intended to supply energy which will be absorbed by the human body, except for devices intended to illuminate the patient's body, in the visible spectrum, in which case they are classified as class I;

– if they are intended to image *in vivo* distribution of radiopharmaceuticals; or

– if they are intended to allow direct diagnosis or monitoring of vital physiological processes, unless they are specifically intended for monitoring of vital physiological parameters and the nature of variations of those parameters is such that it could result in immediate danger to the patient, for instance variations in cardiac performance, respiration, activity of the central nervous system, or they are intended for diagnosis in clinical situations where the patient is in immediate danger, in which cases they are classified as class IIb.

Active devices intended to emit ionizing radiation and intended for diagnostic or therapeutic radiology, including interventional radiology devices and devices which control or monitor such devices, or which directly influence their performance, are classified as class IIb.

6.3. Rule 11

Software **which is** intended to **generate an output that confers a clinical benefit and is used for diagnosis, treatment, prevention, monitoring, prediction, prognosis, compensation or alleviation of a disease or condition** is classified as class I, unless the output is intended for a disease or condition [provide information which is used to take decisions with diagnosis or therapeutic purposes is classified as class IIa, except if such decisions have an impact that may cause]:

– **in a critical situation with a risk of causing death or an irreversible deterioration of a person's state of health, in which case it is classified as class III; ~~or~~**

– **in a serious situation with a risk of causing a serious deterioration of a person's state of health or a surgical intervention, or to drive clinical management in a critical situation** in which cases it is classified as class IIb;

– **in a non-serious situation, or to drive clinical management in a serious situation or to inform clinical management in a critical or serious situation** in which cases it is classified as class IIa.

~~Software intended to monitor physiological processes is classified as class IIa, except if it is intended for monitoring of vital physiological parameters, where the nature of variations of those parameters is such that it could result in immediate danger to the patient, in which case it is classified as class IIb.~~

~~All other software is classified as class I.~~

6.4. Rule 12

All active devices intended to administer and/or remove medicinal products, body liquids or other substances to or from the body are classified as class IIa, unless this is done in a manner that is potentially hazardous, taking account of the nature of the substances involved, of the part of the body concerned and of the mode of application in which case they are classified as class IIb.

6.5. Rule 13

All other active devices are classified as class I.

7. SPECIAL RULES

7.1. Rule 14

All devices incorporating, as an integral part, a substance which, if used separately, can be considered to be a medicinal product, as defined in point 2 of Article 1 of Directive 2001/83/EC, including a medicinal product derived from human blood or human plasma, as defined in point 10 of Article 1 of that Directive, and that has an action ancillary to that of the devices, are classified as class III.

7.2. Rule 15

Proposal for Regulation (EU) 2017/745 – Annex VIII

All devices used for contraception or prevention of the transmission of sexually transmitted diseases are classified as class IIb, unless they are implantable or long term invasive devices, in which case they are classified as class III.

7.3. Rule 16

All devices intended specifically to be used for disinfecting, cleaning, rinsing or, where appropriate, hydrating contact lenses are classified as class IIb.

All devices intended specifically to be used for disinfecting or sterilising medical devices are classified as class IIa, unless they are disinfecting solutions or washer-disinfectors intended specifically to be used for disinfecting invasive devices, as the end point of processing, in which case they are classified as class IIb.

This rule does not apply to devices that are intended to clean devices other than contact lenses by means of physical action only.

7.4. Rule 17

Devices specifically intended for recording of diagnostic images generated by X-ray radiation are classified as class IIa.

7.5. Rule 18

All devices manufactured utilising tissues or cells of human or animal origin, or their derivatives, which are non-viable or rendered non-viable, are classified as class III, unless such devices are manufactured utilising tissues or cells of animal origin, or their derivatives, which are non-viable or rendered non-viable and are devices intended to come into contact with intact skin only.

7.6. Rule 19

All devices incorporating or consisting of nanomaterial **as defined in Commission Recommendation C/2022/3689** are classified as:

- class III if they present a high or medium potential for internal exposure;
- class IIb if they present a low potential for internal exposure; and
- class IIa if they present a negligible potential for internal exposure.

7.7. Rule 20

All invasive devices with respect to body orifices, other than surgically invasive devices, which are intended to administer medicinal products by inhalation are classified as class IIa, unless their mode of action has an essential impact on the efficacy and safety of the administered medicinal product or they are intended to treat life-threatening conditions, in which case they are classified as class IIb.

7.8. Rule 21

Devices that are composed of substances or of combinations of substances that are intended to be introduced into the human body via a body orifice or applied to the skin and that are absorbed by or locally dispersed **on or** in the human body are classified as:

- class III if they, or their products of metabolism, are systemically absorbed by the human body in order to achieve the intended purpose;
- class III if they achieve their intended purpose in the stomach or lower gastrointestinal tract and they, or their products of metabolism, are systemically absorbed by the human body;
- class IIa if they are applied to the skin or if they are applied in the nasal or oral cavity as far as the pharynx, and achieve their intended purpose on those cavities; and
- class IIb in all other cases.

7.9. Rule 22

Proposal for Regulation (EU) 2017/745 – Annex VIII

Active therapeutic devices with an integrated or incorporated diagnostic function which significantly determines the patient management by the device, such as closed loop systems or automated external defibrillators, are classified as class III.

ANNEX IX CONFORMITY ASSESSMENT BASED ON A QUALITY MANAGEMENT SYSTEM AND ON ASSESSMENT OF TECHNICAL DOCUMENTATION

Chapter I Quality Management System

1. The manufacturer shall establish, document and implement a quality management system as described in Article 10(9) and maintain its effectiveness throughout the life cycle of the devices concerned. The manufacturer shall ensure the application of the quality management system as specified in Section 2 and shall be subject to audit, as laid down in Sections 2.3 and 2.4, and to surveillance as specified in Section 3.

2. Quality management system assessment

2.1. The manufacturer shall lodge an application for assessment of its quality management system with a notified body. The application shall include:

- the name of the manufacturer and address of its registered place of business and any additional manufacturing site covered by the quality management system, and, if the manufacturer's application is lodged by its authorised representative, the name of the authorised representative and the address of the authorised representative's registered place of business,
- all relevant information on the device or group of devices covered by the quality management system,
- a written declaration that no application has been lodged with any other notified body for the same device-related quality management system, or information about any previous application for the same device-related quality management system,
- a draft of an EU declaration of conformity in accordance with Article 19 and Annex IV for the device model covered by the conformity assessment procedure,
- the documentation on the manufacturer's quality management system,
- a documented description of the procedures in place to fulfil the obligations arising from the quality management system and required under this Regulation and the undertaking by the manufacturer in question to apply those procedures,
- a description of the procedures in place to ensure that the quality management system remains adequate and effective, and the undertaking by the manufacturer to apply those procedures,
- the documentation on the manufacturer's post-market surveillance system and, where applicable, on the PMCF plan, and the procedures put in place to ensure compliance with the obligations resulting from the provisions on vigilance set out in Articles 87 to 92,
- a description of the procedures in place to keep up to date the post-market surveillance system, and, where applicable, the PMCF plan, and the procedures ensuring compliance with the obligations resulting from the provisions on vigilance set out in Articles 87 to 92, as well as the undertaking by the manufacturer to apply those procedures,
- documentation on the clinical evaluation plan, and
- a description of the procedures in place to keep up to date the clinical evaluation plan, taking into account the state of the art.

2.2. Implementation of the quality management system shall ensure compliance with this Regulation. All the elements, requirements and provisions adopted by the manufacturer for its quality management system shall be documented in a systematic and orderly manner in the form of a quality manual and written policies and procedures such as quality programmes, quality plans and quality records.

Proposal for Regulation (EU) 2017/745 – Annex IX

Moreover, the documentation to be submitted for the assessment of the quality management system shall include an adequate description of, in particular:

(a) the manufacturer's quality objectives;

(b) the organisation of the business and in particular:

- the organisational structures with the assignment of staff responsibilities in relation to critical procedures, the responsibilities of the managerial staff and their organisational authority,
- the methods of monitoring whether the operation of the quality management system is efficient and in particular the ability of that system to achieve the desired design and device quality, including control of devices which fail to conform,
- where the design, manufacture and/or final verification and testing of the devices, or parts of any of those processes, is carried out by another party, the methods of monitoring the efficient operation of the quality management system and in particular the type and extent of control applied to the other party, and
- where the manufacturer does not have a registered place of business in a Member State, the draft mandate for the designation of an authorised representative and a letter of intention from the authorised representative to accept the mandate;

(c) the procedures and techniques for monitoring, verifying, validating and controlling the design of the devices and the corresponding documentation as well as the data and records arising from those procedures and techniques. Those procedures and techniques shall specifically cover:

- the strategy for regulatory compliance, including processes for identification of relevant legal requirements, qualification, classification, handling of equivalence, choice of and compliance with conformity assessment procedures,
- identification of applicable general safety and performance requirements and solutions to fulfil those requirements, taking applicable CS and, where opted for, harmonised standards or other adequate solutions into account,
- risk management as referred to in Section 3 of Annex I,
- the clinical evaluation, pursuant to Article 61 and Annex XIV, including post-market clinical follow-up,
- solutions for fulfilling the applicable specific requirements regarding design and construction, including appropriate pre-clinical evaluation, in particular the requirements of Chapter II of Annex I,
- solutions for fulfilling the applicable specific requirements regarding the information to be supplied with the device, in particular the requirements of Chapter III of Annex I,
- the device identification procedures drawn up and kept up to date from drawings, specifications or other relevant documents at every stage of manufacture, and
- management of design or quality management system changes; and

(d) the verification and quality assurance techniques at the manufacturing stage and in particular the processes and procedures which are to be used, particularly as regards sterilisation and the relevant documents; and

(e) the appropriate tests and trials which are to be carried out before, during and after manufacture, the frequency with which they are to take place, and the test equipment to be used; it shall be possible to trace back adequately the calibration of that test equipment.

In addition, the manufacturer shall grant the notified body access to the technical documentation referred to in Annexes II and III.

2.3. Audit

The notified body shall audit the quality management system to determine whether it meets the requirements referred to in Section 2.2. Where the manufacturer uses a harmonised standard or CS related to a quality management system, the notified body shall assess conformity with those standards or CS. The notified body shall assume that a quality management system which satisfies the relevant harmonised standards or CS conforms to the requirements covered by those standards or CS, unless it duly substantiates not doing so.

The audit team of the notified body shall include at least one member with past experience of assessments of the technology concerned in accordance with Sections 4.3. to 4.5. of Annex VII. In circumstances where such experience is not immediately obvious or applicable, the notified body shall provide a documented rationale for the composition of that team. The assessment procedure shall include an audit on the manufacturer's premises and, if appropriate, on the premises of the manufacturer's suppliers and/or subcontractors to verify the manufacturing and other relevant processes.

Moreover, in the case of class IIa and class IIb devices, the quality management system assessment shall be accompanied by the assessment of the technical documentation, **as referred to in Annex II and III, as specified in Sections 4.3 to 4.8 for a representative device [for devices selected on a representative basis as specified in Section 4]. However, in case of class IIa devices, Section 3(a) of Annex II shall be excluded from the assessment.** In choosing the representative **devices [samples]**, the notified body shall **apply a risk-based approach, taking [take] into account the principle of proportionality [the published guidance developed by the MDCG pursuant to Article 105]** and in particular **the physical, chemical, biological characteristics of the device**, the novelty of the technology, similarities in design, technology, manufacturing and sterilisation methods, the intended purpose, **the application by the manufacturer of harmonised standards or CS for the device** and the results of any previous relevant assessments such as with regard to physical, chemical, biological or clinical properties, that have been carried out in accordance with this Regulation. The notified body in question shall document its rationale for the **representative devices [samples]** taken.

For class IIa and class IIb devices, the notified body may include a 'for-cause' assessment of the technical documentation of additional representative devices on duly justified grounds identified during the quality management system assessment.

If the quality management system **and the technical documentation of the assessed representative device** conforms to the relevant provisions of this Regulation, the notified body shall issue an EU quality management system certificate. The notified body shall notify the manufacturer of its decision to issue the certificate. The decision shall contain the conclusions of the audit and a reasoned report.

2.4. The manufacturer in question shall inform the notified body which approved the quality management system of any plan for substantial changes to the quality management system, or the device-range covered. The notified body shall assess the changes proposed, determine the need for additional audits and verify whether after those changes the quality management system still meets the requirements referred to in Section 2.2. It shall notify the manufacturer of its decision which shall contain the conclusions of the assessment, and where applicable, conclusions of additional audits. The approval of any substantial change to the quality management system or the device-range covered shall take the form of a supplement to the EU quality management system certificate.

3. Surveillance assessment

3.1. The aim of surveillance is to ensure that the manufacturer duly fulfils the obligations arising from the approved quality management system.

3.2. The manufacturer shall give authorisation to the notified body to carry out all the necessary audits, including on-site audits, and supply it with all relevant information, in particular:

- the documentation on its quality management system,

Proposal for Regulation (EU) 2017/745 – Annex IX

- documentation on any findings and conclusions resulting from the application of the post-market surveillance plan, including the PMCF plan, for a representative ~~sample of~~ devices, and of the provisions on vigilance set out in Articles 87 to 92,
- the data stipulated in the part of the quality management system relating to design, such as the results of analyses, calculations, tests and the solutions adopted regarding the risk-management as referred to in Section 4 of Annex I, and
- the data stipulated in the part of the quality management system relating to manufacture, such as quality control reports and test data, calibration data, and records on the qualifications of the personnel concerned.

3.3. Notified bodies shall periodically, ~~at least once every 12 months,~~ carry out appropriate audits and assessments to make sure that the manufacturer in question applies the approved quality management system and the post-market surveillance plan. Those audits and assessments shall include audits on the premises of the manufacturer and, if appropriate, of the manufacturer's suppliers and/or subcontractors. ~~On justified grounds, the audit may be conducted remotely instead of on-site. At the time of such on-site audits,~~ The notified body shall, where necessary, carry out or ask for tests in order to check that the quality management system is working properly. It shall provide the manufacturer with a surveillance audit report and, if a test has been carried out, with a test report.

The notified body shall carry out the surveillance audits and assessments once every 12 months. However, where justified in light of the results of previous surveillance audits and assessments, and in the absence of any concerns resulting from data from post-market surveillance or vigilance, the notified body shall carry out the surveillance audits and assessments only once every 24 months.

3.4. The notified body shall ~~randomly~~ perform ~~at least once every five years short-notice~~ or unannounced audits on the site of the manufacturer and, where appropriate, of the manufacturer's suppliers and/or subcontractors ~~when justified based on concerns related to post-market surveillance or vigilance data or at the request of a competent authority.~~ The short-notice or unannounced audit ~~which~~ may be combined with the periodic surveillance assessment referred to in Section 3.3. or be performed in addition to that surveillance assessment. ~~The notified body shall establish a plan for such unannounced on-site audits but shall not disclose it to the manufacturer.~~

Within the context of such unannounced on-site audits, the notified body ~~may~~ [shall] test an adequate sample of the devices produced or an adequate sample from the manufacturing process to verify that the manufactured device is in conformity with the technical documentation, with the exception of the devices referred to in the second subparagraph of Article 52(8). Prior to unannounced on-site audits, the notified body shall specify the relevant sampling criteria and testing procedure.

Instead of, or in addition to, the sampling referred to in the second paragraph, the notified body ~~may~~ [shall] take samples of devices from the market to verify that the manufactured device is in conformity with the technical documentation, with the exception of the devices referred to in the second subparagraph of Article 52(8). Prior to the sampling, the notified body in question shall specify the relevant sampling criteria and testing procedure.

The notified body shall provide the manufacturer in question with an on-site audit report which shall include, if applicable, the result of the sample test.

3.5. In the case of class IIa and class IIb devices, ~~and of class III devices that are well-established technology devices, during the surveillance assessment the notified body may [shall also] include an 'for-cause' assessment of the technical documentation of as specified in Section 4 for the device or devices concerned on the basis of further~~ representative ~~devices [samples] chosen in accordance with the rationale documented by the notified body in accordance with the third paragraph of Section 2.3 where the notified body has identified potential concerns on the basis of post-market surveillance data or other duly justified grounds.~~

In the case of class III devices, ~~with the exception of well-established technology devices,~~ the surveillance assessment shall also include a test of the approved parts and/or materials that are essential for the

Proposal for Regulation (EU) 2017/745 – Annex IX

integrity of the device, including, where appropriate, a check that the quantities of produced or purchased parts and/or materials correspond to the quantities of finished devices.

3.6. The notified body shall ensure that the composition of the assessment team is such that there is sufficient experience with the evaluation of the devices, systems and processes concerned, continuous objectivity and neutrality; this shall include a rotation of the members of the assessment team at appropriate intervals. As a general rule, a lead auditor shall neither lead nor attend audits for more than three consecutive years in respect of the same manufacturer.

3.7. If the notified body finds a divergence between the sample taken from the devices produced or from the market and the specifications laid down in the technical documentation or the approved design, it shall suspend or withdraw the relevant certificate or impose **conditions or limitations** [restrictions] on it.

Chapter II

Assessment of the technical documentation

4. Assessment of the technical documentation ~~applicable to class III devices and to the class IIb devices referred to in the second subparagraph of Article 52(4)~~

4.1. In addition to the obligations laid down in Section 2, the manufacturer shall lodge with the notified body an application for assessment of the technical documentation relating to the device which it plans to place on the market or put into service and which is covered by the quality management system referred to in Section 2.

4.2. The application shall describe the design, manufacture and performance of the device in question. It shall include the technical documentation as referred to in Annexes II and III **or a plan and related timelines for submission of such technical documentation**.

4.3. The notified body shall assess the technical documentation using staff with proven knowledge and experience regarding the technology concerned and its clinical application. The notified body may require the application to be completed by having further tests carried out or requesting further evidence to be provided to allow assessment of conformity with the relevant requirements of the Regulation. The notified body shall carry out adequate physical or laboratory tests in relation to the device or request the manufacturer to carry out such tests.

4.4. The notified body shall review the clinical evidence presented by the manufacturer in the clinical evaluation report and the related clinical evaluation that was conducted. The notified body shall **use** [employ] device reviewers with sufficient clinical expertise and, if necessary, use external clinical experts with direct and current experience relating to the device in question or the clinical condition in which it is utilised, for the purposes of that review.

4.5. The notified body shall, in circumstances in which the clinical evidence is based partly or totally on data from devices which are claimed to be equivalent to the device under assessment, assess the suitability of using such data, taking into account factors such as new indications and innovation. The notified body shall clearly document its conclusions on the claimed equivalence, and on the relevance and adequacy of the data for demonstrating conformity. For any characteristic of the device claimed as innovative by the manufacturer or for new indications, the notified body shall assess to what extent specific claims are supported by specific pre-clinical and clinical data and risk analysis.

4.6. The notified body shall verify that the clinical evidence and the clinical evaluation are adequate and shall verify the conclusions drawn by the manufacturer on the conformity with the relevant general safety and performance requirements. That verification shall include consideration of the adequacy of the benefit-risk determination, the risk management, the instructions for use, the user training and the manufacturer's post-market surveillance plan, and include a review of the need for, and the adequacy of, the PMCF plan proposed, where applicable.

4.7. Based on its assessment of the clinical evidence, the notified body shall consider the clinical evaluation and the benefit-risk determination and whether specific milestones need to be defined to allow

Proposal for Regulation (EU) 2017/745 – Annex IX

the notified body to review updates to the clinical evidence that result from post-market surveillance and PMCF data.

4.8. The notified body shall **provide the manufacturer with a report on the technical documentation assessment, including a [clearly document the outcome of its assessment in the]** clinical evaluation assessment report.

4.9. ~~The notified body shall provide the manufacturer with a report on the technical documentation assessment, including a clinical evaluation assessment report.~~ If the device conforms to the relevant provisions of this Regulation, the notified body shall issue an EU technical documentation assessment certificate. ~~The certificate shall contain the conclusions of the technical documentation assessment, the conditions of the certificate's validity, the data needed for identification of the approved design, and, where appropriate, a description of the intended purpose of the device.~~

4.10. Changes to the approved device shall require approval from the notified body which issued the EU technical documentation assessment certificate where such changes could affect the safety and performance of the device or the conditions prescribed for use of the device. Where the manufacturer plans to introduce any of the above-mentioned changes it shall inform the notified body which issued the EU technical documentation assessment certificate thereof. The notified body shall assess the planned changes and decide whether the planned changes require a new conformity assessment in accordance with Article 52 or whether they could be addressed by means of a supplement to the EU technical documentation assessment certificate. In the latter case, the notified body shall assess the changes, notify the manufacturer of its decision and, where the changes are approved, provide it with a supplement to the EU technical documentation assessment certificate.

5. Specific additional procedures

5.1. Assessment procedure for **devices covered by Article 54 [certain class III and class IIb devices]**

(a) **For devices covered by Article 54 [For class III implantable devices, and for class IIb active devices intended to administer and/or remove a medicinal product as referred to in Section 6.4. of Annex VIII (Rule 12)],** the notified body shall, having verified the quality of clinical data supporting the clinical evaluation report of the manufacturer referred to in Article 61(1) [61(12)], prepare a clinical evaluation assessment report which sets out its conclusions concerning the clinical evidence provided by the manufacturer, in particular concerning the benefit-risk determination, the consistency of that evidence with the intended purpose, including the medical indication or indications and the PMCF plan referred to in Article 10(3) and Part B of Annex XIV.

The notified body shall transmit its clinical evaluation assessment report, along with the manufacturer's clinical evaluation **report [documentation] and, where applicable, the PMCF plan referred to in points (c) and (d) of Section 6.1 of Annex II,** to the Commission.

The Commission shall immediately transmit those documents to the relevant expert panel referred to in Article 106.

(b) The notified body may be requested to present its conclusions as referred to in point (a) to the expert panel concerned. ~~Where appropriate, the expert panel may also invite the manufacturer to present the conclusions of its clinical evaluation.~~

(c) The expert panel shall decide, under the supervision of the Commission, on the basis of all of the following criteria:

(i) the novelty of the device or of the related clinical procedure involved, and the possible major clinical or health impact thereof;

(ii) a significantly adverse change in the benefit-risk profile of a specific category or group of devices due to scientifically valid health concerns in respect of components or source material or in respect of the impact on health in the case of failure of the device;

Proposal for Regulation (EU) 2017/745 – Annex IX

(iii) a significantly increased rate of serious incidents reported in accordance with Article 87 in respect of a specific category or group of devices,

whether to provide a scientific opinion on the clinical evaluation assessment report of the notified body based on the clinical evidence provided by the manufacturer, in particular concerning the benefit-risk determination, the consistency of that evidence with the medical indication or indications and the PMCF plan. That scientific opinion shall be provided within a period of 60 days, starting on the day of receipt of the documents from the Commission as referred to in point (a). The reasons for the decision to provide a scientific opinion on the basis of the criteria in points (i), (ii) and (iii) shall be included in the scientific opinion. Where the information submitted is not sufficient for the expert panel to reach a conclusion, this shall be stated in the scientific opinion.

(d) The expert panel may decide, under the supervision of the Commission, on the basis of the criteria laid down in point (c) not to provide a scientific opinion, in which case it shall inform the notified body as soon as possible and in any event within 21 days of receipt of the documents as referred to in point (a) from the Commission. The expert panel shall within that time limit provide the notified body and the Commission with the reasons for its decision, whereupon the notified body may proceed with the certification procedure of that device.

(e) The expert panel shall within 21 days of receipt of the documents from the Commission notify the Commission, through Eudamed whether it intends to provide a scientific opinion, pursuant to point (c), or whether it intends not to provide a scientific opinion, pursuant to point (d).

(f) Where no opinion has been delivered within a period of 60 days, the notified body may proceed with the certification procedure of the device in question.

(g) The notified body shall give **utmost [due]** consideration to the views expressed in the scientific opinion of the expert panel **and, where appropriate, update its clinical evaluation assessment report**. Where the expert panel finds that the level of clinical evidence is not sufficient or otherwise gives rise to serious concerns about the benefit-risk determination, the consistency of that evidence with the intended purpose, including the medical indication(s), and with the PMCF plan, the notified body shall, if necessary, advise the manufacturer **to gather additional clinical data and update its clinical evaluation, to limit [restrict]** the intended purpose of the device to certain groups of patients or certain medical indications and/or to impose a limit on the duration of validity of the certificate, to undertake specific PMCF studies, to adapt the instructions for use or the summary of safety and **clinical** performance, or to impose other **conditions [restrictions]** in its conformity assessment report, as appropriate.

The notified body shall provide **an explanation of how it has addressed the views and recommendations expressed in the expert panel opinion and [a full justification]**, where it has not followed the **views and recommendations [advice]** of the expert panel, **provide a substantiated justification for it**. The scientific opinion of the expert panel and the explanation of how it has been addressed or, if applicable, the substantiated justification provided by the notified body shall be publicly available via Eudamed without any confidential information as referred to in Article 109. **[in its conformity assessment report and the Commission shall without prejudice to Article 109 make both the scientific opinion of the expert panel and the written justification provided by the notified body publicly available via Eudamed.]**

(h) The Commission, after consultation with the Member States and relevant scientific experts shall provide guidance for expert panels for consistent interpretation of the criteria in point (c) before 26 May 2021.

5.2. Procedure in the case of devices incorporating a medicinal substance

(a) Where a device incorporates, as an integral part, a substance which, if used separately, may be considered to be a medicinal product within the meaning of point 2 of Article 1 of Directive 2001/83/EC, including a medicinal product derived from human blood or human plasma and that has an action ancillary to that of the device, the quality, safety and usefulness of the substance shall be verified by analogy with the methods specified in Annex I to Directive 2001/83/EC.

Proposal for Regulation (EU) 2017/745 – Annex IX

(b) Before issuing an EU technical documentation assessment certificate, the notified body shall, having verified the usefulness of the substance as part of the device and taking account of the intended purpose of the device, seek a scientific opinion from one of the competent authorities designated by the Member States in accordance with Directive 2001/83/EC or from the EMA, either of which to be referred to in this Section as 'the medicinal products authority consulted' depending on which has been consulted under this point, on the quality and safety of the substance including the benefit or risk of the incorporation of the substance into the device. Where the device incorporates a human blood or plasma derivative or a substance that, if used separately, may be considered to be a medicinal product falling exclusively within the scope of the annex to Regulation (EC) No 726/2004, the notified body shall seek the opinion of the EMA.

(c) When issuing its opinion, the medicinal products authority consulted shall take into account the manufacturing process and the data relating to the usefulness of incorporation of the substance into the device as determined by the notified body.

(d) The medicinal products authority consulted shall provide its opinion to the notified body within 90 [210] days of receipt of all the necessary documentation. This 90-day period may be extended once for a further 30 days on justified grounds. Where the medicinal substance is not previously authorised in the Union, the medicinal products authority consulted shall provide its opinion within 180 days. When preparing its opinion, the medicinal products authority consulted may request the notified body, or the manufacturer, to provide within a specific time period additional information necessary for its assessment referred to in point (b). In case of such a request, the time-limit set out in this paragraph shall be suspended until the additional information requested is provided.

(e) The scientific opinion of the medicinal products authority consulted, and any possible update of that opinion, shall be included in the documentation of the notified body concerning the device. The notified body shall give due consideration to the views expressed in the scientific opinion when making its decision. The notified body shall not deliver the certificate if the scientific opinion is unfavourable and shall convey its final decision to the medicinal products authority consulted.

(f) Before any change is made with respect to an ancillary substance incorporated in a device, in particular related to its manufacturing process, the manufacturer shall inform the notified body of the changes. That notified body shall seek the opinion of the medicinal products authority consulted, in order to confirm that the quality and safety of the ancillary substance remain unchanged. The medicinal products authority consulted shall take into account the data relating to the usefulness of incorporation of the substance into the device as determined by the notified body, in order to ensure that the changes have no negative impact on the risk or benefit previously established concerning the incorporation of the substance into the device. The medicinal products authority consulted shall provide its opinion within 60 days after receipt of all the necessary documentation regarding the changes. The notified body shall not deliver the supplement to the EU technical documentation assessment certificate if the scientific opinion provided by the medicinal products authority consulted is unfavourable. The notified body shall convey its final decision to the medicinal products authority consulted.

(g) Where the medicinal products authority consulted obtains information on the ancillary substance, which could have an impact on the risk or benefit previously established concerning the incorporation of the substance into the device, it shall advise the notified body as to whether this information has an impact on the risk or benefit previously established concerning the incorporation of the substance into the device. In the case of information relating to serious incidents obtained via the notification outlined in the first subparagraph of Article 89(6) of this Regulation, the medicinal products authority consulted shall review the data, which may be presented in aggregated form and may engage with the notified body to obtain further information where necessary. The notified body shall take that advice into account in reconsidering its assessment of the conformity assessment procedure.

5.3. Procedure in the case of devices manufactured utilising, or incorporating, tissues or cells of human or animal origin, or their derivatives, that are non-viable or rendered non-viable

5.3.1. Substances [Tissues or cells] of human origin or their derivatives

Proposal for Regulation (EU) 2017/745 – Annex IX

(a) For devices manufactured utilising derivatives of **substances [tissues or cells]** of human origin that are covered by this Regulation in accordance with point (g) of Article 1(6) and for devices that incorporate, as an integral part, **substances [tissues or cells]** of human origin, or their derivatives, covered by **Regulation (EU) 2024/1938 [Directive 2004/23/EC]**, that have an action ancillary to that of the device, the notified body shall, prior to issuing an EU technical documentation assessment certificate, seek a scientific opinion from one of the competent authorities designated by the Member States in accordance with **Regulation (EU) 2024/1938 [Directive 2004/23/EC]** ('**SoHO [human tissues and cells]** competent authority') on the aspects relating to **donor registration, donor review, collection and testing of the substances [the donation, procurement and testing of tissues or cells]** of human origin or their derivatives. The notified body shall submit a summary of the preliminary conformity assessment which provides, among other things, information about the non-viability of the **substances [tissues or cells]** of human origin in question, **donor registration, donor review, collection [their donation, procurement]** and testing and the risk or benefit of the incorporation of the **substances [tissues or cells]** of human origin or their derivatives into the device.

(b) Within **90 [120]** days of receipt of all the necessary documentation, the **SoHO [human tissues and cells]** competent authority shall provide to the notified body its opinion. **This 90-day period may be extended once for a further 30 days on justified grounds.**

(c) The scientific opinion of the **SoHO [human tissues and cells]** competent authority, and any possible update, shall be included in the documentation of the notified body concerning the device. The notified body shall give due consideration to the views expressed in the scientific opinion of the **SoHO [human tissues and cells]** competent authority when making its decision. The notified body shall not deliver the certificate if that scientific opinion is unfavourable. It shall convey its final decision to the **SoHO [human tissues and cells]** competent authority concerned.

(d) Before any change is made with respect to non-viable **substances [tissues or cells]** of human origin or their derivatives incorporated in a device, in particular relating to **donor registration, donor review, collection and testing [their donation, testing or procurement]**, the manufacturer shall inform the notified body of the intended changes. The notified body shall consult the authority that was involved in the initial consultation, in order to confirm that the quality and safety of the **substances [tissues or cells]** of human origin or their derivatives incorporated in the device are maintained. The **SoHO [human tissues and cells]** competent authority concerned shall take into account the data relating to the usefulness of incorporation of the **substances [tissues or cells]** of human origin or their derivatives into the device as determined by the notified body, in order to ensure that the changes have no negative impact on the established benefit-risk ratio of the addition of the **substances [tissues or cells]** of human origin or their derivatives in the device. It shall provide its opinion within 60 days of receipt of all the necessary documentation regarding the intended changes. The notified body shall not deliver a supplement to the EU technical documentation assessment certificate if the scientific opinion is unfavourable and shall convey its final decision to the **SoHO [human tissues and cells]** competent authority concerned.

(e) In the case of information relating to serious incidents obtained via the notification outlined in the second subparagraph of Article 89(6), the SoHO competent authority consulted shall review the data and may engage with the notified body to obtain further information where necessary.

5.3.2. Tissues or cells of animal origin or their derivatives

~~In the case of devices manufactured utilising animal tissue which is rendered non-viable or utilising non-viable products derived from animal tissue, as referred to in Regulation (EU) No 722/2012, the notified body shall apply the relevant requirements laid down in that Regulation.~~

5.4. Procedure in the case of devices that are composed of substances or of combinations of substances that are absorbed by or locally dispersed in the human body

(a) The quality and safety of devices that are composed of substances or of combinations of substances that are intended to be introduced into the human body via a body orifice or applied to the skin and that are absorbed by, or locally dispersed in, the human body, shall be verified where applicable and only in respect of the requirements not covered by this Regulation, in accordance with the relevant requirements

Proposal for Regulation (EU) 2017/745 – Annex IX

laid down in Annex I to Directive 2001/83/EC for the evaluation of absorption, distribution, metabolism, excretion, local tolerance, toxicity, interaction with other devices, medicinal products or other substances and potential for adverse reactions.

~~(b) In addition, for devices, or their products of metabolism, that are systemically absorbed by the human body in order to achieve their intended purpose, the notified body shall seek a scientific opinion from one of the competent authorities designated by the Member States in accordance with Directive 2001/83/EC or from the EMA, either of which to be referred to in this Section as 'the medicinal products authority consulted' depending on which has been consulted under this point, on the compliance of the device with the relevant requirements laid down in Annex I to Directive 2001/83/EC.~~

~~(c) The opinion of the medicinal products authority consulted shall be drawn up within 150 days of receipt of all the necessary documentation.~~

~~(d) The scientific opinion of the medicinal products authority consulted, and any possible update, shall be included in the documentation of the notified body concerning the device. The notified body shall give due consideration to the views expressed in the scientific opinion when making its decision and shall convey its final decision to the medicinal products authority consulted.~~

6. Batch verification in the case of devices incorporating, as an integral part, a medicinal substance which, if used separately, would be considered to be a medicinal product derived from human blood or human plasma as referred to in Article 1(8)

Upon completing the manufacture of each batch of devices that incorporate, as an integral part, a medicinal substance which, if used separately, would be considered to be a medicinal product derived from human blood or human plasma as referred to in the first subparagraph of Article 1(8), the manufacturer shall inform the notified body of the release of the batch of devices and send it the official certificate concerning the release of the batch of human blood or plasma derivative used in the device, issued by a Member State laboratory or a laboratory designated for that purpose by a Member State in accordance with Article 114(2) of Directive 2001/83/EC.

Chapter III Administrative provisions

7. The manufacturer or, where the manufacturer does not have a registered place of business in a Member State, its authorised representative shall, for a period ending no sooner than ten years, and in the case of implantable devices no sooner than 15 years, after the last device has been placed on the market, keep at the disposal of the competent authorities:

- the EU declaration of conformity,
- the documentation referred to in the fifth indent of Section 2.1 and in particular the data and records arising from the procedures referred to in point (c) of the second paragraph of Section 2.2,
- information on the changes referred to in Section 2.4,
- the documentation referred to in Section 4.2, and
- the decisions and reports from the notified body as referred to in this Annex.

8. Each Member State shall require that the documentation referred to in Section 7 is kept at the disposal of competent authorities for the period indicated in that Section in case a manufacturer, or its authorised representative, established within its territory goes bankrupt or ceases its business activity prior to the end of that period.

ANNEX X CONFORMITY ASSESSMENT BASED ON TYPE EXAMINATION

1. EU type-examination is the procedure whereby a notified body ascertains and certifies that a device, including its technical documentation and relevant life cycle processes and a corresponding representative sample of the device production envisaged, fulfil the relevant provisions of this Regulation.

2. Application

The manufacturer shall lodge an application for assessment with a notified body. The application shall include:

- the name of the manufacturer and address of the registered place of business of the manufacturer and, if the application is lodged by the authorised representative, the name of the authorised representative and the address of its registered place of business,
- the technical documentation referred to in Annexes II and III. The applicant shall make a representative sample of the device production envisaged ('type') available to the notified body. The notified body may request other samples as necessary, and
- a written declaration that no application has been lodged with any other notified body for the same type, or information about any previous application for the same type that was refused by another notified body or was withdrawn by the manufacturer or its authorised representative before that other notified body made its final assessment.

3. Assessment

The notified body shall:

- (a) examine the application by using staff with proven knowledge and experience regarding the technology concerned and its clinical application. The notified body may require the application to be completed by having further tests carried out or requesting further evidence to be provided to allow assessment of conformity with the relevant requirements of this Regulation. The notified body shall carry out adequate physical or laboratory tests in relation to the device or request the manufacturer to carry out such tests;
- (b) examine and assess the technical documentation for conformity with the requirements of this Regulation that are applicable to the device and verify that the type has been manufactured in conformity with that documentation; it shall also record the items designed in conformity with the applicable standards referred to in Article 8 or with applicable CS, and record the items not designed on the basis of the relevant standards referred to in Article 8 or of the relevant CS;
- (c) review the clinical evidence presented by the manufacturer in the clinical evaluation report in accordance with Section 4 of Annex XIV. The notified body shall use [employ] device reviewers with sufficient clinical expertise and, if necessary, use external clinical experts with direct and current experience relating to the device in question or to the clinical condition in which it is utilised, for the purposes of that review;
- (d) in circumstances in which the clinical evidence is based partly or totally on data from devices which are claimed to be similar or equivalent to the device under assessment, assess the suitability of using such data, taking into account factors such as new indications and innovation. The notified body shall clearly document its conclusions on the claimed equivalence, and on the relevance and adequacy of the data for demonstrating conformity;
- (e) clearly document the outcome of its assessment in a pre-clinical and clinical evaluation assessment report as part of the EU type examination report referred to in point (i);
- (f) carry out or arrange for the appropriate assessments and the physical or laboratory tests necessary to verify whether the solutions adopted by the manufacturer meet the general safety and performance requirements laid down in this Regulation in the event that the standards referred to in Article 8 or the CS

Proposal for Regulation (EU) 2017/745 – Annex X

have not been applied. Where the device has to be connected to another device or devices in order to operate as intended, proof shall be provided that it conforms to the general safety and performance requirements when connected to any such device or devices having the characteristics specified by the manufacturer;

(g) carry out or arrange for the appropriate assessments and the physical or laboratory tests necessary to verify whether, in the event that the manufacturer has chosen to apply the relevant harmonised standards, those standards have actually been applied;

(h) agree with the applicant on the place where the necessary assessments and tests are to be carried out; and

(i) draw up an EU type-examination report on the results of the assessments and tests carried out under points (a) to (g).

4. Certificate

If the type conforms to this Regulation, the notified body shall issue an EU type-examination certificate.

~~The certificate shall contain the name and address of the manufacturer, the conclusions of the type examination assessment, the conditions of the certificate's validity and the data needed for identification of the type approved. The certificate shall be drawn up in accordance with Annex XII.~~ The relevant parts of the documentation shall be annexed to the certificate and a copy kept by the notified body.

5. Changes to the type

5.1. The applicant shall inform the notified body which issued the EU type-examination certificate of any planned change to the approved type or of its intended purpose and conditions of use.

5.2. Changes to the approved device including limitations of its intended purpose and conditions of use shall require approval from the notified body which issued the EU type-examination certificate where such changes may affect conformity with the general safety and performance requirements or with the conditions prescribed for use of the product. The notified body shall examine the planned changes, notify the manufacturer of its decision and provide him with a supplement to the EU type-examination report. The approval of any change to the approved type shall take the form of a supplement to the EU type-examination certificate.

5.3. Changes to the intended purpose and conditions of use of the approved device, with the exception of limitations of the intended purpose and conditions of use, shall necessitate a new application for a conformity assessment.

6. Specific additional procedures

Section 5 of Annex IX shall apply with the proviso that any reference to an EU technical documentation assessment certificate shall be understood as a reference to an EU type-examination certificate.

7. Administrative provisions

The manufacturer or, where the manufacturer does not have a registered place of business in a Member State, its authorised representative shall, for a period ending no sooner than ten years, and in the case of implantable devices no sooner than 15 years, after the last device has been placed on the market, keep at the disposal of the competent authorities:

- the documentation referred to in the second indent of Section 2,
- information on the changes referred to in Section 5, and
- copies of EU type-examination certificates, scientific opinions and reports and their additions/supplements.

Section 8 of Annex IX shall apply.

ANNEX XI

CONFORMITY ASSESSMENT BASED ON PRODUCT CONFORMITY VERIFICATION

1. The objective of the conformity assessment based on product conformity verification is to ensure that devices conform to the type for which an EU type-examination certificate has been issued, and that they meet the provisions of this Regulation which apply to them.
2. Where an EU type-examination certificate has been issued in accordance with Annex X, the manufacturer may either apply the procedure set out in Part A (production quality assurance) or the procedure set out in Part B (product verification) of this Annex.
3. By way of derogation from Sections 1 and 2 ~~above~~, **Section 10 or Section 18 the procedures of [in]** this Annex coupled with the drawing up of technical documentation as set out in Annexes II and III may also be applied by manufacturers of class IIa devices.
- (a) By way of derogation from Sections 1 and 2 above, Section 10a may also be applied by manufacturers of class I devices that are placed on the market in sterile condition, have a measuring function or are reusable surgical instruments.**

PART A

PRODUCTION QUALITY ASSURANCE

4. The manufacturer shall ensure that the quality management system approved for the manufacture of the devices concerned is implemented, shall carry out a final verification, as specified in Section 6, and shall be subject to the surveillance referred to in Section 7.
5. When the manufacturer fulfils the obligations laid down in Section 4, it shall draw up and keep an EU declaration of conformity in accordance with Article 19 and Annex IV for the device covered by the conformity assessment procedure. By issuing an EU declaration of conformity, the manufacturer shall be deemed to ensure and to declare that the device concerned conforms to the type described in the EU type-examination certificate and meets the requirements of this Regulation which apply to the device.
6. Quality management system
 - 6.1. The manufacturer shall lodge an application for assessment of its quality management system with a notified body. The application shall include:
 - all elements listed in Section 2.1 of Annex IX,
 - the technical documentation referred to in Annexes II and III for the types approved, and
 - a copy of the EU-type examination certificates referred to in Section 4 of Annex X; if the EU-type examination certificates have been issued by the same notified body with which the application is lodged, a reference to the technical documentation and its updates and the certificates issued shall also be included in the application.
 - 6.2. Implementation of the quality management system shall be such as to ensure that there is compliance with the type described in the EU type-examination certificate and with the provisions of this Regulation which apply to the devices at each stage. All the elements, requirements and provisions adopted by the manufacturer for its quality management system shall be documented in a systematic and orderly manner in the form of a quality manual and written policies and procedures, such as quality programmes, quality plans and quality records.

That documentation shall, in particular, include an adequate description of all elements listed in points (a), (b), (d) and (e) of Section 2.2 of Annex IX.

- 6.3. The first and second paragraph of Section 2.3 of Annex IX shall apply.

If the quality management system is such that it ensures that the devices conform to the type described in the EU type-examination certificate and that it conforms to the relevant provisions of this Regulation, the

Proposal for Regulation (EU) 2017/745 – Annex XI

notified body shall issue an EU quality assurance certificate. The notified body shall notify the manufacturer of its decision to issue the certificate. That decision shall contain the conclusions of the notified body's audit and a reasoned assessment.

6.4. Section 2.4 of Annex IX shall apply.

7. Surveillance

Section 3.1, the first, second and fourth indents of Section 3.2, Sections 3.3, 3.4, 3.6 and 3.7 of Annex IX shall apply.

In the case of class III devices, **other than well-established technology devices**, surveillance shall also include a check that the quantities of produced or purchased raw material or crucial components approved for the type and correspond to the quantities of finished devices.

8. Batch verification in the case of devices incorporating, as an integral part, a medicinal substance which, if used separately, would be considered to be a medicinal product derived from human blood or human plasma referred to in Article 1(8).

Upon completing the manufacture of each batch of devices that incorporate, as an integral part, a medicinal substance which, if used separately, would be considered to be a medicinal product derived from human blood or human plasma referred to in the first subparagraph of Article 1(8), the manufacturer shall inform the notified body of the release of the batch of devices and send it the official certificate concerning the release of the batch of human blood or plasma derivative used in the device, issued by a Member State laboratory or a laboratory designated for that purpose by a Member State in accordance with Article 114(2) of Directive 2001/83/EC.

9. Administrative provisions

The manufacturer or, where the manufacturer does not have a registered place of business in a Member State, its authorised representative shall, for a period ending no sooner than ten years, and in the case of implantable devices no sooner than 15 years, after the last device has been placed on the market, keep at the disposal of the competent authorities:

- the EU declaration of conformity,
- the documentation referred to in the fifth indent of Section 2.1 of Annex IX,
- the documentation referred to in the eighth indent of Section 2.1 of Annex IX, including the EU type-examination certificate referred to in Annex X,
- information on the changes referred to in Section 2.4 of Annex IX, and
- the decisions and reports from the notified body as referred to in Sections 2.3, 3.3 and 3.4 of Annex IX.

Section 8 of Annex IX shall apply.

10. Application to class IIa devices

10.1. The manufacturer shall ensure that the quality management system approved for the manufacture of the devices concerned is implemented.

10.2 [10.1] ~~By way of derogation from Section 5~~ By virtue of the EU declaration of conformity the manufacturer shall be deemed to ensure and to declare that the class IIa devices in question are manufactured in conformity with the technical documentation referred to in Annexes II and III and meet the requirements of this Regulation which apply to them.

~~10.2. For class IIa devices the notified body shall assess, as part of the assessment referred to in Section 6.3, whether the technical documentation as referred to in Annexes II and III for the devices selected on a representative basis is compliant with this Regulation.~~

Proposal for Regulation (EU) 2017/745 – Annex XI

~~In choosing a representative sample or samples of devices, the notified body shall take into account the novelty of the technology, similarities in design, technology, manufacturing and sterilisation methods, the intended use and the results of any previous relevant assessments (e.g. with regard to physical, chemical, biological or clinical properties) that have been carried out in accordance with this Regulation. The notified body shall document its rationale for the sample or samples of devices taken.~~

10.3 The manufacturer shall lodge an application for assessment of its quality management system with a notified body. The application shall include all elements listed in Section 2.1 of Annex IX. The manufacturer shall ensure that an adequate description of all elements listed in Section 2.2, points (a), (b), (d) and (e) of Annex IX is available for the assessment of the quality management system.

10.4 The first four paragraphs of Section 2.3 of Annex IX shall apply.

10.5 [10.3] Where the assessment under Section 10.4. [10.2.] of this Annex confirms that the ~~class IIa~~ devices in question conform to the technical documentation referred to in Annexes II and III and meet the requirements of this Regulation which apply to them, the notified body shall issue an EU production quality assurance certificate pursuant to Part A [this Part] of this Annex.

~~10.4. Samples additional to those taken for the initial conformity assessment of devices shall be assessed by the notified body as part of the surveillance assessment referred to in Section 7.~~

10.6 Section 2.4 of Annex IX shall apply.

10.7 Section 3.1, Section 3.2, first, second and fourth indents, and Sections 3.3 to 3.7 of Annex IX shall apply.

10.8 [10.5]

10.5. ~~By way of derogation from Section 6,~~ The manufacturer or its authorised representative shall, for a period ending no sooner than 10 years after the last device has been placed on the market, keep at the disposal of the competent authorities:

- the EU declaration of conformity,
- the technical documentation referred to in Annexes II and III, and
- the certificate referred to in Section 10.5 [10.3] of this Annex.

10.9 Section 8 of Annex IX shall apply.

10a Application to class I devices that are placed on the market in sterile condition, have a measuring function or are reusable surgical instruments.

10a.1 The manufacturer shall ensure that the quality management system approved for the manufacture of the devices concerned is implemented.

10a.2 By virtue of the EU declaration of conformity the manufacturer shall be deemed to ensure and to declare that the devices in question are manufactured in conformity with the technical documentation referred to in Annexes II and III and meet the requirements of this Regulation which apply to them.

10a.3 The manufacturer shall lodge an application for assessment of its quality management system with a notified body. The application shall include the elements listed in Section 2.1, first to sixth indents, of Annex IX. The manufacturer shall ensure that an adequate description of all elements listed in Section 2.2, points (d) and (e), of Annex IX is available for the assessment of the quality management system.

10a.4 Section 2.3, first and second paragraphs, of Annex IX shall apply. The assessment of the notified body shall be limited:

- in the case of devices placed on the market in sterile condition, to the aspects relating to establishing, securing and maintaining sterile conditions;
- in the case of devices with a measuring function, to the aspects relating to the conformity of the devices with the metrological requirements;

Proposal for Regulation (EU) 2017/745 – Annex XI

in the case of reusable surgical instruments, to the aspects relating to the reuse of the device, in particular cleaning, disinfection, sterilization, maintenance and functional testing and the related instructions for use.

10a.5 Where the assessment under Section 10a.4. confirms that elements of the quality management system comply with the relevant provisions of this Regulation, the notified body shall issue an EU production quality assurance certificate pursuant to this Part of this Annex.

10a.6 Section 2.4 of Annex IX shall apply.

10a.7 Section 3.1, Section 3.2, first, second and fourth indents, Sections 3.3, 3.4 and 3.6 of Annex IX shall apply.

10a.8 The manufacturer or its authorised representative shall, for a period ending no sooner than 10 years after the last device has been placed on the market, keep at the disposal of the competent authorities:

- the EU declaration of conformity,
 - the documentation referred to in the fifth indent of Section 2.1 of Annex IX,
- the certificate referred to in Section 10a.5.

10a.9 Section 8 of Annex IX shall apply.

PART B PRODUCT VERIFICATION

11. Product verification shall be understood to be the procedure whereby after examination of every manufactured device, the manufacturer, by issuing an EU declaration of conformity in accordance with Article 19 and Annex IV, shall be deemed to ensure and to declare that the devices which have been subject to the procedure set out in Sections 14 and 15 conform to the type described in the EU type-examination certificate and meet the requirements of this Regulation which apply to them.

12. The manufacturer shall take all the measures necessary to ensure that the manufacturing process produces devices which conform to the type described in the EU type-examination certificate and to the requirements of the Regulation which apply to them. Prior to the start of manufacture, the manufacturer shall prepare documents defining the manufacturing process, in particular as regards sterilisation where necessary, together with all routine, pre-established procedures to be implemented to ensure homogeneous production and, where appropriate, conformity of the devices with the type described in the EU type-examination certificate and with the requirements of this Regulation which apply to them.

In addition, for devices placed on the market in a sterile condition, and only for those aspects of the manufacturing process designed to secure and maintain sterility, the manufacturer shall apply the provisions of Sections 6.1, 6.2, 6.3, first paragraph, 6.4 and 7 [6 and 7].

13. The manufacturer shall undertake to institute and keep up to date a post-market surveillance plan, including a PMCF plan, and the procedures ensuring compliance with the obligations of the manufacturer resulting from the provisions on vigilance and post-market surveillance system set out in Chapter VII.

14. The notified body shall carry out the appropriate examinations and tests in order to verify the conformity of the device with the requirements of the Regulation by examining and testing every product as specified in Section 15.

The examinations and tests referred to in the first paragraph of this Section shall not apply to aspects of the manufacturing process designed to secure sterility.

15. Verification by examination and testing of every product

15.1. Every device shall be examined individually and the appropriate physical or laboratory tests as defined in the relevant standard or standards referred to in Article 8, or equivalent tests and assessments, shall be carried out in order to verify, where appropriate, the conformity of the devices with the type described in the EU type-examination certificate and with the requirements of this Regulation which apply to them.

Proposal for Regulation (EU) 2017/745 – Annex XI

15.2. The notified body shall affix, or have affixed, its identification number to each approved device and shall draw up an EU product verification certificate relating to the tests and assessments carried out.

16. Batch verification in the case of devices incorporating, as an integral part, a medicinal substance which, if used separately, would be considered to be a medicinal product derived from human blood or human plasma referred to in Article 1(8).

Upon completing the manufacture of each batch of devices that incorporate, as an integral part, a medicinal substance which, if used separately, would be considered to be a medicinal product derived from human blood or human plasma referred to in the first subparagraph of Article 1(8), the manufacturer shall inform the notified body of the release of the batch of devices and send it the official certificate concerning the release of the batch of human blood or plasma derivative used in the device, issued by a Member State laboratory or a laboratory designated for that purpose by a Member State in accordance with Article 114(2) of Directive 2001/83/EC.

17. Administrative provisions

The manufacturer or its authorised representative shall, for a period ending no sooner than ten years, and in the case of implantable devices no sooner than 15 years, after the last device has been placed on the market, keep at the disposal of the competent authorities:

- the EU declaration of conformity,
- the documentation referred to in Section 12,
- the certificate referred to in Section 15.2, and
- the EU type-examination certificate referred to in Annex X.

Section 8 of Annex IX shall apply.

18. Application to class IIa devices

~~18.1. By way of derogation from Section 11, by virtue of the EU declaration of conformity the manufacturer shall be deemed to ensure and to declare that the class IIa devices in question are manufactured in conformity with the technical documentation referred to in Annexes II and III and meet the requirements of this Regulation which apply to them.~~

Product verification shall be understood to be the procedure whereby after examination of every manufactured device, the manufacturer, by issuing an EU declaration of conformity in accordance with Article 19 and Annex IV, shall be deemed to ensure and to declare that the devices which have been subject to the procedure set out in Sections 14 and 15 conform to the technical documentation referred to in Annexes II and III and meet the requirements of this Regulation which apply to them.

~~18.2. The verification conducted by the notified body in accordance with Section 14 is intended to confirm the conformity of the class IIa devices in question with the technical documentation referred to in Annexes II and III and with the requirements of this Regulation which apply to them.~~

The manufacturer shall take all the measures necessary to ensure that the manufacturing process produces devices which conform to the devices described in the EU declaration of conformity and to the requirements of the Regulation which apply to them. Prior to the start of manufacture, the manufacturer shall prepare documents defining the manufacturing process, in particular as regards sterilisation where necessary, together with all routine, pre-established procedures to be implemented to ensure homogeneous production.

In addition, for devices placed on the market in a sterile condition, and only for those aspects of the manufacturing process designed to secure and maintain sterility, the manufacturer shall apply the provisions of Sections 10.3, 10.4, 10.6 and 10.7.

~~18.3. If the verification referred to in Section 18.2 confirms that the class IIa devices in question conform to the technical documentation referred to in Annexes II and III and meet the requirements of this Regulation which apply to them, the notified body shall issue a certificate pursuant to this Part of this Annex.~~

Proposal for Regulation (EU) 2017/745 – Annex XI

Sections 13 and 14 shall apply. Section 15 shall apply with the aim of verifying the conformity of the devices with those described in the EU declaration of conformity.

18.4. By way of derogation from Section 17, the manufacturer or its authorised representative shall, for a period ending no sooner than ten years after the last device has been placed on the market, keep at the disposal of the competent authorities:

- the EU declaration of conformity,
- the documentation referred to in Section 18.2,
- the technical documentation referred to in Annexes II and III, ~~and~~
- the certificate referred to in Section 15.2 [18.3].

18.5 Section 8 of Annex IX shall apply.

ANNEX XII CERTIFICATES ISSUED BY A NOTIFIED BODY

Chapter I General Requirements

1. Certificates shall be drawn up in one of the official languages of the Union.
2. Each certificate shall refer to only one conformity assessment procedure.
3. Certificates shall only be issued to one manufacturer. The name and address of the manufacturer included in the certificate shall be the same as that registered in the electronic system referred to in Article 30.
4. The scope of the certificates shall unambiguously identify the device or devices covered:
 - (a) EU technical documentation assessment certificates, EU type-examination certificates and EU product verification certificates shall include a clear identification, including the name, model and type, of the device or devices, the intended purpose, as included by the manufacturer in the instructions for use and in relation to which the device has been assessed in the conformity assessment procedure, risk classification and the Basic UDI-DI as referred to in Article 27(6);
 - (b) EU quality management system certificates and EU quality assurance certificates shall include the identification of the devices or groups of devices, the risk classification and, ~~for class IIb devices, the intended purpose the manufacturing site(s) covered.~~
5. The notified body shall be able to demonstrate on request, which (individual) devices are covered by the certificate. The notified body shall set up a system that enables the determination of the devices, including their classification, covered by the certificate.
6. Certificates shall contain, if applicable, a note that, for the placing on the market of the device or devices it covers, another certificate issued in accordance with this Regulation is required.
7. EU quality management system certificates and EU quality assurance certificates for class I devices for which the involvement of a notified body is required pursuant to Article 52(7) shall include a statement that the audit by the notified body of the quality management system was limited to the aspects required under that paragraph.
8. Where a certificate is supplemented, modified or re-issued, the new certificate shall contain a reference to the preceding certificate and its date of issue with identification of the changes.

Chapter II Minimum content of the certificates

1. name, address and identification number of the notified body;
2. name and address of the manufacturer and, if applicable, of the authorised representative;
3. unique number identifying the certificate;
4. if already issued, the SRN of the manufacturer referred to in to Article 31(2);
5. date of issue;
6. ~~where applicable~~, date of expiry;
7. data needed for the unambiguous identification of the device or devices where applicable as specified in Section 4 of [Chapter \[Part\] I](#);
8. if applicable, reference to any previous certificate as specified in Section 8 of Chapter I;

Proposal for Regulation (EU) 2017/745 – Annex XII

9. reference to this Regulation and the relevant Annex in accordance with which the conformity assessment has been carried out;
10. ~~examinations and tests performed, e.g.~~ reference to relevant CS and harmonised standards, ~~test reports and audit report(s);~~
11. if applicable, reference to the relevant parts of the technical documentation or other certificates required for the placing on the market of the device or devices covered;
12. if applicable, information about the surveillance by the notified body;
13. conclusions of the notified body's conformity assessment with regard to the relevant Annex;
14. conditions for or limitations to the validity of the certificate;
15. legally binding signature of the notified body in accordance with the applicable national law.

ANNEX XIII PROCEDURE FOR CUSTOM-MADE DEVICES

1. For custom-made devices, the manufacturer or its authorised representative shall draw up a statement containing all of the following information:

- the name and address of the manufacturer, and of all manufacturing sites,
- if applicable, the name and address of the authorised representative,
- data allowing identification of the device in question,
- a statement that the device is intended for exclusive use by a particular patient or user, identified by name, an acronym or a numerical code,
- the name of the person who made out the prescription and who is authorised by national law by virtue of their professional qualifications to do so, and, where applicable, the name of the health institution concerned,
- the specific characteristics of the product as indicated by the prescription,
- a statement that the device in question conforms to the general safety and performance requirements set out in Annex I and, where applicable, indicating which general safety and performance requirements have not been fully met, together with the grounds,
- where applicable, an indication that the device contains or incorporates a medicinal substance, including a human blood or plasma derivative, or tissues or cells of human origin, or of animal origin as referred to in Regulation (EU) No 722/2012, **or in any subsequent implementing rules adopted pursuant to this Regulation.**

2. The manufacturer shall undertake to keep available for the competent national authorities documentation that indicates its manufacturing site or sites and allows an understanding to be formed of the design, manufacture and performance of the device, including the expected performance, so as to allow assessment of conformity with the requirements of this Regulation.

3. The manufacturer shall take all the measures necessary to ensure that the manufacturing process produces devices which are manufactured in accordance with the documentation referred to in Section 2.

4. The statement referred to in the introductory part of Section 1 shall be kept for a period of at least ten years after the device has been placed on the market. In the case of implantable devices, the period shall be at least 15 years.

Section 8 of Annex IX shall apply.

5. The manufacturer shall review and document experience gained in the post-production phase, including **where applicable** from PMCF as referred to in Part B of Annex XIV, and implement appropriate means to apply any necessary corrective action. In that context, it shall report in accordance with Article 87(1) to the competent authorities any serious incidents or field safety corrective actions or both as soon as it learns of them.

ANNEX XIV CLINICAL EVALUATION AND POST-MARKET CLINICAL FOLLOW-UP

PART A CLINICAL EVALUATION

1. To plan, continuously conduct and document a clinical evaluation, manufacturers shall:

(a) establish and update a clinical evaluation plan, which shall include at least:

- an identification of the general safety and performance requirements that require support from relevant clinical data, ~~or adequate justification that confirmation of safety and performance based on clinical data is not deemed appropriate in accordance with Article 61(10);~~
- a specification of the intended purpose of the device;
- a clear specification of intended target groups with clear indications and contra-indications;
- a detailed description of intended clinical benefits to patients with relevant and specified clinical outcome parameters;
- a specification of methods to be used for examination of qualitative and quantitative aspects of clinical safety with clear reference to the determination of residual risks and side-effects;
- an indicative list and specification of parameters to be used to determine, based on the state of the art in medicine, the acceptability of the benefit-risk ratio for the various indications and for the intended purpose or purposes of the device;
- an indication how benefit-risk issues relating to specific components such as use of pharmaceutical, non-viable ~~substances of human origin or non-viable animal~~ [animal or human] tissues, are to be addressed; ~~and~~
- ~~where applicable~~, a clinical development strategy [plan] indicating progression from exploratory investigations, such as first-in-man studies, feasibility and pilot studies, to confirmatory investigations, such as pivotal clinical investigations, and a PMCF as referred to in Part B ~~of this Annex~~ with an indication of milestones and a description of potential acceptance criteria;

~~Points (b) to (e) shall not apply to devices for which confirmation of safety and performance based on clinical data is not deemed appropriate in accordance with Article 61(10).~~

(b) identify available clinical data relevant to the device and its intended purpose and any gaps in clinical evidence through a systematic scientific literature review;

(c) appraise all relevant clinical data by evaluating their suitability for establishing the safety and performance of the device;

(d) generate, through properly designed clinical investigations in accordance with the clinical development plan, any new or additional clinical data necessary to address outstanding issues; and

(e) analyse all relevant clinical data in order to reach conclusions about the safety and clinical performance of the device including its clinical benefits.

2. The clinical evaluation shall be thorough and objective, and take into account both favourable and unfavourable data. Its depth and extent shall be proportionate and appropriate to the nature, classification, intended purpose and risks of the device in question, as well as to the manufacturer's claims in respect of the device.

3. A clinical evaluation may be based on clinical data relating to a device for which equivalence to the device in question can be demonstrated. The following technical, biological and clinical characteristics shall be taken into consideration for the demonstration of equivalence:

Proposal for Regulation (EU) 2017/745 – Annex XIV

- Technical: the device is of similar design; is used under similar conditions of use; has similar specifications and properties including physicochemical properties such as intensity of energy, tensile strength, viscosity, surface characteristics, wavelength and software algorithms; uses similar deployment methods, where relevant; has similar principles of operation and critical performance requirements;
- Biological: the device uses the same **or similar** materials or substances in contact with the same human tissues or body fluids for a similar kind and duration of contact and similar release characteristics of substances, including degradation products and leachables;
- Clinical: the device is used for the same **or similar** clinical condition or purpose, including similar severity and stage of disease, at the same site in the body, in a similar population, including as regards age, anatomy and physiology; has the same kind of user; has similar relevant critical performance in view of the expected clinical effect for a specific intended purpose.

The characteristics listed in the first paragraph shall be similar to the extent that there would be no clinically significant difference in the safety and clinical performance of the device. The demonstration of equivalence. It shall be clearly demonstrated that manufacturers have sufficient levels of access to the data relating to devices with which they are claiming equivalence in order to justify their claims of equivalence.

4. The results of the clinical evaluation and the clinical evidence on which it is based shall be documented in a clinical evaluation report which shall support the assessment of the conformity of the device.

The clinical evidence together with non-clinical data generated from non-clinical testing methods and other relevant documentation shall allow the manufacturer to demonstrate conformity with the general safety and performance requirements and shall be part of the technical documentation for the device in question.

Both favourable and unfavourable data considered in the clinical evaluation shall be included in the technical documentation.

PART B POST-MARKET CLINICAL FOLLOW-UP

5. PMCF shall be understood to be a continuous process that updates the clinical evaluation referred to in Article 61 and Part A of this Annex and shall be addressed in the manufacturer's post-market surveillance plan. **Where PMCF is deemed not applicable, the manufacturer shall provide a justification in the post-market surveillance plan as referred to in Section 1 of Annex III.** When conducting PMCF, the manufacturer shall proactively collect and evaluate clinical data from the use in or on humans of a device which bears the CE marking and is placed on the market or put into service within its intended purpose as referred to in the relevant conformity assessment procedure, with the aim of confirming the safety and performance throughout the expected lifetime of the device, of ensuring the continued acceptability of identified risks and of detecting emerging risks on the basis of factual evidence.

6. PMCF shall be performed pursuant to a documented method laid down in a PMCF plan.

6.1. The PMCF plan shall specify the methods and procedures for proactively collecting and evaluating clinical data with the aim of:

- (a) confirming the safety and performance of the device throughout its expected lifetime,
- (b) identifying previously unknown side-effects and monitoring the identified side-effects and contraindications,
- (c) identifying and analysing emergent risks on the basis of factual evidence,
- (d) ensuring the continued acceptability of the benefit-risk ratio referred to in Sections 1 and 9 of Annex I, and
- (e) identifying possible systematic misuse or off-label use of the device, with a view to verifying that the intended purpose is correct.

6.2. The PMCF plan shall include at least:

- (a) the general methods and procedures of the PMCF to be applied, such as gathering of clinical experience gained, feedback from users, screening of scientific literature and of other sources of clinical data;
- (b) the specific methods and procedures of PMCF to be applied, such as evaluation of suitable registers or PMCF studies;
- (c) a rationale for the appropriateness of the methods and procedures referred to in points (a) and (b);
- (d) a reference to the relevant parts of the clinical evaluation report referred to in Section 4 and to the risk management referred to in Section 3 of Annex I;
- (e) the specific objectives to be addressed by the PMCF;
- (f) an evaluation of the clinical data relating to equivalent or similar devices;
- (g) reference to any relevant CS, harmonised standards when used by the manufacturer, and relevant guidance on PMCF; and
- (h) a detailed and adequately justified time schedule for PMCF activities (e.g. analysis of PMCF data and reporting) to be undertaken by the manufacturer.

7. The manufacturer shall analyse the findings of the PMCF and document the results in ~~a PMCF evaluation report that shall be part of~~ the clinical evaluation report and the technical documentation.

8. The conclusions of the PMCF ~~evaluation report~~ shall be taken into account for the clinical evaluation referred to in Article 61 and Part A of this Annex and in the risk management referred to in Section 3 of Annex I. If, through the PMCF, the need for preventive and/or corrective measures has been identified, the manufacturer shall implement them.

ANNEX XV CLINICAL INVESTIGATIONS

Chapter I General requirements

1. Ethical principles

Each step in the clinical investigation, from the initial consideration of the need for and justification of the study to the publication of the results, shall be carried out in accordance with recognised ethical principles.

2. Methods

2.1. Clinical investigations shall be performed on the basis of an appropriate plan of investigation reflecting the latest scientific and technical knowledge and defined in such a way as to confirm or refute the manufacturer's claims regarding the safety, performance and aspects relating to benefit-risk of devices as referred to in Article 62(1); the clinical investigations shall include an adequate number of observations to guarantee the scientific validity of the conclusions. The rationale for the design and chosen statistical methodology shall be presented as further described in Section 3.6 of Chapter II of this Annex.

2.2. The procedures used to perform the clinical investigation shall be appropriate to the device under investigation.

2.3. The research methodologies used to perform the clinical investigation shall be appropriate to the device under investigation.

2.4. Clinical investigations shall be performed in accordance with the clinical investigation plan by a sufficient number of intended users and in a clinical environment that is representative of the intended normal conditions of use of the device in the target patient population. Clinical investigations shall be in line with the clinical evaluation plan as referred to in Part A of Annex XIV.

2.5. All the appropriate technical and functional features of the device, in particular those involving safety and performance, and their expected clinical outcomes shall be appropriately addressed in the investigational design. A list of the technical and functional features of the device and the related expected clinical outcomes shall be provided.

2.6. The endpoints of the clinical investigation shall address the intended purpose, clinical benefits, performance and safety of the device. The endpoints shall be determined and assessed using scientifically valid methodologies. The primary endpoint shall be appropriate to the device and clinically relevant. **For combined studies, endpoints reporting on the device and medicinal product together may be used.**

2.7. Investigators shall have access to the technical and clinical data regarding the device. Personnel involved in the conduct of an investigation shall be adequately instructed and trained in the proper use of the investigational device, and as regards the clinical investigation plan and good clinical practice. This training shall be verified and where necessary arranged by the sponsor and documented appropriately.

2.8. The clinical investigation report, signed by the investigator, shall contain a critical evaluation of all the data collected during the clinical investigation, and shall include any negative findings.

Chapter II Documentation regarding the application for clinical investigation

For investigational devices covered by Article 62, the sponsor shall draw up and submit the application in accordance with Article 70 accompanied by the following documents:

1. Application form

The application form shall be duly filled in, containing information regarding:

Proposal for Regulation (EU) 2017/745 – Annex XV

- 1.1. name, address and contact details of the sponsor and, if applicable, name, address and contact details of its contact person or legal representative in accordance with Article 62(2) established in the Union;
- 1.2. if different from those in Section 1.1, name, address and contact details of the manufacturer of the device intended for clinical investigation and, if applicable, of its authorised representative;
- 1.3. title of the clinical investigation;
- 1.4. status of the clinical investigation application (i.e. first submission, resubmission, [substantial modification](#) [\[significant amendment\]](#));
- 1.5. details and/or reference to the clinical evaluation plan;
- 1.6. If the application is a resubmission with regard to a device for which an application has been already submitted, the date or dates and reference number or numbers of the earlier application or, in the case of [substantial modification](#) [\[significant amendment\]](#), reference to the original application. The sponsor shall identify all of the changes from the previous application together with a rationale for those changes, in particular, whether any changes have been made to address conclusions of previous competent authority or ethics committee reviews;
- 1.7. if the application is submitted in parallel with an application for a clinical trial in accordance with Regulation (EU) No 536/2014, reference to the official registration number of the clinical trial;
- 1.8. identification of the Member States and third countries in which the clinical investigation is to be conducted as part of a multicentre or multinational study at the time of application;
- 1.9. a brief description of the investigational device, its classification and other information necessary for the identification of the device and device type;
- 1.10. information as to whether the device incorporates a medicinal substance, including a human blood or plasma derivative or whether it is manufactured utilising non-viable tissues or cells of human or animal origin, or their derivatives;
- 1.11. summary of the clinical investigation plan including the objective or objectives of the clinical investigation, the number and gender of Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operators, criteria for Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operator. selection, whether there are Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the

device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operators under 18 years of age, design of the investigation such as controlled and/or randomised studies, planned dates of commencement and of completion of the clinical investigation;

1.12. if applicable, information regarding a comparator device, its classification and other information necessary for the identification of the comparator device;

1.13. evidence from the sponsor that the clinical investigator and the investigational site are capable of conducting the clinical investigation in accordance with the clinical investigation plan;

1.14. details of the anticipated start date and duration of the investigation;

1.15. details to identify the notified body, if already involved at the stage of application for a clinical investigation;

1.16. confirmation that the sponsor is aware that the competent authority may contact the ethics committee that is assessing or has assessed the application; and

1.17. the statement referred to in Section 4.1.

2. Investigator's Brochure

The investigator's brochure (IB) shall contain the clinical and non-clinical information on the investigational device that is relevant for the investigation and available at the time of application. Any updates to the IB or other relevant information that is newly available shall be brought to the attention of the investigators in a timely manner. The IB shall be clearly identified and contain in particular the following information:

2.1. Identification and description of the device, including information on the intended purpose, the risk classification and applicable classification rule pursuant to Annex VIII, design and manufacturing of the device and reference to previous and similar generations of the device.

2.2. Manufacturer's instructions for installation, maintenance, maintaining hygiene standards and for use, including storage and handling requirements, as well as, to the extent that such information is available, information to be placed on the label, and instructions for use to be provided with the device when placed on the market. In addition, information relating to any relevant training required.

2.3. Pre-clinical evaluation based on relevant pre-clinical testing and experimental data, in particular regarding in-design calculations, *in vitro* tests, *ex vivo* tests, animal tests, mechanical or electrical tests, reliability tests, sterilisation validation, software verification and validation, performance tests, evaluation of biocompatibility and biological safety, as applicable.

2.4. Existing clinical data, in particular:

- from relevant scientific literature available relating to the safety, performance, clinical benefits to patients, design characteristics and intended purpose of the device and/or of equivalent or similar devices;
- other relevant clinical data available relating to the safety, performance, clinical benefits to patients, design characteristics and intended purpose of equivalent or similar devices of the same manufacturer, including length of time on the market and a review of performance, clinical benefit and safety-related issues and any corrective actions taken.

2.5. Summary of the benefit-risk analysis and the risk management, including information regarding known or foreseeable risks, any undesirable side-effects, contraindications and warnings.

2.6. In the case of devices that incorporate a medicinal substance, including a human blood or plasma derivative or devices manufactured utilising non-viable tissues or cells of human or animal origin, or their derivatives, detailed information on the medicinal substance or on the tissues, cells or their derivatives,

Proposal for Regulation (EU) 2017/745 – Annex XV

and on the compliance with the relevant general safety and performance requirements and the specific risk management in relation to the substance or tissues, cells or their derivatives, as well as evidence for the added value of incorporation of such constituents in relation to the clinical benefit and/or safety of the device.

2.7. A list detailing the fulfilment of the relevant general safety and performance requirements set out in Annex I, including the standards and CS applied, in full or in part, as well as a description of the solutions for fulfilling the relevant general safety and performance requirements, in so far as those standards and CS have not or have only been partly fulfilled or are lacking.

2.8. A detailed description of the clinical procedures and diagnostic tests used in the course of the clinical investigation and in particular information on any deviation from normal clinical practice.

3. Clinical investigation plan

The clinical investigation plan (CIP) shall set out the rationale, objectives, design methodology, monitoring, conduct, record-keeping and the method of analysis for the clinical investigation. It shall contain in particular the information as laid down in this Annex. If part of this information is submitted in a separate document, it shall be referenced in the CIP.

3.1. General

3.1.1. Single identification number of the clinical investigation, as referred to in Article 70(1).

3.1.2. Identification of the sponsor – name, address and contact details of the sponsor and, where applicable, the name, address and contact details of the sponsor's contact person or legal representative in accordance with Article 62(2) established in the Union.

3.1.3. Information on the principal investigator at each investigational site, the coordinating investigator for the investigation, the address details for each investigational site and the emergency contact details for the principal investigator at each site. The roles, responsibilities and qualifications of the various kinds of investigators shall be specified in the CIP.

3.1.4. A brief description of how the clinical investigation is financed and a brief description of the agreement between the sponsor and the site.

3.1.5. Overall synopsis of the clinical investigation, in an official Union language determined by the Member State concerned.

3.2. Identification and description of the device, including its intended purpose, its manufacturer, its traceability, the target population, materials coming into contact with the human body, the medical or surgical procedures involved in its use and the necessary training and experience for its use, background literature review, the current state of the art in clinical care in the relevant field of application and the proposed benefits of the new device.

3.3. Risks and clinical benefits of the device to be examined, with justification of the corresponding expected clinical outcomes in the clinical investigation plan.

3.4. Description of the relevance of the clinical investigation in the context of the state of the art of clinical practice.

3.5. Objectives and hypotheses of the clinical investigation.

3.6. Design of the clinical investigation with evidence of its scientific robustness and validity.

3.6.1. General information such as type of investigation with rationale for choosing it, for its endpoints and for its variables as set out in the clinical evaluation plan.

3.6.2. Information on the investigational device, on any comparator and on any other device or medication to be used in the clinical investigation.

3.6.3. Information on Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operators, selection criteria, size of investigation population, representativeness of investigation population in relation to target population and, if applicable, information on vulnerable Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operators involved such as children, pregnant women, immuno-compromised or, elderly Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operators.

3.6.4. Details of measures to be taken to minimise bias, such as randomisation, and management of potential confounding factors.

3.6.5. Description of the clinical procedures and diagnostic methods relating to the clinical investigation and in particular highlighting any deviation from normal clinical practice.

3.6.6. Monitoring plan.

3.7. Statistical considerations, with justification, including a power calculation for the sample size, if applicable.

3.8. Data management.

3.9. Information about any amendments to the CIP.

3.10. Policy regarding follow-up and management of any deviations from the CIP at the investigational site and clear prohibition of use of waivers from the CIP.

3.11. Accountability regarding the device, in particular control of access to the device, follow-up in relation to the device used in the clinical investigation and the return of unused, expired or malfunctioning devices.

3.12. Statement of compliance with the recognised ethical principles for medical research involving humans, and the principles of good clinical practice in the field of clinical investigations of devices, as well as with the applicable regulatory requirements.

3.13. Description of the informed consent process.

3.14. Safety reporting, including definitions of adverse events and serious adverse events, device deficiencies, procedures and timelines for reporting.

3.15. Criteria and procedures for follow-up of Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operators following the end, temporary halt or early termination of an investigation, for follow-up of Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operators who have withdrawn their consent and procedures for Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operators lost to follow-up. Such procedures shall for implantable devices, cover as a minimum traceability.

3.16. A description of the arrangements for taking care of the Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operators after their participation in the clinical investigation has ended, where such additional care is necessary because of the Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and

Proposal for Regulation (EU) 2017/745 – Annex XV

communicated to the relevant economic operators' participation in the clinical investigation and where it differs from that normally expected for the medical condition in question.

3.17. Policy as regards the establishment of the clinical investigation report and publication of results in accordance with the legal requirements and the ethical principles referred to in Section 1 of Chapter I.

3.18. List of the technical and functional features of the device, with specific mention of those covered by the investigation.

3.19. Bibliography.

4. Other information

4.1. A signed statement by the natural or legal person responsible for the manufacture of the investigational device that the device in question conforms to the general safety and performance requirements apart from the aspects covered by the clinical investigation and that, with regard to those aspects, every precaution has been taken to protect the health and safety of the Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operator..

4.2. Where applicable according to national law, copy of the opinion or opinions of the ethics committee or committees concerned. Where according to national law the opinion or opinions of the ethics committee or committees is not required at the time of the submission of the application, a copy of the opinion or opinions shall be submitted as soon as available.

4.3. Proof of insurance cover or indemnification of Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operators in case of injury, pursuant to Article 69 and the corresponding national law.

4.4. Documents to be used to obtain informed consent, including the patient information sheet and the informed consent document.

4.5. Description of the arrangements to comply with the applicable rules on the protection and confidentiality of personal data, in particular:

- organisational and technical arrangements that will be implemented to avoid unauthorised access, disclosure, dissemination, alteration or loss of information and personal data processed;
- a description of measures that will be implemented to ensure confidentiality of records and personal data of Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all

Proposal for Regulation (EU) 2017/745 – Annex XV

appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operators; and

– a description of measures that will be implemented in case of a data security breach in order to mitigate the possible adverse effects.

4.6. Full details of the available technical documentation, for example detailed risk analysis/management documentation or specific test reports shall upon request be submitted to the competent authority reviewing an application.

Chapter III

Other obligations of the sponsor

1. The sponsor shall undertake to keep available for the competent national authorities any documentation necessary to provide evidence for the documentation referred to in Chapter II of this Annex. If the sponsor is not the natural or legal person responsible for the manufacture of the investigational device, that obligation may be fulfilled by that person on behalf of the sponsor.

2. The sponsor shall have an agreement in place to ensure that any serious adverse events or any other event as referred to in Article 80(2) are reported by the investigator or investigators to the sponsor in a timely manner.

3. The documentation mentioned in this Annex shall be kept for a period of time of at least ten years after the clinical investigation with the device in question has ended, or, in the event that the device is subsequently placed on the market, at least ten years after the last device has been placed on the market. In the case of implantable devices, the period shall be at least 15 years.

Each Member State shall require that this documentation is kept at the disposal of the competent authorities for the period referred to in the first subparagraph in case the sponsor, or its contact person or legal representative as referred to in Article 62(2) established within its territory, goes bankrupt or ceases its activity prior to the end of this period.

4. The sponsor shall appoint a monitor that is independent from the investigational site to ensure that the investigation is conducted in accordance with the CIP, the principles of good clinical practice and this Regulation.

5. The sponsor shall complete the follow-up of investigation Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operators.

6. The sponsor shall provide evidence that the investigation is being conducted in line with good clinical practice, for instance through internal or external inspection.

7. The sponsor shall prepare a clinical investigation report which includes at least the following:

Proposal for Regulation (EU) 2017/745 – Annex XV

- Cover/introductory page or pages indicating the title of the investigation, the investigational device, the single identification number, the CIP number and the details with signatures of the coordinating investigators and the principal investigators from each investigational site.
- Details of the author and date of the report.
- A summary of the investigation covering the title, purpose of the investigation, description of the investigation, investigational design and methods used, the results of the investigation and conclusion of the investigation. The completion date of the investigation, and in particular details of early termination, temporary halts or suspensions of investigations.
- investigational device description, in particular clearly defined intended purpose.
- A summary of the clinical investigation plan covering objectives, design, ethical aspects, monitoring and quality measures, selection criteria, target patient populations, sample size, treatment schedules, follow-up duration, concomitant treatments, statistical plan, including hypothesis, sample size calculation and analysis methods, as well as a justification.
- Results of the clinical investigation covering, with rationale and justification, Where, having performed an evaluation pursuant to Article 94, the competent authorities find that the device presents an unacceptable risk to the health or safety of patients, users or other persons, or to other aspects of the protection of public health, they shall without delay require the manufacturer of the devices concerned, its authorised representative and all other relevant economic operators to take all appropriate and duly justified corrective action to bring the device into compliance with the requirements of this Regulation relating to the risk presented by the device and, in a manner that is proportionate to the nature of the risk, to restrict the making available of the device on the market, to subject the making available of the device to specific requirements, to withdraw the device from the market, or to recall it, within a reasonable period that is clearly defined and communicated to the relevant economic operator. demographics, analysis of results related to chosen endpoints, details of subgroup analysis, as well as compliance with the CIP, and covering follow-up of missing data and of patients withdrawing from the clinical investigation, or lost to follow-up.
- Summary of serious adverse events, adverse device effects and device deficiencies and any relevant corrective actions.
- Discussion and overall conclusions covering safety and performance results, assessment of risks and clinical benefits, discussion of clinical relevance in accordance with clinical state of the art, any specific precautions for specific patient populations, implications for the investigational device, limitations of the investigation.

ANNEX XVI

LIST OF GROUPS OF PRODUCTS WITHOUT AN INTENDED MEDICAL PURPOSE REFERRED TO IN
ARTICLE 1(2)

1. Contact lenses or other items intended to be introduced into or onto the eye.
2. Products intended to be totally or partially introduced into the human body through surgically invasive means for the purpose of modifying the anatomy or fixation of body parts with the exception of tattooing products and piercings.
3. Substances, combinations of substances, or items intended to be used for facial or other dermal or mucous membrane filling by subcutaneous, submucous or intradermal injection or other introduction, excluding those for tattooing.
4. Equipment intended to be used to reduce, remove or destroy adipose tissue, such as equipment for liposuction, lipolysis or lipoplasty.
5. High intensity electromagnetic radiation (e.g. infra-red, visible light and ultra-violet) emitting equipment intended for use on the human body, including coherent and non-coherent sources, monochromatic and broad spectrum, such as lasers and intense pulsed light equipment, for skin resurfacing, tattoo or hair removal or other skin treatment.
6. Equipment intended for brain stimulation that apply electrical currents or magnetic or electromagnetic fields that penetrate the cranium to modify neuronal activity in the brain.

ANNEX XVII CORRELATION TABLE

Council Directive 90/385/EEC	Council Directive 93/42/EEC	This Regulation
Article 1(1)	Article 1(1)	Article 1(1)
Article 1(2)	Article 1(2)	Article 2
Article 1(3)	Article 1(3) first subparagraph	Article 1(9) first subparagraph
—	Article 1(3) second subparagraph	Article 1(9) second subparagraph
Article 1(4) and (4a)	Article 1(4) and (4a)	Article 1(8) first subparagraph
Article 1(5)	Article 1(7)	Article 1(11)
Article 1(6)	Article 1(5)	Article 1(6)
—	Article 1(6)	—
—	Article 1(8)	Article 1(13)
Article 2	Article 2	Article 5(1)
Article 3 first paragraph	Article 3 first paragraph	Article 5(2)
Article 3 second paragraph	Article 3 second paragraph	Article 1(12)
Article 4(1)	Article 4(1)	Article 24
Article 4(2)	Article 4(2)	Article 21(1) and (2)
Article 4(3)	Article 4(3)	Article 21(3)
Article 4(4)	Article 4(4)	Article 10(11)
Article 4(5)(a)	Article 4(5) first subparagraph	Article 20(6)
Article 4(5)(b)	Article 4(5) second subparagraph	—
Article 5(1)	Article 5(1)	Article 8(1)
Article 5(2)	Article 5(2)	Article 8(2)
Article 6(1)	Articles 5(3) and 6	—
Article 6(2)	Article 7(1)	Article 114
Article 7	Article 8	Articles 94 to 97
—	Article 9	Article 51
Article 8(1)	Article 10(1)	Articles 87(1) and 89(2)
Article 8(2)	Article 10(2)	Article 87(10) and Article 87(11) first subparagraph
Article 8(3)	Article 10(3)	Article 89(7)
Article 8(4)	Article 10(4)	Article 91
Article 9(1)	Article 11(1)	Article 52(3)
—	Article 11(2)	Article 52(6)
—	Article 11(3)	Article 52(4) and (5)
—	Article 11(4)	—
—	Article 11(5)	Article 52(7)

Proposal for Regulation (EU) 2017/745 – Annex XVII

Council Directive 90/385/EEC	Council Directive 93/42/EEC	This Regulation
Article 9(2)	Article 11 (6)	Article 52(8)
Article 9(3)	Article 11(8)	Article 11(3)
Article 9(4)	Article 11(12)	Article 52(12)
Article 9(5)	Article 11(7)	—
Article 9(6)	Article 11(9)	Article 53(1)
Article 9(7)	Article 11(10)	Article 53(4)
Article 9(8)	Article 11(11)	Article 56(2)
Article 9(9)	Article 11(13)	Article 59
Article 9(10)	Article 11(14)	Article 4(5) and Article 122 third paragraph
—	Article 12	Article 22
—	Article 12a	Article 17
Article 9a(1) first indent	Article 13(1)(c)	—
Article 9a(1) second indent	Article 13(1)(d)	Article 4(1)
—	Article 13(1)(a)	Article 51(3)(a) and Article 51(6)
—	Article 13(1)(b)	Article 51(3)(b) and Article 51(6)
Article 10	Article 15	Articles 62 to 82
Article 10a(1), second sentence of Article 10a(2) and Article 10a(3)	Article 14(1), second sentence of Article 14(2) and Article 14(3)	Articles 29(4), 30 and 31
Article 10a(2), first sentence	Article 14(2) first sentence	Article 11(1)
Article 10b	Article 14a	Articles 33 and 34
Article 10c	Article 14b	Article 98
Article 11(1)	Article 16(1)	Articles 42 and 43
Article 11(2)	Article 16(2)	Article 36
Article 11(3)	Article 16(3)	Article 46(4)
Article 11(4)	Article 16(4)	—
Article 11(5)	Article 16(5)	Article 56(5)
Article 11(6)	Article 16(6)	Article 56(4)
Article 11(7)	Article 16(7)	Articles 38(2) and 44(2)
Article 12	Article 17	Article 20
Article 13	Article 18	Articles 94 to 97
Article 14	Article 19	Article 99
Article 15	Article 20	Article 109
Article 15a	Article 20a	Article 102
Article 16	Article 22	—
Article 17	Article 23	—
—	Article 21	—

--- End of Document ---